

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014163.D
 Acq On : 21 Mar 2023 04:43
 Operator : CG/JU
 Sample : PB151522BL
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK522

Quant Time: Mar 21 05:11:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 21 04:47:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.052	152	206627	20.000	ng/u1	0.00
20) Naphthalene-d8	10.869	136	880510	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.692	164	616843	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.445	188	1466538	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.522	240	1344784	20.000	ng/u1	# 0.00
88) Perylene-d12	24.045	264	1177015	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	26449	4.836	ng/uL	0.00
4) Pyridine-d5	3.846	84	301705	19.925	ng/u1	0.00
7) Phenol-d5	7.211	99	485534	26.628	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.369	67	304877	28.669	ng/u1	0.00
11) 2-Chlorophenol-d4	7.581	132	368146	26.500	ng/u1	0.00
15) 4-Methylphenol-d8	8.763	113	410181	27.539	ng/u1	0.00
21) Nitrobenzene-d5	9.222	128	191229	27.219	ng/u1	0.00
24) 2-Nitrophenol-d4	9.952	143	228007	27.112	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.493	165	396844	25.078	ng/u1	0.00
31) 4-Chloroaniline-d4	11.010	131	337724	16.127	ng/u1	0.00
46) Dimethylphthalate-d6	14.092	166	1408375	26.903	ng/u1	0.00
49) Acenaphthylene-d8	14.387	160	1489810	26.865	ng/u1	0.00
54) 4-Nitrophenol-d4	14.898	143	205972	22.327	ng/u1	0.00
60) Fluorene-d10	15.681	176	1180045	26.598	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	248330	22.404	ng/u1	0.00
73) Anthracene-d10	17.545	188	1969531	27.664	ng/u1	0.00
81) Pyrene-d10	19.769	212	2386208	32.173	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.880	264	1785133	28.513	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
Data File : BP014163.D
Acq On : 21 Mar 2023 04:43
Operator : CG/JU
Sample : PB151522BL
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK522

Quant Time: Mar 21 05:11:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 21 04:47:07 2023
Response via : Initial Calibration

