

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014204.D
 Acq On : 22 Mar 2023 05:09
 Operator : CG/JU
 Sample : 01947-02
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CB931

Quant Time: Mar 22 05:42:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 22 03:05:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	235608	20.000	ng/ul	0.00
20) Naphthalene-d8	10.863	136	1019532	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.681	164	671441	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.439	188	1513841	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.516	240	1320561	20.000	ng/ul	# 0.00
88) Perylene-d12	24.027	264	1150567	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	22325	3.580	ng/uL	0.00
4) Pyridine-d5	3.846	84	195373	11.316	ng/ul	0.00
7) Phenol-d5	7.205	99	106403	5.118	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.363	67	292345	24.109	ng/ul	0.00
11) 2-Chlorophenol-d4	7.575	132	286089	18.060	ng/ul	0.00
15) 4-Methylphenol-d8	8.757	113	212973	12.540	ng/ul	0.00
21) Nitrobenzene-d5	9.216	128	181462	22.307	ng/ul	0.00
24) 2-Nitrophenol-d4	9.940	143	206974	21.255	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.487	165	373158	20.366	ng/ul	0.00
31) 4-Chloroaniline-d4	11.004	131	504490	20.806	ng/ul	0.00
46) Dimethylphthalate-d6	14.086	166	1370915	24.058	ng/ul	0.00
49) Acenaphthylene-d8	14.375	160	1430706	23.701	ng/ul	0.00
54) 4-Nitrophenol-d4	14.910	143	38221	3.806	ng/ul	0.02
60) Fluorene-d10	15.675	176	1123137	23.256	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	187782	16.412	ng/ul	0.00
73) Anthracene-d10	17.533	188	1866343	25.396	ng/ul	0.00
81) Pyrene-d10	19.763	212	2177529	29.897	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.862	264	1560807	25.503	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
Data File : BP014204.D
Acq On : 22 Mar 2023 05:09
Operator : CG/JU
Sample : 01947-02
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CB931

Quant Time: Mar 22 05:42:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 22 03:05:29 2023
Response via : Initial Calibration

