

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032421\
 Data File : BP005038.D
 Acq On : 24 Mar 2021 18:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Mar 24 18:35:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 10:16:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	33134	20.00	ng	# 0.00
21) Naphthalene-d8	10.516	136	124377	20.00	ng	# 0.00
39) Acenaphthene-d10	14.381	164	78888	20.00	ng	0.00
64) Phenanthrene-d10	17.139	188	162536	20.00	ng	# 0.00
76) Chrysene-d12	21.233	240	164160	20.00	ng	0.00
86) Perylene-d12	23.521	264	162053	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.317	112	179445	83.34	ng	0.00
7) Phenol-d6	6.905	99	257633	83.54	ng	0.00
23) Nitrobenzene-d5	8.887	82	240262	82.88	ng	0.00
42) 2,4,6-Tribromophenol	15.881	330	86334	83.99	ng	0.01
45) 2-Fluorobiphenyl	12.999	172	482656	82.76	ng	0.00
79) Terphenyl-d14	19.710	244	756297	83.87	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.258	88	38172	40.71	ng	97
3) Pyridine	3.658	79	102207	43.84	ng	94
4) n-Nitrosodimethylamine	3.570	42	35852	43.01	ng	# 81
6) Aniline	7.064	93	145057	41.51	ng	# 90
8) 2-Chlorophenol	7.293	128	93287	41.08	ng	# 80
9) Benzaldehyde	6.875	77	62538	45.19	ng	# 75
10) Phenol	6.928	94	130315	41.23	ng	84
11) bis(2-Chloroethyl)ether	7.158	93	95908	41.45	ng	95
12) 1,3-Dichlorobenzene	7.616	146	103722	41.11	ng	# 90
13) 1,4-Dichlorobenzene	7.758	146	105945	41.36	ng	# 92
14) 1,2-Dichlorobenzene	8.075	146	101358	41.25	ng	# 94
15) Benzyl Alcohol	7.969	79	96345	40.49	ng	# 88
16) 2,2'-oxybis(1-Chloropr...	8.263	45	71272	41.99	ng	89
17) 2-Methylphenol	8.169	107	84551	41.92	ng	# 90
18) Hexachloroethane	8.799	117	40135	41.12	ng	93
19) n-Nitroso-di-n-propyla...	8.540	70	80744	40.36	ng	# 93
20) 3+4-Methylphenols	8.499	107	114790	41.68	ng	94
22) Acetophenone	8.552	105	144629	40.45	ng	# 94
24) Nitrobenzene	8.928	77	124294	40.69	ng	# 79
25) Isophorone	9.452	82	218430	40.83	ng	# 89
26) 2-Nitrophenol	9.634	139	51086	43.12	ng	# 65
27) 2,4-Dimethylphenol	9.699	122	79032	41.48	ng	88
28) bis(2-Chloroethoxy)met...	9.934	93	112450	41.53	ng	98
29) 2,4-Dichlorophenol	10.163	162	88217	41.38	ng	93
30) 1,2,4-Trichlorobenzene	10.381	180	98479	41.13	ng	96
31) Naphthalene	10.569	128	289231	41.17	ng	99
32) Benzoic acid	9.846	122	59191	41.85	ng	# 64
33) 4-Chloroaniline	10.681	127	116620	41.03	ng	# 84
34) Hexachlorobutadiene	10.852	225	64291	41.72	ng	96
35) Caprolactam	11.481	113	29001	41.46	ng	86
36) 4-Chloro-3-methylphenol	11.816	107	106815	41.48	ng	84
37) 2-Methylnaphthalene	12.187	142	204529	40.32	ng	# 95
38) 1-Methylnaphthalene	12.404	142	193770	40.75	ng	# 98
40) 1,2,4,5-Tetrachloroben...	12.557	216	108163	42.10	ng	# 96
41) Hexachlorocyclopentadiene	12.534	237	61446	43.97	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032421\
 Data File : BP005038.D
 Acq On : 24 Mar 2021 18:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Mar 24 18:35:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 10:16:52 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	12.799	196	73691	41.50	ng	98
44) 2,4,5-Trichlorophenol	12.875	196	81334	42.29	ng	# 94
46) 1,1'-Biphenyl	13.210	154	249364	41.44	ng	96
47) 2-Chloronaphthalene	13.246	162	200890	41.01	ng	96
48) 2-Nitroaniline	13.457	65	66552	42.65	ng	# 60
49) Acenaphthylene	14.098	152	313526	41.16	ng	100
50) Dimethylphthalate	13.846	163	245800	40.45	ng	# 99
51) 2,6-Dinitrotoluene	13.963	165	58213	40.65	ng	# 60
52) Acenaphthene	14.445	154	192988	41.17	ng	98
53) 3-Nitroaniline	14.293	138	55620	42.20	ng	# 61
54) 2,4-Dinitrophenol	14.498	184	37061	42.74	ng	# 76
55) Dibenzofuran	14.781	168	314500	40.94	ng	99
56) 4-Nitrophenol	14.604	139	47664	44.08	ng	# 49
57) 2,4-Dinitrotoluene	14.751	165	80471	42.55	ng	# 63
58) Fluorene	15.434	166	254509	41.03	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.010	232	72207	40.70	ng	98
60) Diethylphthalate	15.216	149	242114	40.74	ng	97
61) 4-Chlorophenyl-phenyle...	15.434	204	133120	40.43	ng	98
62) 4-Nitroaniline	15.463	138	59165	43.69	ng	# 64
63) Azobenzene	15.722	77	272406	40.10	ng	86
65) 4,6-Dinitro-2-methylph...	15.516	198	53841	45.39	ng	87
66) n-Nitrosodiphenylamine	15.645	169	217600	41.65	ng	99
67) 4-Bromophenyl-phenylether	16.328	248	79371	42.15	ng	96
68) Hexachlorobenzene	16.439	284	88084	41.59	ng	95
69) Atrazine	16.610	200	75460	43.49	ng	97
70) Pentachlorophenol	16.787	266	60603	43.34	ng	99
71) Phenanthrene	17.181	178	395985	42.02	ng	99
72) Anthracene	17.269	178	392442	42.63	ng	99
73) Carbazole	17.545	167	365839	42.30	ng	100
74) Di-n-butylphthalate	18.104	149	405523	42.61	ng	# 98
75) Fluoranthene	19.157	202	462718	41.74	ng	98
77) Benzidine	19.345	184	184530	51.23	ng	98
78) Pyrene	19.510	202	475477	42.15	ng	98
80) Butylbenzylphthalate	20.386	149	185026	44.20	ng	# 75
81) Benzo(a)anthracene	21.216	228	465576	41.74	ng	98
82) 3,3'-Dichlorobenzidine	21.151	252	162590	43.38	ng	100
83) Chrysene	21.269	228	457167	41.76	ng	98
84) Bis(2-ethylhexyl)phtha...	21.145	149	268768	43.79	ng	97
85) Di-n-octyl phthalate	22.033	149	450404	43.53	ng	# 94
87) Indeno(1,2,3-cd)pyrene	25.874	276	514571	42.35	ng	97
88) Benzo(b)fluoranthene	22.827	252	492880	42.83	ng	98
89) Benzo(k)fluoranthene	22.874	252	449889	40.80	ng	98
90) Benzo(a)pyrene	23.421	252	431122	41.78	ng	98
91) Dibenzo(a,h)anthracene	25.886	278	444152	42.27	ng	97
92) Benzo(g,h,i)perylene	26.592	276	422715	41.69	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032421\
 Data File : BP005038.D
 Acq On : 24 Mar 2021 18:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Mar 24 18:35:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 10:16:52 2021
 Response via : Initial Calibration









Scan 788 (7.722 min): BP005024.D\data.ms (-779) #1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.728 min Scan# 789
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	152	Resp:	33134
Ion	Ratio	Lower	Upper
152	100		
150	153.2	123.8	185.6
115	66.9	43.8	65.6#



Scan 28 (3.252 min): BP005024.D\data.ms (-23) #2
 1,4-Dioxane
 Concen: 40.71 ng
 RT: 3.258 min Scan# 29
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	88	Resp:	38172
Ion	Ratio	Lower	Upper
88	100		
58	57.7	44.2	66.4
43	20.7	15.2	22.8









Pyridine
Concen: 43.84 ng
RT: 3.658 min Scan# 97
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	79	Resp:	102207
Ion	Ratio	Lower	Upper
79	100		
52	45.7	41.3	61.9
51	26.2	21.2	31.8



n-Nitrosodimethylamine
Concen: 43.01 ng
RT: 3.570 min Scan# 82
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	42	Resp:	35852
Ion	Ratio	Lower	Upper
42	100		
74	157.6	148.6	223.0
44	5.1	5.3	7.9#









2-Fluorophenol
 Concen: 83.34 ng
 RT: 5.317 min Scan# 379
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	112	Resp:	179445
Ion	Ratio	Lower	Upper
112	100		
64	58.5	32.8	49.2#
63	30.9	17.0	25.4#



Aniline
 Concen: 41.51 ng
 RT: 7.064 min Scan# 676
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	93	Resp:	145057
Ion	Ratio	Lower	Upper
93	100		
66	38.6	25.4	38.0#
65	18.2	12.5	18.7









Phenol-d6
 Concen: 83.54 ng
 RT: 6.905 min Scan# 649
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	99	Resp:	257633
Ion	Ratio	Lower	Upper
99	100		
42	14.3	8.6	13.0#
71	41.5	23.4	35.2#



2-Chlorophenol
 Concen: 41.08 ng
 RT: 7.293 min Scan# 715
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	128	Resp:	93287
Ion	Ratio	Lower	Upper
128	100		
130	31.7	13.0	53.0
64	51.2	10.4	50.4#









Scan 643 (6.869 min): BP005024.D\data.ms (-635) #9
 Benzaldehyde
 Concen: 45.19 ng
 RT: 6.875 min Scan# 644
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	77	Resp:	62538
Ion	Ratio	Lower	Upper
77	100		
105	90.5	97.1	137.1#
106	87.1	93.7	133.7#



Scan 652 (6.922 min): BP005024.D\data.ms (-642) #10
 Phenol
 Concen: 41.23 ng
 RT: 6.928 min Scan# 653
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	94	Resp:	130315
Ion	Ratio	Lower	Upper
94	100		
65	27.1	2.3	42.3
66	41.7	10.9	50.9









Scan 691 (7.152 min): BP005024.D\data.ms (-684) #1
 bis(2-Chloroethyl)ether
 Concen: 41.45 ng
 RT: 7.158 min Scan# 692
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleID :
 SSTDCCC040EC


Tgt Ion:	93	Resp:	95908
Ion	Ratio	Lower	Upper
93	100		
63	63.4	39.0	79.0
95	30.9	12.8	52.8



Scan 769 (7.610 min): BP005024.D\data.ms (-760) #12
 1,3-Dichlorobenzene
 Concen: 41.11 ng
 RT: 7.616 min Scan# 770
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	146	Resp:	103722
Ion	Ratio	Lower	Upper
146	100		
148	64.4	52.5	78.7
75	39.6	18.1	27.1#









1,4-Dichlorobenzene
Concen: 41.36 ng
RT: 7.758 min Scan# 794
Delta R.T. -0.000 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	146	Resp:	105945
Ion	Ratio	Lower	Upper
146	100		
148	65.6	51.3	76.9
111	46.4	28.9	43.3#



1,2-Dichlorobenzene
Concen: 41.25 ng
RT: 8.075 min Scan# 848
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	146	Resp:	101358
Ion	Ratio	Lower	Upper
146	100		
148	64.7	51.6	77.4
111	47.7	31.0	46.6#









Benzyl Alcohol
 Concen: 40.49 ng
 RT: 7.969 min Scan# 830
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleID :
 SSTDCCC040EC


Tgt Ion:	79	Resp:	96345
Ion	Ratio	Lower	Upper
79	100		
108	72.3	73.5	110.3#
77	62.5	50.3	75.5



2,2'-oxybis(1-Chloropropane)
 Concen: 41.99 ng
 RT: 8.263 min Scan# 880
 Delta R.T. 0.018 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	45	Resp:	71272
Ion	Ratio	Lower	Upper
45	100		
77	27.6	2.3	42.3
79	21.7	0.0	36.6









2-Methylphenol
 Concen: 41.92 ng
 RT: 8.169 min Scan# 864
 Delta R.T. -0.000 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	107	Resp:	84551
Ion	Ratio	Lower	Upper
107	100		
108	108.5	89.6	134.4
77	52.0	31.8	47.6#
79	52.0	31.1	46.7#



Hexachloroethane
 Concen: 41.12 ng
 RT: 8.799 min Scan# 971
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	117	Resp:	40135
Ion	Ratio	Lower	Upper
117	100		
119	98.7	78.8	118.2
201	101.5	92.3	138.5









#19
 n-Nitroso-di-n-propylamine
 Concen: 40.36 ng
 RT: 8.540 min Scan# 927
 Delta R.T. 0.012 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleID :
 SSTDCCC040EC


Tgt Ion:	70	Resp:	80744
Ion	Ratio	Lower	Upper
70	100		
42	40.7	32.8	49.2
101	11.0	9.6	14.4
130	20.6	24.3	36.5#



#20
 3+4-Methylphenols
 Concen: 41.68 ng
 RT: 8.499 min Scan# 920
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	107	Resp:	114790
Ion	Ratio	Lower	Upper
107	100		
108	89.5	69.0	109.0
77	36.8	8.5	48.5
79	31.0	4.8	44.8









Naphthalene-d8
Concen: 20.00 ng
RT: 10.516 min Scan# 1263
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	136	Resp:	124377
Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.6	13.0
54	6.2	3.6	5.4#
68	6.3	3.3	4.9#




Acetophenone
Concen: 40.45 ng
RT: 8.552 min Scan# 929
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	105	Resp:	144629
Ion	Ratio	Lower	Upper
105	100		
71	2.7	2.6	3.8
51	20.4	13.5	20.3#
120	21.7	19.6	29.4






Abundance Scan 985 (8.881 min): BP005024.D\data.ms (-977) #23




Nitrobenzene-d5
Concen: 82.88 ng
RT: 8.887 min Scan# 986
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	82	Resp:	240262
Ion	Ratio	Lower	Upper
82	100		
128	38.1	45.6	68.4#
54	40.3	30.1	45.1






Abundance Scan 992 (8.922 min): BP005024.D\data.ms (-974) #24




Nitrobenzene
Concen: 40.69 ng
RT: 8.928 min Scan# 993
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	77	Resp:	124294
Ion	Ratio	Lower	Upper
77	100		
123	42.0	48.6	72.8#
65	13.8	10.0	15.0





Abundance Scan 1082 (9.452 min): BP005024.D\data.ms (-1025)




Isophorone
Concen: 40.83 ng
RT: 9.452 min Scan# 1082
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	82	Resp:	218430
Ion	Ratio	Lower	Upper
82	100		
95	7.6	6.4	9.6
138	16.3	18.7	28.1#



Abundance Scan 1082 (9.452 min): BP005038.D\data.ms (-1030) (-)





Abundance Scan 1112 (9.628 min): BP005024.D\data.ms (-1106)




2-Nitrophenol
Concen: 43.12 ng
RT: 9.634 min Scan# 1113
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05


Tgt Ion:	139	Resp:	51086
Ion	Ratio	Lower	Upper
139	100		
109	39.6	22.1	33.1#
65	58.3	26.5	39.7#



Abundance Scan 1113 (9.634 min): BP005038.D\data.ms (-1061) (2)






Abundance Scan 1123 (9.693 min): BP005024.D\data.ms (-1127-)




2,4-Dimethylphenol
Concen: 41.48 ng
RT: 9.699 min Scan# 1124
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	122	Resp:	79032
Ion	Ratio	Lower	Upper
122	100		
107	127.8	87.3	130.9
121	59.4	47.3	70.9







Abundance Scan 1164 (9.934 min): BP005024.D\data.ms (-1128-)



bis(2-Chloroethoxy)methane
Concen: 41.53 ng
RT: 9.934 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	93	Resp:	112450
Ion	Ratio	Lower	Upper
93	100		
95	34.1	26.7	40.1
123	11.1	10.5	15.7









2,4-Dichlorophenol
Concen: 41.38 ng
RT: 10.163 min Scan# 1203
Delta R.T. -0.000 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	162	Resp:	88217
Ion	Ratio	Lower	Upper
162	100		
164	65.2	45.8	85.8
98	41.4	10.2	50.2



1,2,4-Trichlorobenzene
Concen: 41.13 ng
RT: 10.381 min Scan# 1240
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	180	Resp:	98479
Ion	Ratio	Lower	Upper
180	100		
182	100.0	76.2	114.2
145	28.8	23.8	35.8









#31 (-)
Naphthalene
Concen: 41.17 ng
RT: 10.569 min Scan# 1272
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	128	Resp:	289231
Ion	Ratio	Lower	Upper
128	100		
129	10.4	8.8	13.2
127	13.3	10.6	15.8




#32 (-)
Benzoic acid
Concen: 41.85 ng
RT: 9.846 min Scan# 1149
Delta R.T. 0.012 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	122	Resp:	59191
Ion	Ratio	Lower	Upper
122	100		
105	143.4	94.1	134.1#
77	110.6	50.1	90.1#






Abundance Scan 1291 (10.681 min): BP005024.D\data.ms (-1239) (-)




4-Chloroaniline
Concen: 41.03 ng
RT: 10.681 min Scan# 1291
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	127	Resp:	116620
Ion	Ratio	Lower	Upper
127	100		
129	29.9	25.7	38.5
65	36.6	17.3	25.9#
92	23.5	12.9	19.3#







Abundance Scan 1319 (10.846 min): BP005024.D\data.ms (-1239) (-)



Hexachlorobutadiene
Concen: 41.72 ng
RT: 10.852 min Scan# 1320
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	225	Resp:	64291
Ion	Ratio	Lower	Upper
225	100		
223	61.1	50.1	75.1
227	59.4	51.4	77.2









Caprolactam
Concen: 41.46 ng
RT: 11.481 min Scan# 1427
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	113	Resp:	29001
Ion	Ratio	Lower	Upper
113	100		
55	105.0	77.3	117.3
56	94.2	55.9	95.9




4-Chloro-3-methylphenol
Concen: 41.48 ng
RT: 11.816 min Scan# 1484
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	107	Resp:	106815
Ion	Ratio	Lower	Upper
107	100		
144	24.2	22.7	34.1
142	70.9	70.7	106.1





Abundance Scan 1546 (12.181 min): BP005024.D\data.ms (-1537) (-)




2-Methylnaphthalene
Concen: 40.32 ng
RT: 12.187 min Scan# 1547
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	142	Resp:	204529
Ion	Ratio	Lower	Upper
142	100		
141	91.3	71.4	107.2
115	39.5	25.8	38.6#



Abundance Scan 1547 (12.187 min): BP005038.D\data.ms (-1495) (-)





Abundance Scan 1583 (12.398 min): BP005024.D\data.ms (-1574) (-)




1-Methylnaphthalene
Concen: 40.75 ng
RT: 12.404 min Scan# 1584
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05


Tgt Ion:	142	Resp:	193770
Ion	Ratio	Lower	Upper
142	100		
141	91.3	74.4	111.6
116	4.5	2.9	4.3#



Abundance Scan 1584 (12.404 min): BP005038.D\data.ms (-1532) (-)






Abundance Scan 1919 (14.375 min): BP005024.D\data.ms (-1819) (-)




Acenaphthene-d10
Concen: 20.00 ng
RT: 14.381 min Scan# 1920
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	164	Resp:	78888
Ion	Ratio	Lower	Upper
164	100		
162	100.5	79.4	119.2
160	41.9	36.1	54.1






Abundance Scan 1609 (12.551 min): BP005024.D\data.ms (-1609) (-)




1,2,4,5-Tetrachlorobenzene
Concen: 42.10 ng
RT: 12.557 min Scan# 1610
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	216	Resp:	108163
Ion	Ratio	Lower	Upper
216	100		
214	75.2	63.0	94.6
179	21.1	16.9	25.3
108	19.4	12.6	19.0#






Abundance Scan 1605 (12.528 min): BP005024.D\data.ms (-1554) (-)




Hexachlorocyclopentadiene
Concen: 43.97 ng
RT: 12.534 min Scan# 1606
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	237	Resp:	61446
Ion	Ratio	Lower	Upper
237	100		
235	62.3	43.2	83.2
272	11.9	0.0	33.3







Abundance Scan 2174 (15.875 min): BP005024.D\data.ms (-2122) (-)



2,4,6-Tribromophenol
Concen: 83.99 ng
RT: 15.881 min Scan# 2175
Delta R.T. 0.012 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	330	Resp:	86334
Ion	Ratio	Lower	Upper
330	100		
332	95.6	77.1	115.7
141	34.9	23.4	35.2









Scan 1651 (12.798 min): BP005024.D\data.ms (-1643) (-)
 2,4,6-Trichlorophenol
 Concen: 41.50 ng
 RT: 12.799 min Scan# 1651
 Delta R.T. -0.000 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleID :
 SSTDCCC040EC

Tgt Ion:	196	Resp:	73691
Ion	Ratio	Lower	Upper
196	100		
198	98.4	77.0	115.6
200	30.3	25.1	37.7




Scan 1663 (12.869 min): BP005024.D\data.ms (-1657) (-)
 2,4,5-Trichlorophenol
 Concen: 42.29 ng
 RT: 12.875 min Scan# 1664
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	196	Resp:	81334
Ion	Ratio	Lower	Upper
196	100		
198	95.8	77.3	115.9
97	52.8	31.0	46.4#
132	25.0	20.9	31.3






Abundance Scan 1684 (12.993 min): BP005024.D\data.ms (-1675) (-)




2-Fluorobiphenyl
Concen: 82.76 ng
RT: 12.999 min Scan# 1685
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	172	Resp:	482656
Ion	Ratio	Lower	Upper
172	100		
171	34.6	28.5	42.7
170	22.1	19.2	28.8






Abundance Scan 1720 (13.204 min): BP005024.D\data.ms (-1705) (-)




1,1'-Biphenyl
Concen: 41.44 ng
RT: 13.210 min Scan# 1721
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	154	Resp:	249364
Ion	Ratio	Lower	Upper
154	100		
153	39.1	20.4	60.4
76	13.7	0.0	29.3





Abundance Scan 1727 (13.245 min): BP005024.D\data.ms (-1727) (-)




2-Chloronaphthalene
Concen: 41.01 ng
RT: 13.246 min Scan# 1727
Delta R.T. -0.000 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	162	Resp:	200890
Ion	Ratio	Lower	Upper
162	100		
127	40.9	29.9	44.9
164	32.1	26.1	39.1



Abundance Scan 1727 (13.246 min): BP005038.D\data.ms (-1676) (-)





Abundance Scan 1763 (13.457 min): BP005024.D\data.ms (-1763) (-)





2-Nitroaniline
Concen: 42.65 ng
RT: 13.457 min Scan# 1763
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	65	Resp:	66552
Ion	Ratio	Lower	Upper
65	100		
92	73.4	70.6	106.0
138	95.2	133.6	200.4#



Abundance Scan 1763 (13.457 min): BP005038.D\data.ms (-1711) (-)









Acenaphthylene
Concen: 41.16 ng
RT: 14.098 min Scan# 1872
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	152	Resp:	313526
Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	12.9	10.5	15.7




Dimethylphthalate
Concen: 40.45 ng
RT: 13.846 min Scan# 1829
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	163	Resp:	245800
Ion	Ratio	Lower	Upper
163	100		
194	5.4	3.3	4.9#
164	10.3	8.2	12.2



Abundance Scan 1848 (13.957 min): BP005024.D\data.ms (-1850) (-)





2,6-Dinitrotoluene
Concen: 40.65 ng
RT: 13.963 min Scan# 1849
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	165	Resp:	58213
Ion	Ratio	Lower	Upper
165	100		
63	62.9	27.8	41.6#
89	64.4	34.6	52.0#


Abundance Scan 1849 (13.963 min): BP005038.D\data.ms



Abundance Scan 1849 (13.963 min): BP005038.D\data.ms (-1797) (-)





Abundance Scan 1930 (14.439 min): BP005024.D\data.ms (-1950) (-)




Acenaphthene
Concen: 41.17 ng
RT: 14.445 min Scan# 1931
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	154	Resp:	192988
Ion	Ratio	Lower	Upper
154	100		
153	111.9	91.0	136.6
152	53.2	43.0	64.6


Abundance Scan 1931 (14.445 min): BP005038.D\data.ms



Abundance Scan 1931 (14.445 min): BP005038.D\data.ms (-1879) (-)



Abundance Scan 1904 (14.287 min): BP005024.D\data.ms (-1853) (-)





3-Nitroaniline
Concen: 42.20 ng
RT: 14.293 min Scan# 1905
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	138	Resp:	55620
Ion	Ratio	Lower	Upper
138	100		
108	12.6	7.5	11.3#
92	134.1	75.0	112.6#


Abundance Scan 1905 (14.293 min): BP005038.D\data.ms



Abundance Scan 1905 (14.293 min): BP005038.D\data.ms (-1853) (-)





Abundance Scan 1939 (14.492 min): BP005024.D\data.ms (-1851) (-)




2,4-Dinitrophenol
Concen: 42.74 ng
RT: 14.498 min Scan# 1940
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	184	Resp:	37061
Ion	Ratio	Lower	Upper
184	100		
63	73.9	32.3	48.5#
154	64.3	49.1	73.7

Abundance Scan 1940 (14.498 min): BP005038.D\data.ms




Abundance Scan 1940 (14.498 min): BP005038.D\data.ms (-1888) (-)






Abundance Scan 1987 (14.775 min): BP005024.D\data.ms (-1955) (-)

Dibenzofuran
Concen: 40.94 ng
RT: 14.781 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC




Tgt Ion:	168	Resp:	314500
Ion	Ratio	Lower	Upper
168	100		
139	35.4	29.1	43.7
169	13.3	10.4	15.6






Abundance Scan 1957 (14.598 min): BP005024.D\data.ms (-1956) (-)

4-Nitrophenol
Concen: 44.08 ng
RT: 14.604 min Scan# 1958
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05




Tgt Ion:	139	Resp:	47664
Ion	Ratio	Lower	Upper
139	100		
109	92.4	44.6	84.6#
65	113.0	41.6	81.6#






Abundance Scan 1982 (14.745 min): BP005024.D\data.ms (-1952) (-)

2,4-Dinitrotoluene
Concen: 42.55 ng
RT: 14.751 min Scan# 1983
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC







Tgt Ion:	165	Resp:	80471
Ion	Ratio	Lower	Upper
165	100		
63	48.1	20.6	31.0#
89	79.0	42.6	63.8#
182	3.4	2.4	3.6




Abundance Scan 2098 (15.428 min): BP005024.D\data.ms (-2058) (-)

Fluorene
Concen: 41.03 ng
RT: 15.434 min Scan# 2099
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	166	Resp:	254509
Ion	Ratio	Lower	Upper
166	100		
165	95.7	77.8	116.8
167	13.3	11.0	16.4






Abundance Scan 2026 (15.004 min): BP005024.D\data.ms (-2015) (-)




2,3,4,6-Tetrachlorophenol
Concen: 40.70 ng
RT: 15.010 min Scan# 2027
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	232	Resp:	72207
Ion	Ratio	Lower	Upper
232	100		
131	41.5	34.1	51.1
130	2.4	1.8	2.8
166	30.0	24.6	36.8






Abundance Scan 2061 (15.210 min): BP005024.D\data.ms (-2015) (-)




Diethylphthalate
Concen: 40.74 ng
RT: 15.216 min Scan# 2062
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	149	Resp:	242114
Ion	Ratio	Lower	Upper
149	100		
177	24.2	18.0	27.0
150	13.1	10.0	15.0






Abundance Scan 2098 (15.428 min): BP005024.D\data.ms (-2069) (-)




4-Chlorophenyl-phenylether
Concen: 40.43 ng
RT: 15.434 min Scan# 2099
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	204	Resp:	133120
Ion	Ratio	Lower	Upper
204	100		
206	32.9	25.9	38.9
141	56.3	46.5	69.7





Abundance Scan 2103 (15.457 min): BP005024.D\data.ms (-2052) (-)




4-Nitroaniline
Concen: 43.69 ng
RT: 15.463 min Scan# 2104
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05


Tgt Ion:	138	Resp:	59165
Ion	Ratio	Lower	Upper
138	100		
92	58.4	20.7	60.7
108	99.9	47.8	87.8#



Abundance Scan 2104 (15.463 min): BP005038.D\data.ms (-2052) (-)





Abundance Scan 2148 (15.722 min): BP005024.D\data.ms (-2361) (-)




Azobenzene
Concen: 40.10 ng
RT: 15.722 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	77	Resp:	272406
Ion	Ratio	Lower	Upper
77	100		
182	23.8	16.9	56.9
105	20.2	5.6	45.6
51	25.9	2.9	42.9



Abundance Scan 2148 (15.722 min): BP005038.D\data.ms (-2097) (-)





Abundance Scan 2388 (17.133 min): BP005024.D\data.ms (-2361) (-)





Phenanthrene-d10
Concen: 20.00 ng
RT: 17.139 min Scan# 2389
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	188	Resp:	162536
Ion	Ratio	Lower	Upper
188	100		
94	8.7	6.2	9.2
80	10.7	6.2	9.2#



Abundance Scan 2389 (17.139 min): BP005038.D\data.ms (-2337) (-)









4,6-Dinitro-2-methylphenol
 Concen: 45.39 ng
 RT: 15.516 min Scan# 2113
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	198	Resp:	53841
Ion	Ratio	Lower	Upper
198	100		
51	37.1	6.4	46.4
105	43.8	18.9	58.9




n-Nitrosodiphenylamine
 Concen: 41.65 ng
 RT: 15.645 min Scan# 2135
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	169	Resp:	217600
Ion	Ratio	Lower	Upper
169	100		
168	65.6	53.8	80.6
167	35.8	28.9	43.3





Abundance Scan 2250 (16.322 min): BP005024.D\data.ms (-246) (-)




4-Bromophenyl-phenylether
Concen: 42.15 ng
RT: 16.328 min Scan# 2251
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	248	Resp:	79371
Ion	Ratio	Lower	Upper
248	100		
250	97.2	76.7	115.1
141	71.8	53.0	79.6



Abundance Scan 2251 (16.328 min): BP005038.D\data.ms (-2199) (-)





Abundance Scan 2269 (16.433 min): BP005024.D\data.ms (-246) (-)




Hexachlorobenzene
Concen: 41.59 ng
RT: 16.439 min Scan# 2270
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05


Tgt Ion:	284	Resp:	88084
Ion	Ratio	Lower	Upper
284	100		
142	32.6	24.2	36.4
249	33.3	23.8	35.8



Abundance Scan 2270 (16.439 min): BP005038.D\data.ms (-2218) (-)






Abundance Scan 2298 (16.604 min): BP005024.D\data.ms (-2247) (-)




Atrazine
Concen: 43.49 ng
RT: 16.610 min Scan# 2299
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	200	Resp:	75460
Ion	Ratio	Lower	Upper
200	100		
173	26.9	5.9	45.9
215	51.0	28.4	68.4







Abundance Scan 2328 (16.781 min): BP005024.D\data.ms (-2247) (-)



Pentachlorophenol
Concen: 43.34 ng
RT: 16.787 min Scan# 2329
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	266	Resp:	60603
Ion	Ratio	Lower	Upper
266	100		
268	65.4	51.1	76.7
264	63.9	50.6	76.0









Phenanthrene
Concen: 42.02 ng
RT: 17.181 min Scan# 2396
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	178	Resp:	395985
Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.3	22.9
179	14.9	12.5	18.7




Anthracene
Concen: 42.63 ng
RT: 17.269 min Scan# 2411
Delta R.T. -0.000 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	178	Resp:	392442
Ion	Ratio	Lower	Upper
178	100		
176	17.9	15.0	22.4
179	15.3	12.5	18.7






Abundance Scan 2458 (17.545 min): BP005024.D\data.ms (-2406) (-)




Carbazole
Concen: 42.30 ng
RT: 17.545 min Scan# 2458
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	167	Resp:	365839
Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.6	26.4
139	12.8	10.2	15.4






Abundance Scan 2552 (18.098 min): BP005024.D\data.ms (-2501) (-)




Di-n-butylphthalate
Concen: 42.61 ng
RT: 18.104 min Scan# 2553
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	149	Resp:	405523
Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	6.3	3.4	5.0#






Abundance Scan 2731 (19.151 min): BP005024.D\data.ms (-275) (-)




Fluoranthene
Concen: 41.74 ng
RT: 19.157 min Scan# 2732
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	202	Resp:	462718
Ion	Ratio	Lower	Upper
202	100		
101	10.6	0.0	29.4
203	17.5	0.0	37.8






Abundance Scan 3084 (21.227 min): BP005024.D\data.ms (-3075) (-)




Chrysene-d12
Concen: 20.00 ng
RT: 21.233 min Scan# 3085
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	240	Resp:	164160
Ion	Ratio	Lower	Upper
240	100		
120	9.1	7.8	11.6
236	25.1	20.8	31.2





Abundance Scan 2763 (19.339 min): BP005024.D\data.ms (-2712) (-)




Benzidine
Concen: 51.23 ng
RT: 19.345 min Scan# 2764
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	184	Resp:	184530
Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.5	17.3
183	13.4	9.6	14.4



Abundance Scan 2764 (19.345 min): BP005038.D\data.ms (-2712) (-)





Abundance Scan 2791 (19.504 min): BP005024.D\data.ms (-2712) (-)




Pyrene
Concen: 42.15 ng
RT: 19.510 min Scan# 2792
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05


Tgt Ion:	202	Resp:	475477
Ion	Ratio	Lower	Upper
202	100		
200	20.5	17.1	25.7
203	17.3	14.2	21.2



Abundance Scan 2792 (19.510 min): BP005038.D\data.ms (-2740) (-)






Abundance Scan 2825 (19.704 min): BP005024.D\data.ms (-2819) (-)




Terphenyl-d14
Concen: 83.87 ng
RT: 19.710 min Scan# 2826
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	244	Resp:	756297
Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.8	8.6
122	9.4	7.8	11.8






Abundance Scan 2941 (20.386 min): BP005024.D\data.ms (-2829) (-)




Butylbenzylphthalate
Concen: 44.20 ng
RT: 20.386 min Scan# 2941
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	149	Resp:	185026
Ion	Ratio	Lower	Upper
149	100		
91	81.0	45.4	68.0#
206	22.9	17.1	25.7





Abundance Scan 3081 (21.210 min): BP005024.D\data.ms (-3081) (-)




Benzo(a)anthracene
Concen: 41.74 ng
RT: 21.216 min Scan# 3082
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	228	Resp:	465576
Ion	Ratio	Lower	Upper
228	100		
226	26.3	22.1	33.1
229	20.4	15.9	23.9



Abundance Scan 3082 (21.216 min): BP005038.D\data.ms (-3030) (-)





Abundance Scan 3071 (21.151 min): BP005024.D\data.ms (-3081) (-)





3,3'-Dichlorobenzidine
Concen: 43.38 ng
RT: 21.151 min Scan# 3071
Delta R.T. -0.000 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	252	Resp:	162590
Ion	Ratio	Lower	Upper
252	100		
254	63.8	51.0	76.4
126	10.2	7.9	11.9



Abundance Scan 3071 (21.151 min): BP005038.D\data.ms (-3020) (-)









Scan 3090 (21.263 min): BP005024.D\data.ms (-3088) (-)
Chrysene
 Concen: 41.76 ng
 RT: 21.269 min Scan# 3091
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	228	Resp:	457167
Ion	Ratio	Lower	Upper
228	100		
226	29.1	24.4	36.6
229	19.5	15.8	23.6




Scan 3070 (21.145 min): BP005024.D\data.ms (-3082) (-)
Bis(2-ethylhexyl)phthalate
 Concen: 43.79 ng
 RT: 21.145 min Scan# 3070
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	149	Resp:	268768
Ion	Ratio	Lower	Upper
149	100		
167	29.9	22.6	34.0
279	5.4	4.2	6.2






Abundance Scan 3220 (22.027 min): BP005024.D\data.ms (-3185) (-)




Di-n-octyl phthalate
Concen: 43.53 ng
RT: 22.033 min Scan# 3221
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	149	Resp:	450404
Ion	Ratio	Lower	Upper
149	100		
167	1.7	1.3	1.9
43	7.5	3.7	5.5#






Abundance Scan 3473 (23.515 min): BP005024.D\data.ms (-3485) (-)




Perylene-d12
Concen: 20.00 ng
RT: 23.521 min Scan# 3474
Delta R.T. 0.012 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	264	Resp:	162053
Ion	Ratio	Lower	Upper
264	100		
260	24.0	18.8	28.2
265	22.1	17.4	26.2






Abundance Scan 3872 (25.862 min): BP005024.D\data.ms (-3837) (-)




Indeno(1,2,3-cd)pyrene
Concen: 42.35 ng
RT: 25.874 min Scan# 3874
Delta R.T. 0.012 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:276 Resp: 514571
Ion Ratio Lower Upper
276 100
138 20.9 18.7 28.1
277 25.7 20.4 30.6







Abundance Scan 3355 (22.821 min): BP005024.D\data.ms (-3344) (-)



Benzo(b)fluoranthene
Concen: 42.83 ng
RT: 22.827 min Scan# 3356
Delta R.T. 0.006 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:252 Resp: 492880
Ion Ratio Lower Upper
252 100
253 21.4 17.6 26.4
125 7.8 7.3 10.9









Scan 3363 (22.868 min): BP005024.D\data.ms (-3363) (-)
 Benzo(k)fluoranthene
 Concen: 40.80 ng
 RT: 22.874 min Scan# 3364
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	252	Resp:	449889
Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.8	26.8
125	8.6	7.4	11.2




Scan 3456 (23.415 min): BP005024.D\data.ms (-3456) (-)
 Benzo(a)pyrene
 Concen: 41.78 ng
 RT: 23.421 min Scan# 3457
 Delta R.T. 0.006 min
 Lab File: BP005038.D
 Acq: 24 Mar 2021 18:05

Tgt Ion:	252	Resp:	431122
Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.0
125	8.5	8.2	12.4






Abundance Scan 3875 (25.880 min): BP005024.D\data.ms (-3875) (-)




Dibenzo(a,h)anthracene
Concen: 42.27 ng
RT: 25.886 min Scan# 3876
Delta R.T. 0.012 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	278	Resp:	444152
Ion	Ratio	Lower	Upper
278	100		
139	13.2	12.4	18.6
279	24.1	19.0	28.6



Abundance Scan 3995 (26.586 min): BP005024.D\data.ms (-3995) (-)



Benzo(g,h,i)perylene
Concen: 41.69 ng
RT: 26.592 min Scan# 3996
Delta R.T. 0.012 min
Lab File: BP005038.D
Acq: 24 Mar 2021 18:05

Tgt Ion:	276	Resp:	422715
Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.0	28.6
138	19.0	16.6	24.8

