

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032421\
 Data File : BP005031.D
 Acq On : 24 Mar 2021 09:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 10:17:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 10:16:52 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	84	0.00
2	1,4-Dioxane	40.000	40.081	-0.2	91	0.00
3	Pyridine	40.000	42.342	-5.9	84	0.00
4	n-Nitrosodimethylamine	40.000	43.565	-8.9	90	0.00
5 S	2-Fluorophenol	80.000	82.720	-3.4	88	0.00
6	Aniline	40.000	41.257	-3.1	86	0.00
7 S	Phenol-d6	80.000	83.454	-4.3	87	0.00
8	2-Chlorophenol	40.000	41.054	-2.6	86	0.00
9	Benzaldehyde	40.000	44.782	-12.0	88	0.00
10 C	Phenol	40.000	41.326	-3.3	86	0.00
11	bis(2-Chloroethyl)ether	40.000	40.653	-1.6	85	0.00
12	1,3-Dichlorobenzene	40.000	40.523	-1.3	85	0.00
13 C	1,4-Dichlorobenzene	40.000	40.378	-0.9	85	0.00
14	1,2-Dichlorobenzene	40.000	41.296	-3.2	88	0.00
15	Benzyl Alcohol	40.000	43.088	-7.7	88	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.041	-0.1	83	0.00
17	2-Methylphenol	40.000	42.795	-7.0	89	0.00
18	Hexachloroethane	40.000	41.399	-3.5	88	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.527	-3.8	85	0.00
20	3+4-Methylphenols	40.000	41.772	-4.4	85	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	84	0.00
22	Acetophenone	40.000	40.305	-0.8	84	0.00
23 S	Nitrobenzene-d5	80.000	84.632	-5.8	88	0.00
24	Nitrobenzene	40.000	41.712	-4.3	88	0.00
25	Isophorone	40.000	42.020	-5.1	87	0.00
26 C	2-Nitrophenol	40.000	42.895	-7.2	88	0.00
27	2,4-Dimethylphenol	40.000	41.609	-4.0	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.344	-5.9	86	0.00
29 C	2,4-Dichlorophenol	40.000	42.869	-7.2	88	0.00
30	1,2,4-Trichlorobenzene	40.000	41.959	-4.9	88	0.00
31	Naphthalene	40.000	42.237	-5.6	89	0.00
32	Benzoic acid	40.000	39.603	1.0	78	0.00
33	4-Chloroaniline	40.000	42.663	-6.7	89	0.00
34 C	Hexachlorobutadiene	40.000	43.527	-8.8	90	0.00
35	Caprolactam	40.000	42.412	-6.0	87	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.601	-6.5	88	0.00
37	2-Methylnaphthalene	40.000	41.847	-4.6	87	0.00
38	1-Methylnaphthalene	40.000	41.985	-5.0	87	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.770	-1.9	88	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.975	-4.9	88	0.00
42 S	2,4,6-Tribromophenol	80.000	85.953	-7.4	93	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.689	-4.2	88	0.00
44	2,4,5-Trichlorophenol	40.000	41.633	-4.1	89	0.00
45 S	2-Fluorobiphenyl	80.000	80.599	-0.7	88	0.00
46	1,1'-Biphenyl	40.000	40.544	-1.4	88	0.00
47	2-Chloronaphthalene	40.000	41.015	-2.5	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032421\
 Data File : BP005031.D
 Acq On : 24 Mar 2021 09:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 10:17:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 10:16:52 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.389	-6.0	89	0.00
49	Acenaphthylene	40.000	41.151	-2.9	89	0.00
50	Dimethylphthalate	40.000	41.204	-3.0	91	0.00
51	2,6-Dinitrotoluene	40.000	41.042	-2.6	88	0.00
52 C	Acenaphthene	40.000	40.787	-2.0	89	0.00
53	3-Nitroaniline	40.000	42.032	-5.1	87	0.00
54 P	2,4-Dinitrophenol	40.000	41.889	-4.7	88	0.00
55	Dibenzofuran	40.000	41.407	-3.5	91	0.00
56 P	4-Nitrophenol	40.000	42.312	-5.8	87	0.00
57	2,4-Dinitrotoluene	40.000	43.557	-8.9	92	0.00
58	Fluorene	40.000	42.120	-5.3	92	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.163	-5.4	90	0.00
60	Diethylphthalate	40.000	41.839	-4.6	90	0.00
61	4-Chlorophenyl-phenylether	40.000	41.053	-2.6	92	0.00
62	4-Nitroaniline	40.000	44.104	-10.3	91	0.00
63	Azobenzene	40.000	42.077	-5.2	91	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.491	-11.2	92	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.908	-4.8	92	0.00
67	4-Bromophenyl-phenylether	40.000	42.092	-5.2	90	0.00
68	Hexachlorobenzene	40.000	42.266	-5.7	92	0.00
69	Atrazine	40.000	43.775	-9.4	91	0.00
70 C	Pentachlorophenol	40.000	44.641	-11.6	97	0.00
71	Phenanthrene	40.000	42.312	-5.8	92	0.00
72	Anthracene	40.000	42.795	-7.0	93	0.00
73	Carbazole	40.000	42.733	-6.8	93	0.00
74	Di-n-butylphthalate	40.000	43.460	-8.7	92	0.00
75 C	Fluoranthene	40.000	42.950	-7.4	94	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	92	0.00
77	Benzydine	40.000	49.596	-24.0	97	0.00
78	Pyrene	40.000	41.839	-4.6	93	0.00
79 S	Terphenyl-d14	80.000	84.400	-5.5	94	0.00
80	Butylbenzylphthalate	40.000	44.607	-11.5	95	0.00
81	Benzo(a)anthracene	40.000	42.754	-6.9	97	0.00
82	3,3'-Dichlorobenzidine	40.000	43.588	-9.0	95	0.00
83	Chrysene	40.000	42.724	-6.8	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.542	-8.9	92	0.00
85 c	Di-n-octyl phthalate	40.000	43.967	-9.9	95	0.00
86 I	Perylene-d12	20.000	20.000	0.0	94	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.505	-3.8	97	0.00
88	Benzo(b)fluoranthene	40.000	42.167	-5.4	99	0.00
89	Benzo(k)fluoranthene	40.000	40.917	-2.3	94	0.00
90 C	Benzo(a)pyrene	40.000	41.395	-3.5	96	0.00
91	Dibenzo(a,h)anthracene	40.000	42.225	-5.6	98	0.00
92	Benzo(g,h,i)perylene	40.000	41.642	-4.1	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032421\
Data File : BP005031.D
Acq On : 24 Mar 2021 09:17
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 24 10:17:06 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 24 10:16:52 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0