

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032421\
 Data File : BP005038.D
 Acq On : 24 Mar 2021 18:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 18:35:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 10:16:52 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	106	0.01
2	1,4-Dioxane	40.000	40.710	-1.8	116	0.00
3	Pyridine	40.000	43.842	-9.6	110	0.00
4	n-Nitrosodimethylamine	40.000	43.011	-7.5	113	0.00
5 S	2-Fluorophenol	80.000	83.344	-4.2	112	0.00
6	Aniline	40.000	41.506	-3.8	109	0.00
7 S	Phenol-d6	80.000	83.538	-4.4	110	0.01
8	2-Chlorophenol	40.000	41.080	-2.7	108	0.00
9	Benzaldehyde	40.000	45.193	-13.0	112	0.00
10 C	Phenol	40.000	41.230	-3.1	109	0.00
11	bis(2-Chloroethyl)ether	40.000	41.447	-3.6	110	0.00
12	1,3-Dichlorobenzene	40.000	41.111	-2.8	109	0.00
13 C	1,4-Dichlorobenzene	40.000	41.360	-3.4	110	0.00
14	1,2-Dichlorobenzene	40.000	41.250	-3.1	111	0.00
15	Benzyl Alcohol	40.000	40.493	-1.2	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.990	-5.0	110	0.02
17	2-Methylphenol	40.000	41.916	-4.8	110	0.00
18	Hexachloroethane	40.000	41.117	-2.8	111	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.359	-0.9	104	0.01
20	3+4-Methylphenols	40.000	41.676	-4.2	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	40.448	-1.1	104	0.00
23 S	Nitrobenzene-d5	80.000	82.885	-3.6	106	0.00
24	Nitrobenzene	40.000	40.687	-1.7	106	0.00
25	Isophorone	40.000	40.827	-2.1	104	0.00
26 C	2-Nitrophenol	40.000	43.120	-7.8	109	0.00
27	2,4-Dimethylphenol	40.000	41.482	-3.7	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.527	-3.8	104	0.00
29 C	2,4-Dichlorophenol	40.000	41.380	-3.5	104	0.00
30	1,2,4-Trichlorobenzene	40.000	41.129	-2.8	106	0.00
31	Naphthalene	40.000	41.165	-2.9	107	0.00
32	Benzoic acid	40.000	41.847	-4.6	102	0.00
33	4-Chloroaniline	40.000	41.025	-2.6	105	0.00
34 C	Hexachlorobutadiene	40.000	41.723	-4.3	106	0.00
35	Caprolactam	40.000	41.456	-3.6	104	0.01
36 C	4-Chloro-3-methylphenol	40.000	41.476	-3.7	105	0.00
37	2-Methylnaphthalene	40.000	40.315	-0.8	103	0.00
38	1-Methylnaphthalene	40.000	40.746	-1.9	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.104	-5.3	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.973	-9.9	104	0.00
42 S	2,4,6-Tribromophenol	80.000	83.995	-5.0	103	0.01
43 C	2,4,6-Trichlorophenol	40.000	41.501	-3.8	100	0.00
44	2,4,5-Trichlorophenol	40.000	42.290	-5.7	103	0.01
45 S	2-Fluorobiphenyl	80.000	82.764	-3.5	102	0.00
46	1,1'-Biphenyl	40.000	41.442	-3.6	103	0.01
47	2-Chloronaphthalene	40.000	41.014	-2.5	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032421\
 Data File : BP005038.D
 Acq On : 24 Mar 2021 18:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 18:35:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 10:16:52 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.646	-6.6	101	0.00
49	Acenaphthylene	40.000	41.161	-2.9	101	0.00
50	Dimethylphthalate	40.000	40.453	-1.1	101	0.00
51	2,6-Dinitrotoluene	40.000	40.651	-1.6	99	0.00
52 C	Acenaphthene	40.000	41.172	-2.9	102	0.00
53	3-Nitroaniline	40.000	42.199	-5.5	99	0.00
54 P	2,4-Dinitrophenol	40.000	42.736	-6.8	102	0.00
55	Dibenzofuran	40.000	40.938	-2.3	102	0.00
56 P	4-Nitrophenol	40.000	44.085	-10.2	103	0.00
57	2,4-Dinitrotoluene	40.000	42.549	-6.4	102	0.00
58	Fluorene	40.000	41.030	-2.6	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.704	-1.8	98	0.00
60	Diethylphthalate	40.000	40.737	-1.8	99	0.00
61	4-Chlorophenyl-phenylether	40.000	40.434	-1.1	102	0.01
62	4-Nitroaniline	40.000	43.693	-9.2	102	0.00
63	Azobenzene	40.000	40.098	-0.2	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.392	-13.5	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.646	-4.1	100	0.00
67	4-Bromophenyl-phenylether	40.000	42.148	-5.4	99	0.00
68	Hexachlorobenzene	40.000	41.592	-4.0	99	0.00
69	Atrazine	40.000	43.486	-8.7	99	0.00
70 C	Pentachlorophenol	40.000	43.341	-8.4	103	0.00
71	Phenanthrene	40.000	42.024	-5.1	100	0.00
72	Anthracene	40.000	42.626	-6.6	102	0.00
73	Carbazole	40.000	42.296	-5.7	101	0.00
74	Di-n-butylphthalate	40.000	42.606	-6.5	99	0.00
75 C	Fluoranthene	40.000	41.741	-4.4	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	51.234	-28.1#	108	0.00
78	Pyrene	40.000	42.151	-5.4	101	0.00
79 S	Terphenyl-d14	80.000	83.868	-4.8	101	0.00
80	Butylbenzylphthalate	40.000	44.203	-10.5	101	0.00
81	Benzo(a)anthracene	40.000	41.744	-4.4	102	0.00
82	3,3'-Dichlorobenzidine	40.000	43.379	-8.4	101	0.00
83	Chrysene	40.000	41.763	-4.4	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.795	-9.5	100	0.00
85 c	Di-n-octyl phthalate	40.000	43.528	-8.8	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.350	-5.9	103	0.01
88	Benzo(b)fluoranthene	40.000	42.830	-7.1	104	0.00
89	Benzo(k)fluoranthene	40.000	40.798	-2.0	98	0.00
90 C	Benzo(a)pyrene	40.000	41.781	-4.5	101	0.00
91	Dibenzo(a,h)anthracene	40.000	42.271	-5.7	102	0.00
92	Benzo(g,h,i)perylene	40.000	41.692	-4.2	101	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032421\
Data File : BP005038.D
Acq On : 24 Mar 2021 18:05
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 24 18:35:21 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 24 10:16:52 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0