

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032423\
 Data File : BP014293.D
 Acq On : 24 Mar 2023 16:02
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792

Quant Time: Mar 24 22:36:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 24 22:30:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	237780	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.851	136	1098984	20.000	ng/u1	#-0.01
38) Acenaphthene-d10	14.681	164	777983	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.439	188	1860405	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.522	240	1852013	20.000	ng/u1	#-0.01
88) Perylene-d12	24.045	264	1784741	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.399	96	45605	7.247	ng/uL	0.00
4) Pyridine-d5	3.834	84	354922	20.369	ng/u1	0.00
7) Phenol-d5	7.193	99	468892	22.346	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.358	67	286336	23.398	ng/u1	0.00
11) 2-Chlorophenol-d4	7.558	132	342425	21.419	ng/u1	-0.01
15) 4-Methylphenol-d8	8.746	113	392063	22.874	ng/u1	-0.01
21) Nitrobenzene-d5	9.204	128	180377	20.570	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.934	143	217603	20.731	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.475	165	403728	20.442	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.999	131	529621	20.263	ng/u1	-0.01
46) Dimethylphthalate-d6	14.087	166	1307871	19.809	ng/u1	0.00
49) Acenaphthylene-d8	14.375	160	1427801	20.414	ng/u1	0.00
54) 4-Nitrophenol-d4	14.887	143	238684	20.514	ng/u1	0.00
60) Fluorene-d10	15.669	176	1091943	19.514	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.798	200	247439	17.597	ng/u1	0.00
73) Anthracene-d10	17.539	188	1798599	19.915	ng/u1	0.00
81) Pyrene-d10	19.763	212	2170350	21.248	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.874	264	1922187	20.247	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.434	88	51421	7.569	ng/uL	89
5) Pyridine	3.852	79	359166	20.326	ng/u1	94
6) Benzaldehyde	7.169	77	274653	26.306	ng/u1	98
8) Phenol	7.216	94	490280	22.657	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.452	93	377445	22.767	ng/u1	99
12) 2-Chlorophenol	7.593	128	351076	21.796	ng/u1	95
13) 2-Methylphenol	8.481	108	365764	22.731	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.563	45	491543	27.455	ng/u1	97
16) Acetophenone	8.863	105	629415	22.731	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.846	70	342644	24.189	ng/u1	96
18) 4-Methylphenol	8.810	108	408852	23.101	ng/u1	96
19) Hexachloroethane	9.116	117	152352	21.596	ng/u1	95
22) Nitrobenzene	9.252	77	500485	21.337	ng/u1	98
23) Isophorone	9.775	82	982781	21.757	ng/u1	96
25) 2-Nitrophenol	9.963	139	227544	21.033	ng/u1	94
26) 2,4-Dimethylphenol	10.022	107	501862	21.194	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.257	93	568204	21.746	ng/u1	99
29) 2,4-Dichlorophenol	10.499	162	390884	20.063	ng/u1	97
30) Naphthalene	10.904	128	1217774	20.179	ng/u1	99
32) 4-Chloroaniline	11.022	127	539668	20.519	ng/u1	98
33) Hexachlorobutadiene	11.181	225	272706	17.590	ng/u1	95
34) Caprolactam	11.816	113	72772	11.504	ng/u1	88
35) 4-Chloro-3-methylphenol	12.140	107	488475	21.683	ng/u1	98
36) 2-Methylnaphthalene	12.504	142	849683	20.074	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032423\
 Data File : BP014293.D
 Acq On : 24 Mar 2023 16:02
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792

Quant Time: Mar 24 22:36:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 24 22:30:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.728	142	855672	20.112	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.869	216	526828	18.726	ng/ul	98
40) Hexachlorocyclopentadiene	12.845	237	312348	15.504	ng/ul	99
41) 2,4,6-Trichlorophenol	13.116	196	353916	20.026	ng/ul	98
42) 2,4,5-Trichlorophenol	13.187	196	385516	19.877	ng/ul	99
43) 1,1'-Biphenyl	13.510	154	1204702	20.113	ng/ul	99
44) 2-Chloronaphthalene	13.557	162	958901	20.157	ng/ul	98
45) 2-Nitroaniline	13.769	65	334771	23.794	ng/ul	94
47) Dimethylphthalate	14.134	163	1302407	19.929	ng/ul	99
48) 2,6-Dinitrotoluene	14.257	165	273620	21.177	ng/ul	99
50) Acenaphthylene	14.404	152	1513453	20.619	ng/ul	99
51) 3-Nitroaniline	14.592	138	258228	22.572	ng/ul	93
52) Acenaphthene	14.740	153	1048250	20.428	ng/ul	100
53) 2,4-Dinitrophenol	14.798	184	157037	16.275	ng/ul	93
55) 4-Nitrophenol	14.898	109	233022	18.541	ng/ul	86
56) Dibenzofuran	15.075	168	1481132	19.841	ng/ul	100
57) 2,4-Dinitrotoluene	15.045	165	408868	21.180	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.304	232	364294	19.633	ng/ul	98
59) Diethylphthalate	15.492	149	1349617	20.361	ng/ul	99
61) Fluorene	15.728	166	1227989	19.845	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.716	204	651247	19.068	ng/ul	99
63) 4-Nitroaniline	15.763	138	250626	21.313	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.810	198	242375	18.068	ng/ul	92
67) N-Nitrosodiphenylamine	15.934	169	1079398	20.543	ng/ul	98
68) 4-Bromophenyl-phenylether	16.616	248	427146	18.992	ng/ul	99
69) Hexachlorobenzene	16.734	284	492408	18.579	ng/ul	95
70) Atrazine	16.892	200	432827	19.099	ng/ul	98
71) Pentachlorophenol	17.086	266	307476	17.749	ng/ul	98
72) Phenanthrene	17.481	178	2030994	20.016	ng/ul	100
74) Anthracene	17.569	178	2096195	20.374	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.475	216	548195	19.200	ng/ul	99
76) Pentachlorobenzene	14.992	250	555815	18.740	ng/ul	99
77) Carbazole	17.839	167	1876771	20.680	ng/ul	99
78) Di-n-butylphthalate	18.369	149	2411057	21.258	ng/ul	100
80) Fluoranthene	19.439	202	2545328	21.862	ng/ul#	96
82) Pyrene	19.792	202	2625496	21.617	ng/ul#	95
83) Butylbenzylphthalate	20.645	149	1142827	23.336	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.433	252	820977	17.316	ng/ul	99
85) Benzo(a)anthracene	21.504	228	2639339	20.261	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.398	149	1703748	23.270	ng/ul	99
87) Chrysene	21.563	228	2478205	19.962	ng/ul	99
89) Di-n-octyl phthalate	22.363	149	2846285	25.274	ng/ul	100
90) Benzo(b)fluoranthene	23.316	252	2410022	20.265	ng/ul	99
91) Benzo(k)fluoranthene	23.316	252	2414658	21.293	ng/ul	99
93) Benzo(a)pyrene	23.933	252	2081389	20.122	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.686	276	2376422	17.355	ng/ul	98
95) Dibenzo(a,h)anthracene	26.709	278	1993716	17.518	ng/ul	99
96) Benzo(g,h,i)perylene	27.504	276	1863243	17.024	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.569 ng/uL
RT: 3.434 min Scan# 41
Delta R.T. -0.006 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 88 Resp: 51421
Ion Ratio Lower Upper
88 100
43 28.5 23.9 35.9
58 73.9 50.2 75.2



#5
Pyridine
Concen: 20.326 ng/uL
RT: 3.852 min Scan# 113
Delta R.T. -0.006 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion: 79 Resp: 359166
Ion Ratio Lower Upper
79 100
52 75.1 55.1 82.7
51 37.0 28.9 43.3









#6
Benzaldehyde
Concen: 26.306 ng/ul
RT: 7.169 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 77 Resp: 274653
Ion Ratio Lower Upper
77 100
105 91.3 74.2 111.2
106 85.2 69.8 104.8



#8
Phenol
Concen: 22.657 ng/ul
RT: 7.216 min Scan# 685
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion: 94 Resp: 490280
Ion Ratio Lower Upper
94 100
65 31.4 25.2 37.8
66 45.6 37.2 55.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 22.767 ng/ul
 RT: 7.452 min Scan# 71
 Delta R.T. -0.006 min
 Lab File: BP014293.D
 Acq: 24 Mar 2023 16:02




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792


Tgt Ion: 93 Resp: 377445
 Ion Ratio Lower Upper
 93 100
 63 75.0 58.8 88.2
 95 32.0 25.8 38.6



#12
 2-Chlorophenol
 Concen: 21.796 ng/ul
 RT: 7.593 min Scan# 749
 Delta R.T. -0.012 min
 Lab File: BP014293.D
 Acq: 24 Mar 2023 16:02

Tgt Ion: 128 Resp: 351076
 Ion Ratio Lower Upper
 128 100
 64 53.4 38.3 57.5
 130 32.4 25.7 38.5





#13

2-Methylphenol

Concen: 22.731 ng/ul

RT: 8.481 min Scan# 900

Delta R.T. -0.006 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :





SSTD020792

Tgt Ion:108 Resp: 365764

Ion Ratio Lower Upper

108 100

107 90.0 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 27.455 ng/ul

RT: 8.563 min Scan# 914

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion: 45 Resp: 491543


Ion Ratio Lower Upper

45 100

77 16.6 14.2 21.2

79 12.0 10.6 15.8









#16
Acetophenone
Concen: 22.731 ng/ul
RT: 8.863 min Scan# 965
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:105 Resp: 629415
Ion Ratio Lower Upper
105 100
77 84.7 65.4 98.2
51 28.6 21.8 32.8



#17
N-Nitroso-di-n-propylamine
Concen: 24.189 ng/ul
RT: 8.846 min Scan# 962
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion: 70 Resp: 342644
Ion Ratio Lower Upper
70 100
42 53.8 40.9 61.3
101 9.4 7.9 11.9
130 18.4 16.6 25.0









#18
4-Methylphenol
Concen: 23.101 ng/ul
RT: 8.810 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:108 Resp: 408852
Ion Ratio Lower Upper
108 100
107 115.9 89.8 134.6



#19
Hexachloroethane
Concen: 21.596 ng/ul
RT: 9.116 min Scan# 1008
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:117 Resp: 152352
Ion Ratio Lower Upper
117 100
201 109.9 91.7 137.5
199 68.8 59.1 88.7









#22
Nitrobenzene
Concen: 21.337 ng/ul
RT: 9.252 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 77 Resp: 500485
Ion Ratio Lower Upper
77 100
123 39.5 32.6 48.8
65 13.6 11.2 16.8



#23
Isophorone
Concen: 21.757 ng/ul
RT: 9.775 min Scan# 1120
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion: 82 Resp: 982781
Ion Ratio Lower Upper
82 100
95 7.5 7.1 10.7
138 14.8 13.3 19.9









#25
2-Nitrophenol
Concen: 21.033 ng/ul
RT: 9.963 min Scan# 1153
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:139 Resp: 227544
Ion Ratio Lower Upper
139 100
65 62.3 46.6 70.0
109 38.4 34.5 51.7



#26
2,4-Dimethylphenol
Concen: 21.194 ng/ul
RT: 10.022 min Scan# 1162
Delta R.T. -0.006 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:107 Resp: 501862
Ion Ratio Lower Upper
107 100
121 47.6 37.1 55.7
122 78.1 61.2 91.8





#27

Bis(2-Chloroethoxy)methane

Concen: 21.746 ng/ul

RT: 10.257 min Scan# 1202

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion: 93 Resp: 568204

Ion Ratio Lower Upper

93 100

95 32.8 25.6 38.4

123 10.5 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 20.063 ng/ul

RT: 10.499 min Scan# 1243

Delta R.T. -0.018 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:162 Resp: 390884


Ion Ratio Lower Upper

162 100

164 65.7 52.0 78.0

98 40.1 29.1 43.7









#30
Naphthalene
Concen: 20.179 ng/ul
RT: 10.904 min Scan# 1313
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:128 Resp: 1217774
Ion Ratio Lower Upper
128 100
129 11.1 8.5 12.7
127 13.3 11.0 16.4



#32
4-Chloroaniline
Concen: 20.519 ng/ul
RT: 11.022 min Scan# 1332
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:127 Resp: 539668
Ion Ratio Lower Upper
127 100
129 32.0 24.9 37.3









#33
Hexachlorobutadiene
Concen: 17.590 ng/ul
RT: 11.181 min Scan# 111
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:225 Resp: 272706
Ion Ratio Lower Upper
225 100
223 62.2 48.0 72.0
227 68.2 50.3 75.5



#34
Caprolactam
Concen: 11.504 ng/ul
RT: 11.816 min Scan# 1467
Delta R.T. -0.006 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:113 Resp: 72772
Ion Ratio Lower Upper
113 100
55 176.2 124.0 186.0
56 134.7 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 21.683 ng/ul
RT: 12.140 min Scan# 1522
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:107 Resp: 488475
Ion Ratio Lower Upper
107 100
144 23.7 19.4 29.0
142 72.3 59.6 89.4



#36
2-Methylnaphthalene
Concen: 20.074 ng/ul
RT: 12.504 min Scan# 1584
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:142 Resp: 849683
Ion Ratio Lower Upper
142 100
141 89.0 71.7 107.5









#37
1-Methylnaphthalene
Concen: 20.112 ng/ul
RT: 12.728 min Scan# 1622
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:142 Resp: 855672
Ion Ratio Lower Upper
142 100
141 92.6 73.8 110.6
116 4.2 3.7 5.5



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.726 ng/ul
RT: 12.869 min Scan# 1646
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:216 Resp: 526828
Ion Ratio Lower Upper
216 100
214 77.7 61.7 92.5
179 22.0 16.2 24.4
108 18.0 13.2 19.8









#40
Hexachlorocyclopentadiene
Concen: 15.504 ng/ul
RT: 12.845 min Scan# 1642
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:237 Resp: 312348
Ion Ratio Lower Upper
237 100
235 62.4 49.0 73.6
272 12.5 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 20.026 ng/ul
RT: 13.116 min Scan# 1688
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:196 Resp: 353916
Ion Ratio Lower Upper
196 100
198 95.9 78.5 117.7
200 31.2 25.4 38.0









#42
2,4,5-Trichlorophenol
Concen: 19.877 ng/ul
RT: 13.187 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:196 Resp: 385516
Ion Ratio Lower Upper
196 100
198 96.0 77.6 116.4
200 30.9 25.3 37.9



#43
1,1'-Biphenyl
Concen: 20.113 ng/ul
RT: 13.510 min Scan# 1755
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:154 Resp: 1204702
Ion Ratio Lower Upper
154 100
153 40.9 32.6 49.0
76 13.9 9.4 14.0









#44
2-Chloronaphthalene
Concen: 20.157 ng/ul
RT: 13.557 min Scan# 1763
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:162 Resp: 958901
Ion Ratio Lower Upper
162 100
164 32.1 26.0 39.0
127 40.8 31.3 46.9



#45
2-Nitroaniline
Concen: 23.794 ng/ul
RT: 13.769 min Scan# 1799
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion: 65 Resp: 334771
Ion Ratio Lower Upper
65 100
92 68.4 56.3 84.5
138 93.9 81.4 122.2









#47
Dimethylphthalate
Concen: 19.929 ng/ul
RT: 14.134 min Scan# 1861
Delta R.T. -0.006 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:163 Resp: 1302407
Ion Ratio Lower Upper
163 100
194 4.1 3.6 5.4
164 9.8 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 21.177 ng/ul
RT: 14.257 min Scan# 1882
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:165 Resp: 273620
Ion Ratio Lower Upper
165 100
89 64.1 51.7 77.5
121 21.6 17.5 26.3





#50

Acenaphthylene

Concen: 20.619 ng/ul

RT: 14.404 min Scan# 1907

Delta R.T. -0.006 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:152 Resp: 1513453

Ion Ratio Lower Upper

152 100

151 20.6 16.2 24.2

153 12.6 10.4 15.6



#51

3-Nitroaniline

Concen: 22.572 ng/ul

RT: 14.592 min Scan# 1939

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:138 Resp: 258228


Ion Ratio Lower Upper

138 100

108 13.1 10.5 15.7

92 131.8 98.5 147.7





#52

Acenaphthene

Concen: 20.428 ng/ul

RT: 14.740 min Scan# 1965

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:153 Resp: 1048250

Ion Ratio Lower Upper

153 100

152 47.0 37.7 56.5

154 87.7 70.4 105.6



#53

2,4-Dinitrophenol

Concen: 16.275 ng/ul

RT: 14.798 min Scan# 1974

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:184 Resp: 157037


Ion Ratio Lower Upper

184 100

63 71.8 51.9 77.9

154 69.5 58.8 88.2









#55
4-Nitrophenol
Concen: 18.541 ng/ul
RT: 14.898 min Scan# 1991
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:109 Resp: 233022
Ion Ratio Lower Upper
109 100
139 102.6 74.5 111.7
65 126.8 86.6 130.0



#56
Dibenzofuran
Concen: 19.841 ng/ul
RT: 15.075 min Scan# 2021
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:168 Resp: 1481132
Ion Ratio Lower Upper
168 100
139 39.9 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 21.180 ng/ul
RT: 15.045 min Scan# 2016
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:165 Resp: 408868
Ion Ratio Lower Upper
165 100
63 48.1 34.6 52.0
182 2.9 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.633 ng/ul
RT: 15.304 min Scan# 2060
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:232 Resp: 364294
Ion Ratio Lower Upper
232 100
131 44.0 34.1 51.1
130 2.4 2.1 3.1
166 30.1 23.0 34.6









#59
Diethylphthalate
Concen: 20.361 ng/ul
RT: 15.492 min Scan# 2092
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:149 Resp: 1349617
Ion Ratio Lower Upper
149 100
177 22.1 18.2 27.2
150 12.4 10.0 15.0



#61
Fluorene
Concen: 19.845 ng/ul
RT: 15.728 min Scan# 2132
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:166 Resp: 1227989
Ion Ratio Lower Upper
166 100
165 98.2 78.9 118.3
167 13.2 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 19.068 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:204 Resp: 651247
Ion Ratio Lower Upper
204 100
206 32.4 26.6 39.8
141 61.4 49.2 73.8



#63
4-Nitroaniline
Concen: 21.313 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:138 Resp: 250626
Ion Ratio Lower Upper
138 100
92 62.0 44.5 66.7
108 102.2 87.2 130.8









#66
4,6-Dinitro-2-methylphenol
Concen: 18.068 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:198 Resp: 242375
Ion Ratio Lower Upper
198 100
182 4.2 3.0 4.6
77 30.2 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 20.543 ng/ul
RT: 15.934 min Scan# 2167
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:169 Resp: 1079398
Ion Ratio Lower Upper
169 100
168 65.3 53.7 80.5
167 36.3 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 18.992 ng/ul
RT: 16.616 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:248 Resp: 427146
Ion Ratio Lower Upper
248 100
250 97.9 78.8 118.2
141 77.4 60.8 91.2



#69
Hexachlorobenzene
Concen: 18.579 ng/ul
RT: 16.734 min Scan# 2303
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:284 Resp: 492408
Ion Ratio Lower Upper
284 100
142 33.6 23.2 34.8
249 33.1 25.7 38.5









#70
Atrazine
Concen: 19.099 ng/ul
RT: 16.892 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:200 Resp: 432827
Ion Ratio Lower Upper
200 100
173 25.3 19.0 28.4
215 47.5 38.5 57.7



#71
Pentachlorophenol
Concen: 17.749 ng/ul
RT: 17.086 min Scan# 2363
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:266 Resp: 307476
Ion Ratio Lower Upper
266 100
264 62.6 48.4 72.6
268 64.0 50.4 75.6





#72

Phenanthrene

Concen: 20.016 ng/ul

RT: 17.481 min Scan# 2430

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:178 Resp: 2030994

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.2 15.4 23.0



#74

Anthracene

Concen: 20.374 ng/ul

RT: 17.569 min Scan# 2445

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:178 Resp: 2096195


Ion Ratio Lower Upper

178 100

179 15.1 12.5 18.7

176 18.3 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.200 ng/uL

RT: 13.475 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:216 Resp: 548195

Ion Ratio Lower Upper

216 100

214 78.0 61.4 92.2

218 47.8 39.1 58.7



#76

Pentachlorobenzene

Concen: 18.740 ng/uL

RT: 14.992 min Scan# 2007

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:250 Resp: 555815


Ion Ratio Lower Upper

250 100

248 63.6 51.0 76.4

252 64.3 50.8 76.2









#77
Carbazole
Concen: 20.680 ng/ul
RT: 17.839 min Scan# 2491
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:167 Resp: 1876771
Ion Ratio Lower Upper
167 100
166 21.8 17.5 26.3
139 13.5 10.5 15.7



#78
Di-n-butylphthalate
Concen: 21.258 ng/ul
RT: 18.369 min Scan# 2581
Delta R.T. -0.012 min
Lab File: BP014293.D
Acq: 24 Mar 2023 16:02

Tgt Ion:149 Resp: 2411057
Ion Ratio Lower Upper
149 100
150 9.2 7.4 11.2
104 5.8 4.6 6.8





#80

Fluoranthene

Concen: 21.862 ng/ul

RT: 19.439 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:202 Resp: 2545328

Ion Ratio Lower Upper

202 100

101 9.9 6.6 10.0

100 7.5 5.0 7.4#



#82

Pyrene

Concen: 21.617 ng/ul

RT: 19.792 min Scan# 2823

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:202 Resp: 2625496


Ion Ratio Lower Upper

202 100

101 11.9 7.8 11.8#

100 9.8 6.5 9.7#





#83

Butylbenzylphthalate

Concen: 23.336 ng/ul

RT: 20.645 min Scan# 2968

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:149 Resp: 1142827

Ion Ratio Lower Upper

149 100

91 77.8 57.9 86.9

206 21.9 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 17.316 ng/ul

RT: 21.433 min Scan# 3102

Delta R.T. -0.006 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:252 Resp: 820977


Ion Ratio Lower Upper

252 100

254 63.4 51.3 76.9

126 9.1 6.9 10.3





#85

Benzo(a)anthracene

Concen: 20.261 ng/ul

RT: 21.504 min Scan# 3114

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:228 Resp: 2639339

Ion Ratio Lower Upper

228 100

229 19.6 15.5 23.3

226 27.3 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.270 ng/ul

RT: 21.398 min Scan# 3096

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:149 Resp: 1703748


Ion Ratio Lower Upper

149 100

167 27.6 22.3 33.5

279 4.6 4.3 6.5





#87

Chrysene

Concen: 19.962 ng/ul

RT: 21.563 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:228 Resp: 2478205

Ion Ratio Lower Upper

228 100

226 29.9 24.2 36.2

229 19.9 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 25.274 ng/ul




RT: 22.363 min Scan# 3260


Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Tgt Ion:149 Resp: 2846285





#90

Benzo(b)fluoranthene

Concen: 20.265 ng/ul

RT: 23.316 min Scan# 3413

Delta R.T. 0.035 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:252 Resp: 2410022

Ion Ratio Lower Upper

252 100

253 21.7 17.3 25.9

125 7.6 5.4 8.2



#91

Benzo(k)fluoranthene

Concen: 21.293 ng/ul

RT: 23.316 min Scan# 3422

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:252 Resp: 2414658


Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 7.6 5.5 8.3





#93

Benzo(a)pyrene

Concen: 20.122 ng/ul

RT: 23.933 min Scan# 3

Delta R.T. -0.012 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:252 Resp: 2081389

Ion Ratio Lower Upper

252 100

253 21.3 17.4 26.2

125 8.6 6.2 9.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.355 ng/ul

RT: 26.686 min Scan# 3995

Delta R.T. -0.023 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02




Tgt Ion:276 Resp: 2376422


Ion Ratio Lower Upper

276 100

138 18.4 13.1 19.7

277 25.3 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 17.518 ng/ul

RT: 26.709 min Scan# 3998

Delta R.T. -0.018 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:278 Resp: 1993716

Ion Ratio Lower Upper

278 100

139 12.3 8.7 13.1

279 23.7 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 17.024 ng/ul

RT: 27.504 min Scan# 4134

Delta R.T. -0.023 min

Lab File: BP014293.D

Acq: 24 Mar 2023 16:02

Tgt Ion:276 Resp: 1863243

Ion Ratio Lower Upper

276 100

138 17.4 11.9 17.9

277 24.5 18.7 28.1

