

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032923\
 Data File : BP014368.D
 Acq On : 29 Mar 2023 18:11
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020795

Quant Time: Mar 30 01:04:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 00:56:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	165303	20.000	ng/u1	0.00
20) Naphthalene-d8	10.845	136	722171	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.675	164	489994	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.433	188	1119471	20.000	ng/u1	0.00
79) Chrysene-d12	21.521	240	1078475	20.000	ng/u1	0.00
88) Perylene-d12	24.045	264	1013991	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	35803	8.399	ng/uL	0.00
4) Pyridine-d5	3.822	84	235257	21.536	ng/u1	0.00
7) Phenol-d5	7.175	99	317417	20.821	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.346	67	211852	21.659	ng/u1	0.00
11) 2-Chlorophenol-d4	7.546	132	235968	21.415	ng/u1	0.00
15) 4-Methylphenol-d8	8.734	113	262539	21.350	ng/u1	0.00
21) Nitrobenzene-d5	9.199	128	121844	20.900	ng/u1	0.00
24) 2-Nitrophenol-d4	9.922	143	143011	21.423	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.463	165	257819	21.206	ng/u1	0.00
31) 4-Chloroaniline-d4	10.987	131	349086	21.798	ng/u1	0.00
46) Dimethylphthalate-d6	14.081	166	849399	21.379	ng/u1	0.00
49) Acenaphthylene-d8	14.369	160	947040	21.483	ng/u1	0.00
54) 4-Nitrophenol-d4	14.881	143	137908	22.977	ng/u1	0.00
60) Fluorene-d10	15.669	176	717580	21.658	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.792	200	167231	20.302	ng/u1	0.00
73) Anthracene-d10	17.533	188	1140220	21.046	ng/u1	0.00
81) Pyrene-d10	19.763	212	1334155	18.417	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.880	264	1155286	21.445	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	38600	8.240	ng/uL	96
5) Pyridine	3.846	79	247021	21.609	ng/u1	97
6) Benzaldehyde	7.157	77	196299	25.905	ng/u1	99
8) Phenol	7.205	94	331262	20.947	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.440	93	269751	21.226	ng/u1	96
12) 2-Chlorophenol	7.581	128	245619	21.417	ng/u1	98
13) 2-Methylphenol	8.469	108	250910	21.274	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.551	45	360477	21.603	ng/u1	99
16) Acetophenone	8.857	105	440829	21.709	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.834	70	241886	22.050	ng/u1	98
18) 4-Methylphenol	8.799	108	275949	21.305	ng/u1	97
19) Hexachloroethane	9.104	117	108578	21.668	ng/u1	93
22) Nitrobenzene	9.240	77	354656	21.175	ng/u1	98
23) Isophorone	9.769	82	659569	21.364	ng/u1	99
25) 2-Nitrophenol	9.957	139	148688	21.102	ng/u1	98
26) 2,4-Dimethylphenol	10.010	107	335031	21.243	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.246	93	387200	20.919	ng/u1	99
29) 2,4-Dichlorophenol	10.493	162	253545	21.191	ng/u1	96
30) Naphthalene	10.893	128	845573	21.021	ng/u1	99
32) 4-Chloroaniline	11.010	127	361103	22.263	ng/u1	99
33) Hexachlorobutadiene	11.169	225	192507	20.309	ng/u1	97
34) Caprolactam	11.798	113	78163	22.113	ng/u1	91
35) 4-Chloro-3-methylphenol	12.134	107	318932	21.672	ng/u1	99
36) 2-Methylnaphthalene	12.498	142	582768	21.377	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032923\
 Data File : BP014368.D
 Acq On : 29 Mar 2023 18:11
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020795

Quant Time: Mar 30 01:04:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 00:56:20 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.716	142	582196	21.534	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.863	216	358122	20.499	ng/ul	99
40) Hexachlorocyclopentadiene	12.834	237	212350	18.770	ng/ul	99
41) 2,4,6-Trichlorophenol	13.110	196	224985	20.811	ng/ul	98
42) 2,4,5-Trichlorophenol	13.187	196	244127	21.212	ng/ul	96
43) 1,1'-Biphenyl	13.504	154	817610	20.653	ng/ul	99
44) 2-Chloronaphthalene	13.551	162	646097	20.642	ng/ul	99
45) 2-Nitroaniline	13.763	65	216202	22.304	ng/ul	98
47) Dimethylphthalate	14.128	163	853157	21.477	ng/ul	99
48) 2,6-Dinitrotoluene	14.257	165	174703	22.511	ng/ul	94
50) Acenaphthylene	14.398	152	999301	21.285	ng/ul	99
51) 3-Nitroaniline	14.592	138	163719	23.547	ng/ul	96
52) Acenaphthene	14.739	153	699700	21.090	ng/ul	99
53) 2,4-Dinitrophenol	14.804	184	99153	19.729	ng/ul	98
55) 4-Nitrophenol	14.898	109	135557	22.884	ng/ul	93
56) Dibenzofuran	15.075	168	985638	21.134	ng/ul	99
57) 2,4-Dinitrotoluene	15.045	165	255468	23.206	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.298	232	225815	21.831	ng/ul	97
59) Diethylphthalate	15.486	149	868563	22.003	ng/ul	97
61) Fluorene	15.728	166	813631	21.523	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.716	204	430904	21.217	ng/ul	97
63) 4-Nitroaniline	15.757	138	142172	25.167	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.810	198	160288	20.079	ng/ul	98
67) N-Nitrosodiphenylamine	15.933	169	688571	19.769	ng/ul	99
68) 4-Bromophenyl-phenylether	16.616	248	275980	19.812	ng/ul	95
69) Hexachlorobenzene	16.733	284	322226	20.138	ng/ul	98
70) Atrazine	16.886	200	272724	21.509	ng/ul	99
71) Pentachlorophenol	17.086	266	183982	19.847	ng/ul	98
72) Phenanthrene	17.480	178	1304180	20.940	ng/ul	100
74) Anthracene	17.569	178	1331016	21.204	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.475	216	365501	18.548	ng/ul	99
76) Pentachlorobenzene	14.992	250	373846	19.013	ng/ul	99
77) Carbazole	17.845	167	1167977	22.235	ng/ul	98
78) Di-n-butylphthalate	18.375	149	1479893	22.405	ng/ul	99
80) Fluoranthene	19.439	202	1595280	18.625	ng/ul	100
82) Pyrene	19.792	202	1651782	18.472	ng/ul	99
83) Butylbenzylphthalate	20.651	149	673466	21.864	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.433	252	541538	22.704	ng/ul	99
85) Benzo(a)anthracene	21.504	228	1618484	21.178	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.404	149	961493	22.827	ng/ul	99
87) Chrysene	21.563	228	1541389	21.360	ng/ul	99
89) Di-n-octyl phthalate	22.368	149	1568820	23.841	ng/ul	100
90) Benzo(b)fluoranthene	23.268	252	1529119	22.329	ng/ul	99
91) Benzo(k)fluoranthene	23.321	252	1450366	21.806	ng/ul	99
93) Benzo(a)pyrene	23.933	252	1275510	21.638	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.698	276	1421016	17.522	ng/ul	98
95) Dibenzo(a,h)anthracene	26.709	278	1187130	17.484	ng/ul	99
96) Benzo(g,h,i)perylene	27.509	276	1102826	16.683	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032923\
 Data File : BP014368.D
 Acq On : 29 Mar 2023 18:11
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020795

Quant Time: Mar 30 01:04:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 00:56:20 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.240 ng/uL
RT: 3.428 min Scan# 41
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion: 88 Resp: 38600
Ion Ratio Lower Upper
88 100
43 28.2 24.4 36.6
58 74.4 57.4 86.0



#5
Pyridine
Concen: 21.609 ng/uL
RT: 3.846 min Scan# 112
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion: 79 Resp: 247021
Ion Ratio Lower Upper
79 100
52 74.5 61.4 92.2
51 37.6 28.5 42.7









#6
Benzaldehyde
Concen: 25.905 ng/ul
RT: 7.157 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion: 77 Resp: 196299
Ion Ratio Lower Upper
77 100
105 89.4 70.6 105.8
106 83.9 66.5 99.7



#8
Phenol
Concen: 20.947 ng/ul
RT: 7.205 min Scan# 683
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion: 94 Resp: 331262
Ion Ratio Lower Upper
94 100
65 30.6 25.2 37.8
66 44.5 36.2 54.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.226 ng/ul
RT: 7.440 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion: 93 Resp: 269751
Ion Ratio Lower Upper
93 100
63 75.2 63.2 94.8
95 31.1 26.0 39.0



#12
2-Chlorophenol
Concen: 21.417 ng/ul
RT: 7.581 min Scan# 747
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:128 Resp: 245619
Ion Ratio Lower Upper
128 100
64 52.9 41.4 62.0
130 30.7 24.1 36.1





#13

2-Methylphenol

Concen: 21.274 ng/ul

RT: 8.469 min Scan# 89

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :





SSTD020795

Tgt Ion:108 Resp: 250910

Ion Ratio Lower Upper

108 100

107 90.2 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.603 ng/ul

RT: 8.551 min Scan# 912

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion: 45 Resp: 360477


Ion Ratio Lower Upper

45 100

77 16.5 12.8 19.2

79 11.9 9.6 14.4









#16
Acetophenone
Concen: 21.709 ng/ul
RT: 8.857 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:105 Resp: 440829
Ion Ratio Lower Upper
105 100
77 82.3 67.3 100.9
51 28.1 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 22.050 ng/ul
RT: 8.834 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion: 70 Resp: 241886
Ion Ratio Lower Upper
70 100
42 53.3 41.1 61.7
101 8.7 7.0 10.6
130 17.9 14.6 21.8









#18
4-Methylphenol
Concen: 21.305 ng/ul
RT: 8.799 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:108 Resp: 275949
Ion Ratio Lower Upper
108 100
107 114.2 89.0 133.6



#19
Hexachloroethane
Concen: 21.668 ng/ul
RT: 9.104 min Scan# 1006
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:117 Resp: 108578
Ion Ratio Lower Upper
117 100
201 105.3 90.2 135.4
199 62.8 54.2 81.2









#22
Nitrobenzene
Concen: 21.175 ng/ul
RT: 9.240 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion: 77 Resp: 354656
Ion Ratio Lower Upper
77 100
123 39.0 30.3 45.5
65 14.3 11.0 16.4



#23
Isophorone
Concen: 21.364 ng/ul
RT: 9.769 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion: 82 Resp: 659569
Ion Ratio Lower Upper
82 100
95 7.5 6.3 9.5
138 15.5 12.2 18.2









#25
2-Nitrophenol
Concen: 21.102 ng/ul
RT: 9.957 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:139 Resp: 148688
Ion Ratio Lower Upper
139 100
65 63.1 49.7 74.5
109 40.7 31.1 46.7



#26
2,4-Dimethylphenol
Concen: 21.243 ng/ul
RT: 10.010 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:107 Resp: 335031
Ion Ratio Lower Upper
107 100
121 45.5 37.0 55.6
122 76.3 59.8 89.6





#27

Bis(2-Chloroethoxy)methane

Concen: 20.919 ng/ul

RT: 10.246 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion: 93 Resp: 387200

Ion Ratio Lower Upper

93 100

95 32.1 25.8 38.8

123 9.9 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 21.191 ng/ul

RT: 10.493 min Scan# 1242

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion:162 Resp: 253545


Ion Ratio Lower Upper

162 100

164 64.7 50.1 75.1

98 41.3 30.6 45.8









#30
Naphthalene
Concen: 21.021 ng/ul
RT: 10.893 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:128 Resp: 845573
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.0 11.0 16.6



#32
4-Chloroaniline
Concen: 22.263 ng/ul
RT: 11.010 min Scan# 1330
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:127 Resp: 361103
Ion Ratio Lower Upper
127 100
129 32.3 25.4 38.0









#33
Hexachlorobutadiene
Concen: 20.309 ng/ul
RT: 11.169 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:225 Resp: 192507
Ion Ratio Lower Upper
225 100
223 62.7 49.3 73.9
227 66.2 49.6 74.4



#34
Caprolactam
Concen: 22.113 ng/ul
RT: 11.798 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:113 Resp: 78163
Ion Ratio Lower Upper
113 100
55 179.6 131.2 196.8
56 133.1 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 21.672 ng/ul
RT: 12.134 min Scan# 1521
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:107 Resp: 318932
Ion Ratio Lower Upper
107 100
144 21.7 17.3 25.9
142 69.7 54.6 81.8



#36
2-Methylnaphthalene
Concen: 21.377 ng/ul
RT: 12.498 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:142 Resp: 582768
Ion Ratio Lower Upper
142 100
141 88.5 71.6 107.4





#37

1-Methylnaphthalene

Concen: 21.534 ng/ul

RT: 12.716 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:142 Resp: 582196

Ion Ratio Lower Upper

142 100

141 93.3 76.6 114.8

116 4.5 3.6 5.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.499 ng/ul

RT: 12.863 min Scan# 1645

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Tgt Ion:216 Resp: 358122




Ion Ratio Lower Upper


216 100

214 77.4 61.0 91.6

179 21.6 16.6 24.8

108 17.5 13.8 20.6





#40

Hexachlorocyclopentadiene

Concen: 18.770 ng/ul

RT: 12.834 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:237 Resp: 212350

Ion Ratio Lower Upper

237 100

235 63.4 51.0 76.4

272 11.9 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 20.811 ng/ul

RT: 13.110 min Scan# 1687

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion:196 Resp: 224985


Ion Ratio Lower Upper

196 100

198 96.9 76.6 114.8

200 30.4 25.4 38.2









#42
2,4,5-Trichlorophenol
Concen: 21.212 ng/ul
RT: 13.187 min Scan# 1700
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:196 Resp: 244127
Ion Ratio Lower Upper
196 100
198 95.8 79.8 119.8
200 30.2 25.6 38.4



#43
1,1'-Biphenyl
Concen: 20.653 ng/ul
RT: 13.504 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:154 Resp: 817610
Ion Ratio Lower Upper
154 100
153 40.7 32.4 48.6
76 14.3 10.7 16.1





#44

2-Chloronaphthalene

Concen: 20.642 ng/ul

RT: 13.551 min Scan# 1762

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:162 Resp: 646097

Ion Ratio Lower Upper

162 100

164 32.7 26.1 39.1

127 41.2 33.7 50.5



#45

2-Nitroaniline

Concen: 22.304 ng/ul

RT: 13.763 min Scan# 1798

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion: 65 Resp: 216202


Ion Ratio Lower Upper

65 100

92 65.2 53.7 80.5

138 87.5 71.7 107.5





#47

Dimethylphthalate

Concen: 21.477 ng/ul

RT: 14.128 min Scan# 1860

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:163 Resp: 853157

Ion Ratio Lower Upper

163 100

194 4.3 3.1 4.7

164 10.1 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 22.511 ng/ul

RT: 14.257 min Scan# 1882

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion:165 Resp: 174703


Ion Ratio Lower Upper

165 100

89 60.6 52.7 79.1

121 20.1 17.9 26.9









#50
Acenaphthylene
Concen: 21.285 ng/ul
RT: 14.398 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:152 Resp: 999301
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.4
153 13.1 10.2 15.4



#51
3-Nitroaniline
Concen: 23.547 ng/ul
RT: 14.592 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:138 Resp: 163719
Ion Ratio Lower Upper
138 100
108 12.3 9.9 14.9
92 129.6 107.6 161.4









#52
Acenaphthene
Concen: 21.090 ng/ul
RT: 14.739 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:153 Resp: 699700
Ion Ratio Lower Upper
153 100
152 46.4 38.2 57.4
154 88.2 70.2 105.2



#53
2,4-Dinitrophenol
Concen: 19.729 ng/ul
RT: 14.804 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:184 Resp: 99153
Ion Ratio Lower Upper
184 100
63 69.5 57.1 85.7
154 68.0 55.3 82.9









#55
4-Nitrophenol
Concen: 22.884 ng/ul
RT: 14.898 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:109 Resp: 135557
Ion Ratio Lower Upper
109 100
139 107.0 77.4 116.2
65 125.5 97.3 145.9



#56
Dibenzofuran
Concen: 21.134 ng/ul
RT: 15.075 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:168 Resp: 985638
Ion Ratio Lower Upper
168 100
139 39.4 32.1 48.1





#57

2,4-Dinitrotoluene

Concen: 23.206 ng/ul

RT: 15.045 min Scan# 2019

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:165 Resp: 255468

Ion Ratio Lower Upper

165 100

63 49.1 41.6 62.4

182 3.0 2.4 3.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.831 ng/ul

RT: 15.298 min Scan# 2059

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Tgt Ion:232 Resp: 225815




Ion Ratio Lower Upper


232 100

131 45.8 35.0 52.4

130 2.7 1.9 2.9

166 31.8 24.1 36.1









#59
Diethylphthalate
Concen: 22.003 ng/ul
RT: 15.486 min Scan# 2094
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:149 Resp: 868563
Ion Ratio Lower Upper
149 100
177 21.3 18.2 27.2
150 13.0 9.8 14.8



#61
Fluorene
Concen: 21.523 ng/ul
RT: 15.728 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:166 Resp: 813631
Ion Ratio Lower Upper
166 100
165 99.0 79.2 118.8
167 12.9 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 21.217 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:204 Resp: 430904
Ion Ratio Lower Upper
204 100
206 33.1 26.9 40.3
141 60.6 51.3 76.9



#63
4-Nitroaniline
Concen: 25.167 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:138 Resp: 142172
Ion Ratio Lower Upper
138 100
92 61.3 47.8 71.6
108 103.1 82.3 123.5









#66
4,6-Dinitro-2-methylphenol
Concen: 20.079 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:198 Resp: 160288
Ion Ratio Lower Upper
198 100
182 4.2 3.5 5.3
77 29.1 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 19.769 ng/ul
RT: 15.933 min Scan# 2167
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:169 Resp: 688571
Ion Ratio Lower Upper
169 100
168 68.9 54.2 81.2
167 36.9 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 19.812 ng/ul
RT: 16.616 min Scan# 2283
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:248 Resp: 275980
Ion Ratio Lower Upper
248 100
250 97.3 78.1 117.1
141 73.0 66.0 99.0



#69
Hexachlorobenzene
Concen: 20.138 ng/ul
RT: 16.733 min Scan# 2303
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:284 Resp: 322226
Ion Ratio Lower Upper
284 100
142 32.0 26.6 39.8
249 32.6 25.7 38.5









#70
Atrazine
Concen: 21.509 ng/ul
RT: 16.886 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:200 Resp: 272724
Ion Ratio Lower Upper
200 100
173 25.8 20.6 30.8
215 47.5 37.2 55.8



#71
Pentachlorophenol
Concen: 19.847 ng/ul
RT: 17.086 min Scan# 2363
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:266 Resp: 183982
Ion Ratio Lower Upper
266 100
264 65.2 51.0 76.4
268 65.9 51.4 77.2





#72

Phenanthrene

Concen: 20.940 ng/ul

RT: 17.480 min Scan# 2432

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:178 Resp: 1304180

Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.4

176 19.1 15.3 22.9



#74

Anthracene

Concen: 21.204 ng/ul

RT: 17.569 min Scan# 2445

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion:178 Resp: 1331016


Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.4

176 18.5 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.548 ng/uL

RT: 13.475 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:216 Resp: 365501

Ion Ratio Lower Upper

216 100

214 77.6 61.0 91.6

218 47.4 38.7 58.1



#76

Pentachlorobenzene

Concen: 19.013 ng/uL

RT: 14.992 min Scan# 2007

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion:250 Resp: 373846


Ion Ratio Lower Upper

250 100

248 61.7 49.0 73.6

252 64.6 50.7 76.1









#77
Carbazole
Concen: 22.235 ng/ul
RT: 17.845 min Scan# 2494
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:167 Resp: 1167977
Ion Ratio Lower Upper
167 100
166 21.1 17.4 26.2
139 13.4 11.3 16.9



#78
Di-n-butylphthalate
Concen: 22.405 ng/ul
RT: 18.375 min Scan# 2582
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:149 Resp: 1479893
Ion Ratio Lower Upper
149 100
150 9.0 7.4 11.0
104 5.8 4.8 7.2





#80

Fluoranthene

Concen: 18.625 ng/ul

RT: 19.439 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:202 Resp: 1595280

Ion Ratio Lower Upper

202 100

101 10.4 8.2 12.2

100 7.9 6.4 9.6



#82

Pyrene

Concen: 18.472 ng/ul

RT: 19.792 min Scan# 2823

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion:202 Resp: 1651782


Ion Ratio Lower Upper

202 100

101 12.2 9.7 14.5

100 10.7 8.0 12.0





#83

Butylbenzylphthalate

Concen: 21.864 ng/ul

RT: 20.651 min Scan# 2971

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:149 Resp: 673466

Ion Ratio Lower Upper

149 100

91 75.1 61.8 92.8

206 22.3 17.8 26.6



#84

3,3'-Dichlorobenzidine

Concen: 22.704 ng/ul

RT: 21.433 min Scan# 3102

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion:252 Resp: 541538


Ion Ratio Lower Upper

252 100

254 64.6 51.4 77.0

126 10.4 7.8 11.6





#85

Benzo(a)anthracene

Concen: 21.178 ng/ul

RT: 21.504 min Scan# 3114

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:228 Resp: 1618484

Ion Ratio Lower Upper

228 100

229 19.6 15.6 23.4

226 27.1 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.827 ng/ul

RT: 21.404 min Scan# 3097

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion:149 Resp: 961493


Ion Ratio Lower Upper

149 100

167 28.0 21.9 32.9

279 4.5 3.7 5.5









#87
Chrysene
Concen: 21.360 ng/ul
RT: 21.563 min Scan# 3126
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:228 Resp: 1541389
Ion Ratio Lower Upper
228 100
226 29.5 23.8 35.8
229 19.6 16.1 24.1



#89
Di-n-octyl phthalate
Concen: 23.841 ng/ul
RT: 22.368 min Scan# 3261
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:149 Resp: 1568820









#90
Benzo(b)fluoranthene
Concen: 22.329 ng/ul
RT: 23.268 min Scan# 3417
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11




Instrument :
BNA_P
ClientSampleId :
SSTD020795


Tgt Ion:252 Resp: 1529119
Ion Ratio Lower Upper
252 100
253 21.6 18.0 27.0
125 8.3 6.8 10.2



#91
Benzo(k)fluoranthene
Concen: 21.806 ng/ul
RT: 23.321 min Scan# 3423
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:252 Resp: 1450366
Ion Ratio Lower Upper
252 100
253 21.3 17.3 25.9
125 8.1 6.9 10.3





#93

Benzo(a)pyrene

Concen: 21.638 ng/ul

RT: 23.933 min Scan# 3

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11

Instrument :

BNA_P

ClientSampleId :

SSTD020795





Tgt Ion:252 Resp: 1275510

Ion Ratio Lower Upper

252 100

253 21.9 16.9 25.3

125 9.2 6.8 10.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.522 ng/ul

RT: 26.698 min Scan# 3997

Delta R.T. 0.000 min

Lab File: BP014368.D

Acq: 29 Mar 2023 18:11




Tgt Ion:276 Resp: 1421016


Ion Ratio Lower Upper

276 100

138 19.7 14.9 22.3

277 26.0 20.3 30.5









#95
Dibenzo(a,h)anthracene
Concen: 17.484 ng/ul
RT: 26.709 min Scan# 3999
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Instrument :
BNA_P
ClientSampleId :
SSTD020795

Tgt Ion:278 Resp: 1187130
Ion Ratio Lower Upper
278 100
139 13.0 10.4 15.6
279 24.0 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 16.683 ng/ul
RT: 27.509 min Scan# 4135
Delta R.T. 0.000 min
Lab File: BP014368.D
Acq: 29 Mar 2023 18:11

Tgt Ion:276 Resp: 1102826
Ion Ratio Lower Upper
276 100
138 17.9 13.0 19.4
277 23.8 18.4 27.6

