

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032923\
 Data File : BP014361.D
 Acq On : 29 Mar 2023 13:42
 Operator : CG/JU
 Sample : 01982-17
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EYGC6

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/30/2023
 Supervised By : Jagrut Upadhyay 03/30/2023

Quant Time: Mar 30 01:01:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 00:56:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.022	152	131772	20.000	ng/u1	0.00
20) Naphthalene-d8	10.846	136	562907	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.675	164	392550	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.433	188	899196	20.000	ng/u1	0.00
79) Chrysene-d12	21.522	240	834518	20.000	ng/u1	0.00
88) Perylene-d12	24.051	264	763241	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	16062	4.727	ng/uL	0.00
4) Pyridine-d5	3.822	84	191225	21.960	ng/u1	0.00
7) Phenol-d5	7.181	99	287449	23.653	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.346	67	196951	25.259	ng/u1	0.00
11) 2-Chlorophenol-d4	7.546	132	215699	24.557	ng/u1	0.00
15) 4-Methylphenol-d8	8.734	113	221794	22.626	ng/u1	0.00
21) Nitrobenzene-d5	9.199	128	105216	23.154	ng/u1	0.00
24) 2-Nitrophenol-d4	9.922	143	123067	23.651	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.469	165	212451	22.419	ng/u1	0.00
31) 4-Chloroaniline-d4	10.987	131	275944	22.106	ng/u1	0.00
46) Dimethylphthalate-d6	14.081	166	777016	24.412	ng/u1	0.00
49) Acenaphthylene-d8	14.369	160	847540	23.998	ng/u1	0.00
54) 4-Nitrophenol-d4	14.898	143	84923m	17.661	ng/u1	0.02
60) Fluorene-d10	15.669	176	678601	25.566	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.792	200	102697	15.522	ng/u1	0.00
73) Anthracene-d10	17.534	188	1089406	25.034	ng/u1	0.00
81) Pyrene-d10	19.763	212	1277020	22.782	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.880	264	1027571	25.340	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032923\
Data File : BP014361.D
Acq On : 29 Mar 2023 13:42
Operator : CG/JU
Sample : 01982-17
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EYGC6

Quant Time: Mar 30 01:01:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 30 00:56:20 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/30/2023
Supervised By :Jagrut Upadhyay 03/30/2023

