




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033122\
Data File : BN019245.D
Acq On : 31 Mar 2022 07:29
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 31 12:05:24 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019245.D\data.ms

(70) Pentachlorophenol (C)

16.869min (-0.003) 22843.55 ng

response 597867

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.08
264.00	61.60	66.90
0.00	0.00	0.00