

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP040521\  
 Data File : BP005185.D  
 Acq On : 05 Apr 2021 18:47  
 Operator : CG/JU  
 Sample : M1888-04  
 Misc :  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 TP-3D-(4-10)

Manual Integrations  
 APPROVED

mohammad  
 4/6/2021 3:23:08 PM

Quant Time: Apr 06 00:32:54 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP032321.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Apr 01 12:23:57 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|--------|------|----------|--------|-------|----------|
| -----                       |        |      |          |        |       |          |
| Internal Standards          |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.699  | 152  | 22179    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.487 | 136  | 83199    | 20.00  | ng    | #-0.01   |
| 39) Acenaphthene-d10        | 14.357 | 164  | 50565    | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.122 | 188  | 108543   | 20.00  | ng    | # 0.00   |
| 76) Chrysene-d12            | 21.216 | 240  | 126344   | 20.00  | ng    | -0.01    |
| 86) Perylene-d12            | 23.498 | 264  | 128070   | 20.00  | ng    | -0.01    |
|                             |        |      |          |        |       |          |
| System Monitoring Compounds |        |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.299  | 112  | 211874   | 147.01 | ng    | 0.00     |
| 7) Phenol-d6                | 6.881  | 99   | 276333   | 133.86 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 8.863  | 82   | 178251   | 91.93  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 15.857 | 330  | 89797    | 136.30 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 12.975 | 172  | 328985   | 88.01  | ng    | 0.00     |
| 79) Terphenyl-d14           | 19.692 | 244  | 475782   | 68.55  | ng    | 0.00     |
|                             |        |      |          |        |       |          |
| Target Compounds            |        |      |          |        |       | Qvalue   |
| 10) Phenol                  | 6.905  | 94   | 6768     | 3.20   | ng    | # 68     |
| 50) Dimethylphthalate       | 13.816 | 163  | 16389    | 4.21   | ng    | # 98     |
| 52) Acenaphthene            | 14.422 | 154  | 16434    | 5.47   | ng    | 95       |
| 58) Fluorene                | 15.410 | 166  | 15503    | 3.90   | ng    | # 92     |
| 71) Phenanthrene            | 17.163 | 178  | 312392   | 49.64  | ng    | 98       |
| 72) Anthracene              | 17.251 | 178  | 46983    | 7.64   | ng    | 98       |
| 75) Fluoranthene            | 19.139 | 202  | 256120   | 34.60  | ng    | 97       |
| 78) Pyrene                  | 19.498 | 202  | 452499   | 52.12  | ng    | 98       |
| 81) Benzo(a)anthracene      | 21.204 | 228  | 220617   | 25.70  | ng    | 98       |
| 83) Chrysene                | 21.257 | 228  | 236959m  | 28.13  | ng    |          |
| 87) Indeno(1,2,3-cd)pyrene  | 25.833 | 276  | 59409    | 6.19   | ng    | 96       |
| 88) Benzo(b)fluoranthene    | 22.810 | 252  | 150212m  | 16.52  | ng    |          |
| 89) Benzo(k)fluoranthene    | 22.851 | 252  | 34648m   | 3.98   | ng    |          |
| 90) Benzo(a)pyrene          | 23.398 | 252  | 134466   | 16.49  | ng    | 99       |
| 91) Dibenzo(a,h)anthracene  | 25.833 | 278  | 22068m   | 2.66   | ng    |          |
| 92) Benzo(g,h,i)perylene    | 26.556 | 276  | 70701    | 8.82   | ng    | 98       |
| -----                       |        |      |          |        |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP040521\  
Data File : BP005185.D  
Acq On : 05 Apr 2021 18:47  
Operator : CG/JU  
Sample : M1888-04  
Misc :  
ALS Vial : 17 Sample Multiplier: 1


Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)

Manual Integrations  
APPROVED

mohammad  
4/6/2021 3:23:08 PM

Quant Time: Apr 06 00:32:54 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP032321.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Apr 01 12:23:57 2021  
Response via : Initial Calibration









#1  
1,4-Dichlorobenzene-d4  
Concen: 20.00 ng  
RT: 7.699 min Scan# 784  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

Tgt Ion:152 Resp: 22179  
Ion Ratio Lower Upper  
152 100  
150 159.5 123.8 185.6  
115 65.5 43.8 65.6

Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#5  
2-Fluorophenol  
Concen: 147.01 ng  
RT: 5.299 min Scan# 376  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47


Tgt Ion:112 Resp: 211874  
Ion Ratio Lower Upper  
112 100  
64 58.0 32.8 49.2#  
63 29.8 17.0 25.4#





#7  
Phenol-d6  
Concen: 133.86 ng  
RT: 6.881 min Scan# 645  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47




Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)




Tgt Ion: 99 Resp: 276333  
Ion Ratio Lower Upper  
99 100  
42 13.5 8.6 13.0#  
71 37.9 23.4 35.2#

Manual Integrations  
APPROVED



mohammad  
4/6/2021 3:23:08 PM




#10  
Phenol  
Concen: 3.20 ng  
RT: 6.905 min Scan# 649  
Delta R.T. -0.012 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47



Tgt Ion: 94 Resp: 6768  
Ion Ratio Lower Upper  
94 100  
65 31.4 2.3 42.3  
66 54.0 10.9 50.9#









#21  
Naphthalene-d8  
Concen: 20.00 ng  
RT: 10.487 min Scan# 1258  
Delta R.T. -0.012 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 136   | Resp: | 83199 |
| Ion Ratio | Lower | Upper |       |
| 136       | 100   |       |       |
| 137       | 11.0  | 8.6   | 13.0  |
| 54        | 6.5   | 3.6   | 5.4#  |
| 68        | 7.4   | 3.3   | 4.9#  |

Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#23  
Nitrobenzene-d5  
Concen: 91.93 ng  
RT: 8.863 min Scan# 982  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 82    | Resp: | 178251 |
| Ion Ratio | Lower | Upper |        |
| 82        | 100   |       |        |
| 128       | 39.0  | 45.6  | 68.4#  |
| 54        | 37.7  | 30.1  | 45.1   |









#39  
Acenaphthene-d10  
Concen: 20.00 ng  
RT: 14.357 min Scan# 1916  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 164   | Resp: | 50565 |
| Ion Ratio | Lower | Upper |       |
| 164       | 100   |       |       |
| 162       | 96.0  | 79.4  | 119.2 |
| 160       | 39.4  | 36.1  | 54.1  |

Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#42  
2,4,6-Tribromophenol  
Concen: 136.30 ng  
RT: 15.857 min Scan# 2171  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47


|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 330   | Resp: | 89797 |
| Ion Ratio | Lower | Upper |       |
| 330       | 100   |       |       |
| 332       | 96.2  | 77.1  | 115.7 |
| 141       | 36.1  | 23.4  | 35.2# |





#45  
2-Fluorobiphenyl  
Concen: 88.01 ng  
RT: 12.975 min Scan# 1681  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47




Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)



Tgt Ion:172 Resp: 328985  
Ion Ratio Lower Upper  
172 100  
171 35.7 28.5 42.7  
170 22.8 19.2 28.8




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#50  
Dimethylphthalate  
Concen: 4.21 ng  
RT: 13.816 min Scan# 1824  
Delta R.T. -0.012 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47


Tgt Ion:163 Resp: 16389  
Ion Ratio Lower Upper  
163 100  
194 5.8 3.3 4.9#  
164 9.9 8.2 12.2





#52  
Acenaphthene  
Concen: 5.47 ng  
RT: 14.422 min Scan# 1927  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47




Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)



Tgt Ion:154 Resp: 16434  
Ion Ratio Lower Upper  
154 100  
153 106.9 91.0 136.6  
152 52.5 43.0 64.6




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#58  
Fluorene  
Concen: 3.90 ng  
RT: 15.410 min Scan# 2095  
Delta R.T. -0.012 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

Tgt Ion:166 Resp: 15503  
Ion Ratio Lower Upper  
166 100  
165 91.2 77.8 116.8  
167 21.3 11.0 16.4#









#64  
Phenanthrene-d10  
Concen: 20.00 ng  
RT: 17.122 min Scan# 2386  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

Tgt Ion:188 Resp: 108543  
Ion Ratio Lower Upper  
188 100  
94 8.8 6.2 9.2  
80 11.4 6.2 9.2#

Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#71  
Phenanthrene  
Concen: 49.64 ng  
RT: 17.163 min Scan# 2393  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

Tgt Ion:178 Resp: 312392  
Ion Ratio Lower Upper  
178 100  
176 18.3 15.3 22.9  
179 15.1 12.5 18.7









#72  
 Anthracene  
 Concen: 7.64 ng  
 RT: 17.251 min Scan# 2408  
 Delta R.T. -0.006 min  
 Lab File: BP005185.D  
 Acq: 05 Apr 2021 18:47

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 178   | Resp: | 46983 |
| Ion Ratio | Lower | Upper |       |
| 178       | 100   |       |       |
| 176       | 18.1  | 15.0  | 22.4  |
| 179       | 16.4  | 12.5  | 18.7  |

Instrument :  
 BNA\_P  
 ClientSampleId :  
 TP-3D-(4-10)




Manual Integrations  
 APPROVED


mohammad  
 4/6/2021 3:23:08 PM




#75  
 Fluoranthene  
 Concen: 34.60 ng  
 RT: 19.139 min Scan# 2729  
 Delta R.T. -0.006 min  
 Lab File: BP005185.D  
 Acq: 05 Apr 2021 18:47

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 202   | Resp: | 256120 |
| Ion Ratio | Lower | Upper |        |
| 202       | 100   |       |        |
| 101       | 12.0  | 0.0   | 29.4   |
| 203       | 18.3  | 0.0   | 37.8   |







#76  
Chrysene-d12  
Concen: 20.00 ng  
RT: 21.216 min Scan# 3082  
Delta R.T. -0.012 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47




Tgt Ion:240 Resp: 126344  
Ion Ratio Lower Upper  
240 100  
120 10.3 7.8 11.6  
236 22.9 20.8 31.2




Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)

Manual Integrations  
APPROVED



mohammad  
4/6/2021 3:23:08 PM




#78  
Pyrene  
Concen: 52.12 ng  
RT: 19.498 min Scan# 2790  
Delta R.T. -0.000 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47



Tgt Ion:202 Resp: 452499  
Ion Ratio Lower Upper  
202 100  
200 20.0 17.1 25.7  
203 17.1 14.2 21.2









#79  
Terphenyl-d14  
Concen: 68.55 ng  
RT: 19.692 min Scan# 2823  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 244   | Resp: | 475782 |
| Ion Ratio | Lower | Upper |        |
| 244       | 100   |       |        |
| 212       | 6.8   | 5.8   | 8.6    |
| 122       | 10.3  | 7.8   | 11.8   |

Instrument :  
BNA\_P  
ClientSampled :  
TP-3D-(4-10)




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#81  
Benzo(a)anthracene  
Concen: 25.70 ng  
RT: 21.204 min Scan# 3080  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 228   | Resp: | 220617 |
| Ion Ratio | Lower | Upper |        |
| 228       | 100   |       |        |
| 226       | 26.0  | 22.1  | 33.1   |
| 229       | 19.6  | 15.9  | 23.9   |









#83  
Chrysene  
Concen: 28.13 ng m  
RT: 21.257 min Scan# 3089  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 228   | Resp: | 236959 |
| Ion Ratio | Lower | Upper |        |
| 228       | 100   |       |        |
| 226       | 29.0  | 24.4  | 36.6   |
| 229       | 20.0  | 15.8  | 23.6   |

Instrument :  
BNA\_P  
ClientSampled :  
TP-3D-(4-10)




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#86  
Perylene-d12  
Concen: 20.00 ng  
RT: 23.498 min Scan# 3470  
Delta R.T. -0.012 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 264   | Resp: | 128070 |
| Ion Ratio | Lower | Upper |        |
| 264       | 100   |       |        |
| 260       | 23.4  | 18.8  | 28.2   |
| 265       | 22.2  | 17.4  | 26.2   |









#87  
Indeno(1,2,3-cd)pyrene  
Concen: 6.19 ng  
RT: 25.833 min Scan# 3867  
Delta R.T. -0.030 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 276   | Resp: | 59409 |
| Ion Ratio | Lower | Upper |       |
| 276       | 100   |       |       |
| 138       | 20.0  | 18.7  | 28.1  |
| 277       | 24.9  | 20.4  | 30.6  |

Instrument :  
BNA\_P  
ClientSampled :  
TP-3D-(4-10)




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#88  
Benzo(b)fluoranthene  
Concen: 16.52 ng m  
RT: 22.810 min Scan# 3353  
Delta R.T. -0.006 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 252   | Resp: | 150212 |
| Ion Ratio | Lower | Upper |        |
| 252       | 100   |       |        |
| 253       | 23.9  | 17.6  | 26.4   |
| 125       | 10.1  | 7.3   | 10.9   |









#89  
Benzo(k)fluoranthene  
Concen: 3.98 ng m  
RT: 22.851 min Scan# 3360  
Delta R.T. -0.012 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 252   | Resp: | 34648 |
| Ion Ratio | Lower | Upper |       |
| 252       | 100   |       |       |
| 253       | 23.7  | 17.8  | 26.8  |
| 125       | 9.4   | 7.4   | 11.2  |

Instrument :  
BNA\_P  
ClientSampleId :  
TP-3D-(4-10)




Manual Integrations  
APPROVED


mohammad  
4/6/2021 3:23:08 PM



#90  
Benzo(a)pyrene  
Concen: 16.49 ng  
RT: 23.398 min Scan# 3453  
Delta R.T. -0.012 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 252   | Resp: | 134466 |
| Ion Ratio | Lower | Upper |        |
| 252       | 100   |       |        |
| 253       | 22.6  | 17.4  | 26.0   |
| 125       | 10.3  | 8.2   | 12.4   |





#91  
Dibenzo(a,h)anthracene  
Concen: 2.66 ng m  
RT: 25.833 min Scan# 3867  
Delta R.T. -0.041 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47




Instrument :  
BNA\_P  
ClientSampled :  
TP-3D-(4-10)



Tgt Ion: 278 Resp: 22068  
Ion Ratio Lower Upper  
278 100  
139 0.0 12.4 18.6#  
279 29.0 19.0 28.6#

Manual Integrations  
APPROVED

mohammad  
4/6/2021 3:23:08 PM



#92  
Benzo(g,h,i)perylene  
Concen: 8.82 ng  
RT: 26.556 min Scan# 3990  
Delta R.T. -0.024 min  
Lab File: BP005185.D  
Acq: 05 Apr 2021 18:47

Tgt Ion: 276 Resp: 70701  
Ion Ratio Lower Upper  
276 100  
277 25.6 19.0 28.6  
138 21.1 16.6 24.8

