

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040722\
 Data File : BP009702.D
 Acq On : 07 Apr 2022 16:55
 Operator : CG/JU
 Sample : SSTD04019
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040619

Quant Time: Apr 07 18:17:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 18:14:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	118743	20.000	ng/ul	0.00
20) Naphthalene-d8	10.793	136	529728	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.610	164	375069	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.363	188	835886	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.445	240	879515	20.000	ng/ul	0.00
88) Perylene-d12	23.927	264	879698	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	41802	14.149	ng/uL	0.00
4) Pyridine-d5	3.811	84	292142	39.631	ng/ul	0.00
7) Phenol-d5	7.134	99	416602	43.310	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.305	67	246373	42.678	ng/ul	0.00
11) 2-Chlorophenol-d4	7.510	132	320166	42.573	ng/ul	0.00
15) 4-Methylphenol-d8	8.687	113	350162	47.622	ng/ul	0.00
21) Nitrobenzene-d5	9.140	128	160066	51.813	ng/ul	0.00
24) 2-Nitrophenol-d4	9.863	143	183194	62.565	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.404	165	363557	46.600	ng/ul	0.00
31) 4-Chloroaniline-d4	10.916	131	494885	44.859	ng/ul	0.00
46) Dimethylphthalate-d6	14.022	166	1182497	44.986	ng/ul	0.00
49) Acenaphthylene-d8	14.304	160	1435528	41.333	ng/ul	0.00
54) 4-Nitrophenol-d4	14.792	143	217746	52.308	ng/ul	0.00
60) Fluorene-d10	15.604	176	996296	43.560	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.716	200	203575	60.002	ng/ul	0.00
73) Anthracene-d10	17.463	188	1614865	41.107	ng/ul	0.00
81) Pyrene-d10	19.692	212	1961907	40.581	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.768	264	1891418	42.510	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	41101	13.913	ng/uL#	16
5) Pyridine	3.828	79	306092	39.133	ng/ul#	42
6) Benzaldehyde	7.110	77	267312	47.234	ng/ul	97
8) Phenol	7.157	94	413286	42.681	ng/ul#	90
10) Bis(2-Chloroethyl)ether	7.399	93	322881	41.775	ng/ul#	75
12) 2-Chlorophenol	7.540	128	321559	42.251	ng/ul	94
13) 2-Methylphenol	8.416	108	327731	45.367	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.516	45	479961	51.642	ng/ul#	86
16) Acetophenone	8.804	105	553531	49.275	ng/ul#	80
17) N-Nitroso-di-n-propyla...	8.799	70	290915	51.975	ng/ul#	74
18) 4-Methylphenol	8.752	108	356679	46.969	ng/ul	90
19) Hexachloroethane	9.069	117	134890	43.409	ng/ul	84
22) Nitrobenzene	9.181	77	410298	51.000	ng/ul#	84
23) Isophorone	9.716	82	823718	46.310	ng/ul	97
25) 2-Nitrophenol	9.899	139	193750	56.552	ng/ul#	86
26) 2,4-Dimethylphenol	9.957	107	434247	46.953	ng/ul#	81
27) Bis(2-Chloroethoxy)met...	10.199	93	471698	43.373	ng/ul#	98
29) 2,4-Dichlorophenol	10.434	162	347872	45.286	ng/ul#	84
30) Naphthalene	10.840	128	1085336	40.326	ng/ul	97
32) 4-Chloroaniline	10.946	127	489438	44.831	ng/ul	98
33) Hexachlorobutadiene	11.140	225	245167	44.042	ng/ul	94
34) Caprolactam	11.716	113	57671	26.017	ng/ul#	1
35) 4-Chloro-3-methylphenol	12.063	107	420154	54.026	ng/ul#	70
36) 2-Methylnaphthalene	12.445	142	798359	44.265	ng/ul	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040722\
 Data File : BP009702.D
 Acq On : 07 Apr 2022 16:55
 Operator : CG/JU
 Sample : SST04019
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST040619


Quant Time: Apr 07 18:17:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 18:14:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	798423	44.275	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.810	216	466540	39.113	ng/ul#	93
40) Hexachlorocyclopentadiene	12.798	237	320186	42.739	ng/ul	97
41) 2,4,6-Trichlorophenol	13.045	196	304073	44.022	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.116	196	319394	43.706	ng/ul#	87
43) 1,1'-Biphenyl	13.451	154	1100955	38.458	ng/ul	93
44) 2-Chloronaphthalene	13.492	162	858615	39.437	ng/ul	93
45) 2-Nitroaniline	13.681	65	275107	71.886	ng/ul#	79
47) Dimethylphthalate	14.069	163	1144943	45.005	ng/ul#	92
48) 2,6-Dinitrotoluene	14.181	165	228342	65.684	ng/ul	94
50) Acenaphthylene	14.334	152	1403984	40.773	ng/ul	95
51) 3-Nitroaniline	14.504	138	235469	53.646	ng/ul#	83
52) Acenaphthene	14.675	153	902340	40.322	ng/ul	97
53) 2,4-Dinitrophenol	14.710	184	125403	59.005	ng/ul#	82
55) 4-Nitrophenol	14.810	109	201677	61.308	ng/ul#	58
56) Dibenzofuran	15.010	168	1324009	41.826	ng/ul	96
57) 2,4-Dinitrotoluene	14.963	165	336647	67.394	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.234	232	304483	53.287	ng/ul#	86
59) Diethylphthalate	15.439	149	1175654	47.751	ng/ul	93
61) Fluorene	15.663	166	1099472	44.308	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.657	204	581086	46.149	ng/ul	96
63) 4-Nitroaniline	15.669	138	249159	58.565	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.734	198	204554	57.345	ng/ul#	94
67) N-Nitrosodiphenylamine	15.869	169	963236	39.380	ng/ul	94
68) 4-Bromophenyl-phenylether	16.551	248	361404	41.446	ng/ul	97
69) Hexachlorobenzene	16.669	284	414453	41.397	ng/ul#	91
70) Atrazine	16.828	200	394549	45.243	ng/ul#	91
71) Pentachlorophenol	17.010	266	281983	47.710	ng/ul#	85
72) Phenanthrene	17.410	178	1796686	40.317	ng/ul	99
74) Anthracene	17.498	178	1825582	41.036	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.416	216	494828	35.128	ng/ul#	86
76) Pentachlorobenzene	14.934	250	479452	38.213	ng/ul	89
77) Carbazole	17.763	167	1641138	41.446	ng/ul#	95
78) Di-n-butylphthalate	18.322	149	2063782	47.011	ng/ul#	93
80) Fluoranthene	19.369	202	2249284	40.377	ng/ul#	95
82) Pyrene	19.722	202	2267951	39.822	ng/ul#	93
83) Butylbenzylphthalate	20.592	149	971891	52.857	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.357	252	835733	45.091	ng/ul#	95
85) Benzo(a)anthracene	21.433	228	2250539	40.480	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.363	149	1430233	48.931	ng/ul#	81
87) Chrysene	21.486	228	2180394	40.469	ng/ul	99
89) Di-n-octyl phthalate	22.321	149	2412868	50.668	ng/ul	100
90) Benzo(b)fluoranthene	23.168	252	2312726	41.714	ng/ul	99
91) Benzo(k)fluoranthene	23.221	252	2197034	43.721	ng/ul#	98
93) Benzo(a)pyrene	23.821	252	2215722	41.704	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.533	276	2422162	38.504	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.551	278	2096588	39.479	ng/ul#	95
96) Benzo(g,h,i)perylene	27.321	276	2006203	37.241	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 07 18:17:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 07 18:14:16 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 13.913 ng/uL
 RT: 3.428 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: BP009702.D
 Acq: 07 Apr 2022 16:55




Instrument :
 BNA_P
 ClientSampleId :
 SST040619


Tgt Ion: 88 Resp: 41101
 Ion Ratio Lower Upper
 88 100
 43 34.7 6.8 10.2#
 58 85.8 28.4 42.6#



#5
 Pyridine
 Concen: 39.133 ng/uL
 RT: 3.828 min Scan# 143
 Delta R.T. 0.000 min
 Lab File: BP009702.D
 Acq: 07 Apr 2022 16:55

Tgt Ion: 79 Resp: 306092
 Ion Ratio Lower Upper
 79 100
 52 90.7 35.5 53.3#
 51 42.5 20.2 30.4#









#6
Benzaldehyde
Concen: 47.234 ng/ul
RT: 7.110 min Scan# 701
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion: 77 Resp: 267312
Ion Ratio Lower Upper
77 100
105 95.7 77.0 115.4
106 91.0 77.0 115.4



#8
Phenol
Concen: 42.681 ng/ul
RT: 7.157 min Scan# 709
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion: 94 Resp: 413286
Ion Ratio Lower Upper
94 100
65 32.8 19.0 28.6
66 46.3 34.6 51.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 41.775 ng/ul
 RT: 7.399 min Scan# 750
 Delta R.T. 0.000 min
 Lab File: BP009702.D
 Acq: 07 Apr 2022 16:55




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040619


Tgt Ion: 93 Resp: 322881
 Ion Ratio Lower Upper
 93 100
 63 84.5 45.0 67.4#
 95 33.1 26.3 39.5



#12
 2-Chlorophenol
 Concen: 42.251 ng/ul
 RT: 7.540 min Scan# 774
 Delta R.T. 0.000 min
 Lab File: BP009702.D
 Acq: 07 Apr 2022 16:55

Tgt Ion:128 Resp: 321559
 Ion Ratio Lower Upper
 128 100
 64 53.1 45.3 67.9
 130 31.7 29.0 43.6









#13
2-Methylphenol
Concen: 45.367 ng/ul
RT: 8.416 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:108 Resp: 327731
Ion Ratio Lower Upper
108 100
107 90.6 70.8 106.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 51.642 ng/ul
RT: 8.516 min Scan# 940
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion: 45 Resp: 479961
Ion Ratio Lower Upper
45 100
77 14.4 17.1 25.7#
79 10.4 13.2 19.8#









#16
Acetophenone
Concen: 49.275 ng/ul
RT: 8.804 min Scan# 989
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:105 Resp: 553531
Ion Ratio Lower Upper
105 100
77 80.5 53.4 80.2#
51 31.1 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 51.975 ng/ul
RT: 8.799 min Scan# 988
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion: 70 Resp: 290915
Ion Ratio Lower Upper
70 100
42 61.5 31.8 47.8#
101 9.6 15.6 23.4#
130 21.7 13.7 20.5#









#18
4-Methylphenol
Concen: 46.969 ng/ul
RT: 8.752 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:108 Resp: 356679
Ion Ratio Lower Upper
108 100
107 115.0 83.6 125.4



#19
Hexachloroethane
Concen: 43.409 ng/ul
RT: 9.069 min Scan# 1034
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:117 Resp: 134890
Ion Ratio Lower Upper
117 100
201 107.9 106.5 159.7
199 68.1 59.0 88.6









#22
Nitrobenzene
Concen: 51.000 ng/ul
RT: 9.181 min Scan# 1053
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion: 77 Resp: 410298
Ion Ratio Lower Upper
77 100
123 42.2 42.3 63.5#
65 14.9 5.8 8.8#



#23
Isophorone
Concen: 46.310 ng/ul
RT: 9.716 min Scan# 1144
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion: 82 Resp: 823718
Ion Ratio Lower Upper
82 100
95 8.2 7.8 11.8
138 17.2 12.9 19.3









#25
2-Nitrophenol
Concen: 56.552 ng/ul
RT: 9.899 min Scan# 1175
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:139 Resp: 193750
Ion Ratio Lower Upper
139 100
65 59.6 49.4 74.2
109 36.8 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 46.953 ng/ul
RT: 9.957 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:107 Resp: 434247
Ion Ratio Lower Upper
107 100
121 47.1 34.1 51.1
122 78.1 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 43.373 ng/ul
 RT: 10.199 min Scan# 1225
 Delta R.T. 0.006 min
 Lab File: BP009702.D
 Acq: 07 Apr 2022 16:55




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040619


Tgt Ion: 93 Resp: 471698
 Ion Ratio Lower Upper
 93 100
 95 33.0 26.2 39.4
 123 11.2 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 45.286 ng/ul
 RT: 10.434 min Scan# 1266
 Delta R.T. 0.000 min
 Lab File: BP009702.D
 Acq: 07 Apr 2022 16:55

Tgt Ion: 162 Resp: 347872
 Ion Ratio Lower Upper
 162 100
 164 64.7 37.0 55.6#
 98 37.5 31.0 46.6









#30
Naphthalene
Concen: 40.326 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:128 Resp: 1085336
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.0 12.2 18.4



#32
4-Chloroaniline
Concen: 44.831 ng/ul
RT: 10.946 min Scan# 1353
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:127 Resp: 489438
Ion Ratio Lower Upper
127 100
129 32.2 24.6 37.0









#33
Hexachlorobutadiene
Concen: 44.042 ng/ul
RT: 11.140 min Scan# 1386
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:225 Resp: 245167
Ion Ratio Lower Upper
225 100
223 63.3 53.4 80.0
227 61.0 53.8 80.8



#34
Caprolactam
Concen: 26.017 ng/ul
RT: 11.716 min Scan# 1484
Delta R.T. 0.012 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:113 Resp: 57671
Ion Ratio Lower Upper
113 100
55 187.8 58.2 87.2#
56 149.6 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 54.026 ng/ul
RT: 12.063 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:107 Resp: 420154
Ion Ratio Lower Upper
107 100
144 23.8 15.8 23.8#
142 75.3 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 44.265 ng/ul
RT: 12.445 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:142 Resp: 798359
Ion Ratio Lower Upper
142 100
141 89.3 64.3 96.5









#37
1-Methylnaphthalene
Concen: 44.275 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:142 Resp: 798423
Ion Ratio Lower Upper
142 100
141 92.2 69.8 104.8
116 4.5 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 39.113 ng/ul
RT: 12.810 min Scan# 1670
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:216 Resp: 466540
Ion Ratio Lower Upper
216 100
214 79.4 58.2 87.2
179 21.4 14.2 21.2#
108 18.4 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 42.739 ng/ul
RT: 12.798 min Scan# 1668
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:237 Resp: 320186
Ion Ratio Lower Upper
237 100
235 64.0 53.2 79.8
272 13.0 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 44.022 ng/ul
RT: 13.045 min Scan# 1710
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:196 Resp: 304073
Ion Ratio Lower Upper
196 100
198 96.6 64.1 96.1#
200 31.5 19.9 29.9#





#42

2,4,5-Trichlorophenol

Concen: 43.706 ng/ul

RT: 13.116 min Scan# 1721

Delta R.T. 0.006 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:196 Resp: 319394

Ion Ratio Lower Upper

196 100

198 96.6 68.6 103.0

200 31.2 32.8 49.2#



#43

1,1'-Biphenyl

Concen: 38.458 ng/ul

RT: 13.451 min Scan# 1779

Delta R.T. 0.000 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55




Tgt Ion:154 Resp: 1100955


Ion Ratio Lower Upper

154 100

153 40.7 28.6 43.0

76 11.8 8.5 12.7









#44
2-Chloronaphthalene
Concen: 39.437 ng/ul
RT: 13.492 min Scan# 1786
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:162 Resp: 858615
Ion Ratio Lower Upper
162 100
164 32.5 27.5 41.3
127 40.4 37.1 55.7



#45
2-Nitroaniline
Concen: 71.886 ng/ul
RT: 13.681 min Scan# 1818
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion: 65 Resp: 275107
Ion Ratio Lower Upper
65 100
92 65.6 77.9 116.9#
138 91.2 81.5 122.3









#47
Dimethylphthalate
Concen: 45.005 ng/ul
RT: 14.069 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:163 Resp: 1144943
Ion Ratio Lower Upper
163 100
194 4.3 4.1 6.1
164 10.2 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 65.684 ng/ul
RT: 14.181 min Scan# 1903
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:165 Resp: 228342
Ion Ratio Lower Upper
165 100
89 58.7 51.7 77.5
121 22.2 18.3 27.5









#50
Acenaphthylene
Concen: 40.773 ng/ul
RT: 14.334 min Scan# 1929
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:152 Resp: 1403984
Ion Ratio Lower Upper
152 100
151 19.9 18.3 27.5
153 13.0 11.0 16.6



#51
3-Nitroaniline
Concen: 53.646 ng/ul
RT: 14.504 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:138 Resp: 235469
Ion Ratio Lower Upper
138 100
108 13.4 15.8 23.8#
92 124.0 117.1 175.7









#52
Acenaphthene
Concen: 40.322 ng/ul
RT: 14.675 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:153 Resp: 902340
Ion Ratio Lower Upper
153 100
152 46.6 34.3 51.5
154 88.6 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 59.005 ng/ul
RT: 14.710 min Scan# 1993
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:184 Resp: 125403
Ion Ratio Lower Upper
184 100
63 67.8 34.6 52.0#
154 67.8 57.7 86.5









#55
4-Nitrophenol
Concen: 61.308 ng/ul
RT: 14.810 min Scan# 2010
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:109 Resp: 201677
Ion Ratio Lower Upper
109 100
139 109.5 61.4 92.2#
65 123.2 65.3 97.9#



#56
Dibenzofuran
Concen: 41.826 ng/ul
RT: 15.010 min Scan# 2044
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:168 Resp: 1324009
Ion Ratio Lower Upper
168 100
139 39.7 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 67.394 ng/ul
RT: 14.963 min Scan# 2036
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:165 Resp: 336647
Ion Ratio Lower Upper
165 100
63 50.1 29.5 44.3#
182 3.3 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 53.287 ng/ul
RT: 15.234 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:232 Resp: 304483
Ion Ratio Lower Upper
232 100
131 44.6 25.4 38.2#
130 2.4 4.2 6.2#
166 31.2 28.1 42.1









#59
Diethylphthalate
Concen: 47.751 ng/ul
RT: 15.439 min Scan# 2116
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:149 Resp: 1175654
Ion Ratio Lower Upper
149 100
177 22.3 22.0 33.0
150 12.6 10.2 15.2



#61
Fluorene
Concen: 44.308 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:166 Resp: 1099472
Ion Ratio Lower Upper
166 100
165 96.8 70.3 105.5
167 13.3 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 46.149 ng/ul
RT: 15.657 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:204 Resp: 581086
Ion Ratio Lower Upper
204 100
206 32.0 27.6 41.4
141 60.9 46.6 69.8



#63
4-Nitroaniline
Concen: 58.565 ng/ul
RT: 15.669 min Scan# 2156
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:138 Resp: 249159
Ion Ratio Lower Upper
138 100
92 54.2 41.6 62.4
108 90.7 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 57.345 ng/ul
RT: 15.734 min Scan# 2166
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:198 Resp: 204554
Ion Ratio Lower Upper
198 100
182 3.6 6.0 9.0#
77 23.9 17.7 26.5



#67
N-Nitrosodiphenylamine
Concen: 39.380 ng/ul
RT: 15.869 min Scan# 2190
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:169 Resp: 963236
Ion Ratio Lower Upper
169 100
168 65.6 48.3 72.5
167 34.9 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 41.446 ng/ul
RT: 16.551 min Scan# 2306
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:248 Resp: 361404
Ion Ratio Lower Upper
248 100
250 100.8 79.4 119.0
141 79.5 67.2 100.8



#69
Hexachlorobenzene
Concen: 41.397 ng/ul
RT: 16.669 min Scan# 2326
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:284 Resp: 414453
Ion Ratio Lower Upper
284 100
142 33.0 19.2 28.8#
249 33.1 27.4 41.2





#70

Atrazine

Concen: 45.243 ng/ul

RT: 16.828 min Scan# 2352

Delta R.T. 0.006 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:200 Resp: 394549

Ion Ratio Lower Upper

200 100

173 25.0 19.2 28.8

215 48.5 32.2 48.4#



#71

Pentachlorophenol

Concen: 47.710 ng/ul

RT: 17.010 min Scan# 2384

Delta R.T. 0.000 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55




Tgt Ion:266 Resp: 281983


Ion Ratio Lower Upper

266 100

264 63.0 57.3 85.9

268 64.7 65.7 98.5#









#72
Phenanthrene
Concen: 40.317 ng/u1
RT: 17.410 min Scan# 2451
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:178 Resp: 1796686
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 18.3 14.5 21.7



#74
Anthracene
Concen: 41.036 ng/u1
RT: 17.498 min Scan# 2467
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:178 Resp: 1825582
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 18.0 13.4 20.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 35.128 ng/uL

RT: 13.416 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:216 Resp: 494828

Ion Ratio Lower Upper

216 100

214 77.5 50.4 75.6#

218 49.3 35.2 52.8



#76

Pentachlorobenzene

Concen: 38.213 ng/uL

RT: 14.934 min Scan# 2031

Delta R.T. 0.000 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55




Tgt Ion:250 Resp: 479452


Ion Ratio Lower Upper

250 100

248 63.5 43.5 65.3

252 65.0 46.5 69.7









#77
Carbazole
Concen: 41.446 ng/ul
RT: 17.763 min Scan# 2511
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:167 Resp: 1641138
Ion Ratio Lower Upper
167 100
166 21.4 15.8 23.8
139 13.2 8.3 12.5#



#78
Di-n-butylphthalate
Concen: 47.011 ng/ul
RT: 18.322 min Scan# 2607
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:149 Resp: 2063782
Ion Ratio Lower Upper
149 100
150 9.3 9.6 14.4#
104 5.6 6.2 9.4#









#80
Fluoranthene
Concen: 40.377 ng/ul
RT: 19.369 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:202 Resp: 2249284
Ion Ratio Lower Upper
202 100
101 9.2 9.6 14.4#
100 7.2 6.3 9.5



#82
Pyrene
Concen: 39.822 ng/ul
RT: 19.722 min Scan# 2845
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:202 Resp: 2267951
Ion Ratio Lower Upper
202 100
101 10.9 10.6 15.8
100 9.1 9.8 14.6#





#83

Butylbenzylphthalate

Concen: 52.857 ng/ul

RT: 20.592 min Scan# 2993

Delta R.T. 0.000 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:149 Resp: 971891

Ion Ratio Lower Upper

149 100

91 73.4 71.1 106.7

206 21.7 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 45.091 ng/ul

RT: 21.357 min Scan# 3123

Delta R.T. 0.000 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55




Tgt Ion:252 Resp: 835733


Ion Ratio Lower Upper

252 100

254 65.8 49.8 74.6

126 9.4 4.3 6.5#









#85
Benzo(a)anthracene
Concen: 40.480 ng/ul
RT: 21.433 min Scan# 3135
Delta R.T. 0.006 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:228 Resp: 2250539
Ion Ratio Lower Upper
228 100
229 19.3 18.2 27.2
226 26.6 22.8 34.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 48.931 ng/ul
RT: 21.363 min Scan# 3124
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:149 Resp: 1430233
Ion Ratio Lower Upper
149 100
167 28.0 32.8 49.2#
279 5.0 5.6 8.4#









#87
Chrysene
Concen: 40.469 ng/ul
RT: 21.486 min Scan# 3145
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:228 Resp: 2180394
Ion Ratio Lower Upper
228 100
226 29.1 22.7 34.1
229 19.6 15.2 22.8



#89
Di-n-octyl phthalate
Concen: 50.668 ng/ul
RT: 22.321 min Scan# 3287
Delta R.T. 0.000 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:149 Resp: 2412868





#90

Benzo(b)fluoranthene

Concen: 41.714 ng/ul

RT: 23.168 min Scan# 3431

Delta R.T. 0.000 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:252 Resp: 2312726

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 8.5 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 43.721 ng/ul

RT: 23.221 min Scan# 3440

Delta R.T. 0.006 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55




Tgt Ion:252 Resp: 2197034


Ion Ratio Lower Upper

252 100

253 21.6 17.9 26.9

125 7.9 5.1 7.7#





#93

Benzo(a)pyrene

Concen: 41.704 ng/ul

RT: 23.821 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:252 Resp: 2215722

Ion Ratio Lower Upper

252 100

253 22.0 17.5 26.3

125 8.7 6.4 9.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 38.504 ng/ul

RT: 26.533 min Scan# 4003

Delta R.T. 0.018 min

Lab File: BP009702.D

Acq: 07 Apr 2022 16:55




Tgt Ion:276 Resp: 2422162


Ion Ratio Lower Upper

276 100

138 19.2 12.3 18.5#

277 25.4 19.3 28.9









#95
Dibenzo(a,h)anthracene
Concen: 39.479 ng/ul
RT: 26.551 min Scan# 4006
Delta R.T. 0.018 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Instrument :
BNA_P
ClientSampleId :
SSTD040619

Tgt Ion:278 Resp: 2096588
Ion Ratio Lower Upper
278 100
139 12.1 6.6 10.0#
279 24.3 18.4 27.6



#96
Benzo(g,h,i)perylene
Concen: 37.241 ng/ul
RT: 27.321 min Scan# 4137
Delta R.T. 0.012 min
Lab File: BP009702.D
Acq: 07 Apr 2022 16:55

Tgt Ion:276 Resp: 2006203
Ion Ratio Lower Upper
276 100
138 16.2 11.6 17.4
277 23.4 15.9 23.9

