

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041621\
 Data File : BP005310.D
 Acq On : 16 Apr 2021 10:07
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Apr 16 11:03:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 11:02:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	23447	20.00	ng	0.00
21) Naphthalene-d8	10.451	136	57950	20.00	ng	0.00
39) Acenaphthene-d10	14.328	164	46433	20.00	ng	0.00
64) Phenanthrene-d10	17.092	188	121400	20.00	ng	0.00
76) Chrysene-d12	21.198	240	173679	20.00	ng	# 0.00
86) Perylene-d12	23.463	264	161171	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.263	112	162185	84.34	ng	0.00
7) Phenol-d6	6.852	99	192887	86.73	ng	0.00
23) Nitrobenzene-d5	8.828	82	165566	88.24	ng	0.00
42) 2,4,6-Tribromophenol	15.834	330	64962	105.25	ng	0.00
45) 2-Fluorobiphenyl	12.940	172	283160	81.00	ng	0.00
79) Terphenyl-d14	19.669	244	731452	83.19	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.199	88	48144	49.23	ng	95
3) Pyridine	3.599	79	119119	51.17	ng	97
4) n-Nitrosodimethylamine	3.511	42	35069	52.28	ng	92
6) Aniline	6.999	93	91041	43.14	ng	99
8) 2-Chlorophenol	7.234	128	67556	44.51	ng	95
9) Benzaldehyde	6.816	77	46429	39.61	ng	92
10) Phenol	6.875	94	99406	44.47	ng	97
11) bis(2-Chloroethyl)ether	7.099	93	70599	43.43	ng	99
12) 1,3-Dichlorobenzene	7.552	146	77645	42.56	ng	98
13) 1,4-Dichlorobenzene	7.699	146	75641	42.25	ng	99
14) 1,2-Dichlorobenzene	8.010	146	71868	41.84	ng	98
15) Benzyl Alcohol	7.916	79	43268	38.28	ng	94
16) 2,2'-oxybis(1-Chloropr...	8.193	45	36081	36.48	ng	93
17) 2-Methylphenol	8.116	107	44811	40.49	ng	99
18) Hexachloroethane	8.728	117	29434	39.24	ng	96
19) n-Nitroso-di-n-propyla...	8.475	70	47496	41.03	ng	97
20) 3+4-Methylphenols	8.452	107	55379	39.13	ng	97
22) Acetophenone	8.487	105	102821	45.31	ng	98
24) Nitrobenzene	8.869	77	83951	43.92	ng	100
25) Isophorone	9.393	82	100430	38.84	ng	99
26) 2-Nitrophenol	9.575	139	21698	46.14	ng	96
27) 2,4-Dimethylphenol	9.646	122	33820	42.82	ng	99
28) bis(2-Chloroethoxy)met...	9.875	93	51689	41.98	ng	98
29) 2,4-Dichlorophenol	10.110	162	40395	45.14	ng	97
30) 1,2,4-Trichlorobenzene	10.316	180	53697	45.80	ng	99
31) Naphthalene	10.504	128	142671	42.45	ng	99
32) Benzoic acid	9.804	122	20643	44.13	ng	# 84
33) 4-Chloroaniline	10.628	127	49388	44.63	ng	97
34) Hexachlorobutadiene	10.781	225	44380	47.16	ng	96
35) Caprolactam	11.428	113	10363	40.54	ng	# 75
36) 4-Chloro-3-methylphenol	11.769	107	46534	44.87	ng	97
37) 2-Methylnaphthalene	12.128	142	93030	45.06	ng	92
38) 1-Methylnaphthalene	12.345	142	91539	45.85	ng	96
40) 1,2,4,5-Tetrachloroben...	12.498	216	66622	40.72	ng	99
41) Hexachlorocyclopentadiene	12.469	237	20378	38.00	ng	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041621\
 Data File : BP005310.D
 Acq On : 16 Apr 2021 10:07
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Apr 16 11:03:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 11:02:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	12.751	196	35304	39.70	ng	98
44) 2,4,5-Trichlorophenol	12.834	196	40363	40.38	ng	98
46) 1,1'-Biphenyl	13.151	154	133351	39.49	ng	98
47) 2-Chloronaphthalene	13.192	162	111062	40.08	ng	99
48) 2-Nitroaniline	13.410	65	31430	35.51	ng	92
49) Acenaphthylene	14.045	152	176475	43.03	ng	98
50) Dimethylphthalate	13.792	163	130202	42.74	ng	97
51) 2,6-Dinitrotoluene	13.916	165	32990	43.04	ng	97
52) Acenaphthene	14.392	154	115073	43.28	ng	99
53) 3-Nitroaniline	14.251	138	26972	39.06	ng	94
54) 2,4-Dinitrophenol	14.469	184	16783	45.54	ng	95
55) Dibenzofuran	14.728	168	209909	44.80	ng	99
56) 4-Nitrophenol	14.581	139	21841	39.90	ng	98
57) 2,4-Dinitrotoluene	14.716	165	50335	42.06	ng	96
58) Fluorene	15.386	166	172933	46.42	ng	96
59) 2,3,4,6-Tetrachlorophenol	14.963	232	44472	48.36	ng	# 90
60) Diethylphthalate	15.163	149	143636	41.75	ng	99
61) 4-Chlorophenyl-phenyle...	15.381	204	95860	49.03	ng	98
62) 4-Nitroaniline	15.422	138	30478	42.90	ng	94
63) Azobenzene	15.675	77	222913	44.34	ng	98
65) 4,6-Dinitro-2-methylph...	15.481	198	30116	42.11	ng	94
66) n-Nitrosodiphenylamine	15.598	169	147616	41.19	ng	99
67) 4-Bromophenyl-phenylether	16.281	248	60872	40.63	ng	98
68) Hexachlorobenzene	16.392	284	72508	43.14	ng	97
69) Atrazine	16.563	200	49887	45.27	ng	95
70) Pentachlorophenol	16.745	266	35505	44.62	ng	92
71) Phenanthrene	17.133	178	300614	41.41	ng	99
72) Anthracene	17.228	178	294002	43.71	ng	99
73) Carbazole	17.504	167	275589	44.07	ng	98
74) Di-n-butylphthalate	18.063	149	237770	43.52	ng	99
75) Fluoranthene	19.116	202	372896	43.71	ng	98
77) Benzidine	19.310	184	92575	43.10	ng	98
78) Pyrene	19.474	202	403982	40.76	ng	100
80) Butylbenzylphthalate	20.351	149	88468	37.49	ng	93
81) Benzo(a)anthracene	21.180	228	443698	40.04	ng	99
82) 3,3'-Dichlorobenzidine	21.116	252	111814	40.81	ng	99
83) Chrysene	21.233	228	432422	39.43	ng	99
84) Bis(2-ethylhexyl)phtha...	21.104	149	143523	36.20	ng	98
85) Di-n-octyl phthalate	21.986	149	269983	36.74	ng	98
87) Indeno(1,2,3-cd)pyrene	25.798	276	499246	41.45	ng	99
88) Benzo(b)fluoranthene	22.780	252	447883	41.46	ng	99
89) Benzo(k)fluoranthene	22.821	252	448957	42.38	ng	99
90) Benzo(a)pyrene	23.368	252	407373	41.14	ng	98
91) Dibenzo(a,h)anthracene	25.804	278	435758	41.59	ng	99
92) Benzo(g,h,i)perylene	26.509	276	408166	41.06	ng	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



Scan 779 (7.669 min): BP005303.D\data.ms (-771) (#1)
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.663 min Scan# 778
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	152	Resp:	23447
Ion Ratio	Lower	Upper	
152	100		
150	148.5	122.9	184.3
115	69.4	52.2	78.2



Scan 20 (3.205 min): BP005303.D\data.ms (-14) (#2)
 1,4-Dioxane
 Concen: 49.23 ng
 RT: 3.199 min Scan# 19
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	88	Resp:	48144
Ion Ratio	Lower	Upper	
88	100		
58	54.4	41.7	62.5
43	22.2	21.3	31.9









Pyridine
Concen: 51.17 ng
RT: 3.599 min Scan# 87
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	79	Resp:	119119
Ion	Ratio	Lower	Upper
79	100		
52	48.6	37.7	56.5
51	25.4	18.8	28.2



n-Nitrosodimethylamine
Concen: 52.28 ng
RT: 3.511 min Scan# 72
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	42	Resp:	35069
Ion	Ratio	Lower	Upper
42	100		
74	161.6	139.0	208.6
44	179.8	135.8	203.8









2-Fluorophenol
Concen: 84.34 ng
RT: 5.263 min Scan# 370
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	112	Resp:	162185
Ion	Ratio	Lower	Upper
112	100		
64	46.8	37.6	56.4
63	25.1	20.4	30.6



Aniline
Concen: 43.14 ng
RT: 6.999 min Scan# 665
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	93	Resp:	91041
Ion	Ratio	Lower	Upper
93	100		
66	37.3	30.1	45.1
65	20.0	15.4	23.2









Phenol-d6
 Concen: 86.73 ng
 RT: 6.852 min Scan# 640
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	99	Resp:	192887
Ion	Ratio	Lower	Upper
99	100		
42	16.6	12.0	18.0
71	37.6	31.4	47.2



2-Chlorophenol
 Concen: 44.51 ng
 RT: 7.234 min Scan# 705
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	128	Resp:	67556
Ion	Ratio	Lower	Upper
128	100		
130	31.3	13.8	53.8
64	44.7	28.8	68.8









Scan 635 (6.822 min): BP005303.D\data.ms (-625)
 Benzaldehyde
 Concen: 39.61 ng
 RT: 6.816 min Scan# 634
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	77	Resp:	46429
Ion	Ratio	Lower	Upper
77	100		
105	100.4	73.2	113.2
106	97.5	69.1	109.1



Scan 646 (6.887 min): BP005303.D\data.ms (-637)
 Phenol
 Concen: 44.47 ng
 RT: 6.875 min Scan# 644
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	94	Resp:	99406
Ion	Ratio	Lower	Upper
94	100		
65	29.2	8.3	48.3
66	41.9	19.4	59.4









bis(2-Chloroethyl)ether
 Concen: 43.43 ng
 RT: 7.099 min Scan# 682
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	93	Resp:	70599
Ion	Ratio	Lower	Upper
93	100		
63	70.9	49.4	89.4
95	32.6	12.6	52.6



1,3-Dichlorobenzene
 Concen: 42.56 ng
 RT: 7.552 min Scan# 759
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	146	Resp:	77645
Ion	Ratio	Lower	Upper
146	100		
148	64.5	50.7	76.1
75	41.2	31.3	46.9









1,4-Dichlorobenzene
Concen: 42.25 ng
RT: 7.699 min Scan# 784
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	146	Resp:	75641
Ion	Ratio	Lower	Upper
146	100		
148	66.1	52.5	78.7
111	46.4	38.6	57.8



1,2-Dichlorobenzene
Concen: 41.84 ng
RT: 8.010 min Scan# 837
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	146	Resp:	71868
Ion	Ratio	Lower	Upper
146	100		
148	62.8	51.4	77.2
111	49.3	40.8	61.2









Benzyl Alcohol
 Concen: 38.28 ng
 RT: 7.916 min Scan# 821
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	79	Resp:	43268
Ion	Ratio	Lower	Upper
79	100		
108	67.8	53.1	79.7
77	87.5	63.2	94.8



2,2'-oxybis(1-Chloropropane)
 Concen: 36.48 ng
 RT: 8.193 min Scan# 868
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	45	Resp:	36081
Ion	Ratio	Lower	Upper
45	100		
77	34.5	9.3	49.3
79	22.0	0.7	40.7









2-Methylphenol
 Concen: 40.49 ng
 RT: 8.116 min Scan# 855
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	107	Resp:	44811
Ion Ratio	Lower	Upper	
107	100		
108	109.7	86.7	130.1
77	61.6	49.0	73.4
79	57.4	45.7	68.5




Hexachloroethane
 Concen: 39.24 ng
 RT: 8.728 min Scan# 959
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	117	Resp:	29434
Ion Ratio	Lower	Upper	
117	100		
119	93.0	78.3	117.5
201	98.2	81.8	122.6






Abundance Scan 917 (8.481 min): BP005303.D\data.ms (-909) #19




n-Nitroso-di-n-propylamine
Concen: 41.03 ng
RT: 8.475 min Scan# 916
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
70	100		
42	42.2	31.8	47.8
101	9.6	7.4	11.0
130	17.3	13.4	20.0







Abundance Scan 912 (8.452 min): BP005303.D\data.ms (-904) #20



3+4-Methylphenols
Concen: 39.13 ng
RT: 8.452 min Scan# 912
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion	Ratio	Lower	Upper
107	100		
108	87.2	66.0	106.0
77	47.3	21.5	61.5
79	37.6	17.1	57.1









Naphthalene-d8
 Concen: 20.00 ng
 RT: 10.451 min Scan# 1252
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTCCCC040

Tgt Ion:	136	Resp:	57950
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.4	12.6
54	7.6	5.8	8.8
68	6.2	5.2	7.8



Acetophenone
 Concen: 45.31 ng
 RT: 8.487 min Scan# 918
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07


Tgt Ion:	105	Resp:	102821
Ion	Ratio	Lower	Upper
105	100		
71	3.0	2.4	3.6
51	24.4	18.2	27.4
120	21.6	17.4	26.2






Abundance Scan 977 (8.834 min): BP005303.D\data.ms (-968) #23

Nitrobenzene-d5
Concen: 88.24 ng
RT: 8.828 min Scan# 976
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	82	Resp:	165566
Ion	Ratio	Lower	Upper
82	100		
128	32.1	25.3	37.9
54	41.3	31.7	47.5






Abundance Scan 984 (8.875 min): BP005303.D\data.ms (-970) #24


Nitrobenzene
Concen: 43.92 ng
RT: 8.869 min Scan# 983
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07



Tgt Ion:	77	Resp:	83951
Ion	Ratio	Lower	Upper
77	100		
123	34.8	27.8	41.8
65	13.7	11.0	16.4






Abundance Scan 1073 (9.399 min): BP005303.D\data.ms (-1025-)




Isophorone
Concen: 38.84 ng
RT: 9.393 min Scan# 1072
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	82	Resp:	100430
Ion	Ratio	Lower	Upper
82	100		
95	9.0	6.6	10.0
138	14.0	11.0	16.6






Abundance Scan 1104 (9.581 min): BP005303.D\data.ms (-1026-)




2-Nitrophenol
Concen: 46.14 ng
RT: 9.575 min Scan# 1103
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	139	Resp:	21698
Ion	Ratio	Lower	Upper
139	100		
109	48.6	39.1	58.7
65	67.3	57.9	86.9






Abundance Scan 1116 (9.652 min): BP005303.D\data.ms (-11097-)




2,4-Dimethylphenol
Concen: 42.82 ng
RT: 9.646 min Scan# 1115
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	122	Resp:	33820
Ion	Ratio	Lower	Upper
122	100		
107	140.0	111.9	167.9
121	64.0	49.9	74.9






Abundance Scan 1155 (9.881 min): BP005303.D\data.ms (-11428-)



bis(2-Chloroethoxy)methane
Concen: 41.98 ng
RT: 9.875 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07


Tgt Ion:	93	Resp:	51689
Ion	Ratio	Lower	Upper
93	100		
95	33.0	25.1	37.7
123	10.6	8.7	13.1






Abundance Scan 1195 (10.116 min): BP005303.D\data.ms (-1129) (-)

2,4-Dichlorophenol
Concen: 45.14 ng
RT: 10.110 min Scan# 1194
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	162	Resp:	40395
Ion	Ratio	Lower	Upper
162	100		
164	67.1	46.1	86.1
98	41.2	25.7	65.7







Abundance Scan 1230 (10.322 min): BP005303.D\data.ms (-1230) (-)

1,2,4-Trichlorobenzene
Concen: 45.80 ng
RT: 10.316 min Scan# 1229
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07



Tgt Ion:	180	Resp:	53697
Ion	Ratio	Lower	Upper
180	100		
182	97.3	78.6	118.0
145	29.3	24.6	36.8









Naphthalene
 Concen: 42.45 ng
 RT: 10.504 min Scan# 1261
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	128	Resp:	142671
Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.6	13.0
127	12.6	10.3	15.5



Benzoic acid
 Concen: 44.13 ng
 RT: 9.804 min Scan# 1142
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	122	Resp:	20643
Ion	Ratio	Lower	Upper
122	100		
105	129.2	133.2	173.2#
77	116.6	112.8	152.8









4-Chloroaniline
 Concen: 44.63 ng
 RT: 10.628 min Scan# 1282
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	127	Resp:	49388
Ion	Ratio	Lower	Upper
127	100		
129	32.2	24.6	36.8
65	41.2	35.0	52.6
92	27.3	21.4	32.0



Hexachlorobutadiene
 Concen: 47.16 ng
 RT: 10.781 min Scan# 1308
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	225	Resp:	44380
Ion	Ratio	Lower	Upper
225	100		
223	64.2	47.9	71.9
227	65.0	51.0	76.4









Caprolactam
Concen: 40.54 ng
RT: 11.428 min Scan# 1418
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	113	Resp:	10363
Ion	Ratio	Lower	Upper
113	100		
55	96.5	104.7	144.7#
56	87.4	93.5	133.5#



4-Chloro-3-methylphenol
Concen: 44.87 ng
RT: 11.769 min Scan# 1476
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	107	Resp:	46534
Ion	Ratio	Lower	Upper
107	100		
144	22.0	16.7	25.1
142	63.5	49.2	73.8









2-Methylnaphthalene
Concen: 45.06 ng
RT: 12.128 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	142	Resp:	93030
Ion	Ratio	Lower	Upper
142	100		
141	94.6	69.8	104.8
115	54.3	38.4	57.6



1-Methylnaphthalene
Concen: 45.85 ng
RT: 12.345 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07


Tgt Ion:	142	Resp:	91539
Ion	Ratio	Lower	Upper
142	100		
141	89.0	74.0	111.0
116	5.5	4.2	6.2






Abundance Scan 1912 (14.334 min): BP005303.D\data.ms (-1860) (-)

Acenaphthene-d10
Concen: 20.00 ng
RT: 14.328 min Scan# 1911
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion	Ratio	Lower	Upper
164	100		
162	101.6	84.8	127.2
160	55.6	44.2	66.4







Abundance Scan 1601 (12.504 min): BP005303.D\data.ms (-1549) (-)

1,2,4,5-Tetrachlorobenzene
Concen: 40.72 ng
RT: 12.498 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07



Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.9	62.4	93.6
179	19.2	16.8	25.2
108	20.1	16.2	24.4









Scan 1596 (12.475 min): BP005303.D\data.ms (-1544) (-)
 Hexachlorocyclopentadiene
 Concen: 38.00 ng
 RT: 12.469 min Scan# 1595
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	237	Resp:	20378
Ion	Ratio	Lower	Upper
237	100		
235	62.9	44.0	84.0
272	13.2	0.0	31.7



Scan 2167 (15.834 min): BP005303.D\data.ms (-2116) (-)
 2,4,6-Tribromophenol
 Concen: 105.25 ng
 RT: 15.834 min Scan# 2167
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	330	Resp:	64962
Ion	Ratio	Lower	Upper
330	100		
332	95.4	78.6	118.0
141	38.3	31.6	47.4









Scan 1644 (12.757 min): BP005303.D\data.ms (-1643) (-)
 2,4,6-Trichlorophenol
 Concen: 39.70 ng
 RT: 12.751 min Scan# 1643
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	196	Resp:	35304
Ion	Ratio	Lower	Upper
196	100		
198	97.0	75.8	113.8
200	31.6	24.8	37.2




Scan 1658 (12.840 min): BP005303.D\data.ms (-1652) (-)
 2,4,5-Trichlorophenol
 Concen: 40.38 ng
 RT: 12.834 min Scan# 1657
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	196	Resp:	40363
Ion	Ratio	Lower	Upper
196	100		
198	96.9	75.5	113.3
97	61.4	47.9	71.9
132	22.9	17.8	26.6





Abundance Scan 1676 (12.945 min): BP005303.D\data.ms (-1645) (-)




2-Fluorobiphenyl
Concen: 81.00 ng
RT: 12.940 min Scan# 1675
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	172	Resp:	283160
Ion	Ratio	Lower	Upper
172	100		
171	35.8	27.8	41.8
170	25.0	18.7	28.1



Abundance Scan 1675 (12.940 min): BP005310.D\data.ms (-1624) (-)





Abundance Scan 1712 (13.157 min): BP005303.D\data.ms (-1705) (-)




1,1'-Biphenyl
Concen: 39.49 ng
RT: 13.151 min Scan# 1711
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07


Tgt Ion:	154	Resp:	133351
Ion	Ratio	Lower	Upper
154	100		
153	41.2	19.7	59.7
76	14.1	0.0	35.1



Abundance Scan 1711 (13.151 min): BP005310.D\data.ms (-1660) (-)





Abundance Scan 1719 (13.198 min): BP005303.D\data.ms (-1707) (-)




2-Chloronaphthalene
Concen: 40.08 ng
RT: 13.192 min Scan# 1718
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	162	Resp:	111062
Ion Ratio	Lower	Upper	
162	100		
127	42.3	34.6	51.8
164	32.8	25.8	38.8



Abundance Scan 1718 (13.192 min): BP005310.D\data.ms (-1667) (-)





Abundance Scan 1756 (13.416 min): BP005303.D\data.ms (-1748) (-)





2-Nitroaniline
Concen: 35.51 ng
RT: 13.410 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	65	Resp:	31430
Ion Ratio	Lower	Upper	
65	100		
92	61.2	54.6	82.0
138	75.1	55.4	83.0



Abundance Scan 1755 (13.410 min): BP005310.D\data.ms (-1704) (-)









Acenaphthylene
 Concen: 43.03 ng
 RT: 14.045 min Scan# 1863
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	152	Resp:	176475
Ion	Ratio	Lower	Upper
152	100		
151	18.4	15.8	23.6
153	13.3	10.7	16.1



Dimethylphthalate
 Concen: 42.74 ng
 RT: 13.792 min Scan# 1820
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:	163	Resp:	130202
Ion	Ratio	Lower	Upper
163	100		
194	5.2	3.9	5.9
164	10.3	7.2	10.8









2,6-Dinitrotoluene
 Concen: 43.04 ng
 RT: 13.916 min Scan# 1841
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTCCCC040

Tgt Ion:	165	Resp:	32990
Ion	Ratio	Lower	Upper
165	100		
63	89.7	67.6	101.4
89	88.2	69.8	104.6



Acenaphthene
 Concen: 43.28 ng
 RT: 14.392 min Scan# 1922
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07


Tgt Ion:	154	Resp:	115073
Ion	Ratio	Lower	Upper
154	100		
153	111.7	88.3	132.5
152	58.1	46.2	69.4






Abundance Scan 1898 (14.251 min): BP005303.D\data.ms (-1851) (-)

3-Nitroaniline
Concen: 39.06 ng
RT: 14.251 min Scan# 1898
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	138	Resp:	26972
Ion	Ratio	Lower	Upper
138	100		
108	14.7	11.8	17.8
92	169.3	129.0	193.4







Abundance Scan 1936 (14.475 min): BP005303.D\data.ms (-1951) (-)

2,4-Dinitrophenol
Concen: 45.54 ng
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07



Tgt Ion:	184	Resp:	16783
Ion	Ratio	Lower	Upper
184	100		
63	79.1	68.1	102.1
154	70.6	54.9	82.3









Dibenzofuran
Concen: 44.80 ng
RT: 14.728 min Scan# 1979
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	168	Resp:	209909
Ion	Ratio	Lower	Upper
168	100		
139	41.2	33.1	49.7
169	13.5	10.3	15.5



4-Nitrophenol
Concen: 39.90 ng
RT: 14.581 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	139	Resp:	21841
Ion	Ratio	Lower	Upper
139	100		
109	126.4	101.8	141.8
65	140.2	121.3	161.3









2,4-Dinitrotoluene
 Concen: 42.06 ng
 RT: 14.716 min Scan# 1977
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	165	Resp:	50335
Ion	Ratio	Lower	Upper
165	100		
63	73.8	58.3	87.5
89	103.2	87.0	130.4
182	2.9	2.6	4.0



Fluorene
 Concen: 46.42 ng
 RT: 15.386 min Scan# 2091
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07


Tgt Ion:	166	Resp:	172933
Ion	Ratio	Lower	Upper
166	100		
165	99.7	76.9	115.3
167	14.1	10.6	16.0






Abundance Scan 2020 (14.969 min): BP005303.D\data.ms (-2059) (-)

2,3,4,6-Tetrachlorophenol
Concen: 48.36 ng
RT: 14.963 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	232	Resp:	44472
Ion	Ratio	Lower	Upper
232	100		
131	42.0	29.5	44.3
130	2.7	1.8	2.8
166	29.8	18.6	27.8#







Abundance Scan 2054 (15.169 min): BP005303.D\data.ms (-2060) (-)

Diethylphthalate
Concen: 41.75 ng
RT: 15.163 min Scan# 2053
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07



Tgt Ion:	149	Resp:	143636
Ion	Ratio	Lower	Upper
149	100		
177	22.6	18.6	27.8
150	13.1	9.8	14.8









4-Chlorophenyl-phenylether
Concen: 49.03 ng
RT: 15.381 min Scan# 2090
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	204	Resp:	95860
Ion	Ratio	Lower	Upper
204	100		
206	34.2	26.2	39.4
141	60.0	46.9	70.3



4-Nitroaniline
Concen: 42.90 ng
RT: 15.422 min Scan# 2097
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	138	Resp:	30478
Ion	Ratio	Lower	Upper
138	100		
92	70.3	57.0	97.0
108	141.4	115.8	155.8









Azobenzene
Concen: 44.34 ng
RT: 15.675 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	77	Resp:	222913
Ion	Ratio	Lower	Upper
77	100		
182	20.4	0.4	40.4
105	16.4	0.0	36.9
51	27.1	5.3	45.3




Phenanthrene-d10
Concen: 20.00 ng
RT: 17.092 min Scan# 2381
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	188	Resp:	121400
Ion	Ratio	Lower	Upper
188	100		
94	8.3	7.0	10.6
80	9.9	8.4	12.6





Abundance Scan 2109 (15.492 min): BP005303.D\data.ms (-2056) (-)




4,6-Dinitro-2-methylphenol
Concen: 42.11 ng
RT: 15.481 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	198	Resp:	30116
Ion Ratio	Lower	Upper	
198	100		
51	43.3	16.1	56.1
105	46.1	25.3	65.3



Abundance Scan 2107 (15.481 min): BP005310.D\data.ms (-2056) (-)





Abundance Scan 2128 (15.604 min): BP005303.D\data.ms (-2108) (-)




n-Nitrosodiphenylamine
Concen: 41.19 ng
RT: 15.598 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	169	Resp:	147616
Ion Ratio	Lower	Upper	
169	100		
168	64.9	52.2	78.4
167	36.5	29.9	44.9




Abundance Scan 2127 (15.598 min): BP005310.D\data.ms (-2076) (-)






Abundance Scan 2244 (16.286 min): BP005303.D\data.ms (-2257) (-)

4-Bromophenyl-phenylether
Concen: 40.63 ng
RT: 16.281 min Scan# 2243
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	248	Resp:	60872
Ion	Ratio	Lower	Upper
248	100		
250	96.2	79.0	118.6
141	72.9	59.0	88.4






Abundance Scan 2263 (16.398 min): BP005303.D\data.ms (-2258) (-)


Hexachlorobenzene
Concen: 43.14 ng
RT: 16.392 min Scan# 2262
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07



Tgt Ion:	284	Resp:	72508
Ion	Ratio	Lower	Upper
284	100		
142	32.6	27.8	41.6
249	33.0	25.5	38.3





Abundance Scan 2292 (16.569 min): BP005303.D\data.ms (-2292) (-)




Atrazine
Concen: 45.27 ng
RT: 16.563 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.4	6.2	46.2
215	48.4	34.1	74.1



Abundance Scan 2291 (16.563 min): BP005310.D\data.ms (-2240) (-)





Abundance Scan 2323 (16.751 min): BP005303.D\data.ms (-2319) (-)




Pentachlorophenol
Concen: 44.62 ng
RT: 16.745 min Scan# 2322
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07


Tgt Ion	Ratio	Lower	Upper
266	100		
268	68.1	49.5	74.3
264	60.2	53.3	79.9



Abundance Scan 2322 (16.745 min): BP005310.D\data.ms (-2271) (-)






Abundance Scan 2389 (17.139 min): BP005303.D\data.ms (-2378) (-)




Phenanthrene
Concen: 41.41 ng
RT: 17.133 min Scan# 2388
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	178	Resp:	300614
Ion Ratio	Lower	Upper	
178	100		
176	18.5	14.9	22.3
179	15.6	11.8	17.6







Abundance Scan 2405 (17.233 min): BP005303.D\data.ms (-2378) (-)



Anthracene
Concen: 43.71 ng
RT: 17.228 min Scan# 2404
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	178	Resp:	294002
Ion Ratio	Lower	Upper	
178	100		
176	18.2	14.3	21.5
179	15.1	12.0	18.0









Carbazole
Concen: 44.07 ng
RT: 17.504 min Scan# 2451
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	167	Resp:	275589
Ion	Ratio	Lower	Upper
167	100		
166	21.0	17.8	26.8
139	12.8	10.2	15.2




Di-n-butylphthalate
Concen: 43.52 ng
RT: 18.063 min Scan# 2546
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	149	Resp:	237770
Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.3	10.9
104	10.5	8.1	12.1





Abundance Scan 2726 (19.122 min): BP005303.D\data.ms (-2725) (-)




Fluoranthene
Concen: 43.71 ng
RT: 19.116 min Scan# 2725
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
202	100		
101	8.6	0.0	29.6
203	17.3	0.0	37.6



Abundance Scan 2725 (19.116 min): BP005310.D\data.ms (-2674) (-)





Abundance Scan 3079 (21.198 min): BP005303.D\data.ms (-3078) (-)




Chrysene-d12
Concen: 20.00 ng
RT: 21.198 min Scan# 3079
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion	Ratio	Lower	Upper
240	100		
120	6.4	6.5	9.7#
236	25.6	19.8	29.6




Abundance Scan 3079 (21.198 min): BP005310.D\data.ms (-3028) (-)






Abundance Scan 2758 (19.310 min): BP005303.D\data.ms (-2757) (-)

Benzidine
Concen: 43.10 ng
RT: 19.310 min Scan# 2758
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	184	Resp:	92575
Ion Ratio	Lower	Upper	
184	100		
185	14.2	11.7	17.5
183	13.8	9.7	14.5






Abundance Scan 2786 (19.474 min): BP005303.D\data.ms (-2778) (-)


Pyrene
Concen: 40.76 ng
RT: 19.474 min Scan# 2786
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07



Tgt Ion:	202	Resp:	403982
Ion Ratio	Lower	Upper	
202	100		
200	20.7	16.5	24.7
203	17.3	13.9	20.9






Abundance Scan 2820 (19.674 min): BP005303.D\data.ms (-2819) (-)




Terphenyl-d14
Concen: 83.19 ng
RT: 19.669 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	244	Resp:	731452
Ion Ratio	Lower	Upper	
244	100		
212	7.2	5.8	8.6
122	8.3	6.7	10.1







Abundance Scan 2935 (20.351 min): BP005303.D\data.ms (-2884) (-)



Butylbenzylphthalate
Concen: 37.49 ng
RT: 20.351 min Scan# 2935
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	149	Resp:	88468
Ion Ratio	Lower	Upper	
149	100		
91	81.0	71.2	106.8
206	25.0	18.6	28.0









Benzo(a)anthracene
Concen: 40.04 ng
RT: 21.180 min Scan# 3076
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	228	Resp:	443698
Ion	Ratio	Lower	Upper
228	100		
226	26.1	21.5	32.3
229	20.0	15.6	23.4



3,3'-Dichlorobenzidine
Concen: 40.81 ng
RT: 21.116 min Scan# 3065
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	252	Resp:	111814
Ion	Ratio	Lower	Upper
252	100		
254	64.3	50.5	75.7
126	9.2	7.2	10.8









Scan 3085 (21.233 min): BP005303.D\data.ms (-3085) (-)
 Chrysene
 Concen: 39.43 ng
 RT: 21.233 min Scan# 3085
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	228	Resp:	432422
Ion	Ratio	Lower	Upper
228	100		
226	29.5	24.0	36.0
229	19.6	15.6	23.4



Scan 3064 (21.110 min): BP005303.D\data.ms (-3064) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 36.20 ng
 RT: 21.104 min Scan# 3063
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07


Tgt Ion:	149	Resp:	143523
Ion	Ratio	Lower	Upper
149	100		
167	28.6	23.8	35.8
279	4.6	4.6	6.8






Abundance Scan 3214 (21.992 min): BP005303.D\data.ms (-3205) (-)

Di-n-octyl phthalate
Concen: 36.74 ng
RT: 21.986 min Scan# 3213
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	149	Resp:	269983
Ion	Ratio	Lower	Upper
149	100		
167	1.7	1.5	2.3
43	7.7	6.9	10.3






Abundance Scan 3465 (23.468 min): BP005303.D\data.ms (-3456) (-)


Perylene-d12
Concen: 20.00 ng
RT: 23.463 min Scan# 3464
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07



Tgt Ion:	264	Resp:	161171
Ion	Ratio	Lower	Upper
264	100		
260	24.9	19.0	28.6
265	22.1	17.0	25.6






Abundance Scan 3862 (25.804 min): BP005303.D\data.ms (-3862) (-)




Indeno(1,2,3-cd)pyrene
Concen: 41.45 ng
RT: 25.798 min Scan# 3861
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	276	Resp:	499246
Ion	Ratio	Lower	Upper
276	100		
138	15.1	12.8	19.2
277	25.7	20.8	31.2







Abundance Scan 3348 (22.780 min): BP005303.D\data.ms (-3348) (-)



Benzo(b)fluoranthene
Concen: 41.46 ng
RT: 22.780 min Scan# 3348
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	252	Resp:	447883
Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.1	25.7
125	5.7	5.3	7.9









Scan 3356 (22.827 min): BP005303.D\data.ms (-3356) (-)
 Benzo(k)fluoranthene
 Concen: 42.38 ng
 RT: 22.821 min Scan# 3355
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTCCCC040

Tgt Ion:	252	Resp:	448957
Ion Ratio	Lower	Upper	
252	100		
253	22.3	17.4	26.0
125	6.0	5.4	8.2



Scan 3448 (23.368 min): BP005303.D\data.ms (-3448) (-)
 Benzo(a)pyrene
 Concen: 41.14 ng
 RT: 23.368 min Scan# 3448
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07


Tgt Ion:	252	Resp:	407373
Ion Ratio	Lower	Upper	
252	100		
253	21.2	17.8	26.6
125	6.3	6.0	9.0






Abundance Scan 3863 (25.809 min): BP005303.D\data.ms (-3863) (-)

Dibenzo(a,h)anthracene
Concen: 41.59 ng
RT: 25.804 min Scan# 3862
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	278	Resp:	435758
Ion	Ratio	Lower	Upper
278	100		
139	10.7	8.1	12.1
279	23.7	18.9	28.3



Abundance Scan 3983 (26.515 min): BP005303.D\data.ms (-3983) (-)

Benzo(g,h,i)perylene
Concen: 41.06 ng
RT: 26.509 min Scan# 3982
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07



Tgt Ion:	276	Resp:	408166
Ion	Ratio	Lower	Upper
276	100		
277	22.7	19.7	29.5
138	12.1	9.8	14.6

