

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041621\
 Data File : BP005310.D
 Acq On : 16 Apr 2021 10:07
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Apr 16 11:03:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 11:02:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	23447	20.00	ng	0.00
21) Naphthalene-d8	10.451	136	57950	20.00	ng	0.00
39) Acenaphthene-d10	14.328	164	46433	20.00	ng	0.00
64) Phenanthrene-d10	17.092	188	121400	20.00	ng	0.00
76) Chrysene-d12	21.198	240	173679	20.00	ng	# 0.00
86) Perylene-d12	23.463	264	161171	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.263	112	162185	84.34	ng	0.00
7) Phenol-d6	6.852	99	192887	86.73	ng	0.00
23) Nitrobenzene-d5	8.828	82	165566	88.24	ng	0.00
42) 2,4,6-Tribromophenol	15.834	330	64962	105.25	ng	0.00
45) 2-Fluorobiphenyl	12.940	172	283160	81.00	ng	0.00
79) Terphenyl-d14	19.669	244	731452	83.19	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.199	88	48144	49.23	ng	95
3) Pyridine	3.599	79	119119	51.17	ng	97
4) n-Nitrosodimethylamine	3.511	42	35069	52.28	ng	92
6) Aniline	6.999	93	91041	43.14	ng	99
8) 2-Chlorophenol	7.234	128	67556	44.51	ng	95
9) Benzaldehyde	6.816	77	46429	39.61	ng	92
10) Phenol	6.875	94	99406	44.47	ng	97
11) bis(2-Chloroethyl)ether	7.099	93	70599	43.43	ng	99
12) 1,3-Dichlorobenzene	7.552	146	77645	42.56	ng	98
13) 1,4-Dichlorobenzene	7.699	146	75641	42.25	ng	99
14) 1,2-Dichlorobenzene	8.010	146	71868	41.84	ng	98
15) Benzyl Alcohol	7.916	79	43268	38.28	ng	94
16) 2,2'-oxybis(1-Chloropr...	8.193	45	36081	36.48	ng	93
17) 2-Methylphenol	8.116	107	44811	40.49	ng	99
18) Hexachloroethane	8.728	117	29434	39.24	ng	96
19) n-Nitroso-di-n-propyla...	8.475	70	47496	41.03	ng	97
20) 3+4-Methylphenols	8.452	107	55379	39.13	ng	97
22) Acetophenone	8.487	105	102821	45.31	ng	98
24) Nitrobenzene	8.869	77	83951	43.92	ng	100
25) Isophorone	9.393	82	100430	38.84	ng	99
26) 2-Nitrophenol	9.575	139	21698	46.14	ng	96
27) 2,4-Dimethylphenol	9.646	122	33820	42.82	ng	99
28) bis(2-Chloroethoxy)met...	9.875	93	51689	41.98	ng	98
29) 2,4-Dichlorophenol	10.110	162	40395	45.14	ng	97
30) 1,2,4-Trichlorobenzene	10.316	180	53697	45.80	ng	99
31) Naphthalene	10.504	128	142671	42.45	ng	99
32) Benzoic acid	9.804	122	20643	44.13	ng	# 84
33) 4-Chloroaniline	10.628	127	49388	44.63	ng	97
34) Hexachlorobutadiene	10.781	225	44380	47.16	ng	96
35) Caprolactam	11.428	113	10363	40.54	ng	# 75
36) 4-Chloro-3-methylphenol	11.769	107	46534	44.87	ng	97
37) 2-Methylnaphthalene	12.128	142	93030	45.06	ng	92
38) 1-Methylnaphthalene	12.345	142	91539	45.85	ng	96
40) 1,2,4,5-Tetrachloroben...	12.498	216	66622	40.72	ng	99
41) Hexachlorocyclopentadiene	12.469	237	20378	38.00	ng	98
43) 2,4,6-Trichlorophenol	12.751	196	35304	39.70	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041621\
 Data File : BP005310.D
 Acq On : 16 Apr 2021 10:07
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Quant Time: Apr 16 11:03:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 11:02:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.834	196	40363	40.38	ng	98
46) 1,1'-Biphenyl	13.151	154	133351	39.49	ng	98
47) 2-Chloronaphthalene	13.192	162	111062	40.08	ng	99
48) 2-Nitroaniline	13.410	65	31430	35.51	ng	92
49) Acenaphthylene	14.045	152	176475	43.03	ng	98
50) Dimethylphthalate	13.792	163	130202	42.74	ng	97
51) 2,6-Dinitrotoluene	13.916	165	32990	43.04	ng	97
52) Acenaphthene	14.392	154	115073	43.28	ng	99
53) 3-Nitroaniline	14.251	138	26972	39.06	ng	94
54) 2,4-Dinitrophenol	14.469	184	16783	45.54	ng	95
55) Dibenzofuran	14.728	168	209909	44.80	ng	99
56) 4-Nitrophenol	14.581	139	21841	39.90	ng	98
57) 2,4-Dinitrotoluene	14.716	165	50335	42.06	ng	96
58) Fluorene	15.386	166	172933	46.42	ng	96
59) 2,3,4,6-Tetrachlorophenol	14.963	232	44472	48.36	ng	# 90
60) Diethylphthalate	15.163	149	143636	41.75	ng	99
61) 4-Chlorophenyl-phenyle...	15.381	204	95860	49.03	ng	98
62) 4-Nitroaniline	15.422	138	30478	42.90	ng	94
63) Azobenzene	15.675	77	222913	44.34	ng	98
65) 4,6-Dinitro-2-methylph...	15.481	198	30116	42.11	ng	94
66) n-Nitrosodiphenylamine	15.598	169	147616	41.19	ng	99
67) 4-Bromophenyl-phenylether	16.281	248	60872	40.63	ng	98
68) Hexachlorobenzene	16.392	284	72508	43.14	ng	97
69) Atrazine	16.563	200	49887	45.27	ng	95
70) Pentachlorophenol	16.745	266	35505	44.62	ng	92
71) Phenanthrene	17.133	178	300614	41.41	ng	99
72) Anthracene	17.228	178	294002	43.71	ng	99
73) Carbazole	17.504	167	275589	44.07	ng	98
74) Di-n-butylphthalate	18.063	149	237770	43.52	ng	99
75) Fluoranthene	19.116	202	372896	43.71	ng	98
77) Benzidine	19.310	184	92575	43.10	ng	98
78) Pyrene	19.474	202	403982	40.76	ng	100
80) Butylbenzylphthalate	20.351	149	88468	37.49	ng	93
81) Benzo(a)anthracene	21.180	228	443698	40.04	ng	99
82) 3,3'-Dichlorobenzidine	21.116	252	111814	40.81	ng	99
83) Chrysene	21.233	228	432422	39.43	ng	99
84) Bis(2-ethylhexyl)phtha...	21.104	149	143523	36.20	ng	98
85) Di-n-octyl phthalate	21.986	149	269983	36.74	ng	98
87) Indeno(1,2,3-cd)pyrene	25.798	276	499246	41.45	ng	99
88) Benzo(b)fluoranthene	22.780	252	447883	41.46	ng	99
89) Benzo(k)fluoranthene	22.821	252	448957	42.38	ng	99
90) Benzo(a)pyrene	23.368	252	407373	41.14	ng	98
91) Dibenzo(a,h)anthracene	25.804	278	435758	41.59	ng	99
92) Benzo(g,h,i)perylene	26.509	276	408166	41.06	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 16 11:03:19 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 16 11:02:58 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.663 min Scan# 778
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:152 Resp: 23447
 Ion Ratio Lower Upper
 152 100
 150 148.5 122.9 184.3
 115 69.4 52.2 78.2



#2
 1,4-Dioxane
 Concen: 49.23 ng
 RT: 3.199 min Scan# 19
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion: 88 Resp: 48144
 Ion Ratio Lower Upper
 88 100
 58 54.4 41.7 62.5
 43 22.2 21.3 31.9









#3
Pyridine
Concen: 51.17 ng
RT: 3.599 min Scan# 87
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 119119
Ion Ratio Lower Upper
79 100
52 48.6 37.7 56.5
51 25.4 18.8 28.2



#4
n-Nitrosodimethylamine
Concen: 52.28 ng
RT: 3.511 min Scan# 72
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion: 42 Resp: 35069
Ion Ratio Lower Upper
42 100
74 161.6 139.0 208.6
44 179.8 135.8 203.8









#5
2-Fluorophenol
Concen: 84.34 ng
RT: 5.263 min Scan# 370
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 162185
Ion Ratio Lower Upper
112 100
64 46.8 37.6 56.4
63 25.1 20.4 30.6



#6
Aniline
Concen: 43.14 ng
RT: 6.999 min Scan# 665
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion: 93 Resp: 91041
Ion Ratio Lower Upper
93 100
66 37.3 30.1 45.1
65 20.0 15.4 23.2









#7
Phenol-d6
Concen: 86.73 ng
RT: 6.852 min Scan# 640
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 99 Resp: 192887
Ion Ratio Lower Upper
99 100
42 16.6 12.0 18.0
71 37.6 31.4 47.2



#8
2-Chlorophenol
Concen: 44.51 ng
RT: 7.234 min Scan# 705
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion: 128 Resp: 67556
Ion Ratio Lower Upper
128 100
130 31.3 13.8 53.8
64 44.7 28.8 68.8









#9
Benzaldehyde
Concen: 39.61 ng
RT: 6.816 min Scan# 634
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 46429
Ion Ratio Lower Upper
77 100
105 100.4 73.2 113.2
106 97.5 69.1 109.1



#10
Phenol
Concen: 44.47 ng
RT: 6.875 min Scan# 644
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion: 94 Resp: 99406
Ion Ratio Lower Upper
94 100
65 29.2 8.3 48.3
66 41.9 19.4 59.4









#11
bis(2-Chloroethyl)ether
Concen: 43.43 ng
RT: 7.099 min Scan# 682
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 70599
Ion Ratio Lower Upper
93 100
63 70.9 49.4 89.4
95 32.6 12.6 52.6



#12
1,3-Dichlorobenzene
Concen: 42.56 ng
RT: 7.552 min Scan# 759
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:146 Resp: 77645
Ion Ratio Lower Upper
146 100
148 64.5 50.7 76.1
75 41.2 31.3 46.9









#13
1,4-Dichlorobenzene
Concen: 42.25 ng
RT: 7.699 min Scan# 784
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 75641
Ion Ratio Lower Upper
146 100
148 66.1 52.5 78.7
111 46.4 38.6 57.8



#14
1,2-Dichlorobenzene
Concen: 41.84 ng
RT: 8.010 min Scan# 837
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:146 Resp: 71868
Ion Ratio Lower Upper
146 100
148 62.8 51.4 77.2
111 49.3 40.8 61.2









#15
Benzyl Alcohol
Concen: 38.28 ng
RT: 7.916 min Scan# 821
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 43268
Ion Ratio Lower Upper
79 100
108 67.8 53.1 79.7
77 87.5 63.2 94.8



#16
2,2'-oxybis(1-Chloropropane)
Concen: 36.48 ng
RT: 8.193 min Scan# 868
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion: 45 Resp: 36081
Ion Ratio Lower Upper
45 100
77 34.5 9.3 49.3
79 22.0 0.7 40.7









#17
2-Methylphenol
Concen: 40.49 ng
RT: 8.116 min Scan# 855
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	107	Resp:	44811
Ion	Ratio	Lower	Upper
107	100		
108	109.7	86.7	130.1
77	61.6	49.0	73.4
79	57.4	45.7	68.5



#18
Hexachloroethane
Concen: 39.24 ng
RT: 8.728 min Scan# 959
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	117	Resp:	29434
Ion	Ratio	Lower	Upper
117	100		
119	93.0	78.3	117.5
201	98.2	81.8	122.6









#19
n-Nitroso-di-n-propylamine
Concen: 41.03 ng
RT: 8.475 min Scan# 916
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	70	Resp:	47496
Ion	Ratio	Lower	Upper
70	100		
42	42.2	31.8	47.8
101	9.6	7.4	11.0
130	17.3	13.4	20.0



#20
3+4-Methylphenols
Concen: 39.13 ng
RT: 8.452 min Scan# 912
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	107	Resp:	55379
Ion	Ratio	Lower	Upper
107	100		
108	87.2	66.0	106.0
77	47.3	21.5	61.5
79	37.6	17.1	57.1









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.451 min Scan# 1252
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	57950
Ion Ratio	Lower	Upper	
136	100		
137	11.1	8.4	12.6
54	7.6	5.8	8.8
68	6.2	5.2	7.8



#22
Acetophenone
Concen: 45.31 ng
RT: 8.487 min Scan# 918
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	105	Resp:	102821
Ion Ratio	Lower	Upper	
105	100		
71	3.0	2.4	3.6
51	24.4	18.2	27.4
120	21.6	17.4	26.2





#23

Nitrobenzene-d5

Concen: 88.24 ng

RT: 8.828 min Scan# 976

Delta R.T. 0.000 min

Lab File: BP005310.D

Acq: 16 Apr 2021 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion: 82 Resp: 165566

Ion Ratio Lower Upper

82 100

128 32.1 25.3 37.9

54 41.3 31.7 47.5



#24

Nitrobenzene

Concen: 43.92 ng

RT: 8.869 min Scan# 983

Delta R.T. 0.000 min

Lab File: BP005310.D

Acq: 16 Apr 2021 10:07




Tgt Ion: 77 Resp: 83951


Ion Ratio Lower Upper

77 100

123 34.8 27.8 41.8

65 13.7 11.0 16.4









#25
Isophorone
Concen: 38.84 ng
RT: 9.393 min Scan# 1072
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 100430
Ion Ratio Lower Upper
82 100
95 9.0 6.6 10.0
138 14.0 11.0 16.6



#26
2-Nitrophenol
Concen: 46.14 ng
RT: 9.575 min Scan# 1103
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion: 139 Resp: 21698
Ion Ratio Lower Upper
139 100
109 48.6 39.1 58.7
65 67.3 57.9 86.9









#27
2,4-Dimethylphenol
Concen: 42.82 ng
RT: 9.646 min Scan# 1115
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 33820
Ion Ratio Lower Upper
122 100
107 140.0 111.9 167.9
121 64.0 49.9 74.9



#28
bis(2-Chloroethoxy)methane
Concen: 41.98 ng
RT: 9.875 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion: 93 Resp: 51689
Ion Ratio Lower Upper
93 100
95 33.0 25.1 37.7
123 10.6 8.7 13.1









#29
2,4-Dichlorophenol
Concen: 45.14 ng
RT: 10.110 min Scan# 1194
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 40395
Ion Ratio Lower Upper
162 100
164 67.1 46.1 86.1
98 41.2 25.7 65.7



#30
1,2,4-Trichlorobenzene
Concen: 45.80 ng
RT: 10.316 min Scan# 1229
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:180 Resp: 53697
Ion Ratio Lower Upper
180 100
182 97.3 78.6 118.0
145 29.3 24.6 36.8









#31
Naphthalene
Concen: 42.45 ng
RT: 10.504 min Scan# 1261
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	128	Resp:	142671
Ion Ratio	Lower	Upper	
128	100		
129	11.2	8.6	13.0
127	12.6	10.3	15.5



#32
Benzoic acid
Concen: 44.13 ng
RT: 9.804 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	122	Resp:	20643
Ion Ratio	Lower	Upper	
122	100		
105	129.2	133.2	173.2#
77	116.6	112.8	152.8









#33
4-Chloroaniline
Concen: 44.63 ng
RT: 10.628 min Scan# 1282
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	49388
Ion	Ratio	Lower	Upper
127	100		
129	32.2	24.6	36.8
65	41.2	35.0	52.6
92	27.3	21.4	32.0



#34
Hexachlorobutadiene
Concen: 47.16 ng
RT: 10.781 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	225	Resp:	44380
Ion	Ratio	Lower	Upper
225	100		
223	64.2	47.9	71.9
227	65.0	51.0	76.4









#35
Caprolactam
Concen: 40.54 ng
RT: 11.428 min Scan# 1418
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 10363
Ion Ratio Lower Upper
113 100
55 96.5 104.7 144.7#
56 87.4 93.5 133.5#



#36
4-Chloro-3-methylphenol
Concen: 44.87 ng
RT: 11.769 min Scan# 1476
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:107 Resp: 46534
Ion Ratio Lower Upper
107 100
144 22.0 16.7 25.1
142 63.5 49.2 73.8









#37
2-Methylnaphthalene
Concen: 45.06 ng
RT: 12.128 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:142	Resp:	93030
Ion Ratio	Lower	Upper
142	100	
141	94.6	69.8 104.8
115	54.3	38.4 57.6



#38
1-Methylnaphthalene
Concen: 45.85 ng
RT: 12.345 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:142	Resp:	91539
Ion Ratio	Lower	Upper
142	100	
141	89.0	74.0 111.0
116	5.5	4.2 6.2









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.328 min Scan# 1911
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 46433
Ion Ratio Lower Upper
164 100
162 101.6 84.8 127.2
160 55.6 44.2 66.4



#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.72 ng
RT: 12.498 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:216 Resp: 66622
Ion Ratio Lower Upper
216 100
214 78.9 62.4 93.6
179 19.2 16.8 25.2
108 20.1 16.2 24.4









#41
Hexachlorocyclopentadiene
Concen: 38.00 ng
RT: 12.469 min Scan# 1595
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	237	Resp:	20378
Ion Ratio	Lower	Upper	
237	100		
235	62.9	44.0	84.0
272	13.2	0.0	31.7



#42
2,4,6-Tribromophenol
Concen: 105.25 ng
RT: 15.834 min Scan# 2167
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	330	Resp:	64962
Ion Ratio	Lower	Upper	
330	100		
332	95.4	78.6	118.0
141	38.3	31.6	47.4









#43
2,4,6-Trichlorophenol
Concen: 39.70 ng
RT: 12.751 min Scan# 1643
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 35304
Ion Ratio Lower Upper
196 100
198 97.0 75.8 113.8
200 31.6 24.8 37.2



#44
2,4,5-Trichlorophenol
Concen: 40.38 ng
RT: 12.834 min Scan# 1657
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:196 Resp: 40363
Ion Ratio Lower Upper
196 100
198 96.9 75.5 113.3
97 61.4 47.9 71.9
132 22.9 17.8 26.6









#45
2-Fluorobiphenyl
Concen: 81.00 ng
RT: 12.940 min Scan# 1675
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:172	Resp: 283160		
Ion Ratio	Lower	Upper	
172	100		
171	35.8	27.8	41.8
170	25.0	18.7	28.1



#46
1,1'-Biphenyl
Concen: 39.49 ng
RT: 13.151 min Scan# 1711
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:154	Resp: 133351
Ion Ratio	Lower Upper
154 100	
153 41.2	19.7 59.7
76 14.1	0.0 35.1









#47
2-Chloronaphthalene
Concen: 40.08 ng
RT: 13.192 min Scan# 1718
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 111062
Ion Ratio Lower Upper
162 100
127 42.3 34.6 51.8
164 32.8 25.8 38.8



#48
2-Nitroaniline
Concen: 35.51 ng
RT: 13.410 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion: 65 Resp: 31430
Ion Ratio Lower Upper
65 100
92 61.2 54.6 82.0
138 75.1 55.4 83.0









#49
Acenaphthylene
Concen: 43.03 ng
RT: 14.045 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 176475
Ion Ratio Lower Upper
152 100
151 18.4 15.8 23.6
153 13.3 10.7 16.1



#50
Dimethylphthalate
Concen: 42.74 ng
RT: 13.792 min Scan# 1820
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:163 Resp: 130202
Ion Ratio Lower Upper
163 100
194 5.2 3.9 5.9
164 10.3 7.2 10.8









#51
2,6-Dinitrotoluene
Concen: 43.04 ng
RT: 13.916 min Scan# 1841
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 32990
Ion Ratio Lower Upper
165 100
63 89.7 67.6 101.4
89 88.2 69.8 104.6



#52
Acenaphthene
Concen: 43.28 ng
RT: 14.392 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:154 Resp: 115073
Ion Ratio Lower Upper
154 100
153 111.7 88.3 132.5
152 58.1 46.2 69.4









#53
3-Nitroaniline
Concen: 39.06 ng
RT: 14.251 min Scan# 1898
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 26972
Ion Ratio Lower Upper
138 100
108 14.7 11.8 17.8
92 169.3 129.0 193.4



#54
2,4-Dinitrophenol
Concen: 45.54 ng
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:184 Resp: 16783
Ion Ratio Lower Upper
184 100
63 79.1 68.1 102.1
154 70.6 54.9 82.3









#55
Dibenzofuran
Concen: 44.80 ng
RT: 14.728 min Scan# 1979
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 209909
Ion Ratio Lower Upper
168 100
139 41.2 33.1 49.7
169 13.5 10.3 15.5



#56
4-Nitrophenol
Concen: 39.90 ng
RT: 14.581 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:139 Resp: 21841
Ion Ratio Lower Upper
139 100
109 126.4 101.8 141.8
65 140.2 121.3 161.3









#57
2,4-Dinitrotoluene
Concen: 42.06 ng
RT: 14.716 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	165	Resp:	50335
Ion Ratio	Lower	Upper	
165	100		
63	73.8	58.3	87.5
89	103.2	87.0	130.4
182	2.9	2.6	4.0



#58
Fluorene
Concen: 46.42 ng
RT: 15.386 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:	166	Resp:	172933
Ion Ratio	Lower	Upper	
166	100		
165	99.7	76.9	115.3
167	14.1	10.6	16.0









#59
2,3,4,6-Tetrachlorophenol
Concen: 48.36 ng
RT: 14.963 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:232	Resp:	44472
Ion Ratio	Lower	Upper
232	100	
131	42.0	29.5 44.3
130	2.7	1.8 2.8
166	29.8	18.6 27.8#



#60
Diethylphthalate
Concen: 41.75 ng
RT: 15.163 min Scan# 2053
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:149	Resp:	143636
Ion Ratio	Lower	Upper
149	100	
177	22.6	18.6 27.8
150	13.1	9.8 14.8









#61
4-Chlorophenyl-phenylether
Concen: 49.03 ng
RT: 15.381 min Scan# 2090
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 95860
Ion Ratio Lower Upper
204 100
206 34.2 26.2 39.4
141 60.0 46.9 70.3



#62
4-Nitroaniline
Concen: 42.90 ng
RT: 15.422 min Scan# 2097
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:138 Resp: 30478
Ion Ratio Lower Upper
138 100
92 70.3 57.0 97.0
108 141.4 115.8 155.8









#63
Azobenzene
Concen: 44.34 ng
RT: 15.675 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77	Resp: 222913
Ion Ratio	Lower Upper
77 100	
182 20.4	0.4 40.4
105 16.4	0.0 36.9
51 27.1	5.3 45.3



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.092 min Scan# 2381
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:188	Resp: 121400
Ion Ratio	Lower Upper
188 100	
94 8.3	7.0 10.6
80 9.9	8.4 12.6









#65
4,6-Dinitro-2-methylphenol
Concen: 42.11 ng
RT: 15.481 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 30116
Ion Ratio Lower Upper
198 100
51 43.3 16.1 56.1
105 46.1 25.3 65.3



#66
n-Nitrosodiphenylamine
Concen: 41.19 ng
RT: 15.598 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:169 Resp: 147616
Ion Ratio Lower Upper
169 100
168 64.9 52.2 78.4
167 36.5 29.9 44.9









#67
4-Bromophenyl-phenylether
Concen: 40.63 ng
RT: 16.281 min Scan# 2243
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 60872
Ion Ratio Lower Upper
248 100
250 96.2 79.0 118.6
141 72.9 59.0 88.4



#68
Hexachlorobenzene
Concen: 43.14 ng
RT: 16.392 min Scan# 2262
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:284 Resp: 72508
Ion Ratio Lower Upper
284 100
142 32.6 27.8 41.6
249 33.0 25.5 38.3









#69
Atrazine
Concen: 45.27 ng
RT: 16.563 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 49887
Ion Ratio Lower Upper
200 100
173 26.4 6.2 46.2
215 48.4 34.1 74.1



#70
Pentachlorophenol
Concen: 44.62 ng
RT: 16.745 min Scan# 2322
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:266 Resp: 35505
Ion Ratio Lower Upper
266 100
268 68.1 49.5 74.3
264 60.2 53.3 79.9









#71
Phenanthrene
Concen: 41.41 ng
RT: 17.133 min Scan# 2388
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 300614
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.6 11.8 17.6



#72
Anthracene
Concen: 43.71 ng
RT: 17.228 min Scan# 2404
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:178 Resp: 294002
Ion Ratio Lower Upper
178 100
176 18.2 14.3 21.5
179 15.1 12.0 18.0









#73
 Carbazole
 Concen: 44.07 ng
 RT: 17.504 min Scan# 2451
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:167 Resp: 275589
 Ion Ratio Lower Upper
 167 100
 166 21.0 17.8 26.8
 139 12.8 10.2 15.2



#74
 Di-n-butylphthalate
 Concen: 43.52 ng
 RT: 18.063 min Scan# 2546
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:149 Resp: 237770
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.3 10.9
 104 10.5 8.1 12.1









#75
Fluoranthene
Concen: 43.71 ng
RT: 19.116 min Scan# 2725
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
202	100		
101	8.6	0.0	29.6
203	17.3	0.0	37.6



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.198 min Scan# 3079
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion	Ratio	Lower	Upper
240	100		
120	6.4	6.5	9.7#
236	25.6	19.8	29.6









#77
Benzidine
Concen: 43.10 ng
RT: 19.310 min Scan# 2758
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:184	Resp:	92575
Ion Ratio	Lower	Upper
184	100	
185	14.2	11.7 17.5
183	13.8	9.7 14.5



#78
Pyrene
Concen: 40.76 ng
RT: 19.474 min Scan# 2786
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:202	Resp:	403982
Ion Ratio	Lower	Upper
202	100	
200	20.7	16.5 24.7
203	17.3	13.9 20.9









#79
 Terphenyl-d14
 Concen: 83.19 ng
 RT: 19.669 min Scan# 2819
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:244 Resp: 731452
 Ion Ratio Lower Upper
 244 100
 212 7.2 5.8 8.6
 122 8.3 6.7 10.1



#80
 Butylbenzylphthalate
 Concen: 37.49 ng
 RT: 20.351 min Scan# 2935
 Delta R.T. 0.000 min
 Lab File: BP005310.D
 Acq: 16 Apr 2021 10:07

Tgt Ion:149 Resp: 88468
 Ion Ratio Lower Upper
 149 100
 91 81.0 71.2 106.8
 206 25.0 18.6 28.0









#81
Benzo(a)anthracene
Concen: 40.04 ng
RT: 21.180 min Scan# 3076
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 443698
Ion Ratio Lower Upper
228 100
226 26.1 21.5 32.3
229 20.0 15.6 23.4



#82
3,3'-Dichlorobenzidine
Concen: 40.81 ng
RT: 21.116 min Scan# 3065
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:252 Resp: 111814
Ion Ratio Lower Upper
252 100
254 64.3 50.5 75.7
126 9.2 7.2 10.8









#83
Chrysene
Concen: 39.43 ng
RT: 21.233 min Scan# 3085
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 432422
Ion Ratio Lower Upper
228 100
226 29.5 24.0 36.0
229 19.6 15.6 23.4



#84
Bis(2-ethylhexyl)phthalate
Concen: 36.20 ng
RT: 21.104 min Scan# 3063
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:149 Resp: 143523
Ion Ratio Lower Upper
149 100
167 28.6 23.8 35.8
279 4.6 4.6 6.8









#85
Di-n-octyl phthalate
Concen: 36.74 ng
RT: 21.986 min Scan# 3213
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 269983
Ion Ratio Lower Upper
149 100
167 1.7 1.5 2.3
43 7.7 6.9 10.3



#86
Perylene-d12
Concen: 20.00 ng
RT: 23.463 min Scan# 3464
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:264 Resp: 161171
Ion Ratio Lower Upper
264 100
260 24.9 19.0 28.6
265 22.1 17.0 25.6









#87
Indeno(1,2,3-cd)pyrene
Concen: 41.45 ng
RT: 25.798 min Scan# 3861
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 499246
Ion Ratio Lower Upper
276 100
138 15.1 12.8 19.2
277 25.7 20.8 31.2



#88
Benzo(b)fluoranthene
Concen: 41.46 ng
RT: 22.780 min Scan# 3348
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:252 Resp: 447883
Ion Ratio Lower Upper
252 100
253 21.5 17.1 25.7
125 5.7 5.3 7.9









#89
Benzo(k)fluoranthene
Concen: 42.38 ng
RT: 22.821 min Scan# 3355
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 448957
Ion Ratio Lower Upper
252 100
253 22.3 17.4 26.0
125 6.0 5.4 8.2



#90
Benzo(a)pyrene
Concen: 41.14 ng
RT: 23.368 min Scan# 3448
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:252 Resp: 407373
Ion Ratio Lower Upper
252 100
253 21.2 17.8 26.6
125 6.3 6.0 9.0









#91
Dibenzo(a,h)anthracene
Concen: 41.59 ng
RT: 25.804 min Scan# 3862
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 435758
Ion Ratio Lower Upper
278 100
139 10.7 8.1 12.1
279 23.7 18.9 28.3



#92
Benzo(g,h,i)perylene
Concen: 41.06 ng
RT: 26.509 min Scan# 3982
Delta R.T. 0.000 min
Lab File: BP005310.D
Acq: 16 Apr 2021 10:07

Tgt Ion:276 Resp: 408166
Ion Ratio Lower Upper
276 100
277 22.7 19.7 29.5
138 12.1 9.8 14.6

