

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042221\
 Data File : BP005366.D
 Acq On : 22 Apr 2021 09:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 22 12:50:56 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 21 17:40:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	62	0.00
2	1,4-Dioxane	40.000	42.740	-6.9	65	0.00
3	Pyridine	40.000	47.840	-19.6	70	0.00
4	n-Nitrosodimethylamine	40.000	49.861	-24.7	79	0.00
5 S	2-Fluorophenol	80.000	81.847	-2.3	63	0.00
6	Aniline	40.000	43.099	-7.7	66	0.00
7 S	Phenol-d6	80.000	84.660	-5.8	66	0.00
8	2-Chlorophenol	40.000	39.371	1.6	62	0.00
9	Benzaldehyde	40.000	44.659	-11.6	68	0.00
10 C	Phenol	40.000	42.099	-5.2	65	0.00
11	bis(2-Chloroethyl)ether	40.000	41.243	-3.1	63	0.00
12	1,3-Dichlorobenzene	40.000	39.858	0.4	63	0.00
13 C	1,4-Dichlorobenzene	40.000	40.254	-0.6	64	0.00
14	1,2-Dichlorobenzene	40.000	39.793	0.5	63	0.00
15	Benzyl Alcohol	40.000	48.306	-20.8	73	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.314	-3.3	64	0.00
17	2-Methylphenol	40.000	42.945	-7.4	65	0.00
18	Hexachloroethane	40.000	45.173	-12.9	69	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	48.519	-21.3	71	0.00
20	3+4-Methylphenols	40.000	43.120	-7.8	65	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	64	0.00
22	Acetophenone	40.000	42.439	-6.1	66	0.00
23 S	Nitrobenzene-d5	80.000	98.267	-22.8	72	0.00
24	Nitrobenzene	40.000	47.394	-18.5	72	0.00
25	Isophorone	40.000	42.409	-6.0	69	0.00
26 C	2-Nitrophenol	40.000	45.832	-14.6	68	0.00
27	2,4-Dimethylphenol	40.000	42.098	-5.2	65	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.112	-7.8	67	0.00
29 C	2,4-Dichlorophenol	40.000	42.350	-5.9	66	0.00
30	1,2,4-Trichlorobenzene	40.000	41.287	-3.2	66	0.00
31	Naphthalene	40.000	40.314	-0.8	64	0.00
32	Benzoic acid	40.000	41.048	-2.6	70	0.00
33	4-Chloroaniline	40.000	41.690	-4.2	64	0.00
34 C	Hexachlorobutadiene	40.000	43.350	-8.4	68	0.00
35	Caprolactam	40.000	43.465	-8.7	67	0.00
36 C	4-Chloro-3-methylphenol	40.000	44.987	-12.5	68	0.00
37	2-Methylnaphthalene	40.000	41.988	-5.0	65	0.00
38	1-Methylnaphthalene	40.000	41.494	-3.7	65	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	67	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.575	-3.9	67	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.500	-6.3	74	0.00
42 S	2,4,6-Tribromophenol	80.000	86.185	-7.7	67	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.271	-8.2	66	0.00
44	2,4,5-Trichlorophenol	40.000	42.957	-7.4	67	0.00
45 S	2-Fluorobiphenyl	80.000	82.575	-3.2	66	0.00
46	1,1'-Biphenyl	40.000	41.101	-2.8	66	0.00
47	2-Chloronaphthalene	40.000	41.156	-2.9	66	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042221\
 Data File : BP005366.D
 Acq On : 22 Apr 2021 09:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 22 12:50:56 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 21 17:40:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.104	-12.8	79	0.00
49	Acenaphthylene	40.000	41.767	-4.4	65	0.00
50	Dimethylphthalate	40.000	43.175	-7.9	68	0.00
51	2,6-Dinitrotoluene	40.000	43.063	-7.7	65	0.00
52 C	Acenaphthene	40.000	41.761	-4.4	67	0.00
53	3-Nitroaniline	40.000	40.804	-2.0	65	0.00
54 P	2,4-Dinitrophenol	40.000	41.597	-4.0	71	0.00
55	Dibenzofuran	40.000	41.689	-4.2	67	0.00
56 P	4-Nitrophenol	40.000	38.492	3.8	64	0.00
57	2,4-Dinitrotoluene	40.000	40.265	-0.7	66	0.00
58	Fluorene	40.000	42.081	-5.2	67	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.980	-12.4	69	0.00
60	Diethylphthalate	40.000	44.763	-11.9	70	0.00
61	4-Chlorophenyl-phenylether	40.000	43.114	-7.8	69	0.00
62	4-Nitroaniline	40.000	40.320	-0.8	66	0.00
63	Azobenzene	40.000	49.976	-24.9	76	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	69	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.933	-7.3	70	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.307	-3.3	68	0.00
67	4-Bromophenyl-phenylether	40.000	41.612	-4.0	69	0.00
68	Hexachlorobenzene	40.000	40.075	-0.2	68	0.00
69	Atrazine	40.000	44.915	-12.3	70	0.00
70 C	Pentachlorophenol	40.000	41.261	-3.2	69	0.00
71	Phenanthrene	40.000	40.000	0.0	67	0.00
72	Anthracene	40.000	41.046	-2.6	68	0.00
73	Carbazole	40.000	41.015	-2.5	68	0.00
74	Di-n-butylphthalate	40.000	40.621	-1.6	72	0.00
75 C	Fluoranthene	40.000	41.866	-4.7	69	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	70	0.00
77	Benzydine	40.000	50.718	-26.8#	82	0.00
78	Pyrene	40.000	40.186	-0.5	69	0.00
79 S	Terphenyl-d14	80.000	81.458	-1.8	71	0.00
80	Butylbenzylphthalate	40.000	46.727	-16.8	76	0.00
81	Benzo(a)anthracene	40.000	40.943	-2.4	71	0.00
82	3,3'-Dichlorobenzidine	40.000	43.983	-10.0	74	0.00
83	Chrysene	40.000	40.358	-0.9	71	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.055	-5.1	76	0.00
85 c	Di-n-octyl phthalate	40.000	43.770	-9.4	77	0.00
86 I	Perylene-d12	20.000	20.000	0.0	74	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.739	-1.8	74	0.00
88	Benzo(b)fluoranthene	40.000	40.330	-0.8	72	0.00
89	Benzo(k)fluoranthene	40.000	40.000	0.0	71	0.00
90 C	Benzo(a)pyrene	40.000	40.684	-1.7	73	0.00
91	Dibenzo(a,h)anthracene	40.000	40.748	-1.9	74	0.00
92	Benzo(g,h,i)perylene	40.000	40.842	-2.1	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042221\
Data File : BP005366.D
Acq On : 22 Apr 2021 09:34
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Apr 22 12:50:56 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 21 17:40:15 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0