

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042623\
 Data File : BP014835.D
 Acq On : 26 Apr 2023 12:09
 Operator : CG/JU
 Sample : 02418-21MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EW8Z5MSD

Quant Time: Apr 26 22:32:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 26 02:53:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.199	152	319682	20.000	ng/u1	0.00
20) Naphthalene-d8	11.034	136	1359091	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.834	164	813581	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.581	188	1609941	20.000	ng/u1	0.00
79) Chrysene-d12	21.657	240	935829	20.000	ng/u1	0.00
88) Perylene-d12	24.257	264	1104770	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.505	96	24033	2.956	ng/uL	0.00
4) Pyridine-d5	3.964	84	10561	0.448	ng/u1	0.02
7) Phenol-d5	7.340	99	485826	16.717	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.516	67	352114	20.162	ng/u1	0.00
11) 2-Chlorophenol-d4	7.728	132	430318	20.053	ng/u1	0.00
15) 4-Methylphenol-d8	8.910	113	357643	15.322	ng/u1	0.00
21) Nitrobenzene-d5	9.375	128	224288	22.046	ng/u1	0.00
24) 2-Nitrophenol-d4	10.105	143	235833	22.553	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.646	165	469657	21.608	ng/u1	0.00
31) 4-Chloroaniline-d4	11.163	131	404573	12.972	ng/u1	0.00
46) Dimethylphthalate-d6	14.234	166	1375018	21.781	ng/u1	0.00
49) Acenaphthylene-d8	14.528	160	1615416	22.320	ng/u1	0.00
54) 4-Nitrophenol-d4	15.010	143	169182	14.710	ng/u1	0.00
60) Fluorene-d10	15.822	176	1185714	22.061	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.934	200	141239	15.810	ng/u1	0.00
73) Anthracene-d10	17.681	188	1696195	22.437	ng/u1	0.00
81) Pyrene-d10	19.892	212	1590503	27.954	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.092	264	1291608	22.552	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.546	88	79277	9.035	ng/uL	98
6) Benzaldehyde	7.328	77	490575	35.153	ng/u1	95
8) Phenol	7.369	94	902319	29.855	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.611	93	648368	26.762	ng/u1	98
12) 2-Chlorophenol	7.758	128	638519	28.528	ng/u1	99
13) 2-Methylphenol	8.640	108	628460	27.568	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.728	45	803708	26.618	ng/u1	99
16) Acetophenone	9.034	105	1223326	33.502	ng/u1	99
17) N-Nitroso-di-n-propyla...	9.016	70	495941	26.540	ng/u1	99
18) 4-Methylphenol	8.975	108	696941	27.309	ng/u1	98
19) Hexachloroethane	9.299	117	251711	27.879	ng/u1	96
22) Nitrobenzene	9.416	77	727751	28.120	ng/u1	100
23) Isophorone	9.952	82	1480001	27.683	ng/u1	99
25) 2-Nitrophenol	10.140	139	351169	30.084	ng/u1	98
26) 2,4-Dimethylphenol	10.187	107	560572	21.642	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.428	93	894226	26.569	ng/u1	98
29) 2,4-Dichlorophenol	10.675	162	645207	29.790	ng/u1	98
30) Naphthalene	11.087	128	2200061	29.266	ng/u1	99
32) 4-Chloroaniline	11.187	127	595490	18.950	ng/u1	100
33) Hexachlorobutadiene	11.369	225	396888	28.812	ng/u1	98
34) Caprolactam	11.969	113	198854	27.848	ng/u1	93
35) 4-Chloro-3-methylphenol	12.293	107	683373	28.495	ng/u1	97
36) 2-Methylnaphthalene	12.675	142	1447016	28.350	ng/u1	99
37) 1-Methylnaphthalene	12.893	142	1470105	28.891	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042623\
 Data File : BP014835.D
 Acq On : 26 Apr 2023 12:09
 Operator : CG/JU
 Sample : 02418-21MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EW8Z5MSD


Quant Time: Apr 26 22:32:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 26 02:53:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) 1,2,4,5-Tetrachloroben...	13.040	216	760490	29.836	ng/ul	98
40) Hexachlorocyclopentadiene	13.016	237	364177	24.657	ng/ul	97
41) 2,4,6-Trichlorophenol	13.269	196	510042	31.527	ng/ul	100
42) 2,4,5-Trichlorophenol	13.340	196	562842	31.481	ng/ul	99
43) 1,1'-Biphenyl	13.669	154	1877761	29.361	ng/ul	100
44) 2-Chloronaphthalene	13.716	162	1470993	29.381	ng/ul	99
45) 2-Nitroaniline	13.916	65	405624	32.349	ng/ul	97
47) Dimethylphthalate	14.281	163	1817060	28.719	ng/ul	100
48) 2,6-Dinitrotoluene	14.404	165	359192	31.685	ng/ul	97
50) Acenaphthylene	14.557	152	2479854	31.727	ng/ul	99
51) 3-Nitroaniline	14.728	138	338260	27.203	ng/ul	99
52) Acenaphthene	14.898	153	1740145	32.388	ng/ul	100
53) 2,4-Dinitrophenol	14.934	184	157813	23.631	ng/ul	96
55) 4-Nitrophenol	15.022	109	240769	27.979	ng/ul	99
56) Dibenzofuran	15.228	168	2350855	31.070	ng/ul	99
57) 2,4-Dinitrotoluene	15.187	165	509065	30.476	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.451	232	456437	30.275	ng/ul	97
59) Diethylphthalate	15.640	149	1811512	29.565	ng/ul	99
61) Fluorene	15.875	166	1968766	32.985	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.863	204	866639	28.641	ng/ul	98
63) 4-Nitroaniline	15.892	138	351972	31.381	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.945	198	236844	26.377	ng/ul	96
67) N-Nitrosodiphenylamine	16.081	169	1501231	31.706	ng/ul	99
68) 4-Bromophenyl-phenylether	16.763	248	553019	31.745	ng/ul	99
69) Hexachlorobenzene	16.881	284	625815	30.977	ng/ul	99
70) Atrazine	17.028	200	439340	27.150	ng/ul	99
71) Pentachlorophenol	17.228	266	346018	29.197	ng/ul	98
72) Phenanthrene	17.628	178	6854052	77.742	ng/ul	99
74) Anthracene	17.716	178	3349459	37.681	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.640	216	762124	31.799	ng/uL	99
76) Pentachlorobenzene	15.145	250	759243	32.087	ng/uL	100
77) Carbazole	17.975	167	2768201	36.590	ng/ul	100
78) Di-n-butylphthalate	18.510	149	3039875	33.855	ng/ul	100
80) Fluoranthene	19.575	202	7636332	113.886	ng/ul	100
82) Pyrene	19.928	202	6939722	100.069	ng/ul	98
83) Butylbenzylphthalate	20.775	149	1205907	51.826	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.563	252	232807	11.922	ng/ul	99
85) Benzo(a)anthracene	21.639	228	3902912	59.950	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.539	149	6497907	188.118	ng/ul#	97
87) Chrysene	21.698	228	3878416	62.842	ng/ul	99
89) Di-n-octyl phthalate	22.527	149	2101836	28.617	ng/ul	100
90) Benzo(b)fluoranthene	23.463	252	4699446	62.113	ng/ul	98
91) Benzo(k)fluoranthene	23.510	252	3272067	44.591	ng/ul	99
93) Benzo(a)pyrene	24.145	252	3730403	58.212	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.033	276	4053021	56.086	ng/ul	100
95) Dibenzo(a,h)anthracene	27.051	278	2464164	40.888	ng/ul	99
96) Benzo(g,h,i)perylene	27.874	276	1442321	25.192	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 26 22:32:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 26 02:53:24 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.035 ng/uL
RT: 3.546 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion: 88 Resp: 79277
Ion Ratio Lower Upper
88 100
43 27.2 21.3 31.9
58 67.3 55.4 83.0



#6
Benzaldehyde
Concen: 35.153 ng/uL
RT: 7.328 min Scan# 670
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion: 77 Resp: 490575
Ion Ratio Lower Upper
77 100
105 103.0 79.9 119.9
106 99.8 75.2 112.8





#8

Phenol

Concen: 29.855 ng/ul

RT: 7.369 min Scan# 6

Delta R.T. 0.006 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion: 94 Resp: 902319

Ion Ratio Lower Upper

94 100

65 25.9 21.0 31.6

66 32.8 27.0 40.6



#10

Bis(2-Chloroethyl)ether

Concen: 26.762 ng/ul

RT: 7.611 min Scan# 718

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion: 93 Resp: 648368


Ion Ratio Lower Upper

93 100

63 72.5 56.3 84.5

95 32.6 25.9 38.9





#12

2-Chlorophenol

Concen: 28.528 ng/ul

RT: 7.758 min Scan# 743

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:128 Resp: 638519

Ion Ratio Lower Upper

128 100

64 45.4 35.6 53.4

130 32.2 25.0 37.6



#13

2-Methylphenol

Concen: 27.568 ng/ul

RT: 8.640 min Scan# 893

Delta R.T. 0.000 min

Lab File: BP014835.D




Acq: 26 Apr 2023 12:09


Tgt Ion:108 Resp: 628460

Ion Ratio Lower Upper

108 100

107 87.9 71.0 106.4









#14
2,2'-oxybis(1-Chloropropane)
Concen: 26.618 ng/ul
RT: 8.728 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion: 45 Resp: 803708
Ion Ratio Lower Upper
45 100
77 15.3 12.5 18.7
79 11.4 9.4 14.2



#16
Acetophenone
Concen: 33.502 ng/ul
RT: 9.034 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:105 Resp: 1223326
Ion Ratio Lower Upper
105 100
77 75.3 59.8 89.8
51 23.6 19.6 29.4









#17
N-Nitroso-di-n-propylamine
Concen: 26.540 ng/ul
RT: 9.016 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion: 70 Resp: 495941
Ion Ratio Lower Upper
70 100
42 53.8 43.5 65.3
101 10.0 7.9 11.9
130 21.2 16.4 24.6



#18
4-Methylphenol
Concen: 27.309 ng/ul
RT: 8.975 min Scan# 950
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion: 108 Resp: 696941
Ion Ratio Lower Upper
108 100
107 108.1 88.6 132.8









#19
Hexachloroethane
Concen: 27.879 ng/ul
RT: 9.299 min Scan# 1005
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:117 Resp: 251711
Ion Ratio Lower Upper
117 100
201 103.6 80.3 120.5
199 65.2 49.0 73.6



#22
Nitrobenzene
Concen: 28.120 ng/ul
RT: 9.416 min Scan# 1025
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion: 77 Resp: 727751
Ion Ratio Lower Upper
77 100
123 44.7 35.8 53.6
65 12.7 10.2 15.2









#23
Isophorone
Concen: 27.683 ng/ul
RT: 9.952 min Scan# 1116
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion: 82 Resp: 1480001
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.2
138 17.4 14.1 21.1



#25
2-Nitrophenol
Concen: 30.084 ng/ul
RT: 10.140 min Scan# 1148
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:139 Resp: 351169
Ion Ratio Lower Upper
139 100
65 48.8 40.7 61.1
109 29.3 23.4 35.0









#26
2,4-Dimethylphenol
Concen: 21.642 ng/ul
RT: 10.187 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:107 Resp: 560572
Ion Ratio Lower Upper
107 100
121 49.8 41.3 61.9
122 84.4 70.4 105.6



#27
Bis(2-Chloroethoxy)methane
Concen: 26.569 ng/ul
RT: 10.428 min Scan# 1197
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion: 93 Resp: 894226
Ion Ratio Lower Upper
93 100
95 32.9 25.2 37.8
123 11.0 8.5 12.7









#29
2,4-Dichlorophenol
Concen: 29.790 ng/ul
RT: 10.675 min Scan# 1239
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:162 Resp: 645207
Ion Ratio Lower Upper
162 100
164 65.3 50.6 75.8
98 35.0 28.5 42.7



#30
Naphthalene
Concen: 29.266 ng/ul
RT: 11.087 min Scan# 1309
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:128 Resp: 2200061
Ion Ratio Lower Upper
128 100
129 11.1 9.1 13.7
127 13.0 10.6 16.0









#32
4-Chloroaniline
Concen: 18.950 ng/ul
RT: 11.187 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:127 Resp: 595490
Ion Ratio Lower Upper
127 100
129 32.1 25.8 38.8



#33
Hexachlorobutadiene
Concen: 28.812 ng/ul
RT: 11.369 min Scan# 1357
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:225 Resp: 396888
Ion Ratio Lower Upper
225 100
223 64.8 49.9 74.9
227 63.4 51.5 77.3









#34
Caprolactam
Concen: 27.848 ng/ul
RT: 11.969 min Scan# 1457
Delta R.T. 0.018 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:113 Resp: 198854
Ion Ratio Lower Upper
113 100
55 153.4 129.0 193.6
56 114.6 99.4 149.2



#35
4-Chloro-3-methylphenol
Concen: 28.495 ng/ul
RT: 12.293 min Scan# 1514
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:107 Resp: 683373
Ion Ratio Lower Upper
107 100
144 25.8 21.4 32.0
142 78.7 65.0 97.6









#36
2-Methylnaphthalene
Concen: 28.350 ng/ul
RT: 12.675 min Scan# 1579
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:142 Resp: 1447016
Ion Ratio Lower Upper
142 100
141 89.7 70.6 106.0



#37
1-Methylnaphthalene
Concen: 28.891 ng/ul
RT: 12.893 min Scan# 1616
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:142 Resp: 1470105
Ion Ratio Lower Upper
142 100
141 91.4 73.7 110.5
116 3.9 3.0 4.4









#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.836 ng/ul
RT: 13.040 min Scan# 1641
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:	216	Resp:	760490
Ion	Ratio	Lower	Upper
216	100		
214	78.1	63.3	94.9
179	20.1	16.7	25.1
108	16.4	14.5	21.7



#40
Hexachlorocyclopentadiene
Concen: 24.657 ng/ul
RT: 13.016 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:	237	Resp:	364177
Ion	Ratio	Lower	Upper
237	100		
235	64.9	50.0	75.0
272	13.6	10.1	15.1





#41

2,4,6-Trichlorophenol

Concen: 31.527 ng/ul

RT: 13.269 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:196 Resp: 510042

Ion Ratio Lower Upper

196 100

198 95.9 76.7 115.1

200 31.7 25.3 37.9



#42

2,4,5-Trichlorophenol

Concen: 31.481 ng/ul

RT: 13.340 min Scan# 1692

Delta R.T. 0.006 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:196 Resp: 562842


Ion Ratio Lower Upper

196 100

198 94.3 75.3 112.9

200 31.4 24.2 36.4









#43
1,1'-Biphenyl
Concen: 29.361 ng/ul
RT: 13.669 min Scan# 1748
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:154 Resp: 1877761
Ion Ratio Lower Upper
154 100
153 39.8 31.9 47.9
76 13.6 11.2 16.8



#44
2-Chloronaphthalene
Concen: 29.381 ng/ul
RT: 13.716 min Scan# 1756
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:162 Resp: 1470993
Ion Ratio Lower Upper
162 100
164 33.3 26.2 39.2
127 37.9 29.7 44.5









#45
2-Nitroaniline
Concen: 32.349 ng/ul
RT: 13.916 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion: 65 Resp: 405624
Ion Ratio Lower Upper
65 100
92 72.9 58.5 87.7
138 114.6 87.6 131.4



#47
Dimethylphthalate
Concen: 28.719 ng/ul
RT: 14.281 min Scan# 1852
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion: 163 Resp: 1817060
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.2 8.0 12.0





#48

2,6-Dinitrotoluene

Concen: 31.685 ng/ul

RT: 14.404 min Scan# 1873

Delta R.T. 0.006 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:165 Resp: 359192

Ion Ratio Lower Upper

165 100

89 57.3 47.6 71.4

121 20.3 17.5 26.3



#50

Acenaphthylene

Concen: 31.727 ng/ul

RT: 14.557 min Scan# 1899

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:152 Resp: 2479854


Ion Ratio Lower Upper

152 100

151 20.0 15.8 23.6

153 13.2 10.4 15.6









#51
3-Nitroaniline
Concen: 27.203 ng/ul
RT: 14.728 min Scan# 1928
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:138 Resp: 338260
Ion Ratio Lower Upper
138 100
108 11.1 8.4 12.6
92 117.4 93.2 139.8



#52
Acenaphthene
Concen: 32.388 ng/ul
RT: 14.898 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:153 Resp: 1740145
Ion Ratio Lower Upper
153 100
152 47.9 38.2 57.4
154 88.2 71.0 106.4









#53
2,4-Dinitrophenol
Concen: 23.631 ng/ul
RT: 14.934 min Scan# 1963
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:184 Resp: 157813
Ion Ratio Lower Upper
184 100
63 61.2 53.6 80.4
154 60.7 49.1 73.7



#55
4-Nitrophenol
Concen: 27.979 ng/ul
RT: 15.022 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:109 Resp: 240769
Ion Ratio Lower Upper
109 100
139 140.1 112.8 169.2
65 131.5 103.6 155.4









#56
Dibenzofuran
Concen: 31.070 ng/ul
RT: 15.228 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:168 Resp: 2350855
Ion Ratio Lower Upper
168 100
139 35.2 28.7 43.1



#57
2,4-Dinitrotoluene
Concen: 30.476 ng/ul
RT: 15.187 min Scan# 2006
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:165 Resp: 509065
Ion Ratio Lower Upper
165 100
63 39.7 33.8 50.6
182 2.9 2.2 3.4









#58
2,3,4,6-Tetrachlorophenol
Concen: 30.275 ng/ul
RT: 15.451 min Scan# 2051
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:232	Resp:	456437
Ion Ratio	Lower	Upper
232	100	
131	41.1	35.0 52.4
130	2.2	1.8 2.6
166	30.1	25.0 37.6



#59
Diethylphthalate
Concen: 29.565 ng/ul
RT: 15.640 min Scan# 2083
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:149	Resp:	1811512
Ion Ratio	Lower	Upper
149	100	
177	23.0	17.9 26.9
150	12.3	9.8 14.6









#61
Fluorene
Concen: 32.985 ng/ul
RT: 15.875 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:166 Resp: 1968766
Ion Ratio Lower Upper
166 100
165 98.7 79.1 118.7
167 13.9 10.7 16.1



#62
4-Chlorophenyl-phenylether
Concen: 28.641 ng/ul
RT: 15.863 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:204 Resp: 866639
Ion Ratio Lower Upper
204 100
206 32.4 26.3 39.5
141 55.7 45.9 68.9









#63
4-Nitroaniline
Concen: 31.381 ng/ul
RT: 15.892 min Scan# 2127
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:138 Resp: 351972
Ion Ratio Lower Upper
138 100
92 50.4 39.0 58.6
108 73.5 59.2 88.8



#66
4,6-Dinitro-2-methylphenol
Concen: 26.377 ng/ul
RT: 15.945 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:198 Resp: 236844
Ion Ratio Lower Upper
198 100
182 4.4 3.5 5.3
77 26.0 18.7 28.1









#67
N-Nitrosodiphenylamine
Concen: 31.706 ng/ul
RT: 16.081 min Scan# 2159
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:169 Resp: 1501231
Ion Ratio Lower Upper
169 100
168 66.4 53.6 80.4
167 36.4 29.4 44.0



#68
4-Bromophenyl-phenylether
Concen: 31.745 ng/ul
RT: 16.763 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:248 Resp: 553019
Ion Ratio Lower Upper
248 100
250 96.9 79.1 118.7
141 66.9 53.8 80.6





#69

Hexachlorobenzene

Concen: 30.977 ng/ul

RT: 16.881 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:284 Resp: 625815

Ion Ratio Lower Upper

284 100

142 33.4 26.5 39.7

249 31.2 24.4 36.6



#70

Atrazine

Concen: 27.150 ng/ul

RT: 17.028 min Scan# 2319

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:200 Resp: 439340


Ion Ratio Lower Upper

200 100

173 26.0 21.5 32.3

215 48.3 38.4 57.6





#71

Pentachlorophenol

Concen: 29.197 ng/ul

RT: 17.228 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:266 Resp: 346018

Ion Ratio Lower Upper

266 100

264 63.9 49.0 73.4

268 63.7 51.4 77.2



#72

Phenanthrene

Concen: 77.742 ng/ul

RT: 17.628 min Scan# 2421

Delta R.T. 0.006 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:178 Resp: 6854052


Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.4

176 20.1 15.7 23.5





#74

Anthracene

Concen: 37.681 ng/uI

RT: 17.716 min Scan# 2437

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:178 Resp: 3349459

Ion Ratio Lower Upper

178 100

179 15.7 12.1 18.1

176 18.9 14.9 22.3



#75

1,2,3,4-Tetrachlorobenzene

Concen: 31.799 ng/uI

RT: 13.640 min Scan# 1743

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:216 Resp: 762124


Ion Ratio Lower Upper

216 100

214 79.2 62.9 94.3

218 48.4 39.5 59.3









#76
 Pentachlorobenzene
 Concen: 32.087 ng/uL
 RT: 15.145 min Scan# 1999
 Delta R.T. 0.000 min
 Lab File: BP014835.D
 Acq: 26 Apr 2023 12:09




Instrument :
 BNA_P
 ClientSampleId :
 EW8Z5MSD


Tgt Ion:250 Resp: 759243
 Ion Ratio Lower Upper
 250 100
 248 62.9 50.4 75.6
 252 62.7 50.3 75.5



#77
 Carbazole
 Concen: 36.590 ng/uL
 RT: 17.975 min Scan# 2480
 Delta R.T. 0.000 min
 Lab File: BP014835.D
 Acq: 26 Apr 2023 12:09

Tgt Ion:167 Resp: 2768201
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.2 25.8
 139 12.6 9.8 14.8





#78

Di-n-butylphthalate

Concen: 33.855 ng/ul

RT: 18.510 min Scan# 2571

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:149 Resp: 3039875

Ion Ratio Lower Upper

149 100

150 9.2 7.4 11.2

104 4.9 3.9 5.9



#80

Fluoranthene

Concen: 113.886 ng/ul

RT: 19.575 min Scan# 2752

Delta R.T. 0.006 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:202 Resp: 7636332


Ion Ratio Lower Upper

202 100

101 11.3 9.1 13.7

100 8.5 7.0 10.4





#82

Pyrene

Concen: 100.069 ng/ul

RT: 19.928 min Scan# 2812

Delta R.T. 0.006 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:202 Resp: 6939722

Ion Ratio Lower Upper

202 100

101 13.3 11.1 16.7

100 10.7 9.0 13.6



#83

Butylbenzylphthalate

Concen: 51.826 ng/ul

RT: 20.775 min Scan# 2956

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:149 Resp: 1205907


Ion Ratio Lower Upper

149 100

91 72.9 58.5 87.7

206 21.4 16.7 25.1





#84

3,3'-Dichlorobenzidine

Concen: 11.922 ng/ul

RT: 21.563 min Scan# 3090

Delta R.T. 0.006 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:252 Resp: 232807

Ion Ratio Lower Upper

252 100

254 65.4 51.4 77.0

126 11.2 9.2 13.8



#85

Benzo(a)anthracene

Concen: 59.950 ng/ul

RT: 21.639 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:228 Resp: 3902912


Ion Ratio Lower Upper

228 100

229 20.5 15.7 23.5

226 28.0 22.0 33.0









#86
 Bis(2-ethylhexyl)phthalate
 Concen: 188.118 ng/ul
 RT: 21.539 min Scan# 3086
 Delta R.T. 0.000 min
 Lab File: BP014835.D
 Acq: 26 Apr 2023 12:09




Instrument :
 BNA_P
 ClientSampleId :
 EW8Z5MSD


Tgt Ion:149 Resp: 6497907
 Ion Ratio Lower Upper
 149 100
 167 28.6 21.8 32.6
 279 5.2 3.4 5.0#



#87
 Chrysene
 Concen: 62.842 ng/ul
 RT: 21.698 min Scan# 3113
 Delta R.T. 0.006 min
 Lab File: BP014835.D
 Acq: 26 Apr 2023 12:09

Tgt Ion:228 Resp: 3878416
 Ion Ratio Lower Upper
 228 100
 226 30.3 24.5 36.7
 229 20.8 16.0 24.0





#89

Di-n-octyl phthalate

Concen: 28.617 ng/ul

RT: 22.527 min Scan# 31

Delta R.T. 0.000 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09





Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD

Tgt Ion:149 Resp: 2101836



#90

Benzo(b)fluoranthene

Concen: 62.113 ng/ul

RT: 23.463 min Scan# 3413

Delta R.T. 0.012 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:252 Resp: 4699446


Ion Ratio Lower Upper

252 100

253 22.5 17.2 25.8

125 10.5 7.8 11.8









#91
Benzo(k)fluoranthene
Concen: 44.591 ng/ul
RT: 23.510 min Scan# 3422
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09




Instrument :
BNA_P
ClientSampleId :
EW8Z5MSD


Tgt Ion:252 Resp: 3272067
Ion Ratio Lower Upper
252 100
253 22.4 17.5 26.3
125 10.7 7.9 11.9



#93
Benzo(a)pyrene
Concen: 58.212 ng/ul
RT: 24.145 min Scan# 3529
Delta R.T. 0.006 min
Lab File: BP014835.D
Acq: 26 Apr 2023 12:09

Tgt Ion:252 Resp: 3730403
Ion Ratio Lower Upper
252 100
253 22.3 17.5 26.3
125 11.7 8.6 12.8





#94

Indeno(1,2,3-cd)pyrene

Concen: 56.086 ng/ul

RT: 27.033 min Scan# 4020

Delta R.T. 0.018 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD





Tgt Ion:276 Resp: 4053021

Ion Ratio Lower Upper

276 100

138 22.2 17.7 26.5

277 24.8 19.6 29.4



#95

Dibenzo(a,h)anthracene

Concen: 40.888 ng/ul

RT: 27.051 min Scan# 4023

Delta R.T. 0.018 min

Lab File: BP014835.D

Acq: 26 Apr 2023 12:09




Tgt Ion:278 Resp: 2464164


Ion Ratio Lower Upper

278 100

139 15.3 12.0 18.0

279 24.5 18.9 28.3





#96

Benzo(g,h,i)perylene

Concen: 25.192 ng/ul

RT: 27.874 min Scan# 41

Delta R.T. 0.012 min

Lab File: BP014835.D


Acq: 26 Apr 2023 12:09

Instrument :

BNA_P

ClientSampleId :

EW8Z5MSD



Tgt Ion:276 Resp: 1442321

Ion Ratio Lower Upper

276 100

138 19.2 16.2 24.2

277 23.8 18.9 28.3

