

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042623\
 Data File : BP014843.D
 Acq On : 26 Apr 2023 16:37
 Operator : CG/JU
 Sample : 02418-08
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EW8Z4

Quant Time: Apr 26 23:29:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 26 02:53:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.199	152	208098	20.000	ng/u1	0.00
20) Naphthalene-d8	11.034	136	904882	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.834	164	560805	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.581	188	1140312	20.000	ng/u1	0.00
79) Chrysene-d12	21.651	240	695060	20.000	ng/u1	0.00
88) Perylene-d12	24.251	264	671271	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.505	96	9098	1.719	ng/uL	0.00
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	7.340	99	191541	10.125	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.517	67	170367	14.986	ng/u1	0.00
11) 2-Chlorophenol-d4	7.722	132	189622	13.575	ng/u1	0.00
15) 4-Methylphenol-d8	8.905	113	193851	12.758	ng/u1	0.00
21) Nitrobenzene-d5	9.375	128	106595	15.737	ng/u1	0.00
24) 2-Nitrophenol-d4	10.105	143	83800	12.036	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.646	165	168628	11.653	ng/u1	0.00
31) 4-Chloroaniline-d4	11.163	131	141069	6.793	ng/u1	0.00
46) Dimethylphthalate-d6	14.228	166	616088	14.158	ng/u1	0.00
49) Acenaphthylene-d8	14.528	160	698354	13.998	ng/u1	0.00
54) 4-Nitrophenol-d4	15.004	143	11525	1.454	ng/u1	0.00
60) Fluorene-d10	15.822	176	620973	16.762	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.928	200	7716	1.219	ng/u1	0.00
73) Anthracene-d10	17.681	188	869944	16.247	ng/u1	0.00
81) Pyrene-d10	19.887	212	220658	5.222	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.075	264	60147	1.728	ng/u1	0.00
Target Compounds						
					Qvalue	
16) Acetophenone	9.034	105	74659	3.141	ng/u1	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042623\
Data File : BP014843.D
Acq On : 26 Apr 2023 16:37
Operator : CG/JU
Sample : 02418-08
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EW8Z4

Quant Time: Apr 26 23:29:25 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 26 02:53:24 2023
Response via : Initial Calibration

