

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050422\
 Data File : BP010236.D
 Acq On : 04 May 2022 22:29
 Operator : CG/JU
 Sample : PB144568BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS568

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/05/2022
 Supervised By :mohammad ahmed 05/09/2022

Quant Time: May 05 02:42:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	177887	20.000	ng/ul	0.00
20) Naphthalene-d8	10.722	136	788504	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.551	164	525317	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.304	188	1076625	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.386	240	748682	20.000	ng/ul	# 0.00
88) Perylene-d12	23.827	264	630518	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.357	96	26552m	6.509	ng/uL	0.00
4) Pyridine-d5	3.769	84	305977	26.662	ng/ul	0.00
7) Phenol-d5	7.081	99	510859	31.573	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.245	67	344670	35.500	ng/ul	0.00
11) 2-Chlorophenol-d4	7.451	132	417391	34.832	ng/ul	0.00
15) 4-Methylphenol-d8	8.628	113	442871	33.035	ng/ul	0.00
21) Nitrobenzene-d5	9.075	128	217843	34.736	ng/ul	0.00
24) 2-Nitrophenol-d4	9.804	143	242054	35.341	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.345	165	439692	32.976	ng/ul	0.00
31) 4-Chloroaniline-d4	10.857	131	407539	21.317	ng/ul	0.00
46) Dimethylphthalate-d6	13.963	166	1414253	34.609	ng/ul	0.00
49) Acenaphthylene-d8	14.245	160	1616175	32.679	ng/ul	0.00
54) 4-Nitrophenol-d4	14.745	143	226878	30.158	ng/ul	0.00
60) Fluorene-d10	15.545	176	1191288	33.890	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.663	200	229047	33.646	ng/ul	0.00
73) Anthracene-d10	17.404	188	1799490	34.548	ng/ul	0.00
81) Pyrene-d10	19.633	212	1881844	44.198	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.668	264	1238149	37.457	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.387	88	52472	12.378	ng/uL#	16
5) Pyridine	3.787	79	377689	31.661	ng/ul#	43
6) Benzaldehyde	7.051	77	344073	39.792	ng/ul	95
8) Phenol	7.110	94	539594	33.190	ng/ul#	94
10) Bis(2-Chloroethyl)ether	7.340	93	428515	34.211	ng/ul#	79
12) 2-Chlorophenol	7.481	128	408208	33.772	ng/ul	97
13) 2-Methylphenol	8.363	108	411179	32.977	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.445	45	661558	35.594	ng/ul#	83
16) Acetophenone	8.739	105	669360	32.002	ng/ul#	78
17) N-Nitroso-di-n-propyla...	8.734	70	363121	33.136	ng/ul#	75
18) 4-Methylphenol	8.692	108	445951	32.484	ng/ul	91
19) Hexachloroethane	8.998	117	174983	34.452	ng/ul#	80
22) Nitrobenzene	9.122	77	524859	34.159	ng/ul#	85
23) Isophorone	9.651	82	1008807	33.241	ng/ul#	97
25) 2-Nitrophenol	9.834	139	249525	34.398	ng/ul#	84
26) 2,4-Dimethylphenol	9.898	107	507775	32.332	ng/ul#	78
27) Bis(2-Chloroethoxy)met...	10.128	93	604057	33.959	ng/ul#	97
29) 2,4-Dichlorophenol	10.369	162	430538	33.590	ng/ul#	84
30) Naphthalene	10.769	128	1388127	33.636	ng/ul	96
32) 4-Chloroaniline	10.881	127	562813	29.874	ng/ul	97
33) Hexachlorobutadiene	11.063	225	300799	33.113	ng/ul	94
34) Caprolactam	11.657	113	146883	31.859	ng/ul#	1
35) 4-Chloro-3-methylphenol	12.004	107	502310	32.442	ng/ul#	72
36) 2-Methylnaphthalene	12.380	142	977831	32.670	ng/ul	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050422\
 Data File : BP010236.D
 Acq On : 04 May 2022 22:29
 Operator : CG/JU
 Sample : PB144568BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS568

Manual Integrations
 APPROVED




Reviewed By :Jagrut Upadhyay 05/05/2022
 Supervised By :mohammad ahmed 05/09/2022

Quant Time: May 05 02:42:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.598	142	998771	32.864	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.751	216	561295	34.584	ng/ul	94
40) Hexachlorocyclopentadiene	12.733	237	293698	32.111	ng/ul	98
41) 2,4,6-Trichlorophenol	12.986	196	366940	34.944	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.057	196	404809	35.431	ng/ul#	88
43) 1,1'-Biphenyl	13.386	154	1345166	34.718	ng/ul#	93
44) 2-Chloronaphthalene	13.427	162	1026993	34.252	ng/ul	94
45) 2-Nitroaniline	13.627	65	342243	35.201	ng/ul#	81
47) Dimethylphthalate	14.004	163	1348003	33.942	ng/ul#	92
48) 2,6-Dinitrotoluene	14.127	165	288346	35.379	ng/ul	91
50) Acenaphthylene	14.274	152	1669353	34.076	ng/ul	95
51) 3-Nitroaniline	14.451	138	273504	37.045	ng/ul#	79
52) Acenaphthene	14.616	153	1084815	34.056	ng/ul	99
53) 2,4-Dinitrophenol	14.657	184	134680	29.572	ng/ul#	77
55) 4-Nitrophenol	14.763	109	185894	28.442	ng/ul#	40
56) Dibenzofuran	14.951	168	1578910	33.473	ng/ul	100
57) 2,4-Dinitrotoluene	14.910	165	406970	34.181	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.174	232	339983	33.476	ng/ul#	87
59) Diethylphthalate	15.369	149	1380376	34.172	ng/ul	93
61) Fluorene	15.598	166	1281987	33.103	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.592	204	678938	33.311	ng/ul	99
63) 4-Nitroaniline	15.616	138	276059m	38.426	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.680	198	218059	33.200	ng/ul#	92
67) N-Nitrosodiphenylamine	15.804	169	1123203	36.360	ng/ul	94
68) 4-Bromophenyl-phenylether	16.492	248	414634	36.106	ng/ul	92
69) Hexachlorobenzene	16.610	284	465621	34.931	ng/ul#	90
70) Atrazine	16.763	200	441917	35.187	ng/ul#	90
71) Pentachlorophenol	16.957	266	268457	31.624	ng/ul#	84
72) Phenanthrene	17.345	178	2027825	34.653	ng/ul	99
74) Anthracene	17.439	178	2031417	34.465	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.351	216	573658	36.498	ng/uL#	86
76) Pentachlorobenzene	14.874	250	549850	35.674	ng/uL	91
77) Carbazole	17.704	167	1720433	32.986	ng/ul#	96
78) Di-n-butylphthalate	18.257	149	2335581	36.398	ng/ul#	93
80) Fluoranthene	19.310	202	2226472	46.361	ng/ul	97
82) Pyrene	19.662	202	2190772	44.277	ng/ul	95
83) Butylbenzylphthalate	20.527	149	893092	43.667	ng/ul#	83
84) 3,3'-Dichlorobenzidine	21.298	252	517444	38.411	ng/ul#	97
85) Benzo(a)anthracene	21.368	228	1712719	36.047	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.292	149	1265383	41.744	ng/ul#	81
87) Chrysene	21.421	228	1658024	35.440	ng/ul	99
89) Di-n-octyl phthalate	22.227	149	1841050	37.560	ng/ul	100
90) Benzo(b)fluoranthene	23.080	252	1541094	36.357	ng/ul	99
91) Benzo(k)fluoranthene	23.133	252	1493221	37.189	ng/ul#	97
93) Benzo(a)pyrene	23.721	252	1463848	37.345	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.374	276	1604151	41.139	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.392	278	1361716	40.812	ng/ul#	95
96) Benzo(g,h,i)perylene	27.156	276	1371945	41.698	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Abundance Scan 51 (3.387 min): BP010224.D\data.ms (-47)







#2
1,4-Dioxane
Concen: 12.378 ng/uL
RT: 3.387 min Scan# 51
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Tgt Ion:	88	Resp:	52472
Ion	Ratio	Lower	Upper
88	100		
43	35.0	6.8	10.2
58	86.0	28.4	42.6

Instrument :
BNA_P
ClientSampleId :
SLCS568


Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#5
Pyridine
Concen: 31.661 ng/uL
RT: 3.787 min Scan# 119
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:	79	Resp:	377689
Ion	Ratio	Lower	Upper
79	100		
52	89.9	35.5	53.3
51	42.3	20.2	30.4





#6
Benzaldehyde
Concen: 39.792 ng/ul
RT: 7.051 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion: 77 Resp: 344073
Ion Ratio Lower Upper
77 100
105 93.4 77.0 115.4
106 88.3 77.0 115.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#8
Phenol
Concen: 33.190 ng/ul
RT: 7.110 min Scan# 684
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion: 94 Resp: 539594
Ion Ratio Lower Upper
94 100
65 30.6 19.0 28.6
66 44.0 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 34.211 ng/ul
RT: 7.340 min Scan# 723
Delta R.T. -0.000 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion: 93 Resp: 42851
Ion Ratio Lower Upper
93 100
63 79.9 45.0 67.4
95 31.8 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#12
2-Chlorophenol
Concen: 33.772 ng/ul
RT: 7.481 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion:128 Resp: 408208
Ion Ratio Lower Upper
128 100
64 55.1 45.3 67.9
130 33.3 29.0 43.6





#13
2-Methylphenol
Concen: 32.977 ng/ul
RT: 8.363 min Scan# 897
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:108 Resp: 411179
Ion Ratio Lower Upper
108 100
107 90.8 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 35.594 ng/ul
RT: 8.445 min Scan# 911
Delta R.T. -0.012 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion: 45 Resp: 661558
Ion Ratio Lower Upper
45 100
77 13.1 17.1 25.7#
79 9.8 13.2 19.8#





#16
Acetophenone
Concen: 32.002 ng/ul
RT: 8.739 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:105 Resp: 669360
Ion Ratio Lower Upper
105 100
77 81.9 53.4 80.2
51 31.3 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#17
N-Nitroso-di-n-propylamine
Concen: 33.136 ng/ul
RT: 8.734 min Scan# 960
Delta R.T. -0.000 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29



Tgt Ion: 70 Resp: 363121
Ion Ratio Lower Upper
70 100
42 61.2 31.8 47.8#
101 9.9 15.6 23.4#
130 20.7 13.7 20.5#









#18
4-Methylphenol
Concen: 32.484 ng/ul
RT: 8.692 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Instrument :
BNA_P
ClientSampleId :
SLCS568

Tgt Ion:108 Resp: 44595
Ion Ratio Lower Upper
108 100
107 114.2 83.6 125.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#19
Hexachloroethane
Concen: 34.452 ng/ul
RT: 8.998 min Scan# 1005
Delta R.T. -0.012 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:117 Resp: 174983
Ion Ratio Lower Upper
117 100
201 103.4 106.5 159.7#
199 65.8 59.0 88.6





#22
Nitrobenzene
Concen: 34.159 ng/ul
RT: 9.122 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion: 77 Resp: 524859
Ion Ratio Lower Upper
77 100
123 43.0 42.3 63.5
65 14.8 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#23
Isophorone
Concen: 33.241 ng/ul
RT: 9.651 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion: 82 Resp: 1008807
Ion Ratio Lower Upper
82 100
95 7.4 7.8 11.8#
138 16.6 12.9 19.3





#25
2-Nitrophenol
Concen: 34.398 ng/ul
RT: 9.834 min Scan# 1147
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:139 Resp: 24952
Ion Ratio Lower Upper
139 100
65 58.8 49.4 74.2
109 33.8 44.7 67.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#26
2,4-Dimethylphenol
Concen: 32.332 ng/ul
RT: 9.898 min Scan# 1158
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion:107 Resp: 507775
Ion Ratio Lower Upper
107 100
121 48.8 34.1 51.1
122 80.5 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 33.959 ng/ul
RT: 10.128 min Scan# 1197
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568



Tgt Ion: 93 Resp: 60405
Ion Ratio Lower Upper
93 100
95 32.2 26.2 39.4
123 11.4 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#29
2,4-Dichlorophenol
Concen: 33.590 ng/ul
RT: 10.369 min Scan# 1238
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:162 Resp: 430538
Ion Ratio Lower Upper
162 100
164 65.4 37.0 55.6#
98 39.4 31.0 46.6





#30
Naphthalene
Concen: 33.636 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. -0.012 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:128 Resp: 138812
Ion Ratio Lower Upper
128 100
129 11.3 8.6 13.0
127 12.7 12.2 18.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#32
4-Chloroaniline
Concen: 29.874 ng/ul
RT: 10.881 min Scan# 1325
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion:127 Resp: 562813
Ion Ratio Lower Upper
127 100
129 32.6 24.6 37.0





#33
Hexachlorobutadiene
Concen: 33.113 ng/ul
RT: 11.063 min Scan# 1356
Delta R.T. -0.012 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:225 Resp: 300799
Ion Ratio Lower Upper
225 100
223 61.8 53.4 80.0
227 61.9 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#34
Caprolactam
Concen: 31.859 ng/ul
RT: 11.657 min Scan# 1457
Delta R.T. -0.000 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion:113 Resp: 146883
Ion Ratio Lower Upper
113 100
55 190.8 58.2 87.2#
56 152.3 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 32.442 ng/ul
RT: 12.004 min Scan# 1516
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568



Tgt Ion:107 Resp: 502310
Ion Ratio Lower Upper
107 100
144 24.3 15.8 23.8
142 73.0 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#36
2-Methylnaphthalene
Concen: 32.670 ng/ul
RT: 12.380 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:142 Resp: 977831
Ion Ratio Lower Upper
142 100
141 88.4 64.3 96.5





#37
1-Methylnaphthalene
Concen: 32.864 ng/ul
RT: 12.598 min Scan# 1617
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:142 Resp: 998772
Ion Ratio Lower Upper
142 100
141 91.7 69.8 104.8
116 3.9 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.584 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29



Tgt Ion:216 Resp: 561295
Ion Ratio Lower Upper
216 100
214 76.6 58.2 87.2
179 20.7 14.2 21.2
108 15.5 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 32.111 ng/ul
RT: 12.733 min Scan# 1639
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Instrument :
BNA_P
ClientSampleId :
SLCS568

Tgt Ion:237 Resp: 293698
Ion Ratio Lower Upper
237 100
235 64.5 53.2 79.8
272 13.8 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#41
2,4,6-Trichlorophenol
Concen: 34.944 ng/ul
RT: 12.986 min Scan# 1683
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:196 Resp: 366940
Ion Ratio Lower Upper
196 100
198 97.0 64.1 96.1#
200 30.6 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 35.431 ng/ul
RT: 13.057 min Scan# 1695
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument : BNA_P
ClientSampleId : SLCS568




Tgt Ion:196 Resp: 404809
Ion Ratio Lower Upper
196 100
198 95.0 68.6 103.0
200 30.9 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#43
1,1'-Biphenyl
Concen: 34.718 ng/ul
RT: 13.386 min Scan# 1751
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion:154 Resp: 1345166
Ion Ratio Lower Upper
154 100
153 39.7 28.6 43.0
76 13.1 8.5 12.7





#44
2-Chloronaphthalene
Concen: 34.252 ng/ul
RT: 13.427 min Scan# 1758
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:162 Resp: 102699
Ion Ratio Lower Upper
162 100
164 33.4 27.5 41.3
127 39.9 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#45
2-Nitroaniline
Concen: 35.201 ng/ul
RT: 13.627 min Scan# 1792
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion: 65 Resp: 342243
Ion Ratio Lower Upper
65 100
92 64.5 77.9 116.9#
138 95.6 81.5 122.3





#47
Dimethylphthalate
Concen: 33.942 ng/ul
RT: 14.004 min Scan# 1856
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument : BNA_P
ClientSampleId : SLCS568




Tgt Ion:163 Resp: 1348000
Ion Ratio Lower Upper
163 100
194 4.3 4.1 6.1
164 10.0 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#48
2,6-Dinitrotoluene
Concen: 35.379 ng/ul
RT: 14.127 min Scan# 1877
Delta R.T. -0.000 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion:165 Resp: 288346
Ion Ratio Lower Upper
165 100
89 56.3 51.7 77.5
121 20.4 18.3 27.5





#50
Acenaphthylene
Concen: 34.076 ng/ul
RT: 14.274 min Scan# 1901
Delta R.T. -0.000 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568



Tgt Ion:152 Resp: 166935
Ion Ratio Lower Upper
152 100
151 19.5 18.3 27.5
153 13.2 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#51
3-Nitroaniline
Concen: 37.045 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:138 Resp: 273504
Ion Ratio Lower Upper
138 100
108 11.5 15.8 23.8#
92 119.7 117.1 175.7





#52
Acenaphthene
Concen: 34.056 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:153 Resp: 108481
Ion Ratio Lower Upper
153 100
152 44.3 34.3 51.5
154 88.0 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#53
2,4-Dinitrophenol
Concen: 29.572 ng/ul
RT: 14.657 min Scan# 1967
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29



Tgt Ion:184 Resp: 134680
Ion Ratio Lower Upper
184 100
63 78.7 34.6 52.0#
154 69.0 57.7 86.5









#55
4-Nitrophenol
Concen: 28.442 ng/ul
RT: 14.763 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Instrument :
BNA_P
ClientSampleId :
SLCS568

Tgt Ion:109 Resp: 185894
Ion Ratio Lower Upper
109 100
139 125.5 61.4 92.2
65 138.8 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#56
Dibenzofuran
Concen: 33.473 ng/ul
RT: 14.951 min Scan# 2017
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Tgt Ion:168 Resp: 1578910
Ion Ratio Lower Upper
168 100
139 37.4 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 34.181 ng/ul
RT: 14.910 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Instrument :
BNA_P
ClientSampleId :
SLCS568

Tgt Ion:165 Resp: 406970
Ion Ratio Lower Upper
165 100
63 50.1 29.5 44.3#
182 3.1 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 33.476 ng/ul
RT: 15.174 min Scan# 2055
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:232 Resp: 339983
Ion Ratio Lower Upper
232 100
131 44.2 25.4 38.2#
130 2.3 4.2 6.2#
166 32.0 28.1 42.1





#59
Diethylphthalate
Concen: 34.172 ng/ul
RT: 15.369 min Scan# 2088
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568



Tgt Ion:149 Resp: 1380370
Ion Ratio Lower Upper
149 100
177 22.1 22.0 33.0
150 12.6 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#61
Fluorene
Concen: 33.103 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Tgt Ion:166 Resp: 1281987
Ion Ratio Lower Upper
166 100
165 95.9 70.3 105.5
167 13.4 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 33.311 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Instrument : BNA_P
ClientSampleId : SLCS568

Tgt Ion:204 Resp: 678938
Ion Ratio Lower Upper
204 100
206 33.7 27.6 41.4
141 59.3 46.6 69.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#63
4-Nitroaniline
Concen: 38.426 ng/ul m
RT: 15.616 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:138 Resp: 276059
Ion Ratio Lower Upper
138 100
92 56.1 41.6 62.4
108 86.1 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 33.200 ng/ul
RT: 15.680 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568



Tgt Ion:198 Resp: 218059
Ion Ratio Lower Upper
198 100
182 4.2 6.0 9.0
77 25.7 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#67
N-Nitrosodiphenylamine
Concen: 36.360 ng/ul
RT: 15.804 min Scan# 2162
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:169 Resp: 1123203
Ion Ratio Lower Upper
169 100
168 65.4 48.3 72.5
167 34.4 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 36.106 ng/ul
RT: 16.492 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:248 Resp: 414634
Ion Ratio Lower Upper
248 100
250 97.7 79.4 119.0
141 70.3 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#69
Hexachlorobenzene
Concen: 34.931 ng/ul
RT: 16.610 min Scan# 2299
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion:284 Resp: 465621
Ion Ratio Lower Upper
284 100
142 34.9 19.2 28.8#
249 33.2 27.4 41.2





#70
Atrazine
Concen: 35.187 ng/ul
RT: 16.763 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:200 Resp: 44191
Ion Ratio Lower Upper
200 100
173 26.2 19.2 28.8
215 48.7 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#71
Pentachlorophenol
Concen: 31.624 ng/ul
RT: 16.957 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Tgt Ion:266 Resp: 268457
Ion Ratio Lower Upper
266 100
264 63.8 57.3 85.9
268 63.1 65.7 98.5





#72
Phenanthrene
Concen: 34.653 ng/u1
RT: 17.345 min Scan# 2424
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:178 Resp: 2027825
Ion Ratio Lower Upper
178 100
179 15.3 11.7 17.5
176 18.5 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#74
Anthracene
Concen: 34.465 ng/u1
RT: 17.439 min Scan# 2440
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29



Tgt Ion:178 Resp: 2031417
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 18.1 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.498 ng/uL
RT: 13.351 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Instrument :
BNA_P
ClientSampleId :
SLCS568

Tgt Ion:216 Resp: 573658
Ion Ratio Lower Upper
216 100
214 77.3 50.4 75.6
218 48.3 35.2 52.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#76
Pentachlorobenzene
Concen: 35.674 ng/uL
RT: 14.874 min Scan# 2004
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:250 Resp: 549850
Ion Ratio Lower Upper
250 100
248 62.8 43.5 65.3
252 63.3 46.5 69.7





#77
 Carbazole
 Concen: 32.986 ng/ul
 RT: 17.704 min Scan# 2485
 Delta R.T. -0.006 min
 Lab File: BP010236.D
 Acq: 04 May 2022 22:29




Instrument :
 BNA_P
 ClientSampleId :
 SLCS568




Tgt Ion:167 Resp: 1720433
 Ion Ratio Lower Upper
 167 100
 166 21.1 15.8 23.8
 139 12.6 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
 Supervised By :mohammad ahmed 05/09/2022




#78
 Di-n-butylphthalate
 Concen: 36.398 ng/ul
 RT: 18.257 min Scan# 2579
 Delta R.T. -0.006 min
 Lab File: BP010236.D
 Acq: 04 May 2022 22:29




Tgt Ion:149 Resp: 2335581
 Ion Ratio Lower Upper
 149 100
 150 9.2 9.6 14.4#
 104 5.2 6.2 9.4#





#80
Fluoranthene
Concen: 46.361 ng/u1
RT: 19.310 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:202 Resp: 2226472
Ion Ratio Lower Upper
202 100
101 9.9 9.6 14.4
100 7.9 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#82
Pyrene
Concen: 44.277 ng/u1
RT: 19.662 min Scan# 2818
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29



Tgt Ion:202 Resp: 2190772
Ion Ratio Lower Upper
202 100
101 11.4 10.6 15.8
100 9.9 9.8 14.6









#83
Butylbenzylphthalate
Concen: 43.667 ng/ul
RT: 20.527 min Scan# 2965
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Instrument :
BNA_P
ClientSampleId :
SLCS568

Tgt Ion:149 Resp: 893092
Ion Ratio Lower Upper
149 100
91 75.8 71.1 106.7
206 21.2 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#84
3,3'-Dichlorobenzidine
Concen: 38.411 ng/ul
RT: 21.298 min Scan# 3096
Delta R.T. -0.012 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:252 Resp: 517444
Ion Ratio Lower Upper
252 100
254 63.8 49.8 74.6
126 10.5 4.3 6.5





#85
Benzo(a)anthracene
Concen: 36.047 ng/ul
RT: 21.368 min Scan# 3108
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568



Tgt Ion:228 Resp: 1712719
Ion Ratio Lower Upper
228 100
229 19.4 18.2 27.2
226 26.2 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 41.744 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:149 Resp: 1265383
Ion Ratio Lower Upper
149 100
167 28.0 32.8 49.2#
279 4.9 5.6 8.4#





#87
Chrysene
Concen: 35.440 ng/ul
RT: 21.421 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568



Tgt Ion:228 Resp: 1658024
Ion Ratio Lower Upper
228 100
226 28.9 22.7 34.1
229 19.5 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#89
Di-n-octyl phthalate
Concen: 37.560 ng/ul
RT: 22.227 min Scan# 3254
Delta R.T. -0.012 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:149 Resp: 1841050





#90
Benzo(b)fluoranthene
Concen: 36.357 ng/ul
RT: 23.080 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568



Tgt Ion:252 Resp: 1541094
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 9.2 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#91
Benzo(k)fluoranthene
Concen: 37.189 ng/ul
RT: 23.133 min Scan# 3408
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29


Tgt Ion:252 Resp: 1493221
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 8.7 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 37.345 ng/ul
RT: 23.721 min Scan# 3508
Delta R.T. -0.006 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29




Instrument :
BNA_P
ClientSampleId :
SLCS568




Tgt Ion:252 Resp: 146384
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 9.4 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 41.139 ng/ul
RT: 26.374 min Scan# 3959
Delta R.T. -0.024 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29



Tgt Ion:276 Resp: 1604151
Ion Ratio Lower Upper
276 100
138 20.1 12.3 18.5#
277 25.1 19.3 28.9









#95
Dibenzo(a,h)anthracene
Concen: 40.812 ng/ul
RT: 26.392 min Scan# 3962
Delta R.T. -0.018 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Instrument :
BNA_P
ClientSampleId :
SLCS568

Tgt Ion:278 Resp: 1361710
Ion Ratio Lower Upper
278 100
139 13.7 6.6 10.0
279 23.4 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :mohammad ahmed 05/09/2022



#96
Benzo(g,h,i)perylene
Concen: 41.698 ng/ul
RT: 27.156 min Scan# 4092
Delta R.T. -0.018 min
Lab File: BP010236.D
Acq: 04 May 2022 22:29

Tgt Ion:276 Resp: 1371945
Ion Ratio Lower Upper
276 100
138 19.1 11.6 17.4
277 23.6 15.9 23.9

