

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051822\
 Data File : BP010391.D
 Acq On : 18 May 2022 09:43
 Operator : CG/JU
 Sample : PB144853BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK853

Quant Time: May 19 01:56:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	138894	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	614935	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.528	164	429164	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.280	188	952517	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.363	240	958545	20.000	ng/u1	# 0.00
88) Perylene-d12	23.792	264	834117	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	18451	5.662	ng/uL	0.00
4) Pyridine-d5	3.764	84	212227	22.938	ng/u1	0.00
7) Phenol-d5	7.063	99	339686	29.138	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	236273	30.166	ng/u1	0.00
11) 2-Chlorophenol-d4	7.428	132	271859	30.476	ng/u1	0.00
15) 4-Methylphenol-d8	8.604	113	291559	29.429	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	143829	30.221	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	156561	31.041	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	276464	28.989	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	289272	20.400	ng/u1	0.00
46) Dimethylphthalate-d6	13.939	166	1002583	31.708	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	1125487	30.859	ng/u1	0.00
54) 4-Nitrophenol-d4	14.728	143	137712	24.469	ng/u1	0.00
60) Fluorene-d10	15.522	176	858027	31.192	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	125554	22.538	ng/u1	0.00
73) Anthracene-d10	17.380	188	1411218	32.719	ng/u1	0.00
81) Pyrene-d10	19.610	212	1721358	33.815	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.633	264	1390434	33.197	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051822\
Data File : BP010391.D
Acq On : 18 May 2022 09:43
Operator : CG/JU
Sample : PB144853BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK853

Quant Time: May 19 01:56:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 13 15:04:31 2022
Response via : Initial Calibration

