

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051822\
 Data File : BP010404.D
 Acq On : 18 May 2022 17:02
 Operator : CG/JU
 Sample : SSTDCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020656

Quant Time: May 19 01:59:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	153789	20.000	ng/ul	0.00
20) Naphthalene-d8	10.699	136	678117	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.528	164	460249	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.281	188	1002919	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.363	240	1020231	20.000	ng/ul	# 0.00
88) Perylene-d12	23.792	264	923258	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	29182	8.088	ng/uL	0.00
4) Pyridine-d5	3.758	84	199797	19.503	ng/ul	0.00
7) Phenol-d5	7.058	99	263620	20.423	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	178841	20.622	ng/ul	0.00
11) 2-Chlorophenol-d4	7.428	132	209025	21.163	ng/ul	0.00
15) 4-Methylphenol-d8	8.605	113	225176	20.527	ng/ul	0.00
21) Nitrobenzene-d5	9.122	128	350	0.067	ng/ul	0.06
24) 2-Nitrophenol-d4	9.775	143	121001	21.756	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	222212	21.130	ng/ul	0.00
31) 4-Chloroaniline-d4	10.828	131	318488	20.368	ng/ul	-0.01
46) Dimethylphthalate-d6	13.934	166	712914	21.024	ng/ul	0.00
49) Acenaphthylene-d8	14.222	160	829180	21.199	ng/ul	0.00
54) 4-Nitrophenol-d4	14.722	143	120028	19.887	ng/ul	0.00
60) Fluorene-d10	15.522	176	613764	20.805	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.640	200	113696	19.384	ng/ul	0.00
73) Anthracene-d10	17.381	188	963000	21.205	ng/ul	0.00
81) Pyrene-d10	19.610	212	1146413	21.159	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.633	264	989421	21.342	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.375	88	31036	8.304	ng/uL#	18
5) Pyridine	3.781	79	207597	19.499	ng/ul#	37
6) Benzaldehyde	7.034	77	165505	24.255	ng/ul	96
8) Phenol	7.087	94	280348	20.132	ng/ul#	93
10) Bis(2-Chloroethyl)ether	7.316	93	226784	20.713	ng/ul#	74
12) 2-Chlorophenol	7.458	128	215680	20.917	ng/ul	96
13) 2-Methylphenol	8.340	108	216273	20.805	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.428	45	369470	20.706	ng/ul#	83
16) Acetophenone	8.716	105	361452	20.972	ng/ul#	77
17) N-Nitroso-di-n-propyla...	8.705	70	192342	20.852	ng/ul#	70
18) 4-Methylphenol	8.669	108	238946	20.797	ng/ul	93
19) Hexachloroethane	8.975	117	92480	20.622	ng/ul#	83
22) Nitrobenzene	9.093	77	286595	20.780	ng/ul#	83
23) Isophorone	9.622	82	539602	21.168	ng/ul#	97
25) 2-Nitrophenol	9.810	139	127016	20.798	ng/ul#	86
26) 2,4-Dimethylphenol	9.869	107	273855	20.667	ng/ul#	80
27) Bis(2-Chloroethoxy)met...	10.105	93	321778	21.200	ng/ul#	98
29) 2,4-Dichlorophenol	10.346	162	225780	20.998	ng/ul#	85
30) Naphthalene	10.746	128	731697	20.658	ng/ul	97
32) 4-Chloroaniline	10.857	127	316054	20.366	ng/ul	99
33) Hexachlorobutadiene	11.040	225	155493	20.408	ng/ul	97
34) Caprolactam	11.628	113	75273	20.060	ng/ul#	1
35) 4-Chloro-3-methylphenol	11.981	107	262072	20.835	ng/ul#	71
36) 2-Methylnaphthalene	12.351	142	507865	20.254	ng/ul	89

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051822\
 Data File : BP010404.D
 Acq On : 18 May 2022 17:02
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020656

Quant Time: May 19 01:59:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.575	142	519689	20.527	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.728	216	296798	21.132	ng/ul	94
40) Hexachlorocyclopentadiene	12.704	237	148683	19.472	ng/ul	97
41) 2,4,6-Trichlorophenol	12.963	196	188883	21.574	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.040	196	208011	21.279	ng/ul#	87
43) 1,1'-Biphenyl	13.363	154	714483	21.150	ng/ul	93
44) 2-Chloronaphthalene	13.404	162	553757	21.223	ng/ul	93
45) 2-Nitroaniline	13.604	65	187071	21.784	ng/ul#	75
47) Dimethylphthalate	13.987	163	701054	20.842	ng/ul#	92
48) 2,6-Dinitrotoluene	14.104	165	144936	21.138	ng/ul	94
50) Acenaphthylene	14.245	152	899029	21.247	ng/ul	95
51) 3-Nitroaniline	14.428	138	137953	21.437	ng/ul#	83
52) Acenaphthene	14.593	153	573863	20.790	ng/ul	96
53) 2,4-Dinitrophenol	14.640	184	63394	16.560	ng/ul#	74
55) 4-Nitrophenol	14.740	109	105531	19.664	ng/ul#	46
56) Dibenzofuran	14.928	168	843076	20.841	ng/ul	99
57) 2,4-Dinitrotoluene	14.893	165	210922	21.368	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.151	232	177092	21.067	ng/ul#	85
59) Diethylphthalate	15.345	149	710219	20.693	ng/ul	93
61) Fluorene	15.575	166	687842	20.792	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.569	204	358186	20.815	ng/ul	96
63) 4-Nitroaniline	15.592	138	134830	21.219	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.657	198	114530	19.580	ng/ul#	90
67) N-Nitrosodiphenylamine	15.781	169	600018	21.333	ng/ul	95
68) 4-Bromophenyl-phenylether	16.469	248	217844	21.091	ng/ul	93
69) Hexachlorobenzene	16.587	284	246572	20.754	ng/ul#	89
70) Atrazine	16.739	200	232186	20.607	ng/ul#	90
71) Pentachlorophenol	16.934	266	143933	19.308	ng/ul#	86
72) Phenanthrene	17.322	178	1126706	21.051	ng/ul	98
74) Anthracene	17.416	178	1127450	20.997	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.328	216	315050	20.670	ng/ul#	85
76) Pentachlorobenzene	14.851	250	294171	21.272	ng/ul	92
77) Carbazole	17.681	167	984599	20.865	ng/ul#	96
78) Di-n-butylphthalate	18.234	149	1235527	21.497	ng/ul#	93
80) Fluoranthene	19.286	202	1346482	21.385	ng/ul	98
82) Pyrene	19.639	202	1389622	21.149	ng/ul	95
83) Butylbenzylphthalate	20.510	149	571958	21.769	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.275	252	415678	20.272	ng/ul#	97
85) Benzo(a)anthracene	21.345	228	1332536	20.887	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.269	149	847282	21.975	ng/ul#	81
87) Chrysene	21.398	228	1327636	21.156	ng/ul	100
89) Di-n-octyl phthalate	22.204	149	1444352	20.854	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	1314350	21.358	ng/ul	100
91) Benzo(k)fluoranthene	23.098	252	1237089	21.152	ng/ul#	97
93) Benzo(a)pyrene	23.686	252	1216572	21.278	ng/ul#	99
94) Indeno(1,2,3-cd)pyrene	26.321	276	1143753	21.264	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.333	278	971053	20.806	ng/ul#	95
96) Benzo(g,h,i)perylene	27.098	276	957893	21.328	ng/ul#	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051822\
Data File : BP010404.D
Acq On : 18 May 2022 17:02
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020656

Quant Time: May 19 01:59:06 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 13 15:04:31 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.304 ng/uL
RT: 3.375 min Scan# 49
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:	88	Resp:	31036
Ion	Ratio	Lower	Upper
88	100		
43	32.3	6.8	10.2#
58	85.7	28.4	42.6#



#5
Pyridine
Concen: 19.499 ng/uL
RT: 3.781 min Scan# 118
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:	79	Resp:	207597
Ion	Ratio	Lower	Upper
79	100		
52	95.7	35.5	53.3#
51	42.3	20.2	30.4#









#6
Benzaldehyde
Concen: 24.255 ng/ul
RT: 7.034 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion: 77 Resp: 165505
Ion Ratio Lower Upper
77 100
105 93.9 77.0 115.4
106 90.3 77.0 115.4



#8
Phenol
Concen: 20.132 ng/ul
RT: 7.087 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion: 94 Resp: 280348
Ion Ratio Lower Upper
94 100
65 31.7 19.0 28.6#
66 44.6 34.6 51.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.713 ng/ul
 RT: 7.316 min Scan# 719
 Delta R.T. -0.006 min
 Lab File: BP010404.D
 Acq: 18 May 2022 17:02




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020656


Tgt Ion: 93 Resp: 226784
 Ion Ratio Lower Upper
 93 100
 63 84.9 45.0 67.4#
 95 31.3 26.3 39.5



#12
 2-Chlorophenol
 Concen: 20.917 ng/ul
 RT: 7.458 min Scan# 743
 Delta R.T. -0.006 min
 Lab File: BP010404.D
 Acq: 18 May 2022 17:02

Tgt Ion:128 Resp: 215680
 Ion Ratio Lower Upper
 128 100
 64 55.3 45.3 67.9
 130 31.2 29.0 43.6





#13

2-Methylphenol

Concen: 20.805 ng/ul

RT: 8.340 min Scan# 893

Delta R.T. -0.006 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02

Instrument :

BNA_P

ClientSampleId :





SSTD020656

Tgt Ion:108 Resp: 216273

Ion Ratio Lower Upper

108 100

107 86.7 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.706 ng/ul

RT: 8.428 min Scan# 908

Delta R.T. 0.000 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02




Tgt Ion: 45 Resp: 369470


Ion Ratio Lower Upper

45 100

77 13.1 17.1 25.7#

79 9.7 13.2 19.8#









#16
Acetophenone
Concen: 20.972 ng/ul
RT: 8.716 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:105 Resp: 361452
Ion Ratio Lower Upper
105 100
77 82.1 53.4 80.2#
51 32.8 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 20.852 ng/ul
RT: 8.705 min Scan# 955
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion: 70 Resp: 192342
Ion Ratio Lower Upper
70 100
42 67.9 31.8 47.8#
101 9.7 15.6 23.4#
130 18.6 13.7 20.5









#18
4-Methylphenol
Concen: 20.797 ng/ul
RT: 8.669 min Scan# 949
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:108 Resp: 238946
Ion Ratio Lower Upper
108 100
107 112.0 83.6 125.4



#19
Hexachloroethane
Concen: 20.622 ng/ul
RT: 8.975 min Scan# 1001
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:117 Resp: 92480
Ion Ratio Lower Upper
117 100
201 106.4 106.5 159.7#
199 66.8 59.0 88.6





#22

Nitrobenzene

Concen: 20.780 ng/ul

RT: 9.093 min Scan# 1109

Delta R.T. -0.006 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02

Instrument :

BNA_P

ClientSampleId :

SSTD020656





Tgt Ion: 77 Resp: 286595

Ion Ratio Lower Upper

77 100

123 40.9 42.3 63.5#

65 14.4 5.8 8.8#



#23

Isophorone

Concen: 21.168 ng/ul

RT: 9.622 min Scan# 1111

Delta R.T. -0.006 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02




Tgt Ion: 82 Resp: 539602


Ion Ratio Lower Upper

82 100

95 7.4 7.8 11.8#

138 16.5 12.9 19.3









#25
2-Nitrophenol
Concen: 20.798 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:139 Resp: 127016
Ion Ratio Lower Upper
139 100
65 61.4 49.4 74.2
109 34.9 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 20.667 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:107 Resp: 273855
Ion Ratio Lower Upper
107 100
121 47.2 34.1 51.1
122 79.2 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.200 ng/ul
 RT: 10.105 min Scan# 1193
 Delta R.T. -0.006 min
 Lab File: BP010404.D
 Acq: 18 May 2022 17:02




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020656


Tgt Ion: 93 Resp: 321778
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.2 39.4
 123 11.0 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 20.998 ng/ul
 RT: 10.346 min Scan# 1234
 Delta R.T. -0.006 min
 Lab File: BP010404.D
 Acq: 18 May 2022 17:02

Tgt Ion:162 Resp: 225780
 Ion Ratio Lower Upper
 162 100
 164 63.5 37.0 55.6#
 98 37.9 31.0 46.6









#30
Naphthalene
Concen: 20.658 ng/ul
RT: 10.746 min Scan# 1302
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:128 Resp: 731697
Ion Ratio Lower Upper
128 100
129 11.3 8.6 13.0
127 13.8 12.2 18.4



#32
4-Chloroaniline
Concen: 20.366 ng/ul
RT: 10.857 min Scan# 1321
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:127 Resp: 316054
Ion Ratio Lower Upper
127 100
129 31.2 24.6 37.0









#33
Hexachlorobutadiene
Concen: 20.408 ng/ul
RT: 11.040 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:225 Resp: 155493
Ion Ratio Lower Upper
225 100
223 63.7 53.4 80.0
227 64.8 53.8 80.8



#34
Caprolactam
Concen: 20.060 ng/ul
RT: 11.628 min Scan# 1452
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:113 Resp: 75273
Ion Ratio Lower Upper
113 100
55 198.6 58.2 87.2#
56 150.0 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 20.835 ng/ul
RT: 11.981 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:107 Resp: 262072
Ion Ratio Lower Upper
107 100
144 22.4 15.8 23.8
142 74.5 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 20.254 ng/ul
RT: 12.351 min Scan# 1575
Delta R.T. -0.012 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:142 Resp: 507865
Ion Ratio Lower Upper
142 100
141 90.0 64.3 96.5









#37
1-Methylnaphthalene
Concen: 20.527 ng/ul
RT: 12.575 min Scan# 1613
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:142 Resp: 519689
Ion Ratio Lower Upper
142 100
141 92.9 69.8 104.8
116 4.0 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.132 ng/ul
RT: 12.728 min Scan# 1639
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:216 Resp: 296798
Ion Ratio Lower Upper
216 100
214 76.4 58.2 87.2
179 21.2 14.2 21.2
108 15.8 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 19.472 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:237 Resp: 148683
Ion Ratio Lower Upper
237 100
235 63.9 53.2 79.8
272 12.1 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 21.574 ng/ul
RT: 12.963 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:196 Resp: 188883
Ion Ratio Lower Upper
196 100
198 96.8 64.1 96.1#
200 31.4 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 21.279 ng/ul
RT: 13.040 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:196 Resp: 208011
Ion Ratio Lower Upper
196 100
198 95.5 68.6 103.0
200 29.6 32.8 49.2#



#43
1,1'-Biphenyl
Concen: 21.150 ng/ul
RT: 13.363 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:154 Resp: 714483
Ion Ratio Lower Upper
154 100
153 40.2 28.6 43.0
76 12.5 8.5 12.7









#44
2-Chloronaphthalene
Concen: 21.223 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:162 Resp: 553757
Ion Ratio Lower Upper
162 100
164 32.9 27.5 41.3
127 39.4 37.1 55.7



#45
2-Nitroaniline
Concen: 21.784 ng/ul
RT: 13.604 min Scan# 1788
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion: 65 Resp: 187071
Ion Ratio Lower Upper
65 100
92 62.5 77.9 116.9#
138 86.6 81.5 122.3









#47
Dimethylphthalate
Concen: 20.842 ng/ul
RT: 13.987 min Scan# 1853
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:163 Resp: 701054
Ion Ratio Lower Upper
163 100
194 4.6 4.1 6.1
164 10.1 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 21.138 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:165 Resp: 144936
Ion Ratio Lower Upper
165 100
89 58.7 51.7 77.5
121 22.5 18.3 27.5









#50
Acenaphthylene
Concen: 21.247 ng/ul
RT: 14.245 min Scan# 1898
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:152 Resp: 899029
Ion Ratio Lower Upper
152 100
151 19.4 18.3 27.5
153 13.3 11.0 16.6



#51
3-Nitroaniline
Concen: 21.437 ng/ul
RT: 14.428 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:138 Resp: 137953
Ion Ratio Lower Upper
138 100
108 12.3 15.8 23.8#
92 125.6 117.1 175.7









#52
Acenaphthene
Concen: 20.790 ng/ul
RT: 14.593 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:153 Resp: 573863
Ion Ratio Lower Upper
153 100
152 46.4 34.3 51.5
154 90.7 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 16.560 ng/ul
RT: 14.640 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:184 Resp: 63394
Ion Ratio Lower Upper
184 100
63 80.0 34.6 52.0#
154 66.3 57.7 86.5









#55
4-Nitrophenol
Concen: 19.664 ng/ul
RT: 14.740 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:109 Resp: 105531
Ion Ratio Lower Upper
109 100
139 120.3 61.4 92.2#
65 133.1 65.3 97.9#



#56
Dibenzofuran
Concen: 20.841 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:168 Resp: 843076
Ion Ratio Lower Upper
168 100
139 37.0 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 21.368 ng/ul
RT: 14.893 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:165 Resp: 210922
Ion Ratio Lower Upper
165 100
63 50.6 29.5 44.3#
182 3.1 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.067 ng/ul
RT: 15.151 min Scan# 2051
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:232 Resp: 177092
Ion Ratio Lower Upper
232 100
131 47.2 25.4 38.2#
130 2.4 4.2 6.2#
166 32.1 28.1 42.1









#59
Diethylphthalate
Concen: 20.693 ng/ul
RT: 15.345 min Scan# 2085
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:149 Resp: 710219
Ion Ratio Lower Upper
149 100
177 22.2 22.0 33.0
150 13.1 10.2 15.2



#61
Fluorene
Concen: 20.792 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:166 Resp: 687842
Ion Ratio Lower Upper
166 100
165 95.6 70.3 105.5
167 13.6 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 20.815 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:204 Resp: 358186
Ion Ratio Lower Upper
204 100
206 34.1 27.6 41.4
141 62.2 46.6 69.8



#63
4-Nitroaniline
Concen: 21.219 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:138 Resp: 134830
Ion Ratio Lower Upper
138 100
92 56.6 41.6 62.4
108 88.2 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 19.580 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:198 Resp: 114530
Ion Ratio Lower Upper
198 100
182 4.4 6.0 9.0#
77 26.9 17.7 26.5#



#67
N-Nitrosodiphenylamine
Concen: 21.333 ng/ul
RT: 15.781 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:169 Resp: 600018
Ion Ratio Lower Upper
169 100
168 64.0 48.3 72.5
167 34.3 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 21.091 ng/ul
RT: 16.469 min Scan# 217844
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:248 Resp: 217844
Ion Ratio Lower Upper
248 100
250 98.2 79.4 119.0
141 71.0 67.2 100.8



#69
Hexachlorobenzene
Concen: 20.754 ng/ul
RT: 16.587 min Scan# 2295
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:284 Resp: 246572
Ion Ratio Lower Upper
284 100
142 35.2 19.2 28.8#
249 33.2 27.4 41.2









#70
Atrazine
Concen: 20.607 ng/ul
RT: 16.739 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:200 Resp: 232186
Ion Ratio Lower Upper
200 100
173 26.4 19.2 28.8
215 48.9 32.2 48.4#



#71
Pentachlorophenol
Concen: 19.308 ng/ul
RT: 16.934 min Scan# 2354
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:266 Resp: 143933
Ion Ratio Lower Upper
266 100
264 64.4 57.3 85.9
268 65.4 65.7 98.5#









#72
Phenanthrene
Concen: 21.051 ng/ul
RT: 17.322 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:178 Resp: 1126706
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 19.1 14.5 21.7



#74
Anthracene
Concen: 20.997 ng/ul
RT: 17.416 min Scan# 2436
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:178 Resp: 1127450
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 18.1 13.4 20.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.670 ng/uL

RT: 13.328 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02

Instrument :

BNA_P

ClientSampleId :

SSTD020656





Tgt Ion:216 Resp: 315050

Ion Ratio Lower Upper

216 100

214 79.2 50.4 75.6#

218 48.3 35.2 52.8



#76

Pentachlorobenzene

Concen: 21.272 ng/uL

RT: 14.851 min Scan# 2000

Delta R.T. -0.006 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02




Tgt Ion:250 Resp: 294171


Ion Ratio Lower Upper

250 100

248 62.2 43.5 65.3

252 62.5 46.5 69.7





#77

Carbazole

Concen: 20.865 ng/ul

RT: 17.681 min Scan# 2482

Delta R.T. -0.006 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02

Instrument :

BNA_P

ClientSampleId :

SSTD020656





Tgt Ion:167 Resp: 984599

Ion Ratio Lower Upper

167 100

166 21.0 15.8 23.8

139 13.2 8.3 12.5#



#78

Di-n-butylphthalate

Concen: 21.497 ng/ul

RT: 18.234 min Scan# 2575

Delta R.T. -0.006 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02




Tgt Ion:149 Resp: 1235527


Ion Ratio Lower Upper

149 100

150 9.0 9.6 14.4#

104 5.5 6.2 9.4#









#80
Fluoranthene
Concen: 21.385 ng/ul
RT: 19.286 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:202 Resp: 1346482
Ion Ratio Lower Upper
202 100
101 10.6 9.6 14.4
100 8.1 6.3 9.5



#82
Pyrene
Concen: 21.149 ng/ul
RT: 19.639 min Scan# 2814
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:202 Resp: 1389622
Ion Ratio Lower Upper
202 100
101 11.8 10.6 15.8
100 10.0 9.8 14.6





#83

Butylbenzylphthalate

Concen: 21.769 ng/ul

RT: 20.510 min Scan# 2962

Delta R.T. 0.000 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02

Instrument :

BNA_P

ClientSampleId :

SSTD020656





Tgt Ion:149 Resp: 571958

Ion Ratio Lower Upper

149 100

91 73.4 71.1 106.7

206 22.6 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 20.272 ng/ul

RT: 21.275 min Scan# 3092

Delta R.T. -0.006 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02




Tgt Ion:252 Resp: 415678


Ion Ratio Lower Upper

252 100

254 64.0 49.8 74.6

126 10.8 4.3 6.5#









#85
Benzo(a)anthracene
Concen: 20.887 ng/ul
RT: 21.345 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:228 Resp: 1332536
Ion Ratio Lower Upper
228 100
229 19.7 18.2 27.2
226 26.2 22.8 34.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.975 ng/ul
RT: 21.269 min Scan# 3091
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:149 Resp: 847282
Ion Ratio Lower Upper
149 100
167 27.8 32.8 49.2#
279 4.2 5.6 8.4#





#87

Chrysene

Concen: 21.156 ng/ul

RT: 21.398 min Scan# 3114

Delta R.T. 0.000 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02

Instrument :

BNA_P

ClientSampleId :

SSTD020656





Tgt Ion:228 Resp: 1327636

Ion Ratio Lower Upper

228 100

226 28.4 22.7 34.1

229 19.2 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 20.854 ng/ul




RT: 22.204 min Scan# 3250


Delta R.T. 0.000 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02

Tgt Ion:149 Resp: 1444352









#90
Benzo(b)fluoranthene
Concen: 21.358 ng/ul
RT: 23.051 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:252 Resp: 1314350
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 8.9 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 21.152 ng/ul
RT: 23.098 min Scan# 3402
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:252 Resp: 1237089
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 9.1 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 21.278 ng/ul
RT: 23.686 min Scan# 3502
Delta R.T. 0.006 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02




Instrument :
BNA_P
ClientSampleId :
SSTD020656


Tgt Ion:252 Resp: 1216572
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 9.9 6.4 9.6#



#94
Indeno(1,2,3-cd)pyrene
Concen: 21.264 ng/ul
RT: 26.321 min Scan# 3950
Delta R.T. 0.000 min
Lab File: BP010404.D
Acq: 18 May 2022 17:02

Tgt Ion:276 Resp: 1143753
Ion Ratio Lower Upper
276 100
138 20.8 12.3 18.5#
277 24.9 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 20.806 ng/ul

RT: 26.333 min Scan# 3953

Delta R.T. 0.000 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02

Instrument :

BNA_P

ClientSampleId :

SSTD020656





Tgt Ion:278 Resp: 971053

Ion Ratio Lower Upper

278 100

139 14.2 6.6 10.0#

279 23.6 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 21.328 ng/ul

RT: 27.098 min Scan# 4082

Delta R.T. 0.006 min

Lab File: BP010404.D

Acq: 18 May 2022 17:02

Tgt Ion:276 Resp: 957893

Ion Ratio Lower Upper

276 100

138 18.0 11.6 17.4#

277 23.5 15.9 23.9

