



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052721\
 Data File : BP005623.D
 Acq On : 27 May 2021 10:27
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu May 27 16:11:11 2021






AutoFind: Scans 2500, 2501, 2502; Background Corrected with Scan 2490

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.3	131900	PASS
68	69	0.00	2	1.5	2381	PASS
69	198	0.00	100	39.4	156085	PASS
70	69	0.00	2	0.4	567	PASS
127	198	10	80	47.3	187123	PASS
197	198	0.00	2	0.4	1596	PASS
198	198	100	100	100.0	395733	PASS
199	198	5	9	6.8	26915	PASS
275	198	10	60	27.4	108344	PASS
365	198	1	100	4.1	16103	PASS
441	198	0.01	100	18.1	71555	PASS
442	442	100	100	100.0	475776	PASS
443	442	15	24	19.9	94469	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052721\
 Data File : BP005623.D
 Acq On : 27 May 2021 10:27
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: May 27 18:31:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 16:11:11 2021
 Response via : Initial Calibration



TIC: BP005623.D\data.ms

(77) Benzidine

19.533min (+ 0.006) 1191408.19 ng




response 3280827

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	14.27
183.00	12.30	11.85
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052721\
Data File : BP005623.D
Acq On : 27 May 2021 10:27
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 27 18:31:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 16:11:11 2021
Response via : Initial Calibration



TIC: BP005623.D\data.ms

(70) Pentachlorophenol (C)

17.004min (+ 0.006) 3859008.33 ng

response 416429

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.70	64.51
264.00	67.10	62.99
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/27/2021	BNA_P	<u>BP005623.D</u>
Compound Name	Response	Retention Time
DDT	1452968	20.73
DDD	3570	20.73
DDE	2237	19.81
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5807	1458775	0.40