

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP053023\
 Data File : BP015332.D
 Acq On : 30 May 2023 13:40
 Operator : CG/JU
 Sample : PB153103BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK103

Quant Time: May 31 00:44:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 30 15:23:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	174703	20.000	ng/ul	0.00
20) Naphthalene-d8	10.946	136	754016	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.757	164	522346	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.516	188	1233915	20.000	ng/ul	0.00
79) Chrysene-d12	21.592	240	1177131	20.000	ng/ul	0.00
88) Perylene-d12	24.157	264	1342295	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	27584	6.135	ng/uL	0.00
4) Pyridine-d5	3.858	84	360762	28.305	ng/ul	0.00
7) Phenol-d5	7.269	99	455468	28.079	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.434	67	286515	30.583	ng/ul	0.00
11) 2-Chlorophenol-d4	7.640	132	368009	30.857	ng/ul	0.00
15) 4-Methylphenol-d8	8.828	113	384400	29.740	ng/ul	0.00
21) Nitrobenzene-d5	9.293	128	195882	33.060	ng/ul	0.00
24) 2-Nitrophenol-d4	10.022	143	228730	34.695	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	366898	30.841	ng/ul	0.00
31) 4-Chloroaniline-d4	11.081	131	564248	32.057	ng/ul	0.00
46) Dimethylphthalate-d6	14.163	166	1414199	35.828	ng/ul	0.00
49) Acenaphthylene-d8	14.457	160	1495517	33.128	ng/ul	0.00
54) 4-Nitrophenol-d4	14.963	143	228147	29.767	ng/ul	0.00
60) Fluorene-d10	15.751	176	1140518	34.243	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.881	200	260564	33.508	ng/ul	0.00
73) Anthracene-d10	17.616	188	1905583	33.726	ng/ul	0.00
81) Pyrene-d10	19.839	212	2308378	33.812	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.992	264	2289312	34.357	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP053023\
Data File : BP015332.D
Acq On : 30 May 2023 13:40
Operator : CG/JU
Sample : PB153103BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK103

Quant Time: May 31 00:44:30 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 30 15:23:22 2023
Response via : Initial Calibration

