

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP053023\
 Data File : BPCCC345.D
 Acq On : 30 May 2023 21:03
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732

Quant Time: May 31 00:49:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 30 15:23:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.110	152	152467	20.000	ng/u1	0.00
20) Naphthalene-d8	10.940	136	682289	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.751	164	471756	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.510	188	1106876	20.000	ng/u1	0.00
79) Chrysene-d12	21.586	240	1104426	20.000	ng/u1	-0.01
88) Perylene-d12	24.145	264	1298535	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	31833	8.112	ng/uL	0.00
4) Pyridine-d5	3.858	84	217459	19.550	ng/u1	0.00
7) Phenol-d5	7.263	99	289417	20.445	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.428	67	168873	20.655	ng/u1	0.00
11) 2-Chlorophenol-d4	7.634	132	215902	20.743	ng/u1	0.00
15) 4-Methylphenol-d8	8.828	113	237260	21.033	ng/u1	0.00
21) Nitrobenzene-d5	9.287	128	116306	21.693	ng/u1	0.00
24) 2-Nitrophenol-d4	10.016	143	135504	22.715	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.557	165	240145	22.309	ng/u1	0.00
31) 4-Chloroaniline-d4	11.075	131	355310	22.308	ng/u1	0.00
46) Dimethylphthalate-d6	14.157	166	788002	22.105	ng/u1	0.00
49) Acenaphthylene-d8	14.445	160	880283	21.591	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.951	143	148026	21.384	ng/u1	-0.01
60) Fluorene-d10	15.745	176	678485	22.556	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.869	200	155057	22.228	ng/u1	-0.01
73) Anthracene-d10	17.604	188	1095181	21.608	ng/u1	-0.01
81) Pyrene-d10	19.828	212	1295598	20.226	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.974	264	1361024	21.114	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	33457	7.935	ng/uL#	95
5) Pyridine	3.875	79	230941	19.963	ng/u1	91
6) Benzaldehyde	7.240	77	107822	27.711	ng/u1	97
8) Phenol	7.293	94	295669	20.368	ng/u1#	88
10) Bis(2-Chloroethyl)ether	7.522	93	234424	20.185	ng/u1	97
12) 2-Chlorophenol	7.669	128	228926	20.794	ng/u1	98
13) 2-Methylphenol	8.558	108	231290	20.616	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.634	45	306647	21.075	ng/u1	98
16) Acetophenone	8.946	105	392993	22.322	ng/u1	93
17) N-Nitroso-di-n-propyla...	8.922	70	206545	22.370	ng/u1	93
18) 4-Methylphenol	8.887	108	256502	20.772	ng/u1	97
19) Hexachloroethane	9.199	117	96934	21.564	ng/u1	87
22) Nitrobenzene	9.328	77	299621	22.627	ng/u1	98
23) Isophorone	9.857	82	601972	22.117	ng/u1	99
25) 2-Nitrophenol	10.046	139	146842	22.447	ng/u1#	87
26) 2,4-Dimethylphenol	10.105	107	296550	22.717	ng/u1	94
27) Bis(2-Chloroethoxy)met...	10.340	93	341304	20.610	ng/u1	98
29) 2,4-Dichlorophenol	10.587	162	246670	22.686	ng/u1	99
30) Naphthalene	10.993	128	788279	21.251	ng/u1	100
32) 4-Chloroaniline	11.104	127	370441	22.630	ng/u1	97
33) Hexachlorobutadiene	11.269	225	161455	24.637	ng/u1	98
34) Caprolactam	11.881	113	94276	22.997	ng/u1	88
35) 4-Chloro-3-methylphenol	12.216	107	287125	23.386	ng/u1	98
36) 2-Methylnaphthalene	12.587	142	550052	22.008	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP053023\
 Data File : BPCCC345.D
 Acq On : 30 May 2023 21:03
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732


Quant Time: May 31 00:49:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 30 15:23:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.804	142	554373	22.088	ng/ul#	97
39) 1,2,4,5-Tetrachloroben...	12.951	216	313071	22.189	ng/ul	98
40) Hexachlorocyclopentadiene	12.928	237	154781	16.932	ng/ul	98
41) 2,4,6-Trichlorophenol	13.193	196	217691	21.850	ng/ul	98
42) 2,4,5-Trichlorophenol	13.269	196	234180	21.679	ng/ul	99
43) 1,1'-Biphenyl	13.587	154	774076	21.148	ng/ul	100
44) 2-Chloronaphthalene	13.634	162	611116	21.379	ng/ul	99
45) 2-Nitroaniline	13.840	65	201412	24.626	ng/ul	90
47) Dimethylphthalate	14.204	163	823911	22.192	ng/ul	99
48) 2,6-Dinitrotoluene	14.328	165	174524	23.728	ng/ul	95
50) Acenaphthylene	14.475	152	971584	21.357	ng/ul	99
51) 3-Nitroaniline	14.657	138	143872	21.723	ng/ul	98
52) Acenaphthene	14.816	153	663660	21.360	ng/ul	99
53) 2,4-Dinitrophenol	14.875	184	103737	21.242	ng/ul	87
55) 4-Nitrophenol	14.969	109	138076	27.120	ng/ul#	81
56) Dibenzofuran	15.151	168	949722	21.976	ng/ul	98
57) 2,4-Dinitrotoluene	15.122	165	256311	24.082	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.381	232	214500	23.528	ng/ul	95
59) Diethylphthalate	15.557	149	841814	22.734	ng/ul	99
61) Fluorene	15.804	166	771025	22.201	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.787	204	390763	23.028	ng/ul	95
63) 4-Nitroaniline	15.822	138	135846	23.752	ng/ul#	86
66) 4,6-Dinitro-2-methylph...	15.887	198	155620	21.616	ng/ul	99
67) N-Nitrosodiphenylamine	16.004	169	680105	20.896	ng/ul	99
68) 4-Bromophenyl-phenylether	16.686	248	260879	22.882	ng/ul	99
69) Hexachlorobenzene	16.810	284	310739	23.064	ng/ul	97
70) Atrazine	16.957	200	276650	22.997	ng/ul	99
71) Pentachlorophenol	17.157	266	185848	22.052	ng/ul	99
72) Phenanthrene	17.551	178	1280332	21.242	ng/ul	99
74) Anthracene	17.639	178	1306627	21.438	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.557	216	325284	20.727	ng/uL	99
76) Pentachlorobenzene	15.069	250	344397	21.856	ng/uL	99
77) Carbazole	17.910	167	1192868	22.319	ng/ul	100
78) Di-n-butylphthalate	18.433	149	1493214	21.695	ng/ul	99
80) Fluoranthene	19.504	202	1573001	20.148	ng/ul	96
82) Pyrene	19.857	202	1632095	20.244	ng/ul	95
83) Butylbenzylphthalate	20.710	149	730517	20.439	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.492	252	585965	24.158	ng/ul	98
85) Benzo(a)anthracene	21.569	228	1632565	21.378	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.463	149	1071185	20.817	ng/ul	99
87) Chrysene	21.627	228	1555845	21.313	ng/ul	98
89) Di-n-octyl phthalate	22.433	149	1862906	20.947	ng/ul	100
90) Benzo(b)fluoranthene	23.357	252	1738274	21.116	ng/ul	98
91) Benzo(k)fluoranthene	23.357	252	1738274	21.991	ng/ul	98
93) Benzo(a)pyrene	24.033	252	1509767	20.749	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.857	276	1859996	19.742	ng/ul	95
95) Dibenzo(a,h)anthracene	26.868	278	1543523	19.958	ng/ul	97
96) Benzo(g,h,i)perylene	27.692	276	1471498	19.154	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 31 00:49:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 30 15:23:22 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.935 ng/uL
RT: 3.452 min Scan# 41
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 88 Resp: 33457
Ion Ratio Lower Upper
88 100
43 34.1 20.4 30.6#
58 70.3 55.4 83.2



#5
Pyridine
Concen: 19.963 ng/uL
RT: 3.875 min Scan# 117
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion: 79 Resp: 230941
Ion Ratio Lower Upper
79 100
52 76.8 55.4 83.2
51 37.4 25.4 38.2









#6
Benzaldehyde
Concen: 27.711 ng/ul
RT: 7.240 min Scan# 690
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 77 Resp: 107822
Ion Ratio Lower Upper
77 100
105 93.1 76.9 115.3
106 91.7 75.0 112.6



#8
Phenol
Concen: 20.368 ng/ul
RT: 7.293 min Scan# 698
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion: 94 Resp: 295669
Ion Ratio Lower Upper
94 100
65 29.7 20.6 31.0
66 41.6 26.1 39.1#









#10
Bis(2-Chloroethyl)ether
Concen: 20.185 ng/ul
RT: 7.522 min Scan# 71
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 93 Resp: 234424
Ion Ratio Lower Upper
93 100
63 74.6 58.0 87.0
95 30.8 26.4 39.6



#12
2-Chlorophenol
Concen: 20.794 ng/ul
RT: 7.669 min Scan# 762
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion: 128 Resp: 228926
Ion Ratio Lower Upper
128 100
64 47.1 36.2 54.2
130 31.9 26.3 39.5





#13

2-Methylphenol

Concen: 20.616 ng/ul

RT: 8.558 min Scan# 91

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :





SSTD020732

Tgt Ion:108 Resp: 231290

Ion Ratio Lower Upper

108 100

107 89.4 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.075 ng/ul

RT: 8.634 min Scan# 926

Delta R.T. -0.018 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion: 45 Resp: 306647


Ion Ratio Lower Upper

45 100

77 15.4 12.8 19.2

79 10.7 9.5 14.3









#16
Acetophenone
Concen: 22.322 ng/ul
RT: 8.946 min Scan# 91
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:105 Resp: 392993
Ion Ratio Lower Upper
105 100
77 80.1 59.4 89.2
51 27.6 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 22.370 ng/ul
RT: 8.922 min Scan# 975
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion: 70 Resp: 206545
Ion Ratio Lower Upper
70 100
42 58.7 41.0 61.6
101 9.4 7.9 11.9
130 21.0 16.1 24.1









#18
4-Methylphenol
Concen: 20.772 ng/ul
RT: 8.887 min Scan# 90
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:108 Resp: 256502
Ion Ratio Lower Upper
108 100
107 114.0 88.5 132.7



#19
Hexachloroethane
Concen: 21.564 ng/ul
RT: 9.199 min Scan# 1022
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:117 Resp: 96934
Ion Ratio Lower Upper
117 100
201 108.2 75.8 113.8
199 66.4 46.5 69.7





#22

Nitrobenzene

Concen: 22.627 ng/ul

RT: 9.328 min Scan# 1044

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion: 77 Resp: 299621

Ion Ratio Lower Upper

77 100

123 43.5 36.2 54.2

65 13.1 10.1 15.1



#23

Isophorone

Concen: 22.117 ng/ul

RT: 9.857 min Scan# 1134

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion: 82 Resp: 601972


Ion Ratio Lower Upper

82 100

95 7.8 5.4 8.0

138 17.0 13.6 20.4









#25
2-Nitrophenol
Concen: 22.447 ng/ul
RT: 10.046 min Scan# 1166
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:139 Resp: 146842
Ion Ratio Lower Upper
139 100
65 56.5 38.4 57.6
109 35.1 22.2 33.2#



#26
2,4-Dimethylphenol
Concen: 22.717 ng/ul
RT: 10.105 min Scan# 1176
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:107 Resp: 296550
Ion Ratio Lower Upper
107 100
121 47.6 41.0 61.6
122 80.6 69.5 104.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.610 ng/ul
 RT: 10.340 min Scan# 1217
 Delta R.T. -0.006 min
 Lab File: BPCCC345.D
 Acq: 30 May 2023 21:03




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732


Tgt Ion: 93 Resp: 341304
 Ion Ratio Lower Upper
 93 100
 95 31.9 26.7 40.1
 123 11.4 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 22.686 ng/ul
 RT: 10.587 min Scan# 1258
 Delta R.T. -0.006 min
 Lab File: BPCCC345.D
 Acq: 30 May 2023 21:03

Tgt Ion: 162 Resp: 246670
 Ion Ratio Lower Upper
 162 100
 164 64.4 52.0 78.0
 98 35.0 28.9 43.3









#30
Naphthalene
Concen: 21.251 ng/ul
RT: 10.993 min Scan# 1327
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:128 Resp: 788279
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 13.3 10.6 15.8



#32
4-Chloroaniline
Concen: 22.630 ng/ul
RT: 11.104 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:127 Resp: 370441
Ion Ratio Lower Upper
127 100
129 30.4 25.8 38.8





#33

Hexachlorobutadiene

Concen: 24.637 ng/ul

RT: 11.269 min Scan# 11

Delta R.T. 0.000 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:225 Resp: 161455

Ion Ratio Lower Upper

225 100

223 62.1 51.0 76.6

227 63.9 52.2 78.4



#34

Caprolactam

Concen: 22.997 ng/ul

RT: 11.881 min Scan# 1478

Delta R.T. -0.012 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion:113 Resp: 94276


Ion Ratio Lower Upper

113 100

55 163.6 121.2 181.8

56 128.4 90.2 135.4









#35
4-Chloro-3-methylphenol
Concen: 23.386 ng/ul
RT: 12.216 min Scan# 1535
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:107 Resp: 287125
Ion Ratio Lower Upper
107 100
144 24.8 20.3 30.5
142 77.4 63.2 94.8



#36
2-Methylnaphthalene
Concen: 22.008 ng/ul
RT: 12.587 min Scan# 1598
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:142 Resp: 550052
Ion Ratio Lower Upper
142 100
141 89.3 71.0 106.6









#37
1-Methylnaphthalene
Concen: 22.088 ng/ul
RT: 12.804 min Scan# 1635
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:142 Resp: 554373
Ion Ratio Lower Upper
142 100
141 90.7 74.6 112.0
116 4.5 2.9 4.3#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.189 ng/ul
RT: 12.951 min Scan# 1660
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:216 Resp: 313071
Ion Ratio Lower Upper
216 100
214 78.5 61.4 92.0
179 21.2 15.9 23.9
108 18.0 13.7 20.5





#40

Hexachlorocyclopentadiene

Concen: 16.932 ng/ul

RT: 12.928 min Scan# 1

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:237 Resp: 154781

Ion Ratio Lower Upper

237 100

235 63.2 49.4 74.0

272 13.1 10.4 15.6



#41

2,4,6-Trichlorophenol

Concen: 21.850 ng/ul

RT: 13.193 min Scan# 1701

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion:196 Resp: 217691


Ion Ratio Lower Upper

196 100

198 92.2 75.5 113.3

200 29.2 23.6 35.4









#42
2,4,5-Trichlorophenol
Concen: 21.679 ng/ul
RT: 13.269 min Scan# 1714
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:196 Resp: 234180
Ion Ratio Lower Upper
196 100
198 97.8 77.0 115.4
200 29.9 24.3 36.5



#43
1,1'-Biphenyl
Concen: 21.148 ng/ul
RT: 13.587 min Scan# 1768
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:154 Resp: 774076
Ion Ratio Lower Upper
154 100
153 40.6 32.7 49.1
76 13.5 10.7 16.1





#44

2-Chloronaphthalene

Concen: 21.379 ng/ul

RT: 13.634 min Scan# 11

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:162 Resp: 611116

Ion Ratio Lower Upper

162 100

164 33.1 25.9 38.9

127 39.4 30.8 46.2



#45

2-Nitroaniline

Concen: 24.626 ng/ul

RT: 13.840 min Scan# 1811

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion: 65 Resp: 201412


Ion Ratio Lower Upper

65 100

92 70.0 59.8 89.8

138 100.6 92.3 138.5









#47
Dimethylphthalate
Concen: 22.192 ng/ul
RT: 14.204 min Scan# 1873
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:163 Resp: 823911
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.0 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 23.728 ng/ul
RT: 14.328 min Scan# 1894
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:165 Resp: 174524
Ion Ratio Lower Upper
165 100
89 64.7 48.0 72.0
121 22.2 17.5 26.3









#50
Acenaphthylene
Concen: 21.357 ng/ul
RT: 14.475 min Scan# 1919
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:152 Resp: 971584
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 13.3 10.3 15.5



#51
3-Nitroaniline
Concen: 21.723 ng/ul
RT: 14.657 min Scan# 1950
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:138 Resp: 143872
Ion Ratio Lower Upper
138 100
108 11.6 8.8 13.2
92 119.4 93.4 140.0









#52
Acenaphthene
Concen: 21.360 ng/ul
RT: 14.816 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:153 Resp: 663660
Ion Ratio Lower Upper
153 100
152 46.6 38.0 57.0
154 89.0 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 21.242 ng/ul
RT: 14.875 min Scan# 1987
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:184 Resp: 103737
Ion Ratio Lower Upper
184 100
63 70.1 49.7 74.5
154 71.2 47.5 71.3









#55
4-Nitrophenol
Concen: 27.120 ng/ul
RT: 14.969 min Scan# 2003
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:109 Resp: 138076
Ion Ratio Lower Upper
109 100
139 113.0 115.3 172.9#
65 119.0 107.9 161.9



#56
Dibenzofuran
Concen: 21.976 ng/ul
RT: 15.151 min Scan# 2034
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:168 Resp: 949722
Ion Ratio Lower Upper
168 100
139 37.7 29.4 44.0









#57
2,4-Dinitrotoluene
Concen: 24.082 ng/ul
RT: 15.122 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:165 Resp: 256311
Ion Ratio Lower Upper
165 100
63 44.9 33.1 49.7
182 2.5 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.528 ng/ul
RT: 15.381 min Scan# 2073
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:232 Resp: 214500
Ion Ratio Lower Upper
232 100
131 41.1 36.8 55.2
130 2.1 1.9 2.9
166 29.8 24.6 37.0









#59
Diethylphthalate
Concen: 22.734 ng/ul
RT: 15.557 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:149 Resp: 841814
Ion Ratio Lower Upper
149 100
177 22.3 17.5 26.3
150 12.6 9.8 14.8



#61
Fluorene
Concen: 22.201 ng/ul
RT: 15.804 min Scan# 2145
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:166 Resp: 771025
Ion Ratio Lower Upper
166 100
165 97.8 77.8 116.6
167 12.9 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 23.028 ng/ul
RT: 15.787 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:204 Resp: 390763
Ion Ratio Lower Upper
204 100
206 32.7 26.1 39.1
141 61.0 44.5 66.7



#63
4-Nitroaniline
Concen: 23.752 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:138 Resp: 135846
Ion Ratio Lower Upper
138 100
92 53.6 41.2 61.8
108 92.9 60.1 90.1#









#66
4,6-Dinitro-2-methylphenol
Concen: 21.616 ng/ul
RT: 15.887 min Scan# 2161
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:198 Resp: 155620
Ion Ratio Lower Upper
198 100
182 4.2 3.7 5.5
77 25.9 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 20.896 ng/ul
RT: 16.004 min Scan# 2179
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:169 Resp: 680105
Ion Ratio Lower Upper
169 100
168 66.7 53.5 80.3
167 36.7 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 22.882 ng/ul
RT: 16.686 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:248 Resp: 260879
Ion Ratio Lower Upper
248 100
250 98.4 79.0 118.4
141 72.4 58.6 87.8



#69
Hexachlorobenzene
Concen: 23.064 ng/ul
RT: 16.810 min Scan# 2316
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:284 Resp: 310739
Ion Ratio Lower Upper
284 100
142 31.1 27.3 40.9
249 30.3 24.6 36.8





#70

Atrazine

Concen: 22.997 ng/ul

RT: 16.957 min Scan# 21

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:200 Resp: 276650

Ion Ratio Lower Upper

200 100

173 25.1 21.4 32.2

215 47.5 38.1 57.1



#71

Pentachlorophenol

Concen: 22.052 ng/ul

RT: 17.157 min Scan# 2375

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion:266 Resp: 185848


Ion Ratio Lower Upper

266 100

264 63.7 51.0 76.4

268 63.0 51.8 77.6





#72

Phenanthrene

Concen: 21.242 ng/ul

RT: 17.551 min Scan# 2443

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:178 Resp: 1280332

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.2

176 19.0 15.6 23.4



#74

Anthracene

Concen: 21.438 ng/ul

RT: 17.639 min Scan# 2457

Delta R.T. -0.012 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion:178 Resp: 1306627


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 18.7 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.727 ng/uL
RT: 13.557 min Scan# 11
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:216 Resp: 325284
Ion Ratio Lower Upper
216 100
214 77.2 62.2 93.4
218 47.4 38.2 57.2



#76
Pentachlorobenzene
Concen: 21.856 ng/uL
RT: 15.069 min Scan# 2020
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:250 Resp: 344397
Ion Ratio Lower Upper
250 100
248 63.4 49.8 74.6
252 63.8 51.6 77.4





#77

Carbazole

Concen: 22.319 ng/ul

RT: 17.910 min Scan# 2503

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:167 Resp: 1192868

Ion Ratio Lower Upper

167 100

166 21.8 17.4 26.2

139 12.9 10.2 15.4



#78

Di-n-butylphthalate

Concen: 21.695 ng/ul

RT: 18.433 min Scan# 2592

Delta R.T. -0.012 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion:149 Resp: 1493214


Ion Ratio Lower Upper

149 100

150 9.2 7.3 10.9

104 5.7 3.9 5.9





#80

Fluoranthene

Concen: 20.148 ng/ul

RT: 19.504 min Scan# 21

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:202 Resp: 1573001

Ion Ratio Lower Upper

202 100

101 9.9 9.5 14.3

100 7.9 7.3 10.9



#82

Pyrene

Concen: 20.244 ng/ul

RT: 19.857 min Scan# 2834

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion:202 Resp: 1632095


Ion Ratio Lower Upper

202 100

101 11.6 11.2 16.8

100 9.9 9.0 13.6









#83
Butylbenzylphthalate
Concen: 20.439 ng/ul
RT: 20.710 min Scan# 2980
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:149 Resp: 730517
Ion Ratio Lower Upper
149 100
91 70.7 57.4 86.2
206 22.2 16.6 25.0



#84
3,3'-Dichlorobenzidine
Concen: 24.158 ng/ul
RT: 21.492 min Scan# 3112
Delta R.T. -0.006 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:252 Resp: 585965
Ion Ratio Lower Upper
252 100
254 62.7 50.8 76.2
126 10.4 9.8 14.8





#85

Benzo(a)anthracene

Concen: 21.378 ng/ul

RT: 21.569 min Scan# 3125

Delta R.T. -0.012 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:228 Resp: 1632565

Ion Ratio Lower Upper

228 100

229 20.0 16.1 24.1

226 27.2 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.817 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion:149 Resp: 1071185


Ion Ratio Lower Upper

149 100

167 27.4 21.8 32.6

279 4.5 3.2 4.8





#87

Chrysene

Concen: 21.313 ng/ul

RT: 21.627 min Scan# 3136

Delta R.T. -0.006 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:228 Resp: 1555845

Ion Ratio Lower Upper

228 100

226 29.5 25.0 37.6

229 19.5 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 20.947 ng/ul




RT: 22.433 min Scan# 3272


Delta R.T. -0.012 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Tgt Ion:149 Resp: 1862906









#90
Benzo(b)fluoranthene
Concen: 21.116 ng/ul
RT: 23.357 min Scan# 3429
Delta R.T. -0.012 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:252 Resp: 1738274
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 8.6 8.3 12.5



#91
Benzo(k)fluoranthene
Concen: 21.991 ng/ul
RT: 23.357 min Scan# 3429
Delta R.T. -0.065 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:252 Resp: 1738274
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 8.6 8.5 12.7





#93

Benzo(a)pyrene

Concen: 20.749 ng/ul

RT: 24.033 min Scan# 3

Delta R.T. -0.012 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:252 Resp: 1509767

Ion Ratio Lower Upper

252 100

253 21.9 17.2 25.8

125 9.8 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.742 ng/ul

RT: 26.857 min Scan# 4024

Delta R.T. -0.024 min

Lab File: BPCCC345.D

Acq: 30 May 2023 21:03




Tgt Ion:276 Resp: 1859996


Ion Ratio Lower Upper

276 100

138 21.0 20.1 30.1

277 25.7 19.8 29.8









#95
Dibenzo(a,h)anthracene
Concen: 19.958 ng/ul
RT: 26.868 min Scan# 4030
Delta R.T. -0.024 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Instrument :
BNA_P
ClientSampleId :
SSTD020732

Tgt Ion:278 Resp: 1543523
Ion Ratio Lower Upper
278 100
139 13.9 13.3 19.9
279 23.6 19.1 28.7



#96
Benzo(g,h,i)perylene
Concen: 19.154 ng/ul
RT: 27.692 min Scan# 4166
Delta R.T. -0.024 min
Lab File: BPCCC345.D
Acq: 30 May 2023 21:03

Tgt Ion:276 Resp: 1471498
Ion Ratio Lower Upper
276 100
138 18.1 17.7 26.5
277 23.8 18.8 28.2

