

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060722\
 Data File : BP010715.D
 Acq On : 07 Jun 2022 18:40
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP060722

Quant Time: Jun 07 23:35:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 07 19:29:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	40.566	-1.4	98	0.00
3	Pyridine	40.000	42.220	-5.5	99	0.00
4	n-Nitrosodimethylamine	40.000	41.887	-4.7	98	0.00
5 S	2-Fluorophenol	80.000	83.490	-4.4	99	0.00
6	Aniline	40.000	42.775	-6.9	101	0.00
7 S	Phenol-d6	80.000	85.955	-7.4	101	0.00
8	2-Chlorophenol	40.000	42.402	-6.0	102	0.00
9	Benzaldehyde	40.000	39.867	0.3	108	0.00
10 C	Phenol	40.000	42.776	-6.9	100	0.00
11	bis(2-Chloroethyl)ether	40.000	42.539	-6.3	104	0.00
12	1,3-Dichlorobenzene	40.000	42.184	-5.5	103	0.00
13 C	1,4-Dichlorobenzene	40.000	41.933	-4.8	100	0.00
14	1,2-Dichlorobenzene	40.000	41.323	-3.3	101	0.00
15	Benzyl Alcohol	40.000	42.758	-6.9	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.741	-4.4	101	0.00
17	2-Methylphenol	40.000	43.010	-7.5	102	0.00
18	Hexachloroethane	40.000	41.863	-4.7	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.103	-7.8	103	0.00
20	3+4-Methylphenols	40.000	43.463	-8.7	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	41.267	-3.2	103	0.00
23 S	Nitrobenzene-d5	80.000	82.612	-3.3	102	0.00
24	Nitrobenzene	40.000	41.515	-3.8	103	0.00
25	Isophorone	40.000	41.615	-4.0	102	0.00
26 C	2-Nitrophenol	40.000	42.741	-6.9	103	0.00
27	2,4-Dimethylphenol	40.000	41.064	-2.7	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.976	-2.4	103	0.00
29 C	2,4-Dichlorophenol	40.000	42.725	-6.8	105	0.00
30	1,2,4-Trichlorobenzene	40.000	40.416	-1.0	101	0.00
31	Naphthalene	40.000	41.088	-2.7	102	0.00
32	Benzoic acid	40.000	39.798	0.5	100	0.00
33	4-Chloroaniline	40.000	42.016	-5.0	102	0.00
34 C	Hexachlorobutadiene	40.000	40.607	-1.5	101	0.00
35	Caprolactam	40.000	41.782	-4.5	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.885	-4.7	103	0.00
37	2-Methylnaphthalene	40.000	41.199	-3.0	102	0.00
38	1-Methylnaphthalene	40.000	40.530	-1.3	101	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.286	-3.2	101	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.307	1.7	101	0.00
42 S	2,4,6-Tribromophenol	80.000	82.351	-2.9	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.627	-6.6	103	0.00
44	2,4,5-Trichlorophenol	40.000	42.192	-5.5	101	0.00
45 S	2-Fluorobiphenyl	80.000	82.594	-3.2	103	0.00
46	1,1'-Biphenyl	40.000	41.328	-3.3	103	0.00
47	2-Chloronaphthalene	40.000	41.224	-3.1	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060722\
 Data File : BP010715.D
 Acq On : 07 Jun 2022 18:40
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP060722

Quant Time: Jun 07 23:35:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 07 19:29:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.566	-6.4	103	0.00
49	Acenaphthylene	40.000	41.330	-3.3	102	0.00
50	Dimethylphthalate	40.000	40.735	-1.8	102	0.00
51	2,6-Dinitrotoluene	40.000	41.897	-4.7	103	0.00
52 C	Acenaphthene	40.000	41.491	-3.7	103	0.00
53	3-Nitroaniline	40.000	42.662	-6.7	101	0.00
54 P	2,4-Dinitrophenol	40.000	39.945	0.1	104	0.00
55	Dibenzofuran	40.000	41.142	-2.9	103	0.00
56 P	4-Nitrophenol	40.000	41.802	-4.5	100	0.00
57	2,4-Dinitrotoluene	40.000	41.571	-3.9	100	0.00
58	Fluorene	40.000	41.214	-3.0	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.912	-4.8	102	0.00
60	Diethylphthalate	40.000	40.827	-2.1	102	0.00
61	4-Chlorophenyl-phenylether	40.000	40.699	-1.7	101	0.00
62	4-Nitroaniline	40.000	43.361	-8.4	102	0.00
63	Azobenzene	40.000	42.247	-5.6	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.726	-4.3	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.497	-3.7	102	0.00
67	4-Bromophenyl-phenylether	40.000	41.279	-3.2	101	0.00
68	Hexachlorobenzene	40.000	40.237	-0.6	102	0.00
69	Atrazine	40.000	41.302	-3.3	103	0.00
70 C	Pentachlorophenol	40.000	41.308	-3.3	99	0.00
71	Phenanthrene	40.000	40.332	-0.8	100	0.00
72	Anthracene	40.000	41.127	-2.8	101	0.00
73	Carbazole	40.000	41.296	-3.2	100	0.00
74	Di-n-butylphthalate	40.000	40.929	-2.3	102	0.00
75 C	Fluoranthene	40.000	40.447	-1.1	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	44.363	-10.9	88	0.00
78	Pyrene	40.000	40.668	-1.7	100	0.00
79 S	Terphenyl-d14	80.000	81.371	-1.7	100	0.00
80	Butylbenzylphthalate	40.000	41.273	-3.2	100	0.00
81	Benzo(a)anthracene	40.000	40.764	-1.9	98	0.00
82	3,3'-Dichlorobenzidine	40.000	40.734	-1.8	97	0.00
83	Chrysene	40.000	39.982	0.0	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.671	-1.7	99	0.00
85 c	Di-n-octyl phthalate	40.000	40.583	-1.5	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.258	-3.1	97	-0.01
88	Benzo(b)fluoranthene	40.000	41.302	-3.3	99	0.00
89	Benzo(k)fluoranthene	40.000	40.283	-0.7	94	0.00
90 C	Benzo(a)pyrene	40.000	40.971	-2.4	96	0.00
91	Dibenzo(a,h)anthracene	40.000	41.467	-3.7	98	0.00
92	Benzo(g,h,i)perylene	40.000	41.147	-2.9	96	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060722\
Data File : BP010715.D
Acq On : 07 Jun 2022 18:40
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP060722

Quant Time: Jun 07 23:35:16 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 07 19:29:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0