

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070722\
 Data File : BP010942.D
 Acq On : 07 Jul 2022 16:17
 Operator : CG/JU
 Sample : PB146062BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS062

Quant Time: Jul 08 00:59:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.704	152	376913	20.000	ng/ul	0.00
20) Naphthalene-d8	10.493	136	1476486	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.357	164	808476	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.116	188	1609112	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.233	240	1570943	20.000	ng/ul	# 0.00
88) Perylene-d12	23.586	264	1604662	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	72277	7.254	ng/uL	0.00
4) Pyridine-d5	3.605	84	873882	32.948	ng/ul	0.00
7) Phenol-d5	6.881	99	1050182	35.198	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	678131	34.753	ng/ul	0.00
11) 2-Chlorophenol-d4	7.234	132	881869	35.625	ng/ul	0.00
15) 4-Methylphenol-d8	8.416	113	830469	35.210	ng/ul	0.00
21) Nitrobenzene-d5	8.863	128	396052	42.137	ng/ul	0.00
24) 2-Nitrophenol-d4	9.581	143	383635	49.334	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.122	165	722642	37.610	ng/ul	0.00
31) 4-Chloroaniline-d4	10.640	131	794892	25.344	ng/ul	0.00
46) Dimethylphthalate-d6	13.775	166	2091519	37.766	ng/ul	0.00
49) Acenaphthylene-d8	14.045	160	2537378	35.899	ng/ul	0.00
54) 4-Nitrophenol-d4	14.575	143	402076	46.482	ng/ul	0.00
60) Fluorene-d10	15.357	176	1764648	37.130	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.481	200	275186	51.619	ng/ul	0.00
73) Anthracene-d10	17.222	188	2669755	37.291	ng/ul	0.00
81) Pyrene-d10	19.469	212	3075500	35.935	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.439	264	3116765	39.306	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.234	88	143172	13.474	ng/uL#	89
5) Pyridine	3.628	79	892857	33.232	ng/ul#	74
6) Benzaldehyde	6.852	77	697871	44.330	ng/ul	96
8) Phenol	6.905	94	1093906	34.279	ng/ul#	86
10) Bis(2-Chloroethyl)ether	7.134	93	883501	34.472	ng/ul	89
12) 2-Chlorophenol	7.269	128	889448	34.808	ng/ul	92
13) 2-Methylphenol	8.152	108	790422	33.271	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.234	45	1198947	34.866	ng/ul	98
16) Acetophenone	8.528	105	1250968	34.536	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.522	70	604941	33.028	ng/ul	91
18) 4-Methylphenol	8.481	108	853013	34.478	ng/ul	98
19) Hexachloroethane	8.775	117	357101	34.508	ng/ul#	83
22) Nitrobenzene	8.904	77	928966	39.604	ng/ul	93
23) Isophorone	9.434	82	1672721	34.006	ng/ul	98
25) 2-Nitrophenol	9.616	139	428698	45.524	ng/ul#	82
26) 2,4-Dimethylphenol	9.681	107	882572	34.304	ng/ul	94
27) Bis(2-Chloroethoxy)met...	9.922	93	1109340	35.750	ng/ul	98
29) 2,4-Dichlorophenol	10.146	162	725818	36.354	ng/ul	99
30) Naphthalene	10.546	128	2706736	35.174	ng/ul	100
32) 4-Chloroaniline	10.663	127	1092868	35.169	ng/ul	100
33) Hexachlorobutadiene	10.828	225	424514	33.281	ng/ul	97
34) Caprolactam	11.451	113	232291	37.036	ng/ul	91
35) 4-Chloro-3-methylphenol	11.798	107	799520	36.948	ng/ul	93
36) 2-Methylnaphthalene	12.163	142	1706867	34.253	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070722\
 Data File : BP010942.D
 Acq On : 07 Jul 2022 16:17
 Operator : CG/JU
 Sample : PB146062BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS062

Quant Time: Jul 08 00:59:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.387	142	1705213	34.445	ng/ul#	100
39) 1,2,4,5-Tetrachloroben...	12.534	216	794139	34.182	ng/ul	99
40) Hexachlorocyclopentadiene	12.510	237	468391	31.816	ng/ul	98
41) 2,4,6-Trichlorophenol	12.781	196	514762	38.431	ng/ul	98
42) 2,4,5-Trichlorophenol	12.851	196	567424	39.476	ng/ul	98
43) 1,1'-Biphenyl	13.187	154	2184484	34.322	ng/ul#	98
44) 2-Chloronaphthalene	13.222	162	1688923	35.226	ng/ul	98
45) 2-Nitroaniline	13.439	65	487753	47.949	ng/ul#	82
47) Dimethylphthalate	13.822	163	2086940	37.628	ng/ul	98
48) 2,6-Dinitrotoluene	13.939	165	396023	47.656	ng/ul	95
50) Acenaphthylene	14.075	152	2657304	34.328	ng/ul	99
51) 3-Nitroaniline	14.269	138	461198	46.293	ng/ul	89
52) Acenaphthene	14.422	153	1821090	36.460	ng/ul	98
53) 2,4-Dinitrophenol	14.475	184	173080	54.253	ng/ul	85
55) 4-Nitrophenol	14.586	109	300959	43.756	ng/ul#	76
56) Dibenzofuran	14.757	168	2472088	36.133	ng/ul	97
57) 2,4-Dinitrotoluene	14.728	165	576504	51.919	ng/ul#	85
58) 2,3,4,6-Tetrachlorophenol	14.986	232	444600	43.193	ng/ul#	90
59) Diethylphthalate	15.198	149	2134410	38.710	ng/ul	98
61) Fluorene	15.410	166	2010135	36.847	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.410	204	933758	36.563	ng/ul	97
63) 4-Nitroaniline	15.439	138	491662	53.376	ng/ul#	79
66) 4,6-Dinitro-2-methylph...	15.492	198	282938	49.114	ng/ul	98
67) N-Nitrosodiphenylamine	15.628	169	1718377	35.258	ng/ul	98
68) 4-Bromophenyl-phenylether	16.310	248	564445	35.001	ng/ul	95
69) Hexachlorobenzene	16.416	284	638686	34.193	ng/ul	97
70) Atrazine	16.598	200	572119	35.406	ng/ul	98
71) Pentachlorophenol	16.763	266	353237	36.727	ng/ul	97
72) Phenanthrene	17.163	178	3168988	36.982	ng/ul	98
74) Anthracene	17.257	178	3224150	37.410	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.145	216	819799	30.874	ng/uL	98
76) Pentachlorobenzene	14.675	250	764142	33.555	ng/uL	99
77) Carbazole	17.533	167	3017441	38.896	ng/ul	99
78) Di-n-butylphthalate	18.098	149	3679831	39.806	ng/ul	99
80) Fluoranthene	19.145	202	3647517	34.653	ng/ul#	88
82) Pyrene	19.498	202	3784385	35.601	ng/ul#	83
83) Butylbenzylphthalate	20.392	149	1670587	44.648	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.157	252	1237838	38.159	ng/ul#	97
85) Benzo(a)anthracene	21.216	228	3721769	37.881	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.157	149	2537657	42.828	ng/ul#	95
87) Chrysene	21.268	228	3495031	37.122	ng/ul	99
89) Di-n-octyl phthalate	22.074	149	4276821	44.902	ng/ul	100
90) Benzo(b)fluoranthene	22.874	252	3988878	39.973	ng/ul#	96
91) Benzo(k)fluoranthene	22.921	252	3730153	39.326	ng/ul#	95
93) Benzo(a)pyrene	23.486	252	3444629	35.485	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.033	276	4558760	38.919	ng/ul#	89
95) Dibenzo(a,h)anthracene	26.056	278	3797606	37.797	ng/ul#	92
96) Benzo(g,h,i)perylene	26.786	276	3718281	37.318	ng/ul#	89





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 13.474 ng/uL
 RT: 3.234 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BP010942.D
 Acq: 07 Jul 2022 16:17




Instrument :
 BNA_P
 ClientSampleId :
 SLCS062


Tgt Ion: 88 Resp: 143172
 Ion Ratio Lower Upper
 88 100
 43 29.9 35.8 53.6#
 58 81.2 68.1 102.1



#5
 Pyridine
 Concen: 33.232 ng/uL
 RT: 3.628 min Scan# 92
 Delta R.T. -0.000 min
 Lab File: BP010942.D
 Acq: 07 Jul 2022 16:17

Tgt Ion: 79 Resp: 892857
 Ion Ratio Lower Upper
 79 100
 52 76.0 83.8 125.8#
 51 35.1 41.0 61.4#









#6
Benzaldehyde
Concen: 44.330 ng/ul
RT: 6.852 min Scan# 640
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion: 77 Resp: 697871
Ion Ratio Lower Upper
77 100
105 99.6 76.0 114.0
106 95.2 73.2 109.8



#8
Phenol
Concen: 34.279 ng/ul
RT: 6.905 min Scan# 649
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion: 94 Resp: 1093906
Ion Ratio Lower Upper
94 100
65 27.9 27.4 41.2
66 36.0 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 34.472 ng/ul
RT: 7.134 min Scan# 689
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion: 93 Resp: 883501
Ion Ratio Lower Upper
93 100
63 78.0 72.4 108.6
95 31.8 27.3 40.9



#12
2-Chlorophenol
Concen: 34.808 ng/ul
RT: 7.269 min Scan# 711
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion: 128 Resp: 889448
Ion Ratio Lower Upper
128 100
64 46.8 44.8 67.2
130 32.6 25.8 38.6





#13

2-Methylphenol

Concen: 33.271 ng/ul

RT: 8.152 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17

Instrument :

BNA_P

ClientSampleId :





SLCS062

Tgt Ion:108 Resp: 790422

Ion Ratio Lower Upper

108 100

107 92.5 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 34.866 ng/ul

RT: 8.234 min Scan# 875

Delta R.T. -0.012 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17




Tgt Ion: 45 Resp: 1198947


Ion Ratio Lower Upper

45 100

77 14.0 11.2 16.8

79 10.3 6.9 10.3









#16
Acetophenone
Concen: 34.536 ng/ul
RT: 8.528 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:105 Resp: 1250968
Ion Ratio Lower Upper
105 100
77 75.5 67.5 101.3
51 26.6 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 33.028 ng/ul
RT: 8.522 min Scan# 924
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion: 70 Resp: 604941
Ion Ratio Lower Upper
70 100
42 58.3 53.8 80.8
101 10.0 7.8 11.6
130 22.6 15.4 23.2









#18
4-Methylphenol
Concen: 34.478 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:108 Resp: 853013
Ion Ratio Lower Upper
108 100
107 113.0 88.3 132.5



#19
Hexachloroethane
Concen: 34.508 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:117 Resp: 357101
Ion Ratio Lower Upper
117 100
201 90.8 85.7 128.5
199 56.6 57.4 86.0#









#22
Nitrobenzene
Concen: 39.604 ng/ul
RT: 8.904 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion: 77 Resp: 928966
Ion Ratio Lower Upper
77 100
123 45.2 31.5 47.3
65 13.6 11.4 17.2



#23
Isophorone
Concen: 34.006 ng/ul
RT: 9.434 min Scan# 1079
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion: 82 Resp: 1672721
Ion Ratio Lower Upper
82 100
95 6.9 6.6 10.0
138 17.7 13.8 20.6









#25
2-Nitrophenol
Concen: 45.524 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:139 Resp: 428698
Ion Ratio Lower Upper
139 100
65 52.0 54.2 81.4#
109 28.8 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 34.304 ng/ul
RT: 9.681 min Scan# 1121
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:107 Resp: 882572
Ion Ratio Lower Upper
107 100
121 53.4 39.1 58.7
122 87.0 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 35.750 ng/ul

RT: 9.922 min Scan# 1162

Delta R.T. -0.000 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17

Instrument :

BNA_P

ClientSampleId :

SLCS062





Tgt Ion: 93 Resp: 1109340

Ion Ratio Lower Upper

93 100

95 32.2 26.6 39.8

123 11.2 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 36.354 ng/ul

RT: 10.146 min Scan# 1200

Delta R.T. -0.000 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17




Tgt Ion: 162 Resp: 725818


Ion Ratio Lower Upper

162 100

164 65.2 52.5 78.7

98 38.4 31.4 47.0









#30
Naphthalene
Concen: 35.174 ng/ul
RT: 10.546 min Scan# 1268
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:128 Resp: 2706736
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 12.9 10.2 15.4



#32
4-Chloroaniline
Concen: 35.169 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:127 Resp: 1092868
Ion Ratio Lower Upper
127 100
129 32.7 26.2 39.2









#33
Hexachlorobutadiene
Concen: 33.281 ng/ul
RT: 10.828 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:225 Resp: 424514
Ion Ratio Lower Upper
225 100
223 64.5 53.7 80.5
227 64.0 49.4 74.2



#34
Caprolactam
Concen: 37.036 ng/ul
RT: 11.451 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:113 Resp: 232291
Ion Ratio Lower Upper
113 100
55 176.4 153.8 230.8
56 128.8 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 36.948 ng/ul
RT: 11.798 min Scan# 1481
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:107 Resp: 799520
Ion Ratio Lower Upper
107 100
144 26.0 18.6 27.8
142 81.2 59.6 89.4



#36
2-Methylnaphthalene
Concen: 34.253 ng/ul
RT: 12.163 min Scan# 1543
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:142 Resp: 1706867
Ion Ratio Lower Upper
142 100
141 87.9 70.1 105.1









#37
1-Methylnaphthalene
Concen: 34.445 ng/ul
RT: 12.387 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:142 Resp: 1705213
Ion Ratio Lower Upper
142 100
141 91.8 73.3 109.9
116 3.8 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.182 ng/ul
RT: 12.534 min Scan# 1606
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:216 Resp: 794139
Ion Ratio Lower Upper
216 100
214 78.7 63.0 94.6
179 21.7 18.1 27.1
108 19.7 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 31.816 ng/ul
RT: 12.510 min Scan# 1602
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:237 Resp: 468391
Ion Ratio Lower Upper
237 100
235 64.3 52.8 79.2
272 13.3 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 38.431 ng/ul
RT: 12.781 min Scan# 1648
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:196 Resp: 514762
Ion Ratio Lower Upper
196 100
198 95.2 78.2 117.4
200 30.4 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 39.476 ng/ul
RT: 12.851 min Scan# 1660
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:196 Resp: 567424
Ion Ratio Lower Upper
196 100
198 93.7 77.0 115.4
200 30.6 24.8 37.2



#43
1,1'-Biphenyl
Concen: 34.322 ng/ul
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:154 Resp: 2184484
Ion Ratio Lower Upper
154 100
153 40.5 31.9 47.9
76 14.2 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 35.226 ng/ul
RT: 13.222 min Scan# 1723
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:162 Resp: 1688923
Ion Ratio Lower Upper
162 100
164 32.3 25.5 38.3
127 41.5 31.9 47.9



#45
2-Nitroaniline
Concen: 47.949 ng/ul
RT: 13.439 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion: 65 Resp: 487753
Ion Ratio Lower Upper
65 100
92 70.5 49.0 73.6
138 103.1 66.4 99.6#









#47
Dimethylphthalate
Concen: 37.628 ng/ul
RT: 13.822 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:163 Resp: 2086940
Ion Ratio Lower Upper
163 100
194 3.8 3.4 5.2
164 10.2 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 47.656 ng/ul
RT: 13.939 min Scan# 1845
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:165 Resp: 396023
Ion Ratio Lower Upper
165 100
89 61.8 52.9 79.3
121 23.3 19.4 29.2









#50
Acenaphthylene
Concen: 34.328 ng/ul
RT: 14.075 min Scan# 1868
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:152 Resp: 2657304
Ion Ratio Lower Upper
152 100
151 19.9 16.5 24.7
153 13.0 10.1 15.1



#51
3-Nitroaniline
Concen: 46.293 ng/ul
RT: 14.269 min Scan# 1901
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:138 Resp: 461198
Ion Ratio Lower Upper
138 100
108 11.6 11.0 16.4
92 117.4 104.6 156.8









#52
Acenaphthene
Concen: 36.460 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:153 Resp: 1821090
Ion Ratio Lower Upper
153 100
152 47.3 36.2 54.2
154 89.1 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 54.253 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:184 Resp: 173080
Ion Ratio Lower Upper
184 100
63 74.1 73.8 110.6
154 63.4 57.2 85.8









#55
4-Nitrophenol
Concen: 43.756 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:109 Resp: 300959
Ion Ratio Lower Upper
109 100
139 145.3 75.6 113.4#
65 139.6 107.6 161.4



#56
Dibenzofuran
Concen: 36.133 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:168 Resp: 2472088
Ion Ratio Lower Upper
168 100
139 38.8 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 51.919 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:165 Resp: 576504
Ion Ratio Lower Upper
165 100
63 46.9 47.0 70.6#
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 43.193 ng/ul
RT: 14.986 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:232 Resp: 444600
Ion Ratio Lower Upper
232 100
131 52.7 35.8 53.8
130 2.8 1.8 2.8#
166 33.8 23.7 35.5









#59
Diethylphthalate
Concen: 38.710 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:149 Resp: 2134410
Ion Ratio Lower Upper
149 100
177 21.1 17.6 26.4
150 12.2 10.3 15.5



#61
Fluorene
Concen: 36.847 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:166 Resp: 2010135
Ion Ratio Lower Upper
166 100
165 98.3 77.4 116.2
167 13.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 36.563 ng/ul
RT: 15.410 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:204 Resp: 933758
Ion Ratio Lower Upper
204 100
206 33.0 26.0 39.0
141 65.1 49.3 73.9



#63
4-Nitroaniline
Concen: 53.376 ng/ul
RT: 15.439 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:138 Resp: 491662
Ion Ratio Lower Upper
138 100
92 53.2 46.2 69.2
108 77.5 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 49.114 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:198 Resp: 282938
Ion Ratio Lower Upper
198 100
182 4.8 4.3 6.5
77 28.7 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 35.258 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:169 Resp: 1718377
Ion Ratio Lower Upper
169 100
168 67.5 52.6 78.8
167 36.3 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 35.001 ng/ul
RT: 16.310 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:248 Resp: 564445
Ion Ratio Lower Upper
248 100
250 97.9 80.8 121.2
141 85.7 63.9 95.9



#69
Hexachlorobenzene
Concen: 34.193 ng/ul
RT: 16.416 min Scan# 2266
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:284 Resp: 638686
Ion Ratio Lower Upper
284 100
142 37.9 27.8 41.6
249 31.6 25.4 38.2









#70
Atrazine
Concen: 35.406 ng/ul
RT: 16.598 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:200 Resp: 572119
Ion Ratio Lower Upper
200 100
173 27.8 20.7 31.1
215 47.6 38.6 57.8



#71
Pentachlorophenol
Concen: 36.727 ng/ul
RT: 16.763 min Scan# 2325
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:266 Resp: 353237
Ion Ratio Lower Upper
266 100
264 62.2 51.4 77.2
268 63.9 53.5 80.3









#72
Phenanthrene
Concen: 36.982 ng/ul
RT: 17.163 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:178 Resp: 3168988
Ion Ratio Lower Upper
178 100
179 15.3 12.7 19.1
176 19.2 14.6 22.0



#74
Anthracene
Concen: 37.410 ng/ul
RT: 17.257 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:178 Resp: 3224150
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.2
176 18.5 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 30.874 ng/uL

RT: 13.145 min Scan# 1710

Delta R.T. -0.000 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17

Instrument :

BNA_P

ClientSampleId :

SLCS062





Tgt Ion:216 Resp: 819799

Ion Ratio Lower Upper

216 100

214 79.3 65.0 97.4

218 48.5 39.1 58.7



#76

Pentachlorobenzene

Concen: 33.555 ng/uL

RT: 14.675 min Scan# 1970

Delta R.T. -0.000 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17




Tgt Ion:250 Resp: 764142


Ion Ratio Lower Upper

250 100

248 63.0 49.8 74.6

252 64.0 51.2 76.8









#77
 Carbazole
 Concen: 38.896 ng/ul
 RT: 17.533 min Scan# 2456
 Delta R.T. -0.000 min
 Lab File: BP010942.D
 Acq: 07 Jul 2022 16:17

Instrument :
 BNA_P
 ClientSampleId :
 SLCS062

Tgt Ion:167 Resp: 3017441




Ion	Ratio	Lower	Upper
167	100		
166	21.6	16.9	25.3
139	13.5	10.6	16.0




#78
 Di-n-butylphthalate
 Concen: 39.806 ng/ul
 RT: 18.098 min Scan# 2552
 Delta R.T. -0.006 min
 Lab File: BP010942.D
 Acq: 07 Jul 2022 16:17

Tgt Ion:149 Resp: 3679831

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.1	10.7
104	5.3	4.6	6.8









#80
Fluoranthene
Concen: 34.653 ng/ul
RT: 19.145 min Scan# 2730
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:202 Resp: 3647517
Ion Ratio Lower Upper
202 100
101 13.9 7.1 10.7#
100 10.6 5.7 8.5#



#82
Pyrene
Concen: 35.601 ng/ul
RT: 19.498 min Scan# 2790
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:202 Resp: 3784385
Ion Ratio Lower Upper
202 100
101 16.6 7.9 11.9#
100 13.7 6.7 10.1#





#83

Butylbenzylphthalate

Concen: 44.648 ng/ul

RT: 20.392 min Scan# 2942

Delta R.T. -0.006 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17

Instrument :

BNA_P

ClientSampleId :

SLCS062





Tgt Ion:149 Resp: 1670587

Ion Ratio Lower Upper

149 100

91 72.6 60.0 90.0

206 19.6 17.4 26.2



#84

3,3'-Dichlorobenzidine

Concen: 38.159 ng/ul

RT: 21.157 min Scan# 3072

Delta R.T. -0.006 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17




Tgt Ion:252 Resp: 1237838


Ion Ratio Lower Upper

252 100

254 65.3 51.0 76.4

126 14.0 8.6 12.8#





#85

Benzo(a)anthracene

Concen: 37.881 ng/ul

RT: 21.216 min Scan# 3083

Delta R.T. -0.006 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17

Instrument :

BNA_P

ClientSampleId :

SLCS062





Tgt Ion:228 Resp: 3721769

Ion Ratio Lower Upper

228 100

229 20.2 15.0 22.4

226 27.3 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 42.828 ng/ul

RT: 21.157 min Scan# 3072

Delta R.T. -0.006 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17




Tgt Ion:149 Resp: 2537657


Ion Ratio Lower Upper

149 100

167 25.7 22.6 33.8

279 3.6 4.2 6.4#





#87

Chrysene

Concen: 37.122 ng/ul

RT: 21.268 min Scan# 3091

Delta R.T. -0.006 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17

Instrument :

BNA_P

ClientSampleId :

SLCS062





Tgt Ion:228 Resp: 3495031

Ion Ratio Lower Upper

228 100

226 29.8 23.6 35.4

229 20.2 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 44.902 ng/ul




RT: 22.074 min Scan# 3228


Delta R.T. -0.006 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17

Tgt Ion:149 Resp: 4276821









#90
Benzo(b)fluoranthene
Concen: 39.973 ng/ul
RT: 22.874 min Scan# 3364
Delta R.T. -0.000 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17




Instrument :
BNA_P
ClientSampleId :
SLCS062


Tgt Ion:252 Resp: 3988878
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.6
125 12.8 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 39.326 ng/ul
RT: 22.921 min Scan# 3372
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:252 Resp: 3730153
Ion Ratio Lower Upper
252 100
253 22.0 17.0 25.4
125 12.4 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 35.485 ng/ul

RT: 23.486 min Scan# 34

Delta R.T. -0.006 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17

Instrument :

BNA_P

ClientSampleId :

SLCS062





Tgt Ion:252 Resp: 3444629

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 14.3 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 38.919 ng/ul

RT: 26.033 min Scan# 3901

Delta R.T. -0.006 min

Lab File: BP010942.D

Acq: 07 Jul 2022 16:17




Tgt Ion:276 Resp: 4558760


Ion Ratio Lower Upper

276 100

138 29.3 14.6 22.0#

277 25.0 19.6 29.4









#95
Dibenzo(a,h)anthracene
Concen: 37.797 ng/ul
RT: 26.056 min Scan# 3905
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Instrument :
BNA_P
ClientSampleId :
SLCS062

Tgt Ion:278 Resp: 3797606
Ion Ratio Lower Upper
278 100
139 20.4 9.4 14.2#
279 23.9 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 37.318 ng/ul
RT: 26.786 min Scan# 4029
Delta R.T. -0.006 min
Lab File: BP010942.D
Acq: 07 Jul 2022 16:17

Tgt Ion:276 Resp: 3718281
Ion Ratio Lower Upper
276 100
138 26.2 12.7 19.1#
277 23.4 18.2 27.4

