

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071422\
 Data File : BP010968.D
 Acq On : 14 Jul 2022 12:05
 Operator : CG/JU
 Sample : SST00517
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005617

Quant Time: Jul 14 13:51:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 14 13:50:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.687	152	368873	20.000	ng/u1	0.00
20) Naphthalene-d8	10.481	136	1513662	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.351	164	873253	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.122	188	1746919	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.245	240	1761849	20.000	ng/u1	# 0.00
88) Perylene-d12	23.604	264	1733102	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	20610	2.114	ng/uL	0.00
4) Pyridine-d5	3.599	84	122203	4.708	ng/u1	0.00
7) Phenol-d5	6.864	99	137930	4.724	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.028	67	93512	4.897	ng/u1	0.00
11) 2-Chlorophenol-d4	7.222	132	114839	4.740	ng/u1	0.00
15) 4-Methylphenol-d8	8.405	113	109943	4.763	ng/u1	0.00
21) Nitrobenzene-d5	8.852	128	50440	5.235	ng/u1	0.00
24) 2-Nitrophenol-d4	9.569	143	47664	5.979	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.105	165	93058	4.724	ng/u1	0.00
31) 4-Chloroaniline-d4	10.628	131	154843	4.816	ng/u1	0.00
46) Dimethylphthalate-d6	13.769	166	294288	4.920	ng/u1	0.00
49) Acenaphthylene-d8	14.040	160	323462	4.237	ng/u1	0.00
54) 4-Nitrophenol-d4	14.563	143	46043	4.928	ng/u1	0.00
60) Fluorene-d10	15.351	176	253156	4.932	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.475	200	27474	4.747	ng/u1	0.00
73) Anthracene-d10	17.222	188	362636	4.666	ng/u1	0.00
81) Pyrene-d10	19.475	212	510234	5.316	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.445	264	390395	4.558	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.217	88	22086	2.124	ng/uL	89
5) Pyridine	3.617	79	128285	4.879	ng/u1#	82
6) Benzaldehyde	6.834	77	85972	5.580	ng/u1	97
8) Phenol	6.893	94	148387	4.751	ng/u1#	86
10) Bis(2-Chloroethyl)ether	7.122	93	124266	4.954	ng/u1	92
12) 2-Chlorophenol	7.252	128	120046	4.800	ng/u1	93
13) 2-Methylphenol	8.140	108	109008	4.688	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.228	45	169202	5.028	ng/u1#	98
16) Acetophenone	8.516	105	176781	4.987	ng/u1#	91
17) N-Nitroso-di-n-propyla...	8.499	70	88292	4.926	ng/u1	92
18) 4-Methylphenol	8.463	108	115559	4.773	ng/u1	98
19) Hexachloroethane	8.763	117	47939	4.734	ng/u1#	82
22) Nitrobenzene	8.893	77	123107	5.119	ng/u1	92
23) Isophorone	9.422	82	226812	4.498	ng/u1	97
25) 2-Nitrophenol	9.599	139	54126	5.607	ng/u1#	85
26) 2,4-Dimethylphenol	9.669	107	124830	4.733	ng/u1	95
27) Bis(2-Chloroethoxy)met...	9.910	93	159266	5.007	ng/u1#	97
29) 2,4-Dichlorophenol	10.134	162	99307	4.852	ng/u1	99
30) Naphthalene	10.534	128	388007	4.918	ng/u1	100
32) 4-Chloroaniline	10.652	127	157500	4.944	ng/u1	100
33) Hexachlorobutadiene	10.822	225	60419	4.620	ng/u1	96
34) Caprolactam	11.446	113	29670	4.614	ng/u1	85
35) 4-Chloro-3-methylphenol	11.787	107	111413	5.022	ng/u1	92
36) 2-Methylnaphthalene	12.157	142	247095	4.837	ng/u1	96


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071422\
 Data File : BP010968.D
 Acq On : 14 Jul 2022 12:05
 Operator : CG/JU
 Sample : SST00517
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005617

Quant Time: Jul 14 13:51:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 14 13:50:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.375	142	247718	4.881	ng/ul#	97
39) 1,2,4,5-Tetrachloroben...	12.528	216	115153	4.589	ng/ul	99
40) Hexachlorocyclopentadiene	12.504	237	61049	3.839	ng/ul	94
41) 2,4,6-Trichlorophenol	12.775	196	72297	4.997	ng/ul	97
42) 2,4,5-Trichlorophenol	12.846	196	75333	4.852	ng/ul	97
43) 1,1'-Biphenyl	13.181	154	326800	4.754	ng/ul#	97
44) 2-Chloronaphthalene	13.216	162	247655	4.782	ng/ul	99
45) 2-Nitroaniline	13.434	65	62733	5.710	ng/ul	90
47) Dimethylphthalate	13.816	163	299784	5.004	ng/ul	98
48) 2,6-Dinitrotoluene	13.934	165	44553	4.964	ng/ul	98
50) Acenaphthylene	14.069	152	378717	4.529	ng/ul	99
51) 3-Nitroaniline	14.263	138	52215	4.852	ng/ul	89
52) Acenaphthene	14.416	153	262022	4.857	ng/ul	99
53) 2,4-Dinitrophenol	14.469	184	20583	5.973	ng/ul	93
55) 4-Nitrophenol	14.575	109	38729	5.213	ng/ul#	79
56) Dibenzofuran	14.751	168	365755	4.949	ng/ul	97
57) 2,4-Dinitrotoluene	14.722	165	62061	5.175	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	14.981	232	59518	5.353	ng/ul#	86
59) Diethylphthalate	15.193	149	314430	5.279	ng/ul	98
61) Fluorene	15.410	166	294842	5.004	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.410	204	138445	5.019	ng/ul	96
63) 4-Nitroaniline	15.434	138	48202	4.845	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.487	198	29266	4.679	ng/ul#	90
67) N-Nitrosodiphenylamine	15.628	169	249253	4.711	ng/ul	99
68) 4-Bromophenyl-phenylether	16.310	248	77029	4.400	ng/ul	94
69) Hexachlorobenzene	16.410	284	91206	4.498	ng/ul#	91
70) Atrazine	16.592	200	80169	4.570	ng/ul	99
71) Pentachlorophenol	16.763	266	48190	4.615	ng/ul	89
72) Phenanthrene	17.163	178	465083	4.999	ng/ul	98
74) Anthracene	17.257	178	454077	4.853	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.140	216	139257	4.831	ng/uL	97
76) Pentachlorobenzene	14.669	250	109827	4.442	ng/uL	97
77) Carbazole	17.534	167	459187	5.452	ng/ul	99
78) Di-n-butylphthalate	18.104	149	541266	5.393	ng/ul	100
80) Fluoranthene	19.151	202	541285	4.585	ng/ul#	89
82) Pyrene	19.504	202	581670	4.879	ng/ul#	84
83) Butylbenzylphthalate	20.404	149	243536	5.803	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.169	252	105729	2.906	ng/ul#	97
85) Benzo(a)anthracene	21.227	228	563433	5.113	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.169	149	370495	5.575	ng/ul#	94
87) Chrysene	21.280	228	563969	5.341	ng/ul	100
89) Di-n-octyl phthalate	22.086	149	523116	5.085	ng/ul	100
90) Benzo(b)fluoranthene	22.880	252	518742	4.813	ng/ul#	95
91) Benzo(k)fluoranthene	22.927	252	502289	4.903	ng/ul#	94
93) Benzo(a)pyrene	23.492	252	497600	4.746	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.039	276	573763	4.535	ng/ul#	89
95) Dibenzo(a,h)anthracene	26.068	278	502637	4.632	ng/ul#	92
96) Benzo(g,h,i)perylene	26.792	276	487345	4.529	ng/ul#	86





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 2.124 ng/uL
RT: 3.217 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion: 88 Resp: 22086
Ion Ratio Lower Upper
88 100
43 47.7 35.8 53.6
58 71.3 68.1 102.1



#5
Pyridine
Concen: 4.879 ng/uL
RT: 3.617 min Scan# 90
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion: 79 Resp: 128285
Ion Ratio Lower Upper
79 100
52 82.6 83.8 125.8#
51 43.6 41.0 61.4









#6
Benzaldehyde
Concen: 5.580 ng/ul
RT: 6.834 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion: 77 Resp: 85972
Ion Ratio Lower Upper
77 100
105 96.9 76.0 114.0
106 95.1 73.2 109.8



#8
Phenol
Concen: 4.751 ng/ul
RT: 6.893 min Scan# 647
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion: 94 Resp: 148387
Ion Ratio Lower Upper
94 100
65 29.0 27.4 41.2
66 34.6 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 4.954 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion: 93 Resp: 124266
Ion Ratio Lower Upper
93 100
63 81.1 72.4 108.6
95 32.3 27.3 40.9



#12
2-Chlorophenol
Concen: 4.800 ng/ul
RT: 7.252 min Scan# 708
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:128 Resp: 120046
Ion Ratio Lower Upper
128 100
64 47.9 44.8 67.2
130 32.2 25.8 38.6





#13

2-Methylphenol

Concen: 4.688 ng/ul

RT: 8.140 min Scan# 81

Delta R.T. 0.006 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05

Instrument :

BNA_P

ClientSampleId :





SSTD005617

Tgt Ion:108 Resp: 109008

Ion Ratio Lower Upper

108 100

107 88.3 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 5.028 ng/ul

RT: 8.228 min Scan# 874

Delta R.T. 0.012 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05




Tgt Ion: 45 Resp: 169202


Ion Ratio Lower Upper

45 100

77 14.3 11.2 16.8

79 10.4 6.9 10.3#









#16
Acetophenone
Concen: 4.987 ng/ul
RT: 8.516 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:105 Resp: 176781
Ion Ratio Lower Upper
105 100
77 78.1 67.5 101.3
51 28.6 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 4.926 ng/ul
RT: 8.499 min Scan# 920
Delta R.T. -0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion: 70 Resp: 88292
Ion Ratio Lower Upper
70 100
42 59.5 53.8 80.8
101 9.2 7.8 11.6
130 21.9 15.4 23.2









#18
4-Methylphenol
Concen: 4.773 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:108 Resp: 115559
Ion Ratio Lower Upper
108 100
107 112.5 88.3 132.5



#19
Hexachloroethane
Concen: 4.734 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:117 Resp: 47939
Ion Ratio Lower Upper
117 100
201 88.7 85.7 128.5
199 57.0 57.4 86.0#









#22
 Nitrobenzene
 Concen: 5.119 ng/ul
 RT: 8.893 min Scan# 987
 Delta R.T. 0.000 min
 Lab File: BP010968.D
 Acq: 14 Jul 2022 12:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617


Tgt Ion:	77	Resp:	123107
Ion	Ratio	Lower	Upper
77	100		
123	45.3	31.5	47.3
65	12.6	11.4	17.2



#23
 Isophorone
 Concen: 4.498 ng/ul
 RT: 9.422 min Scan# 1077
 Delta R.T. 0.000 min
 Lab File: BP010968.D
 Acq: 14 Jul 2022 12:05

Tgt Ion:	82	Resp:	226812
Ion	Ratio	Lower	Upper
82	100		
95	6.9	6.6	10.0
138	18.3	13.8	20.6









#25
2-Nitrophenol
Concen: 5.607 ng/ul
RT: 9.599 min Scan# 1107
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion	Ratio	Lower	Upper
139	100		
65	54.3	54.2	81.4
109	30.4	30.6	45.8



#26
2,4-Dimethylphenol
Concen: 4.733 ng/ul
RT: 9.669 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion	Ratio	Lower	Upper
107	100		
121	53.4	39.1	58.7
122	86.9	66.1	99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 5.007 ng/ul

RT: 9.910 min Scan# 1159

Delta R.T. 0.006 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion: 93 Resp: 159266

Ion Ratio Lower Upper

93 100

95 31.8 26.6 39.8

123 12.3 7.8 11.8#



#29

2,4-Dichlorophenol

Concen: 4.852 ng/ul

RT: 10.134 min Scan# 1198

Delta R.T. 0.000 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05




Tgt Ion:162 Resp: 99307


Ion Ratio Lower Upper

162 100

164 64.6 52.5 78.7

98 40.3 31.4 47.0









#30
Naphthalene
Concen: 4.918 ng/ul
RT: 10.534 min Scan# 1266
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:128 Resp: 388007
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.8 10.2 15.4



#32
4-Chloroaniline
Concen: 4.944 ng/ul
RT: 10.652 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:127 Resp: 157500
Ion Ratio Lower Upper
127 100
129 32.5 26.2 39.2









#33
Hexachlorobutadiene
Concen: 4.620 ng/ul
RT: 10.822 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:225 Resp: 60419
Ion Ratio Lower Upper
225 100
223 62.6 53.7 80.5
227 63.7 49.4 74.2



#34
Caprolactam
Concen: 4.614 ng/ul
RT: 11.446 min Scan# 1421
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:113 Resp: 29670
Ion Ratio Lower Upper
113 100
55 171.0 153.8 230.8
56 118.8 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 5.022 ng/ul
RT: 11.787 min Scan# 1479
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:107 Resp: 111413
Ion Ratio Lower Upper
107 100
144 26.0 18.6 27.8
142 81.6 59.6 89.4



#36
2-Methylnaphthalene
Concen: 4.837 ng/ul
RT: 12.157 min Scan# 1542
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:142 Resp: 247095
Ion Ratio Lower Upper
142 100
141 91.7 70.1 105.1









#37
1-Methylnaphthalene
Concen: 4.881 ng/ul
RT: 12.375 min Scan# 1579
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:142 Resp: 247718
Ion Ratio Lower Upper
142 100
141 94.1 73.3 109.9
116 3.7 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.589 ng/ul
RT: 12.528 min Scan# 1605
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:216 Resp: 115153
Ion Ratio Lower Upper
216 100
214 79.0 63.0 94.6
179 23.0 18.1 27.1
108 19.6 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 3.839 ng/ul
RT: 12.504 min Scan# 1601
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:237 Resp: 61049
Ion Ratio Lower Upper
237 100
235 60.6 52.8 79.2
272 14.0 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 4.997 ng/ul
RT: 12.775 min Scan# 1647
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:196 Resp: 72297
Ion Ratio Lower Upper
196 100
198 94.4 78.2 117.4
200 30.1 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 4.852 ng/ul
RT: 12.846 min Scan# 1659
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:196 Resp: 75333
Ion Ratio Lower Upper
196 100
198 92.1 77.0 115.4
200 30.7 24.8 37.2



#43
1,1'-Biphenyl
Concen: 4.754 ng/ul
RT: 13.181 min Scan# 1716
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:154 Resp: 326800
Ion Ratio Lower Upper
154 100
153 40.9 31.9 47.9
76 14.1 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 4.782 ng/ul
RT: 13.216 min Scan# 1722
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:162 Resp: 247655
Ion Ratio Lower Upper
162 100
164 31.3 25.5 38.3
127 40.3 31.9 47.9



#45
2-Nitroaniline
Concen: 5.710 ng/ul
RT: 13.434 min Scan# 1759
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion: 65 Resp: 62733
Ion Ratio Lower Upper
65 100
92 66.4 49.0 73.6
138 94.2 66.4 99.6









#47
Dimethylphthalate
Concen: 5.004 ng/ul
RT: 13.816 min Scan# 1824
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:163 Resp: 299784
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.2
164 10.4 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 4.964 ng/ul
RT: 13.934 min Scan# 1844
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:165 Resp: 44553
Ion Ratio Lower Upper
165 100
89 66.3 52.9 79.3
121 28.2 19.4 29.2





#50

Acenaphthylene

Concen: 4.529 ng/ul

RT: 14.069 min Scan# 1867

Delta R.T. 0.000 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:152 Resp: 378717

Ion Ratio Lower Upper

152 100

151 20.2 16.5 24.7

153 13.4 10.1 15.1



#51

3-Nitroaniline

Concen: 4.852 ng/ul

RT: 14.263 min Scan# 1900

Delta R.T. 0.000 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05




Tgt Ion:138 Resp: 52215


Ion Ratio Lower Upper

138 100

108 14.1 11.0 16.4

92 117.3 104.6 156.8









#52
Acenaphthene
Concen: 4.857 ng/ul
RT: 14.416 min Scan# 1926
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:153 Resp: 262022
Ion Ratio Lower Upper
153 100
152 47.9 36.2 54.2
154 88.5 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 5.973 ng/ul
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:184 Resp: 20583
Ion Ratio Lower Upper
184 100
63 85.3 73.8 110.6
154 66.1 57.2 85.8









#55
4-Nitrophenol
Concen: 5.213 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:109 Resp: 38729
Ion Ratio Lower Upper
109 100
139 133.8 75.6 113.4#
65 143.9 107.6 161.4



#56
Dibenzofuran
Concen: 4.949 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:168 Resp: 365755
Ion Ratio Lower Upper
168 100
139 39.0 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 5.175 ng/ul
RT: 14.722 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion	Ratio	Lower	Upper
165	100		
63	55.3	47.0	70.6
182	2.7	2.2	3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 5.353 ng/ul
RT: 14.981 min Scan# 2022
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion	Ratio	Lower	Upper
232	100		
131	55.3	35.8	53.8#
130	3.6	1.8	2.8#
166	35.6	23.7	35.5#









#59
Diethylphthalate
Concen: 5.279 ng/ul
RT: 15.193 min Scan# 2058
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:149 Resp: 314430
Ion Ratio Lower Upper
149 100
177 20.5 17.6 26.4
150 12.9 10.3 15.5



#61
Fluorene
Concen: 5.004 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:166 Resp: 294842
Ion Ratio Lower Upper
166 100
165 97.8 77.4 116.2
167 13.9 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 5.019 ng/ul
RT: 15.410 min Scan# 2094
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:204 Resp: 138445
Ion Ratio Lower Upper
204 100
206 33.4 26.0 39.0
141 66.0 49.3 73.9



#63
4-Nitroaniline
Concen: 4.845 ng/ul
RT: 15.434 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:138 Resp: 48202
Ion Ratio Lower Upper
138 100
92 55.4 46.2 69.2
108 73.6 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 4.679 ng/ul
RT: 15.487 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:198 Resp: 29266
Ion Ratio Lower Upper
198 100
182 10.1 4.3 6.5#
77 34.9 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 4.711 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:169 Resp: 249253
Ion Ratio Lower Upper
169 100
168 65.6 52.6 78.8
167 35.6 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 4.400 ng/ul
RT: 16.310 min Scan# 2248
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:248 Resp: 77029
Ion Ratio Lower Upper
248 100
250 93.6 80.8 121.2
141 82.7 63.9 95.9



#69
Hexachlorobenzene
Concen: 4.498 ng/ul
RT: 16.410 min Scan# 2265
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:284 Resp: 91206
Ion Ratio Lower Upper
284 100
142 44.0 27.8 41.6#
249 32.2 25.4 38.2





#70

Atrazine

Concen: 4.570 ng/ul

RT: 16.592 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:200 Resp: 80169

Ion Ratio Lower Upper

200 100

173 26.5 20.7 31.1

215 47.3 38.6 57.8



#71

Pentachlorophenol

Concen: 4.615 ng/ul

RT: 16.763 min Scan# 2325

Delta R.T. 0.000 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05




Tgt Ion:266 Resp: 48190


Ion Ratio Lower Upper

266 100

264 54.4 51.4 77.2

268 58.7 53.5 80.3









#72
Phenanthrene
Concen: 4.999 ng/ul
RT: 17.163 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:178 Resp: 465083
Ion Ratio Lower Upper
178 100
179 15.3 12.7 19.1
176 19.5 14.6 22.0



#74
Anthracene
Concen: 4.853 ng/ul
RT: 17.257 min Scan# 2409
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:178 Resp: 454077
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.2
176 18.4 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.831 ng/uL
RT: 13.140 min Scan# 1708
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:216 Resp: 139257
Ion Ratio Lower Upper
216 100
214 77.8 65.0 97.4
218 47.8 39.1 58.7



#76
Pentachlorobenzene
Concen: 4.442 ng/uL
RT: 14.669 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:250 Resp: 109827
Ion Ratio Lower Upper
250 100
248 63.6 49.8 74.6
252 66.7 51.2 76.8









#77
 Carbazole
 Concen: 5.452 ng/ul
 RT: 17.534 min Scan# 2456
 Delta R.T. 0.000 min
 Lab File: BP010968.D
 Acq: 14 Jul 2022 12:05

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617

Tgt Ion:167 Resp: 459187




Ion	Ratio	Lower	Upper
167	100		
166	21.2	16.9	25.3
139	13.7	10.6	16.0




#78
 Di-n-butylphthalate
 Concen: 5.393 ng/ul
 RT: 18.104 min Scan# 2553
 Delta R.T. 0.000 min
 Lab File: BP010968.D
 Acq: 14 Jul 2022 12:05

Tgt Ion:149 Resp: 541266

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.1	10.7
104	5.3	4.6	6.8









#80
Fluoranthene
Concen: 4.585 ng/u1
RT: 19.151 min Scan# 2731
Delta R.T. 0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:202 Resp: 541285
Ion Ratio Lower Upper
202 100
101 13.4 7.1 10.7#
100 10.0 5.7 8.5#



#82
Pyrene
Concen: 4.879 ng/u1
RT: 19.504 min Scan# 2791
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:202 Resp: 581670
Ion Ratio Lower Upper
202 100
101 16.3 7.9 11.9#
100 13.4 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 5.803 ng/ul
RT: 20.404 min Scan# 2944
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:149 Resp: 243536
Ion Ratio Lower Upper
149 100
91 73.3 60.0 90.0
206 20.8 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 2.906 ng/ul
RT: 21.169 min Scan# 3074
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:252 Resp: 105729
Ion Ratio Lower Upper
252 100
254 62.1 51.0 76.4
126 15.1 8.6 12.8#









#85
Benzo(a)anthracene
Concen: 5.113 ng/ul
RT: 21.227 min Scan# 3084
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:228 Resp: 563433
Ion Ratio Lower Upper
228 100
229 19.2 15.0 22.4
226 26.5 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 5.575 ng/ul
RT: 21.169 min Scan# 3074
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:149 Resp: 370495
Ion Ratio Lower Upper
149 100
167 25.3 22.6 33.8
279 3.4 4.2 6.4





#87

Chrysene

Concen: 5.341 ng/ul

RT: 21.280 min Scan# 3093

Delta R.T. 0.000 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:228 Resp: 563969

Ion Ratio Lower Upper

228 100

226 29.4 23.6 35.4

229 19.3 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 5.085 ng/ul




RT: 22.086 min Scan# 3230


Delta R.T. 0.000 min

Lab File: BP010968.D

Acq: 14 Jul 2022 12:05

Tgt Ion:149 Resp: 523116









#90
Benzo(b)fluoranthene
Concen: 4.813 ng/ul
RT: 22.880 min Scan# 3365
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:252 Resp: 518742
Ion Ratio Lower Upper
252 100
253 23.0 17.8 26.6
125 13.4 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 4.903 ng/ul
RT: 22.927 min Scan# 3373
Delta R.T. 0.000 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:252 Resp: 502289
Ion Ratio Lower Upper
252 100
253 22.2 17.0 25.4
125 13.2 6.2 9.4#









#93
Benzo(a)pyrene
Concen: 4.746 ng/ul
RT: 23.492 min Scan# 3470
Delta R.T. -0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:252 Resp: 497600
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.0
125 14.3 6.6 10.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.535 ng/ul
RT: 26.039 min Scan# 3902
Delta R.T. -0.006 min
Lab File: BP010968.D
Acq: 14 Jul 2022 12:05

Tgt Ion:276 Resp: 573763
Ion Ratio Lower Upper
276 100
138 29.2 14.6 22.0#
277 23.6 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 4.632 ng/ul
 RT: 26.068 min Scan# 3907
 Delta R.T. 0.000 min
 Lab File: BP010968.D
 Acq: 14 Jul 2022 12:05

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617

Tgt Ion:278 Resp: 502637
 Ion Ratio Lower Upper
 278 100
 139 20.1 9.4 14.2#
 279 23.3 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 4.529 ng/ul
 RT: 26.792 min Scan# 4030
 Delta R.T. 0.000 min
 Lab File: BP010968.D
 Acq: 14 Jul 2022 12:05

Tgt Ion:276 Resp: 487345
 Ion Ratio Lower Upper
 276 100
 138 27.2 12.7 19.1#
 277 25.0 18.2 27.4

