

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071422\
 Data File : BP010973.D
 Acq On : 14 Jul 2022 14:57
 Operator : CG/JU
 Sample : SSTD16022
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160622

Quant Time: Jul 14 15:23:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 14 13:50:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.687	152	363658	20.000	ng/ul	0.00
20) Naphthalene-d8	10.487	136	1532233	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.351	164	853707	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.122	188	1722985	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.251	240	1534551	20.000	ng/ul	# 0.01
88) Perylene-d12	23.610	264	1612636	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	651895	67.811	ng/uL	0.00
4) Pyridine-d5	3.587	84	4271313	166.913	ng/ul	0.00
7) Phenol-d5	6.875	99	4978892	172.954	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	7.040	67	3018181	160.314	ng/ul	0.01
11) 2-Chlorophenol-d4	7.228	132	4049167	169.536	ng/ul	0.00
15) 4-Methylphenol-d8	8.416	113	3937547	173.028	ng/ul	0.01
21) Nitrobenzene-d5	8.858	128	1971606	202.130	ng/ul	0.01
24) 2-Nitrophenol-d4	9.575	143	2092149	259.256	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.116	165	3418617	171.449	ng/ul	0.01
31) 4-Chloroaniline-d4	10.634	131	5363215	164.777	ng/ul	0.00
46) Dimethylphthalate-d6	13.781	166	9326423	159.482	ng/ul	0.01
49) Acenaphthylene-d8	14.046	160	11200208	150.065	ng/ul	0.00
54) 4-Nitrophenol-d4	14.593	143	2063126	225.870	ng/ul	0.03
60) Fluorene-d10	15.357	176	7671927	152.875	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.493	200	1645005	288.172	ng/ul	0.02
73) Anthracene-d10	17.228	188	11286655	147.234	ng/ul	0.01
81) Pyrene-d10	19.481	212	12390299	148.204	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.469	264	12932662	162.288	ng/ul	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.217	88	676079	65.946	ng/uL#	85
5) Pyridine	3.611	79	4345717	167.643	ng/ul#	73
6) Benzaldehyde	6.840	77	1369556	90.167	ng/ul	96
8) Phenol	6.905	94	5183577	168.357	ng/ul#	86
10) Bis(2-Chloroethyl)ether	7.134	93	3956771	160.010	ng/ul	89
12) 2-Chlorophenol	7.258	128	4110833	166.740	ng/ul	92
13) 2-Methylphenol	8.146	108	3850192	167.970	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.234	45	5266843	158.744	ng/ul#	95
16) Acetophenone	8.528	105	5541639	158.565	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.534	70	2822318	159.707	ng/ul	88
18) 4-Methylphenol	8.493	108	4042270	169.339	ng/ul	100
19) Hexachloroethane	8.763	117	1633612	163.617	ng/ul#	82
22) Nitrobenzene	8.905	77	4273084	175.545	ng/ul#	89
23) Isophorone	9.446	82	8199927	160.635	ng/ul	97
25) 2-Nitrophenol	9.610	139	2225586	227.740	ng/ul#	80
26) 2,4-Dimethylphenol	9.681	107	4268908	159.888	ng/ul	95
27) Bis(2-Chloroethoxy)met...	9.916	93	5075136	157.604	ng/ul	98
29) 2,4-Dichlorophenol	10.146	162	3426540	165.380	ng/ul	98
30) Naphthalene	10.540	128	11888383	148.868	ng/ul	98
32) 4-Chloroaniline	10.663	127	5236137	162.372	ng/ul	99
33) Hexachlorobutadiene	10.822	225	1947599	147.134	ng/ul	95
34) Caprolactam	11.504	113	704380	108.219	ng/ul	84
35) 4-Chloro-3-methylphenol	11.804	107	3975993	177.057	ng/ul	94
36) 2-Methylnaphthalene	12.157	142	7859041	151.977	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071422\
 Data File : BP010973.D
 Acq On : 14 Jul 2022 14:57
 Operator : CG/JU
 Sample : SSTD16022
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD160622


Quant Time: Jul 14 15:23:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 14 13:50:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.381	142	7789488	151.622	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.528	216	3592722	146.448	ng/ul#	98
40) Hexachlorocyclopentadiene	12.504	237	2457610	158.093	ng/ul	97
41) 2,4,6-Trichlorophenol	12.781	196	2551322	180.384	ng/ul	98
42) 2,4,5-Trichlorophenol	12.851	196	2754574	181.486	ng/ul	99
43) 1,1'-Biphenyl	13.187	154	9668613	143.864	ng/ul#	96
44) 2-Chloronaphthalene	13.222	162	7413976	146.443	ng/ul	95
45) 2-Nitroaniline	13.446	65	2594099	241.503	ng/ul#	77
47) Dimethylphthalate	13.828	163	9192962	156.971	ng/ul	98
48) 2,6-Dinitrotoluene	13.951	165	2083585	237.449	ng/ul	90
50) Acenaphthylene	14.081	152	11814022	144.530	ng/ul	98
51) 3-Nitroaniline	14.281	138	2071594	196.922	ng/ul#	89
52) Acenaphthene	14.422	153	7882651	149.455	ng/ul	98
53) 2,4-Dinitrophenol	14.481	184	1304840	387.339	ng/ul#	83
55) 4-Nitrophenol	14.610	109	1535100	211.360	ng/ul#	80
56) Dibenzofuran	14.763	168	10496541	145.291	ng/ul	100
57) 2,4-Dinitrotoluene	14.740	165	2876468	245.326	ng/ul	85
58) 2,3,4,6-Tetrachlorophenol	14.987	232	2153766	198.155	ng/ul#	85
59) Diethylphthalate	15.210	149	9623692	165.288	ng/ul	99
61) Fluorene	15.416	166	8303684	144.145	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.410	204	3828975	141.988	ng/ul	93
63) 4-Nitroaniline	15.416	138	124677	12.818	ng/ul#	21
66) 4,6-Dinitro-2-methylph...	15.510	198	1633281	264.778	ng/ul#	92
67) N-Nitrosodiphenylamine	15.634	169	7593618	145.511	ng/ul	97
68) 4-Bromophenyl-phenylether	16.316	248	2571819	148.938	ng/ul	97
69) Hexachlorobenzene	16.422	284	2878728	143.933	ng/ul	96
70) Atrazine	16.610	200	2741322	158.438	ng/ul	97
71) Pentachlorophenol	16.769	266	1940421	188.418	ng/ul	98
72) Phenanthrene	17.169	178	13141779	143.228	ng/ul	94
74) Anthracene	17.263	178	12725915	137.901	ng/ul#	92
75) 1,2,3,4-Tetrachloroben...	13.146	216	3793484	133.422	ng/uL	98
76) Pentachlorobenzene	14.675	250	3396450	139.286	ng/uL	99
77) Carbazole	17.539	167	12716558	153.086	ng/ul	96
78) Di-n-butylphthalate	18.104	149	13320103	134.564	ng/ul#	87
80) Fluoranthene	19.151	202	13602638	132.294	ng/ul#	76
82) Pyrene	19.510	202	13030914	125.492	ng/ul#	68
83) Butylbenzylphthalate	20.404	149	7694130	210.510	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.175	252	4191884	132.289	ng/ul	99
85) Benzo(a)anthracene	21.233	228	13437870	140.018	ng/ul#	82
86) Bis(2-ethylhexyl)phtha...	21.169	149	10296217	177.892	ng/ul	96
87) Chrysene	21.286	228	12297159	133.712	ng/ul#	83
89) Di-n-octyl phthalate	22.080	149	14768752	154.290	ng/ul	100
90) Benzo(b)fluoranthene	22.898	252	16229723	161.836	ng/ul#	94
91) Benzo(k)fluoranthene	22.945	252	14193073	148.894	ng/ul#	92
93) Benzo(a)pyrene	23.521	252	15281026	156.639	ng/ul#	93
94) Indeno(1,2,3-cd)pyrene	26.086	276	18569716	157.749	ng/ul#	85
95) Dibenzo(a,h)anthracene	26.110	278	15445916	152.970	ng/ul#	91
96) Benzo(g,h,i)perylene	26.845	276	15883870	158.626	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 14 15:23:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 14 13:50:37 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 65.946 ng/uL
RT: 3.217 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion: 88 Resp: 676079
Ion Ratio Lower Upper
88 100
43 27.7 35.8 53.6#
58 76.6 68.1 102.1



#5
Pyridine
Concen: 167.643 ng/uL
RT: 3.611 min Scan# 89
Delta R.T. 0.000 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion: 79 Resp: 4345717
Ion Ratio Lower Upper
79 100
52 75.6 83.8 125.8#
51 34.8 41.0 61.4#









#6
Benzaldehyde
Concen: 90.167 ng/ul
RT: 6.840 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion: 77 Resp: 1369556
Ion Ratio Lower Upper
77 100
105 100.1 76.0 114.0
106 93.6 73.2 109.8



#8
Phenol
Concen: 168.357 ng/ul
RT: 6.905 min Scan# 649
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion: 94 Resp: 5183577
Ion Ratio Lower Upper
94 100
65 27.5 27.4 41.2
66 36.0 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 160.010 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion: 93 Resp: 3956771
Ion Ratio Lower Upper
93 100
63 76.3 72.4 108.6
95 33.0 27.3 40.9



#12
2-Chlorophenol
Concen: 166.740 ng/ul
RT: 7.258 min Scan# 709
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:128 Resp: 4110833
Ion Ratio Lower Upper
128 100
64 46.4 44.8 67.2
130 32.4 25.8 38.6





#13

2-Methylphenol

Concen: 167.970 ng/ul

RT: 8.146 min Scan# 80

Delta R.T. 0.012 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Instrument :

BNA_P

ClientSampleId :





SSTD160622

Tgt Ion:108 Resp: 3850192

Ion Ratio Lower Upper

108 100

107 90.1 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 158.744 ng/ul

RT: 8.234 min Scan# 875

Delta R.T. 0.018 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57




Tgt Ion: 45 Resp: 5266843


Ion Ratio Lower Upper

45 100

77 15.0 11.2 16.8

79 11.7 6.9 10.3#









#16
Acetophenone
Concen: 158.565 ng/ul
RT: 8.528 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:105 Resp: 5541639
Ion Ratio Lower Upper
105 100
77 76.5 67.5 101.3
51 26.2 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 159.707 ng/ul
RT: 8.534 min Scan# 926
Delta R.T. 0.030 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion: 70 Resp: 2822318
Ion Ratio Lower Upper
70 100
42 54.9 53.8 80.8
101 10.2 7.8 11.6
130 22.7 15.4 23.2









#18
4-Methylphenol
Concen: 169.339 ng/ul
RT: 8.493 min Scan# 914
Delta R.T. 0.030 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:108 Resp: 4042270
Ion Ratio Lower Upper
108 100
107 110.0 88.3 132.5



#19
Hexachloroethane
Concen: 163.617 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:117 Resp: 1633612
Ion Ratio Lower Upper
117 100
201 88.2 85.7 128.5
199 56.1 57.4 86.0#









#22
Nitrobenzene
Concen: 175.545 ng/ul
RT: 8.905 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion: 77 Resp: 4273084
Ion Ratio Lower Upper
77 100
123 48.3 31.5 47.3#
65 13.8 11.4 17.2



#23
Isophorone
Concen: 160.635 ng/ul
RT: 9.446 min Scan# 1081
Delta R.T. 0.024 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion: 82 Resp: 8199927
Ion Ratio Lower Upper
82 100
95 7.2 6.6 10.0
138 18.7 13.8 20.6









#25
2-Nitrophenol
Concen: 227.740 ng/ul
RT: 9.610 min Scan# 1109
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:139 Resp: 2225586
Ion Ratio Lower Upper
139 100
65 49.5 54.2 81.4#
109 28.8 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 159.888 ng/ul
RT: 9.681 min Scan# 1121
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:107 Resp: 4268908
Ion Ratio Lower Upper
107 100
121 52.3 39.1 58.7
122 87.3 66.1 99.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 157.604 ng/ul
 RT: 9.916 min Scan# 1159
 Delta R.T. 0.012 min
 Lab File: BP010973.D
 Acq: 14 Jul 2022 14:57




Instrument :
 BNA_P
 ClientSampleId :
 SSTD160622


Tgt Ion: 93 Resp: 5075136
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.6 39.8
 123 11.5 7.8 11.8



#29
 2,4-Dichlorophenol
 Concen: 165.380 ng/ul
 RT: 10.146 min Scan# 1200
 Delta R.T. 0.012 min
 Lab File: BP010973.D
 Acq: 14 Jul 2022 14:57

Tgt Ion: 162 Resp: 3426540
 Ion Ratio Lower Upper
 162 100
 164 64.8 52.5 78.7
 98 37.7 31.4 47.0









#30
Naphthalene
Concen: 148.868 ng/ul
RT: 10.540 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:128 Resp:11888383
Ion Ratio Lower Upper
128 100
129 11.4 8.7 13.1
127 13.7 10.2 15.4



#32
4-Chloroaniline
Concen: 162.372 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:127 Resp: 5236137
Ion Ratio Lower Upper
127 100
129 32.3 26.2 39.2





#33

Hexachlorobutadiene

Concen: 147.134 ng/ul

RT: 10.822 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:225 Resp: 1947599

Ion Ratio Lower Upper

225 100

223 62.3 53.7 80.5

227 64.1 49.4 74.2



#34

Caprolactam

Concen: 108.219 ng/ul

RT: 11.504 min Scan# 1431

Delta R.T. 0.059 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57




Tgt Ion:113 Resp: 704380


Ion Ratio Lower Upper

113 100

55 164.7 153.8 230.8

56 122.4 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 177.057 ng/ul
RT: 11.804 min Scan# 1479
Delta R.T. 0.018 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:107 Resp: 3975993
Ion Ratio Lower Upper
107 100
144 26.0 18.6 27.8
142 79.3 59.6 89.4



#36
2-Methylnaphthalene
Concen: 151.977 ng/ul
RT: 12.157 min Scan# 1542
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:142 Resp: 7859041
Ion Ratio Lower Upper
142 100
141 89.5 70.1 105.1









#37
1-Methylnaphthalene
Concen: 151.622 ng/ul
RT: 12.381 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:142 Resp: 7789488
Ion Ratio Lower Upper
142 100
141 92.1 73.3 109.9
116 3.9 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 146.448 ng/ul
RT: 12.528 min Scan# 1605
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:216 Resp: 3592722
Ion Ratio Lower Upper
216 100
214 78.8 63.0 94.6
179 23.0 18.1 27.1
108 21.0 13.6 20.4#









#40
Hexachlorocyclopentadiene
Concen: 158.093 ng/ul
RT: 12.504 min Scan# 1601
Delta R.T. 0.000 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:237 Resp: 2457610
Ion Ratio Lower Upper
237 100
235 63.5 52.8 79.2
272 12.5 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 180.384 ng/ul
RT: 12.781 min Scan# 1648
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:196 Resp: 2551322
Ion Ratio Lower Upper
196 100
198 95.1 78.2 117.4
200 30.9 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 181.486 ng/ul
RT: 12.851 min Scan# 1660
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:196 Resp: 2754574
Ion Ratio Lower Upper
196 100
198 95.5 77.0 115.4
200 30.5 24.8 37.2



#43
1,1'-Biphenyl
Concen: 143.864 ng/ul
RT: 13.187 min Scan# 1717
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:154 Resp: 9668613
Ion Ratio Lower Upper
154 100
153 41.7 31.9 47.9
76 14.4 9.2 13.8#





#44

2-Chloronaphthalene

Concen: 146.443 ng/ul

RT: 13.222 min Scan# 1722

Delta R.T. 0.006 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:162 Resp: 7413976

Ion Ratio Lower Upper

162 100

164 33.6 25.5 38.3

127 43.8 31.9 47.9



#45

2-Nitroaniline

Concen: 241.503 ng/ul

RT: 13.446 min Scan# 1761

Delta R.T. 0.012 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57




Tgt Ion: 65 Resp: 2594099


Ion Ratio Lower Upper

65 100

92 71.3 49.0 73.6

138 110.4 66.4 99.6#









#47
Dimethylphthalate
Concen: 156.971 ng/ul
RT: 13.828 min Scan# 1826
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:163 Resp: 9192962
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.6 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 237.449 ng/ul
RT: 13.951 min Scan# 1847
Delta R.T. 0.018 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:165 Resp: 2083585
Ion Ratio Lower Upper
165 100
89 57.2 52.9 79.3
121 21.4 19.4 29.2









#50
Acenaphthylene
Concen: 144.530 ng/ul
RT: 14.081 min Scan# 1869
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:152 Resp:11814022
Ion Ratio Lower Upper
152 100
151 21.0 16.5 24.7
153 13.8 10.1 15.1



#51
3-Nitroaniline
Concen: 196.922 ng/ul
RT: 14.281 min Scan# 1903
Delta R.T. 0.018 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:138 Resp: 2071594
Ion Ratio Lower Upper
138 100
108 11.0 11.0 16.4#
92 116.7 104.6 156.8









#52
Acenaphthene
Concen: 149.455 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:153 Resp: 7882651
Ion Ratio Lower Upper
153 100
152 47.9 36.2 54.2
154 88.8 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 387.339 ng/ul
RT: 14.481 min Scan# 1937
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:184 Resp: 1304840
Ion Ratio Lower Upper
184 100
63 71.3 73.8 110.6#
154 62.9 57.2 85.8









#55
4-Nitrophenol
Concen: 211.360 ng/ul
RT: 14.610 min Scan# 1959
Delta R.T. 0.035 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:109 Resp: 1535100
Ion Ratio Lower Upper
109 100
139 141.5 75.6 113.4#
65 133.9 107.6 161.4



#56
Dibenzofuran
Concen: 145.291 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:168 Resp:10496541
Ion Ratio Lower Upper
168 100
139 40.9 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 245.326 ng/ul
RT: 14.740 min Scan# 1981
Delta R.T. 0.018 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:165 Resp: 2876468
Ion Ratio Lower Upper
165 100
63 47.3 47.0 70.6
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 198.155 ng/ul
RT: 14.987 min Scan# 2023
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:232 Resp: 2153766
Ion Ratio Lower Upper
232 100
131 55.9 35.8 53.8#
130 3.0 1.8 2.8#
166 36.2 23.7 35.5#









#59
Diethylphthalate
Concen: 165.288 ng/ul
RT: 15.210 min Scan# 2061
Delta R.T. 0.018 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:149 Resp: 9623692
Ion Ratio Lower Upper
149 100
177 21.7 17.6 26.4
150 12.8 10.3 15.5



#61
Fluorene
Concen: 144.145 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:166 Resp: 8303684
Ion Ratio Lower Upper
166 100
165 98.6 77.4 116.2
167 14.0 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 141.988 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:204 Resp: 3828975
Ion Ratio Lower Upper
204 100
206 32.8 26.0 39.0
141 69.8 49.3 73.9



#63
4-Nitroaniline
Concen: 12.818 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. -0.017 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:138 Resp: 124677
Ion Ratio Lower Upper
138 100
92 24.8 46.2 69.2#
108 5.6 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 264.778 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.024 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:198 Resp: 1633281
Ion Ratio Lower Upper
198 100
182 3.5 4.3 6.5#
77 25.2 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 145.511 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:169 Resp: 7593618
Ion Ratio Lower Upper
169 100
168 67.4 52.6 78.8
167 36.9 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 148.938 ng/ul
RT: 16.316 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:248 Resp: 2571819
Ion Ratio Lower Upper
248 100
250 97.3 80.8 121.2
141 82.7 63.9 95.9



#69
Hexachlorobenzene
Concen: 143.933 ng/ul
RT: 16.422 min Scan# 2267
Delta R.T. 0.012 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:284 Resp: 2878728
Ion Ratio Lower Upper
284 100
142 38.8 27.8 41.6
249 32.1 25.4 38.2





#70

Atrazine

Concen: 158.438 ng/ul

RT: 16.610 min Scan# 21

Delta R.T. 0.018 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:200 Resp: 2741322

Ion Ratio Lower Upper

200 100

173 28.7 20.7 31.1

215 47.2 38.6 57.8



#71

Pentachlorophenol

Concen: 188.418 ng/ul

RT: 16.769 min Scan# 2326

Delta R.T. 0.006 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57




Tgt Ion:266 Resp: 1940421


Ion Ratio Lower Upper

266 100

264 63.9 51.4 77.2

268 63.8 53.5 80.3





#72

Phenanthrene

Concen: 143.228 ng/ul

RT: 17.169 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:178 Resp:13141779

Ion Ratio Lower Upper

178 100

179 17.5 12.7 19.1

176 21.8 14.6 22.0



#74

Anthracene

Concen: 137.901 ng/ul

RT: 17.263 min Scan# 2410

Delta R.T. 0.012 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57




Tgt Ion:178 Resp:12725915


Ion Ratio Lower Upper

178 100

179 18.0 12.2 18.2

176 21.8 14.5 21.7#





#75

1,2,3,4-Tetrachlorobenzene

Concen: 133.422 ng/uL

RT: 13.146 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:216 Resp: 3793484

Ion Ratio Lower Upper

216 100

214 79.4 65.0 97.4

218 48.2 39.1 58.7



#76

Pentachlorobenzene

Concen: 139.286 ng/uL

RT: 14.675 min Scan# 1970

Delta R.T. 0.006 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57




Tgt Ion:250 Resp: 3396450


Ion Ratio Lower Upper

250 100

248 63.8 49.8 74.6

252 63.7 51.2 76.8





#77

Carbazole

Concen: 153.086 ng/ul

RT: 17.539 min Scan# 2456

Delta R.T. 0.006 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:167 Resp:12716558

Ion Ratio Lower Upper

167 100

166 22.8 16.9 25.3

139 14.9 10.6 16.0



#78

Di-n-butylphthalate

Concen: 134.564 ng/ul

RT: 18.104 min Scan# 2553

Delta R.T. 0.000 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57




Tgt Ion:149 Resp:13320103


Ion Ratio Lower Upper

149 100

150 14.9 7.1 10.7#

104 8.4 4.6 6.8#









#80
Fluoranthene
Concen: 132.294 ng/u1
RT: 19.151 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:202 Resp:13602638
Ion Ratio Lower Upper
202 100
101 18.7 7.1 10.7#
100 14.4 5.7 8.5#



#82
Pyrene
Concen: 125.492 ng/u1
RT: 19.510 min Scan# 2792
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:202 Resp:13030914
Ion Ratio Lower Upper
202 100
101 22.7 7.9 11.9#
100 18.5 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 210.510 ng/ul
RT: 20.404 min Scan# 2944
Delta R.T. 0.000 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:149 Resp: 7694130
Ion Ratio Lower Upper
149 100
91 74.6 60.0 90.0
206 20.5 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 132.289 ng/ul
RT: 21.175 min Scan# 3075
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:252 Resp: 4191884
Ion Ratio Lower Upper
252 100
254 64.6 51.0 76.4
126 11.9 8.6 12.8









#85
Benzo(a)anthracene
Concen: 140.018 ng/ul
RT: 21.233 min Scan# 3084
Delta R.T. 0.006 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:228 Resp:13437870
Ion Ratio Lower Upper
228 100
229 25.9 15.0 22.4#
226 35.9 20.6 31.0#



#86
Bis(2-ethylhexyl)phthalate
Concen: 177.892 ng/ul
RT: 21.169 min Scan# 3074
Delta R.T. 0.000 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:149 Resp:10296217
Ion Ratio Lower Upper
149 100
167 30.3 22.6 33.8
279 4.3 4.2 6.4





#87

Chrysene

Concen: 133.712 ng/ul

RT: 21.286 min Scan# 3093

Delta R.T. 0.006 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:228 Resp:12297159

Ion Ratio Lower Upper

228 100

226 39.8 23.6 35.4#

229 26.3 15.6 23.4#



#89

Di-n-octyl phthalate

Concen: 154.290 ng/ul




RT: 22.080 min Scan# 3229


Delta R.T. -0.006 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Tgt Ion:149 Resp:14768752









#90
Benzo(b)fluoranthene
Concen: 161.836 ng/ul
RT: 22.898 min Scan# 3368
Delta R.T. 0.018 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:252 Resp:16229723
Ion Ratio Lower Upper
252 100
253 23.3 17.8 26.6
125 13.5 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 148.894 ng/ul
RT: 22.945 min Scan# 3376
Delta R.T. 0.018 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:252 Resp:14193073
Ion Ratio Lower Upper
252 100
253 23.4 17.0 25.4
125 13.6 6.2 9.4#









#93
Benzo(a)pyrene
Concen: 156.639 ng/ul
RT: 23.521 min Scan# 3470
Delta R.T. 0.024 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:252 Resp:15281026
Ion Ratio Lower Upper
252 100
253 23.2 17.4 26.0
125 14.7 6.6 10.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 157.749 ng/ul
RT: 26.086 min Scan# 3910
Delta R.T. 0.041 min
Lab File: BP010973.D
Acq: 14 Jul 2022 14:57

Tgt Ion:276 Resp:18569716
Ion Ratio Lower Upper
276 100
138 31.8 14.6 22.0#
277 25.9 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 152.970 ng/ul

RT: 26.110 min Scan# 3907

Delta R.T. 0.041 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:278 Resp:15445916

Ion Ratio Lower Upper

278 100

139 20.4 9.4 14.2#

279 24.7 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 158.626 ng/ul

RT: 26.845 min Scan# 4039

Delta R.T. 0.053 min

Lab File: BP010973.D

Acq: 14 Jul 2022 14:57

Tgt Ion:276 Resp:15883870

Ion Ratio Lower Upper

276 100

138 28.7 12.7 19.1#

277 24.2 18.2 27.4

