

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071524\
 Data File : BP020965.D
 Acq On : 15 Jul 2024 19:37
 Operator : MA/JU
 Sample : P3087-02MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A4CB7MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

Quant Time: Jul 15 23:04:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 15 12:22:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	204020	20.000	ng/u1	0.01
20) Naphthalene-d8	10.457	136	832574	20.000	ng/u1	0.01
38) Acenaphthene-d10	14.310	164	549086	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.128	188	1165498	20.000	ng/u1	0.01
79) Chrysene-d12	21.557	240	1097098	20.000	ng/u1	0.00
88) Perylene-d12	24.857	264	1320265	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	14914m	3.327	ng/uL	0.01
4) Pyridine-d5	3.670	84	75094	5.785	ng/u1	0.02
7) Phenol-d5	6.905	99	341832	20.551	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.052	67	202895	20.128	ng/u1	0.02
11) 2-Chlorophenol-d4	7.246	132	284758	20.779	ng/u1	0.01
15) 4-Methylphenol-d8	8.410	113	246494	17.843	ng/u1	0.02
21) Nitrobenzene-d5	8.852	128	144994	22.422	ng/u1	0.02
24) 2-Nitrophenol-d4	9.557	143	165791	22.400	ng/u1	0.02
28) 2,4-Dichlorophenol-d3	10.099	165	316363	23.638	ng/u1	0.01
31) 4-Chloroaniline-d4	10.604	131	248656	14.024	ng/u1	0.02
46) Dimethylphthalate-d6	13.722	166	968578	24.266	ng/u1	0.01
49) Acenaphthylene-d8	13.998	160	1058174	23.610	ng/u1	0.00
54) 4-Nitrophenol-d4	14.569	143	146835	25.855	ng/u1	0.00
60) Fluorene-d10	15.316	176	829590	25.334	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.469	200	166744	23.645	ng/u1	0.00
73) Anthracene-d10	17.228	188	1252784	24.201	ng/u1	0.00
81) Pyrene-d10	19.604	212	1361765	20.077	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.627	264	1070844	16.445	ng/u1	0.02
Target Compounds						
2) 1,4-Dioxane	3.287	88	43394	8.800	ng/uL	98
5) Pyridine	3.687	79	107772	8.050	ng/u1	99
6) Benzaldehyde	6.869	77	129124	15.398	ng/u1	99
8) Phenol	6.928	94	424764	25.237	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.140	93	316682	22.776	ng/u1	99
12) 2-Chlorophenol	7.281	128	340566	24.795	ng/u1	99
13) 2-Methylphenol	8.152	108	298170	22.907	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.222	45	466182m	23.207	ng/u1	
16) Acetophenone	8.522	105	541029	25.375	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.499	70	260692	23.335	ng/u1	98
18) 4-Methylphenol	8.475	108	337311	24.189	ng/u1	96
19) Hexachloroethane	8.752	117	112875	18.287	ng/u1	98
22) Nitrobenzene	8.893	77	383273	24.742	ng/u1	99
23) Isophorone	9.405	82	741332	24.072	ng/u1	100
25) 2-Nitrophenol	9.587	139	202466	26.228	ng/u1	99
26) 2,4-Dimethylphenol	9.657	107	170259	11.185	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.869	93	446849	23.871	ng/u1	99
29) 2,4-Dichlorophenol	10.128	162	353747	26.819	ng/u1	98
30) Naphthalene	10.504	128	1114796	25.455	ng/u1	99
32) 4-Chloroaniline	10.634	127	259012	15.298	ng/u1	99
33) Hexachlorobutadiene	10.775	225	242197	24.489	ng/u1	98
34) Caprolactam	11.434	113	98662	24.282	ng/u1	97
35) 4-Chloro-3-methylphenol	11.763	107	388912	27.126	ng/u1	98
36) 2-Methylnaphthalene	12.104	142	784919	26.110	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071524\
 Data File : BP020965.D
 Acq On : 15 Jul 2024 19:37
 Operator : MA/JU
 Sample : P3087-02MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A4CB7MS


Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

Quant Time: Jul 15 23:04:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 15 12:22:22 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
37) 1-Methylnaphthalene	12.346	142	816357	26.404	ng/ul	99	07/16/2024
39) 1,2,4,5-Tetrachloroben...	12.475	216	466927	26.185	ng/ul	97	Supervised By :mohammad ahmed
40) Hexachlorocyclopentadiene	12.446	237	91334	9.791	ng/ul	99	
41) 2,4,6-Trichlorophenol	12.740	196	297925	26.441	ng/ul	99	
42) 2,4,5-Trichlorophenol	12.822	196	335551	28.146	ng/ul	99	
43) 1,1'-Biphenyl	13.134	154	1041940	25.965	ng/ul	99	
44) 2-Chloronaphthalene	13.181	162	830297	25.918	ng/ul	99	07/18/2024
45) 2-Nitroaniline	13.404	65	256686	28.975	ng/ul	98	
47) Dimethylphthalate	13.775	163	1081605	26.706	ng/ul	100	
48) 2,6-Dinitrotoluene	13.904	165	225476	27.846	ng/ul	99	
50) Acenaphthylene	14.028	152	1330880	26.226	ng/ul	99	
51) 3-Nitroaniline	14.245	138	198712	28.201	ng/ul	98	
52) Acenaphthene	14.369	153	913196	27.113	ng/ul	99	
53) 2,4-Dinitrophenol	14.463	184	107264	25.225	ng/ul	98	
55) 4-Nitrophenol	14.587	109	140551	30.424	ng/ul	96	
56) Dibenzofuran	14.716	168	1288725	27.941	ng/ul	98	
57) 2,4-Dinitrotoluene	14.704	165	333614	30.014	ng/ul	96	
58) 2,3,4,6-Tetrachlorophenol	14.963	232	272214	26.665	ng/ul	99	
59) Diethylphthalate	15.151	149	1088627	26.896	ng/ul	99	
61) Fluorene	15.381	166	1087047	28.882	ng/ul	100	
62) 4-Chlorophenyl-phenyle...	15.381	204	541693	26.523	ng/ul	95	
63) 4-Nitroaniline	15.428	138	204719	35.083	ng/ul	96	
66) 4,6-Dinitro-2-methylph...	15.487	198	203180	27.169	ng/ul	98	
67) N-Nitrosodiphenylamine	15.604	169	895039	26.372	ng/ul	99	
68) 4-Bromophenyl-phenylether	16.292	248	358290	26.227	ng/ul	96	
69) Hexachlorobenzene	16.404	284	337690	21.559	ng/ul	98	
70) Atrazine	16.581	200	237668	19.214	ng/ul	99	
71) Pentachlorophenol	16.769	266	181310	21.312	ng/ul	97	
72) Phenanthrene	17.169	178	2584604	42.748	ng/ul	99	
74) Anthracene	17.257	178	1792140	29.075	ng/ul	100	
75) 1,2,3,4-Tetrachloroben...	13.104	216	482778	25.880	ng/ul	99	
76) Pentachlorobenzene	14.634	250	409135	22.072	ng/ul	99	
77) Carbazole	17.551	167	1532000	30.405	ng/ul	99	
78) Di-n-butylphthalate	18.128	149	1990724	28.636	ng/ul	99	
80) Fluoranthene	19.263	202	3833160	46.796	ng/ul	100	
82) Pyrene	19.639	202	3006395	36.069	ng/ul	99	
83) Butylbenzylphthalate	20.580	149	920592	26.508	ng/ul	98	
84) 3,3'-Dichlorobenzidine	21.463	252	412951	16.472	ng/ul	99	
85) Benzo(a)anthracene	21.539	228	2912409	37.896	ng/ul	99	
86) Bis(2-ethylhexyl)phtha...	21.457	149	2241572	44.852	ng/ul	99	
87) Chrysene	21.604	228	2851028	39.923	ng/ul	99	
89) Di-n-octyl phthalate	22.698	149	2319859	26.599	ng/ul	100	
90) Benzo(b)fluoranthene	23.804	252	3307512	41.282	ng/ul	99	
91) Benzo(k)fluoranthene	23.880	252	2689463	33.636	ng/ul	99	
93) Benzo(a)pyrene	24.698	252	1427290	18.852	ng/ul	99	
94) Indeno(1,2,3-cd)pyrene	28.598	276	2415043	24.754	ng/ul	100	
95) Dibenzo(a,h)anthracene	28.656	278	2375692	29.406	ng/ul	99	
96) Benzo(g,h,i)perylene	29.727	276	253629m	3.192	ng/ul		

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.800 ng/uL
RT: 3.287 min Scan# 51
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion: 88 Resp: 43394
Ion Ratio Lower Upper
88 100
43 31.1 25.4 38.0
58 76.0 59.2 88.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#5
Pyridine
Concen: 8.050 ng/uL
RT: 3.687 min Scan# 119
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion: 79 Resp: 107772
Ion Ratio Lower Upper
79 100
52 81.8 64.9 97.3
51 39.2 30.5 45.7





#6
Benzaldehyde
Concen: 15.398 ng/ul
RT: 6.869 min Scan# 60
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion: 77 Resp: 129124
Ion Ratio Lower Upper
77 100
105 99.8 79.3 118.9
106 94.4 74.6 112.0


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




07/16/2024
Supervised By :mohammad
ahmed



07/18/2024




#8
Phenol
Concen: 25.237 ng/ul
RT: 6.928 min Scan# 670
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37




Tgt Ion: 94 Resp: 424764
Ion Ratio Lower Upper
94 100
65 29.0 23.4 35.0
66 40.7 33.9 50.9





#10
Bis(2-Chloroethyl)ether
Concen: 22.776 ng/ul
RT: 7.140 min Scan# 705
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion: 93 Resp: 31668
Ion Ratio Lower Upper
93 100
63 76.8 61.5 92.3
95 32.5 25.1 37.7


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




07/16/2024
Supervised By :mohammad
ahmed



07/18/2024




#12
2-Chlorophenol
Concen: 24.795 ng/ul
RT: 7.281 min Scan# 730
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Tgt Ion:128 Resp: 340566
Ion Ratio Lower Upper
128 100
64 46.5 37.9 56.9
130 32.8 25.6 38.4







#13
2-Methylphenol
Concen: 22.907 ng/ul
RT: 8.152 min Scan# 81
Delta R.T. 0.018 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Instrument :
BNA_P
ClientSampleId :
A4CB7MS

Tgt Ion:108 Resp: 298170
Ion Ratio Lower Upper
108 100
107 88.2 73.8 110.8



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#14
2,2'-oxybis(1-Chloropropane)
Concen: 23.207 ng/ul m
RT: 8.222 min Scan# 890
Delta R.T. 0.018 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion: 45 Resp: 466182
Ion Ratio Lower Upper
45 100
77 13.8 10.9 16.3
79 10.6 8.1 12.1





#16
Acetophenone
Concen: 25.375 ng/ul
RT: 8.522 min Scan# 941
Delta R.T. 0.023 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:105 Resp: 541029
Ion Ratio Lower Upper
105 100
77 76.1 61.8 92.8
51 27.3 21.5 32.3


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#17
N-Nitroso-di-n-propylamine
Concen: 23.335 ng/ul
RT: 8.499 min Scan# 937
Delta R.T. 0.018 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion: 70 Resp: 260692
Ion Ratio Lower Upper
70 100
42 56.3 46.6 69.8
101 10.0 7.4 11.0
130 22.5 17.2 25.8





#18
4-Methylphenol
Concen: 24.189 ng/ul
RT: 8.475 min Scan# 91
Delta R.T. 0.018 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:108 Resp: 33731
Ion Ratio Lower Upper
108 100
107 114.3 87.8 131.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#19
Hexachloroethane
Concen: 18.287 ng/ul
RT: 8.752 min Scan# 980
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion:117 Resp: 112875
Ion Ratio Lower Upper
117 100
201 111.8 88.2 132.2
199 72.6 55.8 83.6





#22
Nitrobenzene
Concen: 24.742 ng/ul
RT: 8.893 min Scan# 1004
Delta R.T. 0.023 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion: 77 Resp: 38327
Ion Ratio Lower Upper
77 100
123 46.3 36.7 55.1
65 13.7 11.1 16.7


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




07/16/2024
Supervised By :mohammad
ahmed



07/18/2024




#23
Isophorone
Concen: 24.072 ng/ul
RT: 9.405 min Scan# 1091
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Tgt Ion: 82 Resp: 741332
Ion Ratio Lower Upper
82 100
95 7.9 6.1 9.1
138 17.9 14.4 21.6







#25
2-Nitrophenol
Concen: 26.228 ng/ul
RT: 9.587 min Scan# 1122
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Instrument :
BNA_P
ClientSampleId :
A4CB7MS

Tgt Ion:139 Resp: 202460
Ion Ratio Lower Upper
139 100
65 52.6 42.9 64.3
109 32.1 26.2 39.4



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#26
2,4-Dimethylphenol
Concen: 11.185 ng/ul
RT: 9.657 min Scan# 1134
Delta R.T. 0.018 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Tgt Ion:107 Resp: 170259
Ion Ratio Lower Upper
107 100
121 51.3 39.6 59.4
122 84.2 66.3 99.5





#27

Bis(2-Chloroethoxy)methane

Concen: 23.871 ng/ul

RT: 9.869 min Scan# 1170

Delta R.T. 0.006 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37

Instrument :

BNP_P



ClientSampleId :

A4CB7MS

Tgt Ion: 93	Resp: 446849
Ion Ratio	Lower Upper
93 100	
95 32.9	25.5 38.3
123 11.1	9.0 13.4



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel



07/16/2024
Supervised By :mohammad
ahmed

07/18/2024



#29

2,4-Dichlorophenol

Concen: 26.819 ng/ul




RT: 10.128 min Scan# 1214


Delta R.T. 0.018 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37


Tgt Ion:162	Resp: 353747
Ion Ratio	Lower Upper
162 100	
164 65.0	53.3 79.9
98 35.3	28.7 43.1





#30
Naphthalene
Concen: 25.455 ng/ul
RT: 10.504 min Scan# 1114790
Delta R.T. 0.006 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:128 Resp: 1114790
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 13.1 10.8 16.2


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#32
4-Chloroaniline
Concen: 15.298 ng/ul
RT: 10.634 min Scan# 1300
Delta R.T. 0.023 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion:127 Resp: 259012
Ion Ratio Lower Upper
127 100
129 31.9 26.1 39.1





#33
Hexachlorobutadiene
Concen: 24.489 ng/ul
RT: 10.775 min Scan# 11
Delta R.T. 0.018 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:225 Resp: 24219
Ion Ratio Lower Upper
225 100
223 61.1 50.2 75.2
227 62.4 50.6 75.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#34
Caprolactam
Concen: 24.282 ng/ul
RT: 11.434 min Scan# 1436
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion:113 Resp: 98662
Ion Ratio Lower Upper
113 100
55 171.2 140.2 210.4
56 129.2 105.5 158.3





#35
4-Chloro-3-methylphenol
Concen: 27.126 ng/ul
RT: 11.763 min Scan# 1492
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:107 Resp: 388912
Ion Ratio Lower Upper
107 100
144 25.8 19.3 28.9
142 77.3 60.7 91.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




07/16/2024
Supervised By :mohammad
ahmed



07/18/2024




#36
2-Methylnaphthalene
Concen: 26.110 ng/ul
RT: 12.104 min Scan# 1550
Delta R.T. -0.000 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37




Tgt Ion:142 Resp: 784919
Ion Ratio Lower Upper
142 100
141 89.9 72.6 108.8





#37
1-Methylnaphthalene
Concen: 26.404 ng/ul
RT: 12.346 min Scan# 1590
Delta R.T. 0.023 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:142 Resp: 81635
Ion Ratio Lower Upper
142 100
141 92.8 75.0 112.6
116 3.8 3.2 4.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#39
1,2,4,5-Tetrachlorobenzene
Concen: 26.185 ng/ul
RT: 12.475 min Scan# 1613
Delta R.T. -0.000 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion:216 Resp: 466927
Ion Ratio Lower Upper
216 100
214 77.5 64.2 96.4
179 20.3 16.6 24.8
108 14.3 13.4 20.2





#40
Hexachlorocyclopentadiene
Concen: 9.791 ng/ul
RT: 12.446 min Scan# 1608
Delta R.T. -0.000 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:237 Resp: 91334
Ion Ratio Lower Upper
237 100
235 61.4 49.7 74.5
272 12.7 10.2 15.2


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




07/16/2024
Supervised By :mohammad
ahmed



07/18/2024




#41
2,4,6-Trichlorophenol
Concen: 26.441 ng/ul
RT: 12.740 min Scan# 1658
Delta R.T. 0.006 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37




Tgt Ion:196 Resp: 297925
Ion Ratio Lower Upper
196 100
198 95.4 76.6 115.0
200 30.8 26.0 39.0





#42
2,4,5-Trichlorophenol
Concen: 28.146 ng/ul
RT: 12.822 min Scan# 1672
Delta R.T. 0.006 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:196 Resp: 33555
Ion Ratio Lower Upper
196 100
198 96.8 77.1 115.7
200 31.0 25.3 37.9


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#43
1,1'-Biphenyl
Concen: 25.965 ng/ul
RT: 13.134 min Scan# 1725
Delta R.T. -0.000 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Tgt Ion:154 Resp: 1041940
Ion Ratio Lower Upper
154 100
153 41.4 32.7 49.1
76 12.8 9.4 14.2







#44
 2-Chloronaphthalene
 Concen: 25.918 ng/ul
 RT: 13.181 min Scan# 1733
 Delta R.T. 0.006 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37


Instrument : BNA_P
 ClientSampleId : A4CB7MS

Tgt Ion:	162	Resp:	83029
Ion Ratio	Lower	Upper	
162	100		
164	31.9	26.3	39.5
127	39.1	31.8	47.8


Manual Integrations
APPROVED
 Reviewed By :Yogesh Patel



07/16/2024
 Supervised By :mohammad ahmed







07/18/2024



#45
 2-Nitroaniline
 Concen: 28.975 ng/ul
 RT: 13.404 min Scan# 1771
 Delta R.T. 0.006 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37


Tgt Ion:	65	Resp:	256686
Ion Ratio	Lower	Upper	
65	100		
92	67.7	56.7	85.1
138	107.5	85.6	128.4





#47
Dimethylphthalate
Concen: 26.706 ng/ul
RT: 13.775 min Scan# 1834
Delta R.T. 0.018 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:163 Resp: 108160
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 9.9 8.1 12.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




07/16/2024
Supervised By :mohammad
ahmed



07/18/2024




#48
2,6-Dinitrotoluene
Concen: 27.846 ng/ul
RT: 13.904 min Scan# 1856
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Tgt Ion:165 Resp: 225476
Ion Ratio Lower Upper
165 100
89 56.4 44.7 67.1
121 21.1 16.5 24.7







#50
 Acenaphthylene
 Concen: 26.226 ng/ul
 RT: 14.028 min Scan# 1877
 Delta R.T. -0.000 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37

Instrument : BNA_P
 ClientSampleId : A4CB7MS

Tgt Ion:152 Resp: 1330880



Ion	Ratio	Lower	Upper
152	100		
151	19.7	16.2	24.2
153	12.7	10.3	15.5

Manual Integrations
APPROVED
 Reviewed By :Yogesh Patel



07/16/2024
 Supervised By :mohammad ahmed




07/18/2024




#51
 3-Nitroaniline
 Concen: 28.201 ng/ul
 RT: 14.245 min Scan# 1914
 Delta R.T. 0.012 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37

Tgt Ion:138 Resp: 198712


Ion	Ratio	Lower	Upper
138	100		
108	11.3	9.8	14.8
92	120.0	98.0	147.0





#52
Acenaphthene
Concen: 27.113 ng/ul
RT: 14.369 min Scan# 1935
Delta R.T. -0.000 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:153 Resp: 913190
Ion Ratio Lower Upper
153 100
152 47.0 37.8 56.6
154 88.1 71.8 107.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




07/16/2024
Supervised By :mohammad
ahmed



07/18/2024




#53
2,4-Dinitrophenol
Concen: 25.225 ng/ul
RT: 14.463 min Scan# 1951
Delta R.T. 0.006 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37




Tgt Ion:184 Resp: 107264
Ion Ratio Lower Upper
184 100
63 62.1 50.3 75.5
154 64.0 53.0 79.6





#55
4-Nitrophenol
Concen: 30.424 ng/ul
RT: 14.587 min Scan# 1973
Delta R.T. 0.006 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:109 Resp: 14055
Ion Ratio Lower Upper
109 100
139 128.0 98.0 147.0
65 129.2 106.0 159.0


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#56
Dibenzofuran
Concen: 27.941 ng/ul
RT: 14.716 min Scan# 1994
Delta R.T. 0.006 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion:168 Resp: 1288725
Ion Ratio Lower Upper
168 100
139 37.5 31.1 46.7





#57
2,4-Dinitrotoluene
Concen: 30.014 ng/ul
RT: 14.704 min Scan# 1992
Delta R.T. -0.006 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:165 Resp: 333614
Ion Ratio Lower Upper
165 100
63 53.8 41.0 61.4
182 2.9 2.2 3.4


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#58
2,3,4,6-Tetrachlorophenol
Concen: 26.665 ng/ul
RT: 14.963 min Scan# 2036
Delta R.T. 0.018 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Tgt Ion:232 Resp: 272214
Ion Ratio Lower Upper
232 100
131 43.4 34.4 51.6
130 2.4 2.0 3.0
166 30.7 24.4 36.6





#59
Diethylphthalate
Concen: 26.896 ng/ul
RT: 15.151 min Scan# 2068
Delta R.T. -0.000 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Instrument :

BNP_P



ClientSampleId :

A4CB7MS

Tgt Ion:149 Resp: 108862
Ion Ratio Lower Upper
149 100
177 22.9 17.7 26.5
150 13.0 10.2 15.2



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#61
Fluorene
Concen: 28.882 ng/ul
RT: 15.381 min Scan# 2107
Delta R.T. 0.006 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion:166 Resp: 1087047
Ion Ratio Lower Upper
166 100
165 97.3 78.1 117.1
167 13.4 10.9 16.3





#62
4-Chlorophenyl-phenylether
Concen: 26.523 ng/ul
RT: 15.381 min Scan# 2107
Delta R.T. -0.000 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:204 Resp: 54169
Ion Ratio Lower Upper
204 100
206 33.4 26.4 39.6
141 55.5 48.5 72.7


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#63
4-Nitroaniline
Concen: 35.083 ng/ul
RT: 15.428 min Scan# 2115
Delta R.T. 0.006 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion:138 Resp: 204719
Ion Ratio Lower Upper
138 100
92 52.3 40.9 61.3
108 85.6 64.7 97.1





#66
4,6-Dinitro-2-methylphenol
Concen: 27.169 ng/ul
RT: 15.487 min Scan# 2125
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:198 Resp: 203180
Ion Ratio Lower Upper
198 100
182 4.4 3.8 5.6
77 23.8 20.0 30.0


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#67
N-Nitrosodiphenylamine
Concen: 26.372 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37


Tgt Ion:169 Resp: 895039
Ion Ratio Lower Upper
169 100
168 66.1 53.4 80.2
167 36.1 28.4 42.6





#68
4-Bromophenyl-phenylether
Concen: 26.227 ng/ul
RT: 16.292 min Scan# 2262
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:248 Resp: 358290
Ion Ratio Lower Upper
248 100
250 96.2 79.3 118.9
141 67.8 57.4 86.0


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#69
Hexachlorobenzene
Concen: 21.559 ng/ul
RT: 16.404 min Scan# 2281
Delta R.T. 0.018 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Tgt Ion:284 Resp: 337690
Ion Ratio Lower Upper
284 100
142 33.8 25.8 38.8
249 31.9 25.1 37.7





#70

Atrazine

Concen: 19.214 ng/ul

RT: 16.581 min Scan# 2312

Delta R.T. 0.006 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37

Instrument :

BNA_P

ClientSampleId :

A4CB7MS

Tgt Ion:200 Resp: 23766

Ion Ratio Lower Upper

200 100

173 25.1 20.8 31.2



215 47.7 38.7 58.1

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



07/16/2024

Supervised By :mohammad
ahmed

07/18/2024



#71

Pentachlorophenol

Concen: 21.312 ng/ul

RT: 16.769 min Scan# 2343

Delta R.T. 0.006 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37




Tgt Ion:266 Resp: 181310


Ion Ratio Lower Upper

266 100

264 61.5 52.3 78.5

268 64.2 52.3 78.5








#72
 Phenanthrene
 Concen: 42.748 ng/ul
 RT: 17.169 min Scan# 2411
 Delta R.T. 0.006 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37

Instrument : BNA_P
 ClientSampleId : A4CB7MS

Tgt Ion:178 Resp: 2584604


Ion	Ratio	Lower	Upper
178	100		
179	15.6	12.2	18.4
176	19.3	15.0	22.4

Manual Integrations
APPROVED
 Reviewed By :Yogesh Patel



07/16/2024
 Supervised By :mohammad ahmed




07/18/2024




#74
 Anthracene
 Concen: 29.075 ng/ul
 RT: 17.257 min Scan# 2426
 Delta R.T. -0.000 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37

Tgt Ion:178 Resp: 1792140


Ion	Ratio	Lower	Upper
178	100		
179	15.6	12.5	18.7
176	18.3	14.8	22.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 25.880 ng/uL
RT: 13.104 min Scan# 1720
Delta R.T. 0.023 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:216 Resp: 482778
Ion Ratio Lower Upper
216 100
214 77.9 63.2 94.8
218 48.5 38.2 57.4


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#76
Pentachlorobenzene
Concen: 22.072 ng/uL
RT: 14.634 min Scan# 1980
Delta R.T. -0.000 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Tgt Ion:250 Resp: 409135
Ion Ratio Lower Upper
250 100
248 62.1 50.8 76.2
252 64.9 51.6 77.4





#77

Carbazole

Concen: 30.405 ng/ul

RT: 17.551 min Scan# 2477

Delta R.T. 0.012 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37

Instrument :

BNA_P

ClientSampleId :

A4CB7MS

Tgt Ion:167 Resp: 1532000

Ion Ratio Lower Upper

167 100

166 21.3 17.3 25.9



139 12.6 10.2 15.2

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



07/16/2024

Supervised By :mohammad
ahmed

07/18/2024



#78

Di-n-butylphthalate

Concen: 28.636 ng/ul

RT: 18.128 min Scan# 2574

Delta R.T. 0.012 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37




Tgt Ion:149 Resp: 1990724


Ion Ratio Lower Upper

149 100

150 8.9 7.4 11.0


104 4.9 3.9 5.9





#80
Fluoranthene
Concen: 46.796 ng/u1
RT: 19.263 min Scan# 2767
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:202 Resp: 3833160
Ion Ratio Lower Upper
202 100
101 9.8 8.0 12.0
100 7.5 6.1 9.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#82
Pyrene
Concen: 36.069 ng/u1
RT: 19.639 min Scan# 2831
Delta R.T. 0.012 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Tgt Ion:202 Resp: 3006395
Ion Ratio Lower Upper
202 100
101 11.7 9.4 14.2
100 10.0 7.8 11.6





#83

Butylbenzylphthalate

Concen: 26.508 ng/ul

RT: 20.580 min Scan# 2991

Delta R.T. 0.012 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37

Instrument :

BNA_P



ClientSampleId :

A4CB7MS

Tgt Ion:	149	Resp:	920592
Ion Ratio	Lower	Upper	
149	100		
91	66.8	54.1	81.1
206	23.3	17.0	25.6



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel



07/16/2024
Supervised By :mohammad
ahmed

07/18/2024



#84

3,3'-Dichlorobenzidine

Concen: 16.472 ng/ul




RT: 21.463 min Scan# 3141


Delta R.T. 0.006 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37

Tgt Ion:	252	Resp:	412951
Ion Ratio	Lower	Upper	
252	100		
254	65.0	51.5	77.3
126	9.9	8.0	12.0





#85

Benzo(a)anthracene

Concen: 37.896 ng/ul

RT: 21.539 min Scan# 3155

Delta R.T. 0.006 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37

Instrument :

BNP_P

ClientSampleId :

A4CB7MS

Tgt Ion:228 Resp: 2912409

Ion Ratio Lower Upper

228 100

229 19.9 15.5 23.3



226 27.4 21.4 32.2

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



07/16/2024

Supervised By :mohammad
ahmed

07/18/2024



#86

Bis(2-ethylhexyl)phthalate

Concen: 44.852 ng/ul

RT: 21.457 min Scan# 3140

Delta R.T. 0.012 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37




Tgt Ion:149 Resp: 2241572


Ion Ratio Lower Upper

149 100

167 27.9 22.1 33.1

279 5.3 3.7 5.5





#87

Chrysene

Concen: 39.923 ng/ul

RT: 21.604 min Scan# 3165

Delta R.T. 0.006 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37

Instrument :

BNA_P

ClientSampleId :

A4CB7MS

Tgt Ion:228 Resp: 2851028

Ion Ratio Lower Upper

228 100



226 29.5 23.6 35.4

229 20.3 15.5 23.3

Manual Integrations



APPROVED

Reviewed By :Yogesh
Patel



07/16/2024
Supervised By :mohammad
ahmed

07/18/2024



#89

Di-n-octyl phthalate

Concen: 26.599 ng/ul




RT: 22.698 min Scan# 3351


Delta R.T. -0.000 min

Lab File: BP020965.D

Acq: 15 Jul 2024 19:37


Tgt Ion:149 Resp: 2319859





#90
Benzo(b)fluoranthene
Concen: 41.282 ng/ul
RT: 23.804 min Scan# 3539
Delta R.T. 0.023 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37



Instrument :
BNA_P
ClientSampleId :
A4CB7MS



Tgt Ion:252 Resp: 330751
Ion Ratio Lower Upper
252 100
253 22.2 17.1 25.7
125 9.0 7.0 10.4


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






07/16/2024
Supervised By :mohammad
ahmed


07/18/2024



#91
Benzo(k)fluoranthene
Concen: 33.636 ng/ul
RT: 23.880 min Scan# 3552
Delta R.T. 0.029 min
Lab File: BP020965.D
Acq: 15 Jul 2024 19:37

Tgt Ion:252 Resp: 2689463
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 8.5 7.4 11.0







#93
 Benzo(a)pyrene
 Concen: 18.852 ng/ul
 RT: 24.698 min Scan# 30
 Delta R.T. 0.029 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37

Instrument : BNA_P
 ClientSampleId : A4CB7MS

Tgt Ion: 252 Resp: 1427290



Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.4	26.0
125	10.2	7.7	11.5

Manual Integrations
APPROVED
 Reviewed By : Yogesh Patel



07/16/2024
 Supervised By : mohammad ahmed




07/18/2024




#94
 Indeno(1,2,3-cd)pyrene
 Concen: 24.754 ng/ul
 RT: 28.598 min Scan# 4354
 Delta R.T. 0.023 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37

Tgt Ion: 276 Resp: 2415043


Ion	Ratio	Lower	Upper
276	100		
138	18.1	14.3	21.5
277	23.8	19.2	28.8





#95
 Dibenzo(a,h)anthracene
 Concen: 29.406 ng/ul
 RT: 28.656 min Scan# 41
 Delta R.T. 0.023 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37


Instrument :
 BNA_P
 ClientSampleId :
 A4CB7MS



Tgt Ion:278 Resp: 2375692
 Ion Ratio Lower Upper
 278 100
 139 12.9 10.9 16.3
 279 23.9 19.4 29.0



Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel



07/16/2024
 Supervised By :mohammad
 ahmed

07/18/2024



#96
 Benzo(g,h,i)perylene
 Concen: 3.192 ng/ul m
 RT: 29.727 min Scan# 4546
 Delta R.T. -0.000 min
 Lab File: BP020965.D
 Acq: 15 Jul 2024 19:37

Tgt Ion:276 Resp: 253629
 Ion Ratio Lower Upper
 276 100
 138 20.2 15.3 22.9
 277 25.0 18.2 27.2

