

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP073021\
 Data File : BP006438.D
 Acq On : 30 Jul 2021 14:59
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC0.1

Quant Time: Jul 30 17:07:45 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270-SIM-BP073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 30 17:01:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.785	152	2939	0.400	ng	0.00
7) Naphthalene-d8	10.546	136	11635	0.400	ng	0.00
13) Acenaphthene-d10	14.391	164	5723	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	11346	0.400	ng	0.00
29) Chrysene-d12	21.253	240	10334	0.400	ng	0.00
35) Perylene-d12	23.589	264	9602	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.357	112	1067	0.135	ng	0.00
5) Phenol-d6	6.941	99	1518	0.152	ng	0.00
8) Nitrobenzene-d5	8.913	82	1511	0.194	ng	0.00
11) 2-Methylnaphthalene-d10	12.145	152	1665	0.091	ng	0.00
14) 2,4,6-Tribromophenol	15.881	330	323	0.139	ng	0.00
15) 2-Fluorobiphenyl	13.025	172	2941	0.135	ng	0.00
27) Fluoranthene-d10	19.140	212	3238	0.101	ng	0.00
31) Terphenyl-d14	19.730	244	2564	0.098	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.325	88	393	0.382	ng	# 100
3) n-Nitrosodimethylamine	3.615	42	439	0.218	ng	# 75
6) bis(2-Chloroethyl)ether	7.205	93	1204	0.153	ng	96
9) Naphthalene	10.609	128	3579	0.119	ng	# 93
10) Hexachlorobutadiene	10.898	225	593	0.101	ng	# 97
12) 2-Methylnaphthalene	12.216	142	2283	0.112	ng	96
16) Acenaphthylene	14.101	152	2651	0.102	ng	98
17) Acenaphthene	14.453	154	2129	0.131	ng	99
18) Fluorene	15.446	166	2626	0.129	ng	99
20) 4,6-Dinitro-2-methylph...	15.529	198	248	0.294	ng	# 56
21) 4-Bromophenyl-phenylether	16.342	248	745	0.110	ng	# 90
22) Hexachlorobenzene	16.441	284	819	0.125	ng	99
23) Atrazine	16.620	200	671	0.111	ng	# 90
24) Pentachlorophenol	16.798	266	407	0.169	ng	94
25) Phenanthrene	17.175	178	3973	0.127	ng	98
26) Anthracene	17.275	178	3543	0.117	ng	98
28) Fluoranthene	19.164	202	5005	0.140	ng	99
30) Pyrene	19.522	202	5431	0.154	ng	99
32) Benzo(a)anthracene	21.236	228	4383	0.125	ng	95
33) Chrysene	21.288	228	4584	0.138	ng	97
36) Indeno(1,2,3-cd)pyrene	26.003	276	4958	0.142	ng	98
37) Benzo(b)fluoranthene	22.873	252	4725	0.145	ng	# 66
38) Benzo(k)fluoranthene	22.922	252	4456	0.139	ng	# 66
39) Benzo(a)pyrene	23.481	252	3818	0.130	ng	# 61
40) Dibenzo(a,h)anthracene	26.025	278	4202	0.142	ng	# 76
41) Benzo(g,h,i)perylene	26.741	276	4498	0.151	ng	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP073021\
Data File : BP006438.D
Acq On : 30 Jul 2021 14:59
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC0.1

Quant Time: Jul 30 17:07:45 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270-SIM-BP073021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 30 17:01:25 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.785 min Scan# 179
 Delta R.T. 0.000 min
 Lab File: BP006438.D
 Acq: 30 Jul 2021 14:59




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC0.1


Tgt Ion: 152 Resp: 2939
 Ion Ratio Lower Upper
 152 100
 150 129.2 105.7 158.5
 115 52.1 41.6 62.4



#2
 1,4-Dioxane
 Concen: 0.382 ng
 RT: 3.325 min Scan# 10
 Delta R.T. 0.000 min
 Lab File: BP006438.D
 Acq: 30 Jul 2021 14:59

Tgt Ion: 88 Resp: 393
 Ion Ratio Lower Upper
 88 100
 43 0.0 0.0 0.0
 58 0.0 0.0 0.0









#3
n-Nitrosodimethylamine
Concen: 0.218 ng
RT: 3.615 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion: 42 Resp: 439
Ion Ratio Lower Upper
42 100
74 0.0 0.0 0.0
44 0.0 7.4 11.0#



#4
2-Fluorophenol
Concen: 0.135 ng
RT: 5.357 min Scan# 87
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion: 112 Resp: 1067
Ion Ratio Lower Upper
112 100
64 65.9 58.7 88.1
63 34.0 29.8 44.8









#5
Phenol-d6
Concen: 0.152 ng
RT: 6.941 min Scan# 147
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion: 99 Resp: 1518
Ion Ratio Lower Upper
99 100
42 0.0 0.0 0.0
71 32.4 25.0 37.4



#6
bis(2-Chloroethyl)ether
Concen: 0.153 ng
RT: 7.205 min Scan# 157
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion: 93 Resp: 1204
Ion Ratio Lower Upper
93 100
63 68.3 51.4 77.0
95 33.6 27.2 40.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.546 min Scan# 307
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion:	136	Resp:	11635
Ion Ratio	Lower	Upper	
136	100		
137	11.7	9.2	13.8
54	8.2	6.2	9.2
68	8.3	6.1	9.1



#8
Nitrobenzene-d5
Concen: 0.194 ng
RT: 8.913 min Scan# 228
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:	82	Resp:	1511
Ion Ratio	Lower	Upper	
82	100		
128	37.4	25.5	38.3
54	52.1	37.0	55.6









#9
Naphthalene
Concen: 0.119 ng
RT: 10.609 min Scan# 310
Delta R.T. 0.021 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 3579
Ion Ratio Lower Upper
128 100
129 16.2 9.3 13.9#
127 16.4 12.0 18.0



#10
Hexachlorobutadiene
Concen: 0.101 ng
RT: 10.898 min Scan# 324
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:225 Resp: 593
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 65.1 50.0 75.0









#11
2-Methylnaphthalene-d10
Concen: 0.091 ng
RT: 12.145 min Scan# 442
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 1665
Ion Ratio Lower Upper
152 100
151 20.6 17.0 25.4



#12
2-Methylnaphthalene
Concen: 0.112 ng
RT: 12.216 min Scan# 452
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:142 Resp: 2283
Ion Ratio Lower Upper
142 100
141 91.7 71.9 107.9
115 45.1 31.9 47.9









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.391 min Scan# 584
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC0.1


Tgt Ion:	164	Resp:	5723
Ion Ratio	Lower	Upper	
164	100		
162	106.6	84.3	126.5
160	51.6	40.1	60.1



#14
2,4,6-Tribromophenol
Concen: 0.139 ng
RT: 15.881 min Scan# 656
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:	330	Resp:	323
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	46.7	36.5	54.7









#15
2-Fluorobiphenyl
Concen: 0.135 ng
RT: 13.025 min Scan# 518
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDICC0.1


Tgt Ion:	172	Resp:	2941
Ion Ratio	Lower	Upper	
172	100		
171	35.4	27.4	41.0
170	24.0	18.2	27.2



#16
Acenaphthylene
Concen: 0.102 ng
RT: 14.101 min Scan# 570
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:	152	Resp:	2651
Ion Ratio	Lower	Upper	
152	100		
151	20.7	15.7	23.5
153	13.4	10.2	15.4









#17
Acenaphthene
Concen: 0.131 ng
RT: 14.453 min Scan# 587
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion:154 Resp: 2129
Ion Ratio Lower Upper
154 100
153 110.7 88.4 132.6
152 57.0 44.7 67.1



#18
Fluorene
Concen: 0.129 ng
RT: 15.446 min Scan# 635
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:166 Resp: 2626
Ion Ratio Lower Upper
166 100
165 99.7 80.8 121.2
167 14.4 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 719

Delta R.T. 0.000 min

Lab File: BP006438.D

Acq: 30 Jul 2021 14:59





Instrument :

BNA_P

ClientSampleId :

SSTDIC0.1

Tgt Ion:188	Resp:	11346
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	25.2	17.0
		25.4



#20

4,6-Dinitro-2-methylphenol

Concen: 0.294 ng




RT: 15.529 min Scan# 639


Delta R.T. 0.000 min

Lab File: BP006438.D

Acq: 30 Jul 2021 14:59

Tgt Ion:198	Resp:	248
Ion Ratio	Lower	Upper
198	100	
182	36.5	13.8
77	0.0	0.0
		20.8









#21
4-Bromophenyl-phenylether
Concen: 0.110 ng
RT: 16.342 min Scan# 679
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion:248 Resp: 745
Ion Ratio Lower Upper
248 100
250 98.6 77.6 116.4
141 58.1 31.0 46.6#



#22
Hexachlorobenzene
Concen: 0.125 ng
RT: 16.441 min Scan# 684
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:284 Resp: 819
Ion Ratio Lower Upper
284 100
142 47.6 38.4 57.6
249 32.7 27.4 41.2









#23
Atrazine
Concen: 0.111 ng
RT: 16.620 min Scan# 693
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion:200 Resp: 671
Ion Ratio Lower Upper
200 100
173 32.9 18.6 28.0#
215 49.8 37.0 55.4



#24
Pentachlorophenol
Concen: 0.169 ng
RT: 16.798 min Scan# 702
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:266 Resp: 407
Ion Ratio Lower Upper
266 100
264 66.6 49.8 74.8
268 69.0 51.8 77.6









#25
Phenanthrene
Concen: 0.127 ng
RT: 17.175 min Scan# 721
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion:	178	Resp:	3973
Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.5	23.3
179	17.6	12.3	18.5



#26
Anthracene
Concen: 0.117 ng
RT: 17.275 min Scan# 726
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:	178	Resp:	3543
Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.1	22.7
179	16.7	12.0	18.0









#27
Fluoranthene-d10
Concen: 0.101 ng
RT: 19.140 min Scan# 905
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion:212 Resp: 3238
Ion Ratio Lower Upper
212 100
106 19.2 14.6 21.8
104 11.6 8.9 13.3



#28
Fluoranthene
Concen: 0.140 ng
RT: 19.164 min Scan# 908
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:202 Resp: 5005
Ion Ratio Lower Upper
202 100
101 13.3 11.3 16.9
203 17.0 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.253 min Scan# 1086

Delta R.T. 0.000 min

Lab File: BP006438.D

Acq: 30 Jul 2021 14:59





Instrument :

BNA_P

ClientSampleId :

SSTDIC0.1

Tgt Ion:	240	Resp:	10334
Ion Ratio	Lower	Upper	
240	100		
120	8.7	5.8	8.8
236	28.5	22.6	33.8



#30

Pyrene

Concen: 0.154 ng




RT: 19.522 min Scan# 953


Delta R.T. 0.000 min

Lab File: BP006438.D

Acq: 30 Jul 2021 14:59

Tgt Ion:	202	Resp:	5431
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.9	25.3
203	18.3	14.1	21.1









#31
Terphenyl-d14
Concen: 0.098 ng
RT: 19.730 min Scan# 979
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion:244 Resp: 2564
Ion Ratio Lower Upper
244 100
212 12.5 8.5 12.7
122 18.9 11.8 17.6#



#32
Benzo(a)anthracene
Concen: 0.125 ng
RT: 21.236 min Scan# 1085
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:228 Resp: 4383
Ion Ratio Lower Upper
228 100
226 27.6 20.6 30.8
229 23.7 16.6 25.0





#33

Chrysene

Concen: 0.138 ng

RT: 21.288 min Scan# 1088

Delta R.T. 0.000 min

Lab File: BP006438.D

Acq: 30 Jul 2021 14:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC0.1





Tgt Ion:228 Resp: 4584

Ion Ratio Lower Upper

228 100

226 29.9 22.4 33.6

229 23.1 17.4 26.2



#35

Perylene-d12

Concen: 0.400 ng

RT: 23.589 min Scan# 1370

Delta R.T. 0.006 min

Lab File: BP006438.D

Acq: 30 Jul 2021 14:59




Tgt Ion:264 Resp: 9602


Ion Ratio Lower Upper

264 100

260 27.5 22.2 33.4

265 91.8 64.2 96.2









#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.142 ng
 RT: 26.003 min Scan# 1815
 Delta R.T. 0.006 min
 Lab File: BP006438.D
 Acq: 30 Jul 2021 14:59




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:276 Resp: 4958
 Ion Ratio Lower Upper
 276 100
 138 36.4 27.7 41.5
 277 25.5 19.8 29.8



#37
 Benzo(b)fluoranthene
 Concen: 0.145 ng
 RT: 22.873 min Scan# 1238
 Delta R.T. 0.000 min
 Lab File: BP006438.D
 Acq: 30 Jul 2021 14:59

Tgt Ion:252 Resp: 4725
 Ion Ratio Lower Upper
 252 100
 253 38.1 20.8 31.2#
 125 41.9 16.3 24.5#









#38
Benzo(k)fluoranthene
Concen: 0.139 ng
RT: 22.922 min Scan# 1247
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59




Instrument :
BNA_P
ClientSampleId :
SSTDIC0.1


Tgt Ion:252 Resp: 4456
Ion Ratio Lower Upper
252 100
253 38.7 21.0 31.6#
125 43.0 17.1 25.7#



#39
Benzo(a)pyrene
Concen: 0.130 ng
RT: 23.481 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:252 Resp: 3818
Ion Ratio Lower Upper
252 100
253 43.9 22.9 34.3#
125 50.6 20.2 30.2#









#40
Dibenzo(a,h)anthracene
Concen: 0.142 ng
RT: 26.025 min Scan# 1819
Delta R.T. 0.006 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Instrument :
BNA_P
ClientSampleId :
SSTDICC0.1

Tgt Ion:278 Resp: 4202
Ion Ratio Lower Upper
278 100
139 35.4 19.3 28.9#
279 40.0 21.8 32.6#



#41
Benzo(g,h,i)perylene
Concen: 0.151 ng
RT: 26.741 min Scan# 1951
Delta R.T. 0.006 min
Lab File: BP006438.D
Acq: 30 Jul 2021 14:59

Tgt Ion:276 Resp: 4498
Ion Ratio Lower Upper
276 100
277 30.7 20.6 31.0
138 36.5 25.5 38.3

