

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP073021\
 Data File : BP006445.D
 Acq On : 30 Jul 2021 19:33
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP073021

Quant Time: Jul 31 00:49:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270-SIM-BP073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 30 18:59:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	-0.03
2	1,4-Dioxane	0.400	0.368	8.0	81	0.00
3	n-Nitrosodimethylamine	0.400	0.448	-12.0	99	0.00
4 S	2-Fluorophenol	0.400	0.390	2.5	99	0.00
5 S	Phenol-d6	0.400	0.377	5.8	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	92	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.356	11.0	93	0.00
9	Naphthalene	0.400	0.377	5.8	93	0.00
10	Hexachlorobutadiene	0.400	0.386	3.5	93	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.378	5.5	94	0.00
12	2-Methylnaphthalene	0.400	0.374	6.5	94	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	93	0.00
14 S	2,4,6-Tribromophenol	0.400	0.336	16.0	90	0.00
15 S	2-Fluorobiphenyl	0.400	0.368	8.0	93	0.00
16	Acenaphthylene	0.400	0.350	12.5	92	0.00
17	Acenaphthene	0.400	0.356	11.0	91	0.00
18	Fluorene	0.400	0.359	10.3	91	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.322	19.5	86	0.00
21	4-Bromophenyl-phenylether	0.400	0.366	8.5	90	0.00
22	Hexachlorobenzene	0.400	0.380	5.0	95	0.00
23	Atrazine	0.400	0.346	13.5	90	0.00
24	Pentachlorophenol	0.400	0.331	17.3	86	0.00
25	Phenanthrene	0.400	0.378	5.5	94	0.00
26	Anthracene	0.400	0.356	11.0	92	0.00
27 SURR	Fluoranthene-d10	0.400	0.372	7.0	92	0.00
28	Fluoranthene	0.400	0.365	8.8	93	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.356	11.0	91	0.00
31 S	Terphenyl-d14	0.400	0.361	9.8	92	0.00
32	Benzo(a)anthracene	0.400	0.373	6.8	95	0.00
33	Chrysene	0.400	0.360	10.0	91	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.348	13.0	85	0.00
35 I	Perylene-d12	0.400	0.400	0.0	103	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.372	7.0	107	0.00
37	Benzo(b)fluoranthene	0.400	0.362	9.5	101	0.00
38	Benzo(k)fluoranthene	0.400	0.368	8.0	102	0.00
39 C	Benzo(a)pyrene	0.400	0.351	12.3	98	0.00
40	Dibenzo(a,h)anthracene	0.400	0.380	5.0	109	0.00
41	Benzo(g,h,i)perylene	0.400	0.374	6.5	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_P

ClientSampleId :

ICVBP073021