

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080124\
 Data File : BP021334.D
 Acq On : 03 Aug 2024 00:45
 Operator : CG/JU
 Sample : PB162171BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS171

Quant Time: Aug 03 01:37:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 29 12:46:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.616	152	206295	20.000	ng/u1	-0.02
20) Naphthalene-d8	10.375	136	813691	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.263	164	505885	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.075	188	1056315	20.000	ng/u1	-0.02
79) Chrysene-d12	21.527	240	1132348	20.000	ng/u1	0.00
88) Perylene-d12	24.821	264	1404234	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.158	96	22988	5.071	ng/uL	-0.02
4) Pyridine-d5	3.575	84	307696	23.444	ng/u1	-0.02
7) Phenol-d5	6.840	99	413302	24.574	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.969	67	240796	23.625	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.169	132	348169	25.126	ng/u1	-0.01
15) 4-Methylphenol-d8	8.352	113	324496	23.230	ng/u1	-0.01
21) Nitrobenzene-d5	8.775	128	164991	26.106	ng/u1	-0.02
24) 2-Nitrophenol-d4	9.487	143	188991	26.127	ng/u1	-0.02
28) 2,4-Dichlorophenol-d3	10.040	165	331026	25.308	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.546	131	414853	23.941	ng/u1	-0.01
46) Dimethylphthalate-d6	13.669	166	903202	24.560	ng/u1	-0.01
49) Acenaphthylene-d8	13.951	160	1060232	25.676	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.587	143	179076	34.225	ng/u1	0.00
60) Fluorene-d10	15.281	176	779263	25.829	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.451	200	147208	23.033	ng/u1	-0.01
73) Anthracene-d10	17.186	188	1221963	26.046	ng/u1	0.00
81) Pyrene-d10	19.575	212	1598290	22.831	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.598	264	1844251	26.629	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.193	88	46220	9.270	ng/uL	100
5) Pyridine	3.599	79	303424	22.414	ng/u1	98
6) Benzaldehyde	6.793	77	242138	28.557	ng/u1	100
8) Phenol	6.869	94	417451	24.529	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.058	93	315662	22.453	ng/u1	99
12) 2-Chlorophenol	7.205	128	341360	24.579	ng/u1	99
13) 2-Methylphenol	8.081	108	306856	23.314	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.128	45	457775	22.538	ng/u1	95
16) Acetophenone	8.440	105	490982	22.774	ng/u1	92
17) N-Nitroso-di-n-propyla...	8.416	70	239793	21.228	ng/u1	98
18) 4-Methylphenol	8.416	108	328317	23.284	ng/u1	96
19) Hexachloroethane	8.663	117	144208	23.106	ng/u1	99
22) Nitrobenzene	8.816	77	376484	24.868	ng/u1	99
23) Isophorone	9.334	82	679618	22.580	ng/u1	100
25) 2-Nitrophenol	9.516	139	191301	25.357	ng/u1	100
26) 2,4-Dimethylphenol	9.587	107	261078	17.549	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.810	93	412064	22.524	ng/u1	98
29) 2,4-Dichlorophenol	10.069	162	319359	24.774	ng/u1	99
30) Naphthalene	10.428	128	1039161	24.279	ng/u1	99
32) 4-Chloroaniline	10.569	127	366123	22.127	ng/u1	99
33) Hexachlorobutadiene	10.681	225	224458	23.222	ng/u1	99
34) Caprolactam	11.393	113	101687	25.607	ng/u1	94
35) 4-Chloro-3-methylphenol	11.728	107	349673	24.955	ng/u1	98
36) 2-Methylnaphthalene	12.057	142	701307	23.870	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080124\
 Data File : BP021334.D
 Acq On : 03 Aug 2024 00:45
 Operator : CG/JU
 Sample : PB162171BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS171

Quant Time: Aug 03 01:37:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 29 12:46:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.269	142	699347	23.144	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.428	216	411250	25.032	ng/ul	97
40) Hexachlorocyclopentadiene	12.381	237	41239	4.798	ng/ul	99
41) 2,4,6-Trichlorophenol	12.698	196	242568	23.366	ng/ul	99
42) 2,4,5-Trichlorophenol	12.793	196	261035	23.766	ng/ul	99
43) 1,1'-Biphenyl	13.081	154	905648	24.496	ng/ul	99
44) 2-Chloronaphthalene	13.122	162	730223	24.741	ng/ul	99
45) 2-Nitroaniline	13.369	65	227281	27.846	ng/ul	99
47) Dimethylphthalate	13.716	163	883198	23.669	ng/ul	100
48) 2,6-Dinitrotoluene	13.863	165	186114	24.948	ng/ul	99
50) Acenaphthylene	13.981	152	1135453	24.286	ng/ul	99
51) 3-Nitroaniline	14.222	138	189549	29.198	ng/ul	99
52) Acenaphthene	14.328	153	762997	24.588	ng/ul	99
53) 2,4-Dinitrophenol	14.457	184	79979	20.414	ng/ul	96
55) 4-Nitrophenol	14.598	109	135345	31.799	ng/ul	95
56) Dibenzofuran	14.675	168	1055893	24.848	ng/ul	100
57) 2,4-Dinitrotoluene	14.687	165	274935	26.847	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.922	232	207378	22.048	ng/ul	96
59) Diethylphthalate	15.104	149	867866	23.272	ng/ul	99
61) Fluorene	15.334	166	853373	24.610	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.328	204	443843	23.588	ng/ul	99
63) 4-Nitroaniline	15.410	138	200642	37.321	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.463	198	152377	22.482	ng/ul	96
67) N-Nitrosodiphenylamine	15.551	169	734067	23.864	ng/ul	99
68) 4-Bromophenyl-phenylether	16.239	248	285693	23.075	ng/ul	98
69) Hexachlorobenzene	16.363	284	336968	23.736	ng/ul	97
70) Atrazine	16.545	200	119488	10.659	ng/ul	97
71) Pentachlorophenol	16.745	266	157262	20.396	ng/ul	98
72) Phenanthrene	17.128	178	1408320	25.700	ng/ul	100
74) Anthracene	17.222	178	1376927	24.648	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.040	216	400116	23.666	ng/ul	98
76) Pentachlorobenzene	14.587	250	391915	23.328	ng/ul	100
77) Carbazole	17.528	167	1344500	29.441	ng/ul	100
78) Di-n-butylphthalate	18.081	149	1499314	23.797	ng/ul	100
80) Fluoranthene	19.222	202	1786578	21.132	ng/ul	100
82) Pyrene	19.598	202	1898273	22.065	ng/ul	98
83) Butylbenzylphthalate	20.539	149	784746	21.893	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.439	252	689189	26.636	ng/ul	99
85) Benzo(a)anthracene	21.510	228	1974227	24.889	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.422	149	1062276	20.593	ng/ul	98
87) Chrysene	21.574	228	1904751	25.842	ng/ul	99
89) Di-n-octyl phthalate	22.651	149	2001726	21.579	ng/ul	100
90) Benzo(b)fluoranthene	23.774	252	2127646	24.968	ng/ul	99
91) Benzo(k)fluoranthene	23.851	252	2105369	24.756	ng/ul	99
93) Benzo(a)pyrene	24.663	252	1847732	22.945	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.580	276	2585840	24.920	ng/ul	99
95) Dibenzo(a,h)anthracene	28.633	278	2118635	24.656	ng/ul	99
96) Benzo(g,h,i)perylene	29.750	276	2033320	24.060	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.270 ng/uL
RT: 3.193 min Scan# 31
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion: 88	Resp: 46220
Ion Ratio	Lower Upper
88 100	
43 31.0	25.4 38.0
58 73.9	59.2 88.8



#5
Pyridine
Concen: 22.414 ng/uL
RT: 3.599 min Scan# 104
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion: 79	Resp: 303424
Ion Ratio	Lower Upper
79 100	
52 83.3	64.9 97.3
51 37.6	30.5 45.7









#6
Benzaldehyde
Concen: 28.557 ng/ul
RT: 6.793 min Scan# 647
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion: 77 Resp: 242138
Ion Ratio Lower Upper
77 100
105 99.4 79.3 118.9
106 93.6 74.6 112.0



#8
Phenol
Concen: 24.529 ng/ul
RT: 6.869 min Scan# 660
Delta R.T. -0.006 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion: 94 Resp: 417451
Ion Ratio Lower Upper
94 100
65 30.3 23.4 35.0
66 41.5 33.9 50.9





#10

Bis(2-Chloroethyl)ether

Concen: 22.453 ng/ul

RT: 7.058 min Scan# 692

Delta R.T. -0.017 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion: 93 Resp: 315662

Ion Ratio Lower Upper

93 100

63 77.4 61.5 92.3

95 32.0 25.1 37.7



#12

2-Chlorophenol

Concen: 24.579 ng/ul

RT: 7.205 min Scan# 717

Delta R.T. -0.005 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion:128 Resp: 341360


Ion Ratio Lower Upper

128 100

64 47.5 37.9 56.9

130 32.7 25.6 38.4









#13
2-Methylphenol
Concen: 23.314 ng/ul
RT: 8.081 min Scan# 80
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:108 Resp: 306856
Ion Ratio Lower Upper
108 100
107 91.6 73.8 110.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 22.538 ng/ul
RT: 8.128 min Scan# 874
Delta R.T. -0.029 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion: 45 Resp: 457775
Ion Ratio Lower Upper
45 100
77 15.8 10.9 16.3
79 12.1 8.1 12.1









#16
Acetophenone
Concen: 22.774 ng/ul
RT: 8.440 min Scan# 91
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:105 Resp: 490982
Ion Ratio Lower Upper
105 100
77 84.2 61.8 92.8
51 30.3 21.5 32.3



#17
N-Nitroso-di-n-propylamine
Concen: 21.228 ng/ul
RT: 8.416 min Scan# 923
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion: 70 Resp: 239793
Ion Ratio Lower Upper
70 100
42 60.1 46.6 69.8
101 10.8 7.4 11.0
130 21.3 17.2 25.8









#18
4-Methylphenol
Concen: 23.284 ng/ul
RT: 8.416 min Scan# 911
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:108 Resp: 328317
Ion Ratio Lower Upper
108 100
107 114.5 87.8 131.8



#19
Hexachloroethane
Concen: 23.106 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:117 Resp: 144208
Ion Ratio Lower Upper
117 100
201 108.3 88.2 132.2
199 69.1 55.8 83.6









#22
Nitrobenzene
Concen: 24.868 ng/ul
RT: 8.816 min Scan# 991
Delta R.T. -0.023 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion: 77 Resp: 376484
Ion Ratio Lower Upper
77 100
123 44.8 36.7 55.1
65 13.9 11.1 16.7



#23
Isophorone
Concen: 22.580 ng/ul
RT: 9.334 min Scan# 1079
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion: 82 Resp: 679618
Ion Ratio Lower Upper
82 100
95 7.4 6.1 9.1
138 18.1 14.4 21.6









#25
2-Nitrophenol
Concen: 25.357 ng/ul
RT: 9.516 min Scan# 1110
Delta R.T. -0.023 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:139	Resp: 191301
Ion Ratio	Lower Upper
139 100	
65 53.5	42.9 64.3
109 33.3	26.2 39.4



#26
2,4-Dimethylphenol
Concen: 17.549 ng/ul
RT: 9.587 min Scan# 1122
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:107	Resp: 261078
Ion Ratio	Lower Upper
107 100	
121 52.2	39.6 59.4
122 80.8	66.3 99.5









#27
Bis(2-Chloroethoxy)methane
Concen: 22.524 ng/ul
RT: 9.810 min Scan# 1160
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion: 93 Resp: 412064
Ion Ratio Lower Upper
93 100
95 33.0 25.5 38.3
123 10.7 9.0 13.4



#29
2,4-Dichlorophenol
Concen: 24.774 ng/ul
RT: 10.069 min Scan# 1204
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:162 Resp: 319359
Ion Ratio Lower Upper
162 100
164 67.1 53.3 79.9
98 36.8 28.7 43.1









#30
Naphthalene
Concen: 24.279 ng/ul
RT: 10.428 min Scan# 1265
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:128 Resp: 1039161
Ion Ratio Lower Upper
128 100
129 11.2 8.6 13.0
127 13.4 10.8 16.2



#32
4-Chloroaniline
Concen: 22.127 ng/ul
RT: 10.569 min Scan# 1289
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:127 Resp: 366123
Ion Ratio Lower Upper
127 100
129 32.1 26.1 39.1





#33

Hexachlorobutadiene

Concen: 23.222 ng/ul

RT: 10.681 min Scan# 11

Delta R.T. -0.029 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:225 Resp: 224458

Ion Ratio Lower Upper

225 100

223 64.4 50.2 75.2

227 63.7 50.6 75.8



#34

Caprolactam

Concen: 25.607 ng/ul

RT: 11.393 min Scan# 1429

Delta R.T. -0.011 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion:113 Resp: 101687


Ion Ratio Lower Upper

113 100

55 164.9 140.2 210.4

56 127.8 105.5 158.3









#35
4-Chloro-3-methylphenol
Concen: 24.955 ng/ul
RT: 11.728 min Scan# 1486
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:107 Resp: 349673
Ion Ratio Lower Upper
107 100
144 24.5 19.3 28.9
142 73.7 60.7 91.1



#36
2-Methylnaphthalene
Concen: 23.870 ng/ul
RT: 12.057 min Scan# 1542
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:142 Resp: 701307
Ion Ratio Lower Upper
142 100
141 89.1 72.6 108.8









#37
1-Methylnaphthalene
Concen: 23.144 ng/ul
RT: 12.269 min Scan# 1579
Delta R.T. -0.023 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:142 Resp: 699347
Ion Ratio Lower Upper
142 100
141 91.2 75.0 112.6
116 4.0 3.2 4.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 25.032 ng/ul
RT: 12.428 min Scan# 1605
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:216 Resp: 411250
Ion Ratio Lower Upper
216 100
214 77.6 64.2 96.4
179 20.3 16.6 24.8
108 15.5 13.4 20.2





#40

Hexachlorocyclopentadiene

Concen: 4.798 ng/ul

RT: 12.381 min Scan# 1597

Delta R.T. -0.017 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:237 Resp: 41239

Ion Ratio Lower Upper

237 100

235 62.4 49.7 74.5

272 14.5 10.2 15.2



#41

2,4,6-Trichlorophenol

Concen: 23.366 ng/ul

RT: 12.698 min Scan# 1651

Delta R.T. -0.006 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion:196 Resp: 242568


Ion Ratio Lower Upper

196 100

198 94.9 76.6 115.0

200 31.8 26.0 39.0









#42
2,4,5-Trichlorophenol
Concen: 23.766 ng/ul
RT: 12.793 min Scan# 1667
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:196 Resp: 261035
Ion Ratio Lower Upper
196 100
198 95.7 77.1 115.7
200 32.2 25.3 37.9



#43
1,1'-Biphenyl
Concen: 24.496 ng/ul
RT: 13.081 min Scan# 1716
Delta R.T. -0.006 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:154 Resp: 905648
Ion Ratio Lower Upper
154 100
153 40.5 32.7 49.1
76 12.7 9.4 14.2





#44

2-Chloronaphthalene

Concen: 24.741 ng/ul

RT: 13.122 min Scan# 1723

Delta R.T. -0.017 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:162 Resp: 730223

Ion Ratio Lower Upper

162 100

164 33.1 26.3 39.5

127 40.2 31.8 47.8



#45

2-Nitroaniline

Concen: 27.846 ng/ul

RT: 13.369 min Scan# 1765

Delta R.T. 0.000 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion: 65 Resp: 227281


Ion Ratio Lower Upper

65 100

92 71.2 56.7 85.1

138 105.1 85.6 128.4









#47
Dimethylphthalate
Concen: 23.669 ng/ul
RT: 13.716 min Scan# 1824
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:163 Resp: 883198
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.2 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 24.948 ng/ul
RT: 13.863 min Scan# 1849
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:165 Resp: 186114
Ion Ratio Lower Upper
165 100
89 56.5 44.7 67.1
121 21.5 16.5 24.7









#50
Acenaphthylene
Concen: 24.286 ng/ul
RT: 13.981 min Scan# 11
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:152 Resp: 1135453
Ion Ratio Lower Upper
152 100
151 19.4 16.2 24.2
153 12.7 10.3 15.5



#51
3-Nitroaniline
Concen: 29.198 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:138 Resp: 189549
Ion Ratio Lower Upper
138 100
108 12.7 9.8 14.8
92 121.2 98.0 147.0





#52

Acenaphthene

Concen: 24.588 ng/ul

RT: 14.328 min Scan# 1928

Delta R.T. -0.011 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:153 Resp: 762997

Ion Ratio Lower Upper

153 100

152 46.6 37.8 56.6

154 88.4 71.8 107.6



#53

2,4-Dinitrophenol

Concen: 20.414 ng/ul

RT: 14.457 min Scan# 1950

Delta R.T. -0.005 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion:184 Resp: 79979


Ion Ratio Lower Upper

184 100

63 64.5 50.3 75.5

154 71.1 53.0 79.6









#55
4-Nitrophenol
Concen: 31.799 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. 0.006 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:109 Resp: 135345
Ion Ratio Lower Upper
109 100
139 120.0 98.0 147.0
65 124.4 106.0 159.0



#56
Dibenzofuran
Concen: 24.848 ng/ul
RT: 14.675 min Scan# 1987
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:168 Resp: 1055893
Ion Ratio Lower Upper
168 100
139 38.8 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 26.847 ng/ul
RT: 14.687 min Scan# 1990
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:165	Resp:	274935
Ion Ratio	Lower	Upper
165	100	
63	48.0	41.0 61.4
182	3.1	2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.048 ng/ul
RT: 14.922 min Scan# 2029
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:232	Resp:	207378
Ion Ratio	Lower	Upper
232	100	
131	39.6	34.4 51.6
130	2.5	2.0 3.0
166	29.7	24.4 36.6









#59
Diethylphthalate
Concen: 23.272 ng/ul
RT: 15.104 min Scan# 2060
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:149 Resp: 867866
Ion Ratio Lower Upper
149 100
177 22.7 17.7 26.5
150 13.1 10.2 15.2



#61
Fluorene
Concen: 24.610 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:166 Resp: 853373
Ion Ratio Lower Upper
166 100
165 99.2 78.1 117.1
167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 23.588 ng/ul
RT: 15.328 min Scan# 2097
Delta R.T. -0.006 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:204 Resp: 443843
Ion Ratio Lower Upper
204 100
206 33.5 26.4 39.6
141 59.2 48.5 72.7



#63
4-Nitroaniline
Concen: 37.321 ng/ul
RT: 15.410 min Scan# 2112
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:138 Resp: 200642
Ion Ratio Lower Upper
138 100
92 53.6 40.9 61.3
108 79.8 64.7 97.1









#66
4,6-Dinitro-2-methylphenol
Concen: 22.482 ng/ul
RT: 15.463 min Scan# 2121
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:198 Resp: 152377
Ion Ratio Lower Upper
198 100
182 5.3 3.8 5.6
77 27.0 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 23.864 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. -0.017 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:169 Resp: 734067
Ion Ratio Lower Upper
169 100
168 65.5 53.4 80.2
167 35.5 28.4 42.6









#68
4-Bromophenyl-phenylether
Concen: 23.075 ng/ul
RT: 16.239 min Scan# 21
Delta R.T. -0.023 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:248 Resp: 285693
Ion Ratio Lower Upper
248 100
250 100.4 79.3 118.9
141 70.1 57.4 86.0



#69
Hexachlorobenzene
Concen: 23.736 ng/ul
RT: 16.363 min Scan# 2274
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:284 Resp: 336968
Ion Ratio Lower Upper
284 100
142 29.8 25.8 38.8
249 30.2 25.1 37.7









#70
 Atrazine
 Concen: 10.659 ng/ul
 RT: 16.545 min Scan# 21
 Delta R.T. -0.011 min
 Lab File: BP021334.D
 Acq: 03 Aug 2024 00:45




Instrument :
 BNA_P
 ClientSampleId :
 SLCS171


Tgt Ion:200 Resp: 119488
 Ion Ratio Lower Upper
 200 100
 173 26.6 20.8 31.2
 215 51.1 38.7 58.1



#71
 Pentachlorophenol
 Concen: 20.396 ng/ul
 RT: 16.745 min Scan# 2339
 Delta R.T. -0.006 min
 Lab File: BP021334.D
 Acq: 03 Aug 2024 00:45

Tgt Ion:266 Resp: 157262
 Ion Ratio Lower Upper
 266 100
 264 63.7 52.3 78.5
 268 64.0 52.3 78.5





#72

Phenanthrene

Concen: 25.700 ng/u1

RT: 17.128 min Scan# 2403

Delta R.T. -0.006 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:178 Resp: 1408320

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.7 15.0 22.4



#74

Anthracene

Concen: 24.648 ng/u1

RT: 17.222 min Scan# 2420

Delta R.T. -0.011 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion:178 Resp: 1376927


Ion Ratio Lower Upper

178 100

179 15.5 12.5 18.7

176 18.5 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 23.666 ng/uL

RT: 13.040 min Scan# 1710

Delta R.T. -0.017 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:216 Resp: 400116

Ion Ratio Lower Upper

216 100

214 77.6 63.2 94.8

218 49.1 38.2 57.4



#76

Pentachlorobenzene

Concen: 23.328 ng/uL

RT: 14.587 min Scan# 1972

Delta R.T. -0.005 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion:250 Resp: 391915


Ion Ratio Lower Upper

250 100

248 64.0 50.8 76.2

252 64.4 51.6 77.4









#77
 Carbazole
 Concen: 29.441 ng/ul
 RT: 17.528 min Scan# 2470
 Delta R.T. 0.006 min
 Lab File: BP021334.D
 Acq: 03 Aug 2024 00:45




Instrument :
 BNA_P
 ClientSampleId :
 SLCS171


Tgt Ion:167 Resp: 1344500
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.3 25.9
 139 12.8 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 23.797 ng/ul
 RT: 18.081 min Scan# 2566
 Delta R.T. -0.017 min
 Lab File: BP021334.D
 Acq: 03 Aug 2024 00:45

Tgt Ion:149 Resp: 1499314
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 5.0 3.9 5.9









#80
Fluoranthene
Concen: 21.132 ng/u1
RT: 19.222 min Scan# 21
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:202 Resp: 1786578
Ion Ratio Lower Upper
202 100
101 10.2 8.0 12.0
100 7.6 6.1 9.1



#82
Pyrene
Concen: 22.065 ng/u1
RT: 19.598 min Scan# 2824
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:202 Resp: 1898273
Ion Ratio Lower Upper
202 100
101 12.7 9.4 14.2
100 10.2 7.8 11.6





#83

Butylbenzylphthalate

Concen: 21.893 ng/ul

RT: 20.539 min Scan# 2983

Delta R.T. -0.011 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:149 Resp: 784746

Ion Ratio Lower Upper

149 100

91 70.3 54.1 81.1

206 23.0 17.0 25.6



#84

3,3'-Dichlorobenzidine

Concen: 26.636 ng/ul

RT: 21.439 min Scan# 3137

Delta R.T. 0.000 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion:252 Resp: 689189


Ion Ratio Lower Upper

252 100

254 64.8 51.5 77.3

126 8.9 8.0 12.0





#85

Benzo(a)anthracene

Concen: 24.889 ng/ul

RT: 21.510 min Scan# 3149

Delta R.T. 0.000 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:228 Resp: 1974227

Ion Ratio Lower Upper

228 100

229 19.9 15.5 23.3

226 27.3 21.4 32.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.593 ng/ul

RT: 21.422 min Scan# 3134

Delta R.T. 0.000 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion:149 Resp: 1062276


Ion Ratio Lower Upper

149 100

167 28.7 22.1 33.1

279 4.9 3.7 5.5





#87

Chrysene

Concen: 25.842 ng/ul

RT: 21.574 min Scan# 3160

Delta R.T. -0.011 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:228 Resp: 1904751

Ion Ratio Lower Upper

228 100

226 29.8 23.6 35.4

229 19.7 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 21.579 ng/ul




RT: 22.651 min Scan# 3343


Delta R.T. -0.029 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Tgt Ion:149 Resp: 2001726









#90
Benzo(b)fluoranthene
Concen: 24.968 ng/ul
RT: 23.774 min Scan# 3534
Delta R.T. -0.011 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45




Instrument :
BNA_P
ClientSampleId :
SLCS171


Tgt Ion:252 Resp: 2127646
Ion Ratio Lower Upper
252 100
253 22.2 17.1 25.7
125 8.9 7.0 10.4



#91
Benzo(k)fluoranthene
Concen: 24.756 ng/ul
RT: 23.851 min Scan# 3547
Delta R.T. 0.000 min
Lab File: BP021334.D
Acq: 03 Aug 2024 00:45

Tgt Ion:252 Resp: 2105369
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 8.2 7.4 11.0





#93

Benzo(a)pyrene

Concen: 22.945 ng/ul

RT: 24.663 min Scan# 3

Delta R.T. -0.017 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:252 Resp: 1847732

Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.0

125 9.8 7.7 11.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 24.920 ng/ul

RT: 28.580 min Scan# 4351

Delta R.T. -0.017 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45




Tgt Ion:276 Resp: 2585840


Ion Ratio Lower Upper

276 100

138 18.8 14.3 21.5

277 24.3 19.2 28.8





#95

Dibenzo(a,h)anthracene

Concen: 24.656 ng/ul

RT: 28.633 min Scan# 41

Delta R.T. -0.006 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Instrument :

BNA_P

ClientSampleId :

SLCS171





Tgt Ion:278 Resp: 2118635

Ion Ratio Lower Upper

278 100

139 13.0 10.9 16.3

279 23.7 19.4 29.0



#96

Benzo(g,h,i)perylene

Concen: 24.060 ng/ul

RT: 29.750 min Scan# 4550

Delta R.T. -0.011 min

Lab File: BP021334.D

Acq: 03 Aug 2024 00:45

Tgt Ion:276 Resp: 2033320

Ion Ratio Lower Upper

276 100

138 17.4 15.3 22.9

277 24.1 18.2 27.2

