

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080321\
 Data File : BP006486.D
 Acq On : 03 Aug 2021 19:50
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 04 01:18:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	107	0.00
2	1,4-Dioxane	40.000	38.446	3.9	104	0.00
3	Pyridine	40.000	40.320	-0.8	102	0.00
4	n-Nitrosodimethylamine	40.000	42.142	-5.4	108	0.00
5 S	2-Fluorophenol	80.000	80.223	-0.3	104	0.00
6	Aniline	40.000	42.370	-5.9	103	0.00
7 S	Phenol-d6	80.000	83.854	-4.8	104	0.00
8	2-Chlorophenol	40.000	41.183	-3.0	104	0.00
9	Benzaldehyde	40.000	40.694	-1.7	106	0.00
10 C	Phenol	40.000	41.661	-4.2	103	0.00
11	bis(2-Chloroethyl)ether	40.000	41.264	-3.2	104	0.00
12	1,3-Dichlorobenzene	40.000	40.046	-0.1	103	0.00
13 C	1,4-Dichlorobenzene	40.000	40.018	-0.0	103	0.00
14	1,2-Dichlorobenzene	40.000	40.467	-1.2	103	0.00
15	Benzyl Alcohol	40.000	43.509	-8.8	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.609	-4.0	103	0.00
17	2-Methylphenol	40.000	42.860	-7.1	105	0.00
18	Hexachloroethane	40.000	40.400	-1.0	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.097	-10.2	105	0.00
20	3+4-Methylphenols	40.000	43.487	-8.7	104	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	108	0.00
22	Acetophenone	40.000	39.948	0.1	104	0.00
23 S	Nitrobenzene-d5	80.000	80.802	-1.0	105	0.00
24	Nitrobenzene	40.000	40.266	-0.7	105	0.00
25	Isophorone	40.000	41.794	-4.5	106	0.00
26 C	2-Nitrophenol	40.000	40.841	-2.1	104	0.00
27	2,4-Dimethylphenol	40.000	41.122	-2.8	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.242	-0.6	105	0.00
29 C	2,4-Dichlorophenol	40.000	40.907	-2.3	104	0.00
30	1,2,4-Trichlorobenzene	40.000	38.810	3.0	104	0.00
31	Naphthalene	40.000	39.802	0.5	105	0.00
32	Benzoic acid	40.000	41.249	-3.1	107	0.00
33	4-Chloroaniline	40.000	41.373	-3.4	105	0.00
34 C	Hexachlorobutadiene	40.000	38.589	3.5	105	0.00
35	Caprolactam	40.000	44.672	-11.7	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.490	-6.2	105	0.00
37	2-Methylnaphthalene	40.000	40.938	-2.3	106	0.00
38	1-Methylnaphthalene	40.000	40.733	-1.8	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.370	4.1	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.610	1.0	106	0.00
42 S	2,4,6-Tribromophenol	80.000	82.881	-3.6	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.870	-2.2	107	0.00
44	2,4,5-Trichlorophenol	40.000	40.569	-1.4	106	0.00
45 S	2-Fluorobiphenyl	80.000	77.492	3.1	106	0.00
46	1,1'-Biphenyl	40.000	39.038	2.4	106	0.00
47	2-Chloronaphthalene	40.000	39.034	2.4	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080321\
 Data File : BP006486.D
 Acq On : 03 Aug 2021 19:50
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 04 01:18:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.271	-5.7	107	0.00
49	Acenaphthylene	40.000	40.168	-0.4	106	0.00
50	Dimethylphthalate	40.000	40.323	-0.8	106	0.00
51	2,6-Dinitrotoluene	40.000	41.354	-3.4	106	0.00
52 C	Acenaphthene	40.000	40.043	-0.1	106	0.00
53	3-Nitroaniline	40.000	42.140	-5.4	106	0.00
54 P	2,4-Dinitrophenol	40.000	40.720	-1.8	106	0.00
55	Dibenzofuran	40.000	39.667	0.8	106	0.00
56 P	4-Nitrophenol	40.000	42.798	-7.0	105	0.00
57	2,4-Dinitrotoluene	40.000	42.682	-6.7	108	0.00
58	Fluorene	40.000	40.257	-0.6	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.106	-2.8	107	0.00
60	Diethylphthalate	40.000	40.757	-1.9	107	0.00
61	4-Chlorophenyl-phenylether	40.000	39.728	0.7	106	0.00
62	4-Nitroaniline	40.000	42.839	-7.1	106	0.00
63	Azobenzene	40.000	41.585	-4.0	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.940	0.2	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.883	0.3	106	0.00
67	4-Bromophenyl-phenylether	40.000	39.550	1.1	106	0.00
68	Hexachlorobenzene	40.000	39.814	0.5	109	0.00
69	Atrazine	40.000	41.419	-3.5	107	0.00
70 C	Pentachlorophenol	40.000	41.626	-4.1	106	0.00
71	Phenanthrene	40.000	39.931	0.2	107	0.00
72	Anthracene	40.000	40.510	-1.3	107	0.00
73	Carbazole	40.000	40.599	-1.5	105	0.00
74	Di-n-butylphthalate	40.000	41.458	-3.6	108	0.00
75 C	Fluoranthene	40.000	40.484	-1.2	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	41.823	-4.6	96	0.00
78	Pyrene	40.000	40.264	-0.7	105	0.00
79 S	Terphenyl-d14	80.000	79.976	0.0	106	0.00
80	Butylbenzylphthalate	40.000	42.251	-5.6	107	0.00
81	Benzo(a)anthracene	40.000	39.648	0.9	104	0.00
82	3,3'-Dichlorobenzidine	40.000	40.976	-2.4	105	0.00
83	Chrysene	40.000	39.776	0.6	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.753	-1.9	105	0.00
85 c	Di-n-octyl phthalate	40.000	41.308	-3.3	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.989	0.0	103	0.00
88	Benzo(b)fluoranthene	40.000	39.954	0.1	104	0.00
89	Benzo(k)fluoranthene	40.000	40.183	-0.5	103	0.00
90 C	Benzo(a)pyrene	40.000	40.575	-1.4	104	0.00
91	Dibenzo(a,h)anthracene	40.000	39.807	0.5	103	-0.01
92	Benzo(g,h,i)perylene	40.000	40.086	-0.2	104	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080321\
Data File : BP006486.D
Acq On : 03 Aug 2021 19:50
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 04 01:18:11 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 15:41:30 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0