



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080321\
 Data File : BP006471.D
 Acq On : 03 Aug 2021 09:00
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Aug 03 15:41:30 2021






AutoFind: Scans 2467, 2468, 2469; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.1	128307	PASS
68	69	0.00	2	1.9	3074	PASS
69	198	0.00	100	52.3	163017	PASS
70	69	0.00	2	0.5	884	PASS
127	198	10	80	58.3	181807	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	311936	PASS
199	198	5	9	6.8	21223	PASS
275	198	10	60	27.3	85163	PASS
365	198	1	100	4.1	12826	PASS
441	198	0.01	100	11.9	37099	PASS
442	442	100	100	100.0	238436	PASS
443	442	15	24	19.3	46016	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080321\
Data File : BP006471.D
Acq On : 03 Aug 2021 09:00
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Aug 03 11:56:01 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 11:52:20 2021
Response via : Initial Calibration



TIC: BP006471.D\data.ms

(77) Benzidine

19.351min (+ 0.006) 5047.90 ng




response 2366381

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	14.18
183.00	11.90	12.06
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080321\
Data File : BP006471.D
Acq On : 03 Aug 2021 09:00
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Aug 03 11:56:01 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 11:52:20 2021
Response via : Initial Calibration



TIC: BP006471.D\data.ms

(70) Pentachlorophenol (C)

16.787min (+ 0.000) 14318.48 ng

response 375637

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	64.47
264.00	62.90	62.67
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/3/2021	BNA_P	<u>BP006471.D</u>
Compound Name	Response	Retention Time
DDT	1103112	20.55
DDD	5991	20.12
DDE	1831	19.62
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7822	1110934	0.70