

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080822\
 Data File : BP011311.D
 Acq On : 08 Aug 2022 15:07
 Operator : CG/JU
 Sample : SSTD00538
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005638

Quant Time: Aug 08 17:10:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 08 17:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.593	152	174472	20.000	ng/u1	0.00
20) Naphthalene-d8	10.387	136	736236	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.263	164	400150	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.039	188	817678	20.000	ng/u1	0.00
79) Chrysene-d12	21.169	240	757431	20.000	ng/u1	0.00
88) Perylene-d12	23.480	264	724897	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	48	0.022	ng/uL	0.17
4) Pyridine-d5	3.511	84	63280	10.872	ng/u1	0.01
7) Phenol-d5	6.775	99	63110	4.341	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.940	67	49678	4.735	ng/u1	0.00
11) 2-Chlorophenol-d4	7.128	132	51857	4.376	ng/u1	0.00
15) 4-Methylphenol-d8	8.316	113	47544	3.974	ng/u1	0.00
21) Nitrobenzene-d5	8.757	128	18091	3.416	ng/u1	0.00
24) 2-Nitrophenol-d4	9.475	143	15945	3.212	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.010	165	37056	3.915	ng/u1	0.00
31) 4-Chloroaniline-d4	10.540	131	62379	4.260	ng/u1	0.00
46) Dimethylphthalate-d6	13.692	166	121393	4.429	ng/u1	0.00
49) Acenaphthylene-d8	13.957	160	134628	3.904	ng/u1	0.00
54) 4-Nitrophenol-d4	14.498	143	12213	2.621	ng/u1	0.00
60) Fluorene-d10	15.269	176	107074	4.592	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.398	200	9644	2.540	ng/u1	0.00
73) Anthracene-d10	17.139	188	163278	4.544	ng/u1	0.00
81) Pyrene-d10	19.398	212	193298	5.047	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.333	264	159187	4.407	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.122	88	11677	5.868	ng/uL#	72
5) Pyridine	3.528	79	66844	11.341	ng/u1	87
6) Benzaldehyde	6.746	77	41470	5.834	ng/u1	96
8) Phenol	6.805	94	70926	4.549	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.034	93	60041	4.693	ng/u1	96
12) 2-Chlorophenol	7.158	128	54706	4.444	ng/u1	99
13) 2-Methylphenol	8.052	108	48161	4.083	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.128	45	61366	2.815	ng/u1#	93
16) Acetophenone	8.422	105	87835	4.530	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.410	70	39833	4.222	ng/u1	95
18) 4-Methylphenol	8.375	108	53034	4.192	ng/u1	97
19) Hexachloroethane	8.657	117	23416	3.967	ng/u1	99
22) Nitrobenzene	8.799	77	55274	3.809	ng/u1	99
23) Isophorone	9.328	82	104311	4.039	ng/u1	99
25) 2-Nitrophenol	9.504	139	18858	3.283	ng/u1	92
26) 2,4-Dimethylphenol	9.575	107	56434	4.343	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.816	93	73668	4.531	ng/u1	99
29) 2,4-Dichlorophenol	10.040	162	41151	4.160	ng/u1	99
30) Naphthalene	10.434	128	188966	4.862	ng/u1	99
32) 4-Chloroaniline	10.563	127	66585	4.489	ng/u1	94
33) Hexachlorobutadiene	10.716	225	27528	4.455	ng/u1	96
34) Caprolactam	11.351	113	8267	2.821	ng/u1	93
35) 4-Chloro-3-methylphenol	11.704	107	43989	3.932	ng/u1	95
36) 2-Methylnaphthalene	12.063	142	112036	4.613	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080822\
 Data File : BP011311.D
 Acq On : 08 Aug 2022 15:07
 Operator : CG/JU
 Sample : SST00538
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005638

Quant Time: Aug 08 17:10:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 08 17:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.287	142	114761	4.704	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.434	216	49889	4.559	ng/ul	97
40) Hexachlorocyclopentadiene	12.410	237	18261	2.641	ng/ul	99
41) 2,4,6-Trichlorophenol	12.687	196	24838	3.623	ng/ul	97
42) 2,4,5-Trichlorophenol	12.757	196	24998	3.398	ng/ul	95
43) 1,1'-Biphenyl	13.092	154	147328	4.613	ng/ul	99
44) 2-Chloronaphthalene	13.128	162	109560	4.599	ng/ul	99
45) 2-Nitroaniline	13.351	65	17149	2.992	ng/ul	96
47) Dimethylphthalate	13.740	163	124217	4.493	ng/ul	99
48) 2,6-Dinitrotoluene	13.857	165	12506	3.013	ng/ul	95
50) Acenaphthylene	13.987	152	166300	4.268	ng/ul	99
51) 3-Nitroaniline	14.192	138	12744	2.784	ng/ul#	96
52) Acenaphthene	14.328	153	120412	4.778	ng/ul	99
53) 2,4-Dinitrophenol	14.404	184	6495	2.635	ng/ul	92
55) 4-Nitrophenol	14.510	109	11225	2.669	ng/ul	99
56) Dibenzofuran	14.669	168	163160	4.731	ng/ul	100
57) 2,4-Dinitrotoluene	14.651	165	20662	3.161	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	14.904	232	21088	3.796	ng/ul	97
59) Diethylphthalate	15.116	149	125860	4.411	ng/ul	98
61) Fluorene	15.328	166	133314	4.851	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.328	204	58710	4.838	ng/ul	98
63) 4-Nitroaniline	15.363	138	7133	1.862	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.416	198	10164	2.589	ng/ul#	86
67) N-Nitrosodiphenylamine	15.545	169	100029	4.185	ng/ul	97
68) 4-Bromophenyl-phenylether	16.228	248	32705	4.434	ng/ul	87
69) Hexachlorobenzene	16.328	284	40672	4.720	ng/ul	100
70) Atrazine	16.516	200	34287	4.796	ng/ul	96
71) Pentachlorophenol	16.686	266	19361	3.836	ng/ul	98
72) Phenanthrene	17.081	178	216991	4.847	ng/ul	100
74) Anthracene	17.175	178	209006	4.694	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.051	216	60319	4.820	ng/ul	99
76) Pentachlorobenzene	14.587	250	47135	4.419	ng/ul	99
77) Carbazole	17.457	167	186866	4.680	ng/ul	100
78) Di-n-butylphthalate	18.028	149	188787	3.811	ng/ul	98
80) Fluoranthene	19.075	202	233563	4.987	ng/ul	99
82) Pyrene	19.427	202	260660	5.333	ng/ul	98
83) Butylbenzylphthalate	20.327	149	66936	3.587	ng/ul	89
84) 3,3'-Dichlorobenzidine	21.098	252	44186	3.567	ng/ul	96
85) Benzo(a)anthracene	21.151	228	225609	4.766	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.098	149	110256	3.458	ng/ul	99
87) Chrysene	21.204	228	225702	4.844	ng/ul	99
89) Di-n-octyl phthalate	21.998	149	157396	3.185	ng/ul	100
90) Benzo(b)fluoranthene	22.774	252	219467	4.752	ng/ul	98
91) Benzo(k)fluoranthene	22.821	252	218964	5.054	ng/ul	98
93) Benzo(a)pyrene	23.374	252	207753	4.621	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	25.862	276	229334	4.274	ng/ul	98
95) Dibenzo(a,h)anthracene	25.886	278	193649	4.256	ng/ul	99
96) Benzo(g,h,i)perylene	26.592	276	191949	4.175	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 5.868 ng/uL
 RT: 3.122 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BP011311.D
 Acq: 08 Aug 2022 15:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005638


Tgt Ion: 88 Resp: 11677
 Ion Ratio Lower Upper
 88 100
 43 48.8 29.6 44.4#
 58 86.5 49.1 73.7#



#5
 Pyridine
 Concen: 11.341 ng/uL
 RT: 3.528 min Scan# 92
 Delta R.T. 0.012 min
 Lab File: BP011311.D
 Acq: 08 Aug 2022 15:07

Tgt Ion: 79 Resp: 66844
 Ion Ratio Lower Upper
 79 100
 52 84.6 57.8 86.6
 51 44.8 31.5 47.3









#6
Benzaldehyde
Concen: 5.834 ng/ul
RT: 6.746 min Scan# 611
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion: 77 Resp: 41470
Ion Ratio Lower Upper
77 100
105 94.1 72.0 108.0
106 90.1 69.4 104.0



#8
Phenol
Concen: 4.549 ng/ul
RT: 6.805 min Scan# 649
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion: 94 Resp: 70926
Ion Ratio Lower Upper
94 100
65 30.7 24.7 37.1
66 39.9 32.6 49.0









#10
Bis(2-Chloroethyl)ether
Concen: 4.693 ng/ul
RT: 7.034 min Scan# 687
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion: 93 Resp: 60041
Ion Ratio Lower Upper
93 100
63 84.6 71.0 106.6
95 31.6 26.0 39.0



#12
2-Chlorophenol
Concen: 4.444 ng/ul
RT: 7.158 min Scan# 709
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion: 128 Resp: 54706
Ion Ratio Lower Upper
128 100
64 53.6 43.8 65.6
130 31.6 25.4 38.2









#13
2-Methylphenol
Concen: 4.083 ng/ul
RT: 8.052 min Scan# 80
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:108 Resp: 48161
Ion Ratio Lower Upper
108 100
107 93.3 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 2.815 ng/ul
RT: 8.128 min Scan# 874
Delta R.T. -0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion: 45 Resp: 61366
Ion Ratio Lower Upper
45 100
77 15.5 9.9 14.9#
79 10.8 7.2 10.8









#16
Acetophenone
Concen: 4.530 ng/ul
RT: 8.422 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:105 Resp: 87835
Ion Ratio Lower Upper
105 100
77 82.0 65.4 98.0
51 33.3 26.5 39.7



#17
N-Nitroso-di-n-propylamine
Concen: 4.222 ng/ul
RT: 8.410 min Scan# 922
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion: 70 Resp: 39833
Ion Ratio Lower Upper
70 100
42 64.5 55.5 83.3
101 8.5 7.1 10.7
130 19.3 14.8 22.2









#18
4-Methylphenol
Concen: 4.192 ng/ul
RT: 8.375 min Scan# 916
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:108 Resp: 53034
Ion Ratio Lower Upper
108 100
107 113.7 88.2 132.4



#19
Hexachloroethane
Concen: 3.967 ng/ul
RT: 8.657 min Scan# 964
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:117 Resp: 23416
Ion Ratio Lower Upper
117 100
201 78.5 62.2 93.2
199 50.9 39.8 59.6









#22
Nitrobenzene
Concen: 3.809 ng/ul
RT: 8.799 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion: 77 Resp: 55274
Ion Ratio Lower Upper
77 100
123 38.6 31.4 47.0
65 13.0 10.6 15.8



#23
Isophorone
Concen: 4.039 ng/ul
RT: 9.328 min Scan# 1078
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion: 82 Resp: 104311
Ion Ratio Lower Upper
82 100
95 7.6 5.8 8.8
138 16.0 12.6 18.8









#25
2-Nitrophenol
Concen: 3.283 ng/ul
RT: 9.504 min Scan# 1108
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:139 Resp: 18858
Ion Ratio Lower Upper
139 100
65 64.7 46.7 70.1
109 33.4 24.2 36.2



#26
2,4-Dimethylphenol
Concen: 4.343 ng/ul
RT: 9.575 min Scan# 1120
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:107 Resp: 56434
Ion Ratio Lower Upper
107 100
121 51.1 40.3 60.5
122 87.5 67.7 101.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.531 ng/ul
 RT: 9.816 min Scan# 1161
 Delta R.T. -0.000 min
 Lab File: BP011311.D
 Acq: 08 Aug 2022 15:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005638


Tgt Ion: 93	Resp:	73668
Ion	Ratio	Lower Upper
93	100	
95	31.6	25.2 37.8
123	11.1	8.4 12.6



#29
 2,4-Dichlorophenol
 Concen: 4.160 ng/ul
 RT: 10.040 min Scan# 1199
 Delta R.T. 0.006 min
 Lab File: BP011311.D
 Acq: 08 Aug 2022 15:07

Tgt Ion: 162	Resp:	41151
Ion	Ratio	Lower Upper
162	100	
164	64.8	52.6 79.0
98	43.4	34.2 51.4









#30
Naphthalene
Concen: 4.862 ng/ul
RT: 10.434 min Scan# 1266
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:128 Resp: 188966
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.4
127 13.4 10.4 15.6



#32
4-Chloroaniline
Concen: 4.489 ng/ul
RT: 10.563 min Scan# 1288
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:127 Resp: 66585
Ion Ratio Lower Upper
127 100
129 30.0 26.6 40.0









#33
Hexachlorobutadiene
Concen: 4.455 ng/ul
RT: 10.716 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:	225	Resp:	27528
Ion Ratio	Lower	Upper	
225	100		
223	64.1	51.2	76.8
227	69.8	50.6	75.8



#34
Caprolactam
Concen: 2.821 ng/ul
RT: 11.351 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:	113	Resp:	8267
Ion Ratio	Lower	Upper	
113	100		
55	212.8	164.2	246.4
56	144.7	127.0	190.6









#35
4-Chloro-3-methylphenol
Concen: 3.932 ng/ul
RT: 11.704 min Scan# 1481
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:107 Resp: 43989
Ion Ratio Lower Upper
107 100
144 25.4 19.0 28.4
142 79.4 59.8 89.8



#36
2-Methylnaphthalene
Concen: 4.613 ng/ul
RT: 12.063 min Scan# 1543
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:142 Resp: 112036
Ion Ratio Lower Upper
142 100
141 89.9 71.8 107.8









#37
1-Methylnaphthalene
Concen: 4.704 ng/ul
RT: 12.287 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:142 Resp: 114761
Ion Ratio Lower Upper
142 100
141 89.9 74.1 111.1
116 4.0 3.4 5.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.559 ng/ul
RT: 12.434 min Scan# 1606
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:216 Resp: 49889
Ion Ratio Lower Upper
216 100
214 79.6 61.1 91.7
179 25.3 18.6 27.8
108 23.4 19.1 28.7









#40
 Hexachlorocyclopentadiene
 Concen: 2.641 ng/ul
 RT: 12.410 min Scan# 1602
 Delta R.T. 0.006 min
 Lab File: BP011311.D
 Acq: 08 Aug 2022 15:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005638


Tgt Ion:237 Resp: 18261
 Ion Ratio Lower Upper
 237 100
 235 62.7 50.4 75.6
 272 10.3 9.9 14.9



#41
 2,4,6-Trichlorophenol
 Concen: 3.623 ng/ul
 RT: 12.687 min Scan# 1649
 Delta R.T. -0.000 min
 Lab File: BP011311.D
 Acq: 08 Aug 2022 15:07

Tgt Ion:196 Resp: 24838
 Ion Ratio Lower Upper
 196 100
 198 97.4 75.5 113.3
 200 30.9 24.4 36.6





#42

2,4,5-Trichlorophenol

Concen: 3.398 ng/ul

RT: 12.757 min Scan# 1661

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:196 Resp: 24998

Ion Ratio Lower Upper

196 100

198 89.7 76.8 115.2

200 31.0 25.0 37.6



#43

1,1'-Biphenyl

Concen: 4.613 ng/ul

RT: 13.092 min Scan# 1718

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07




Tgt Ion:154 Resp: 147328


Ion Ratio Lower Upper

154 100

153 39.5 31.9 47.9

76 16.1 12.2 18.2





#44

2-Chloronaphthalene

Concen: 4.599 ng/ul

RT: 13.128 min Scan# 1724

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:162 Resp: 109560

Ion Ratio Lower Upper

162 100

164 32.4 25.6 38.4

127 43.6 34.6 51.8



#45

2-Nitroaniline

Concen: 2.992 ng/ul

RT: 13.351 min Scan# 1762

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07




Tgt Ion: 65 Resp: 17149


Ion Ratio Lower Upper

65 100

92 70.0 51.4 77.2

138 85.2 67.4 101.0









#47
Dimethylphthalate
Concen: 4.493 ng/ul
RT: 13.740 min Scan# 1827
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:163 Resp: 124217
Ion Ratio Lower Upper
163 100
194 4.1 2.9 4.3
164 10.9 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 3.013 ng/ul
RT: 13.857 min Scan# 1848
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:165 Resp: 12506
Ion Ratio Lower Upper
165 100
89 87.1 65.8 98.6
121 27.6 24.2 36.2









#50
Acenaphthylene
Concen: 4.268 ng/ul
RT: 13.987 min Scan# 1869
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:152 Resp: 166300
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.8
153 12.4 10.5 15.7



#51
3-Nitroaniline
Concen: 2.784 ng/ul
RT: 14.192 min Scan# 1905
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:138 Resp: 12744
Ion Ratio Lower Upper
138 100
108 16.5 9.9 14.9#
92 137.4 106.8 160.2





#52

Acenaphthene

Concen: 4.778 ng/ul

RT: 14.328 min Scan# 1928

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:153 Resp: 120412

Ion Ratio Lower Upper

153 100

152 49.3 39.5 59.3

154 86.6 68.6 103.0



#53

2,4-Dinitrophenol

Concen: 2.635 ng/ul

RT: 14.404 min Scan# 1941

Delta R.T. 0.006 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07




Tgt Ion:184 Resp: 6495


Ion Ratio Lower Upper

184 100

63 81.2 75.0 112.6

154 65.8 53.1 79.7









#55
4-Nitrophenol
Concen: 2.669 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:109 Resp: 11225
Ion Ratio Lower Upper
109 100
139 128.2 100.6 150.8
65 153.3 122.6 184.0



#56
Dibenzofuran
Concen: 4.731 ng/ul
RT: 14.669 min Scan# 1986
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:168 Resp: 163160
Ion Ratio Lower Upper
168 100
139 42.7 34.2 51.2









#57
2,4-Dinitrotoluene
Concen: 3.161 ng/ul
RT: 14.651 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:165 Resp: 20662
Ion Ratio Lower Upper
165 100
63 71.5 50.6 75.8
182 2.3 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 3.796 ng/ul
RT: 14.904 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:232 Resp: 21088
Ion Ratio Lower Upper
232 100
131 61.9 50.4 75.6
130 3.1 2.6 3.8
166 37.1 31.9 47.9









#59
Diethylphthalate
Concen: 4.411 ng/ul
RT: 15.116 min Scan# 2062
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:149 Resp: 125860
Ion Ratio Lower Upper
149 100
177 19.4 16.1 24.1
150 12.4 9.4 14.0



#61
Fluorene
Concen: 4.851 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:166 Resp: 133314
Ion Ratio Lower Upper
166 100
165 98.5 77.7 116.5
167 12.6 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 4.838 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:204 Resp: 58710
Ion Ratio Lower Upper
204 100
206 34.2 27.3 40.9
141 74.0 57.7 86.5



#63
4-Nitroaniline
Concen: 1.862 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:138 Resp: 7133
Ion Ratio Lower Upper
138 100
92 66.8 55.9 83.9
108 88.4 67.1 100.7









#66
4,6-Dinitro-2-methylphenol
Concen: 2.589 ng/ul
RT: 15.416 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:198 Resp: 10164
Ion Ratio Lower Upper
198 100
182 14.1 7.4 11.0#
77 44.2 28.6 42.8#



#67
N-Nitrosodiphenylamine
Concen: 4.185 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:169 Resp: 100029
Ion Ratio Lower Upper
169 100
168 70.0 53.6 80.4
167 37.2 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 4.434 ng/ul
RT: 16.228 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:248 Resp: 32705
Ion Ratio Lower Upper
248 100
250 105.8 77.3 115.9
141 111.4 76.4 114.6



#69
Hexachlorobenzene
Concen: 4.720 ng/ul
RT: 16.328 min Scan# 2268
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:284 Resp: 40672
Ion Ratio Lower Upper
284 100
142 49.2 39.5 59.3
249 31.8 25.1 37.7









#70
Atrazine
Concen: 4.796 ng/ul
RT: 16.516 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:200 Resp: 34287
Ion Ratio Lower Upper
200 100
173 28.6 22.6 34.0
215 41.8 36.5 54.7



#71
Pentachlorophenol
Concen: 3.836 ng/ul
RT: 16.686 min Scan# 2329
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:266 Resp: 19361
Ion Ratio Lower Upper
266 100
264 63.4 52.2 78.2
268 64.5 50.6 75.8









#72
Phenanthrene
Concen: 4.847 ng/ul
RT: 17.081 min Scan# 2169
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:178 Resp: 216991
Ion Ratio Lower Upper
178 100
179 15.3 12.1 18.1
176 19.2 15.4 23.2



#74
Anthracene
Concen: 4.694 ng/ul
RT: 17.175 min Scan# 2412
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:178 Resp: 209006
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 19.1 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.820 ng/uL
RT: 13.051 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:216 Resp: 60319
Ion Ratio Lower Upper
216 100
214 77.7 62.0 93.0
218 46.8 38.2 57.4



#76
Pentachlorobenzene
Concen: 4.419 ng/uL
RT: 14.587 min Scan# 1972
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:250 Resp: 47135
Ion Ratio Lower Upper
250 100
248 60.1 49.2 73.8
252 63.1 50.8 76.2





#77

Carbazole

Concen: 4.680 ng/ul

RT: 17.457 min Scan# 2459

Delta R.T. 0.006 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:167 Resp: 186866

Ion Ratio Lower Upper

167 100

166 21.2 17.1 25.7

139 14.5 11.5 17.3



#78

Di-n-butylphthalate

Concen: 3.811 ng/ul

RT: 18.028 min Scan# 2557

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07




Tgt Ion:149 Resp: 188787


Ion Ratio Lower Upper

149 100

150 9.6 7.4 11.0

104 7.6 5.1 7.7









#80
Fluoranthene
Concen: 4.987 ng/ul
RT: 19.075 min Scan# 2735
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:202 Resp: 233563
Ion Ratio Lower Upper
202 100
101 15.1 11.9 17.9
100 11.0 9.3 13.9



#82
Pyrene
Concen: 5.333 ng/ul
RT: 19.427 min Scan# 2795
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:202 Resp: 260660
Ion Ratio Lower Upper
202 100
101 16.7 14.3 21.5
100 14.4 11.8 17.6









#83
Butylbenzylphthalate
Concen: 3.587 ng/ul
RT: 20.327 min Scan# 2948
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:149 Resp: 66936
Ion Ratio Lower Upper
149 100
91 91.4 64.2 96.4
206 18.7 14.0 21.0



#84
3,3'-Dichlorobenzidine
Concen: 3.567 ng/ul
RT: 21.098 min Scan# 3079
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:252 Resp: 44186
Ion Ratio Lower Upper
252 100
254 61.4 51.8 77.6
126 16.4 11.4 17.0









#85
Benzo(a)anthracene
Concen: 4.766 ng/ul
RT: 21.151 min Scan# 3088
Delta R.T. -0.000 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:228 Resp: 225609
Ion Ratio Lower Upper
228 100
229 19.9 15.8 23.6
226 27.3 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 3.458 ng/ul
RT: 21.098 min Scan# 3079
Delta R.T. 0.006 min
Lab File: BP011311.D
Acq: 08 Aug 2022 15:07

Tgt Ion:149 Resp: 110256
Ion Ratio Lower Upper
149 100
167 24.9 20.4 30.6
279 3.9 2.9 4.3





#87

Chrysene

Concen: 4.844 ng/ul

RT: 21.204 min Scan# 3097

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:228 Resp: 225702

Ion Ratio Lower Upper

228 100

226 28.9 23.8 35.6

229 19.0 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 3.185 ng/ul




RT: 21.998 min Scan# 3232


Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Tgt Ion:149 Resp: 157396





#90

Benzo(b)fluoranthene

Concen: 4.752 ng/ul

RT: 22.774 min Scan# 3364

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:252 Resp: 219467

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 15.0 10.8 16.2



#91

Benzo(k)fluoranthene

Concen: 5.054 ng/ul

RT: 22.821 min Scan# 3372

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07




Tgt Ion:252 Resp: 218964


Ion Ratio Lower Upper

252 100

253 22.5 17.6 26.4

125 15.4 11.0 16.6





#93

Benzo(a)pyrene

Concen: 4.621 ng/ul

RT: 23.374 min Scan# 3466

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:252 Resp: 207753

Ion Ratio Lower Upper

252 100

253 22.2 17.4 26.2

125 17.7 12.5 18.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.274 ng/ul

RT: 25.862 min Scan# 3889

Delta R.T. 0.006 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07




Tgt Ion:276 Resp: 229334


Ion Ratio Lower Upper

276 100

138 33.0 25.5 38.3

277 25.4 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 4.256 ng/ul

RT: 25.886 min Scan# 3893

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:278 Resp: 193649

Ion Ratio Lower Upper

278 100

139 22.1 18.5 27.7

279 24.5 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 4.175 ng/ul

RT: 26.592 min Scan# 4013

Delta R.T. -0.000 min

Lab File: BP011311.D

Acq: 08 Aug 2022 15:07

Tgt Ion:276 Resp: 191949

Ion Ratio Lower Upper

276 100

138 30.8 22.8 34.2

277 23.6 18.5 27.7

