

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081621\
 Data File : BP006679.D
 Acq On : 16 Aug 2021 10:35
 Operator : CG/JU
 Sample : PB138361BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK361

Quant Time: Aug 16 11:09:16 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP081021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 16 11:01:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.704	152	144450	20.000	ng/u1	0.00
20) Naphthalene-d8	10.493	136	598940	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.351	164	327654	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.110	188	650927	20.000	ng/u1	0.00
79) Chrysene-d12	21.215	240	671710	20.000	ng/u1	0.00
88) Perylene-d12	23.539	264	733710	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.216	96	33214	7.777	ng/uL	0.00
4) Pyridine-d5	3.622	84	386610	31.491	ng/u1	0.00
7) Phenol-d5	6.887	99	492742	33.932	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.052	67	315598	35.364	ng/u1	0.00
11) 2-Chlorophenol-d4	7.240	132	393332	36.369	ng/u1	0.00
15) 4-Methylphenol-d8	8.422	113	379810	32.739	ng/u1	0.00
21) Nitrobenzene-d5	8.863	128	192829	37.928	ng/u1	0.00
24) 2-Nitrophenol-d4	9.581	143	192256	36.139	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.116	165	315723	35.126	ng/u1	0.00
31) 4-Chloroaniline-d4	10.640	131	281210	19.249	ng/u1	0.00
46) Dimethylphthalate-d6	13.775	166	968906	38.560	ng/u1	0.00
49) Acenaphthylene-d8	14.039	160	1224702	39.735	ng/u1	0.00
54) 4-Nitrophenol-d4	14.569	143	170764	32.535	ng/u1	0.00
60) Fluorene-d10	15.351	176	791793	38.745	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.481	200	128296	29.818	ng/u1	0.00
73) Anthracene-d10	17.210	188	1209715	39.278	ng/u1	0.00
81) Pyrene-d10	19.457	212	1338984	37.959	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.386	264	1564125	39.532	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081621\
Data File : BP006679.D
Acq On : 16 Aug 2021 10:35
Operator : CG/JU
Sample : PB138361BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK361

Quant Time: Aug 16 11:09:16 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP081021.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 16 11:01:16 2021
Response via : Initial Calibration

