

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP082620\
 Data File : BP003125.D
 Acq On : 26 Aug 2020 21:26
 Operator : CG/JU
 Sample : L3786-08DL 4X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 1994DL

Quant Time: Aug 27 00:29:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 24 15:32:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	293406	20.00	ng	0.00
21) Naphthalene-d8	10.44	136	1179051	20.00	ng	0.00
39) Acenaphthene-d10	14.30	164	700604	20.00	ng	0.00
64) Phenanthrene-d10	17.06	188	1036494	20.00	ng	0.00
76) Chrysene-d12	21.16	240	1375378	20.00	ng	0.00
86) Perylene-d12	23.39	264	1516218	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.27	112	230696	13.90	ng	0.00
7) Phenol-d6	6.84	99	281482	13.53	ng	0.00
23) Nitrobenzene-d5	8.80	82	160034	9.78	ng	0.00
42) 2,4,6-Tribromophenol	15.80	330	130686	13.79	ng	0.00
45) 2-Fluorobiphenyl	12.92	172	525274	10.98	ng	0.00
79) Terphenyl-d14	19.64	244	578574	9.26	ng	0.00

Target Compounds


						Qvalue
32) Benzoic acid	9.74	122	783	9.054	ng	91
49) Acenaphthylene	14.02	152	266738	4.101	ng	94
51) 2,6-Dinitrotoluene	13.86	165	46705	4.206	ng	# 41
52) Acenaphthene	14.37	154	127358	3.187	ng	98
54) 2,4-Dinitrophenol	14.41	184	27	7.759	ng	# 1
58) Fluorene	15.36	166	197755	4.091	ng	96
62) 4-Nitroaniline	15.20	138	35666	2.758	ng	80
69) Atrazine	16.54	200	22359	2.187	ng	# 1
71) Phenanthrene	17.11	178	2065145	36.770	ng	# 95
72) Anthracene	17.20	178	627529	11.675	ng	# 91
73) Carbazole	17.47	167	215425	4.176	ng	# 70
75) Fluoranthene	19.09	202	3748381	57.745	ng	94
78) Pyrene	19.44	202	3592308	40.243	ng	97
81) Benzo(a)anthracene	21.14	228	2501505	29.036	ng	97
83) Chrysene	21.19	228	2518055	30.527	ng	97
84) Bis(2-ethylhexyl)phthalate	21.09	149	363654	6.851	ng	96
87) Indeno(1,2,3-cd)pyrene	25.66	276	2089898	18.483	ng	95
88) Benzo(b)fluoranthene	22.72	252	3605782	38.263	ng	97
89) Benzo(k)fluoranthene	22.72	252	3605782	40.047	ng	98
90) Benzo(a)pyrene	23.30	252	2534082	28.985	ng	98
91) Dibenzo(a,h)anthracene	25.92	278	382391	4.117	ng	96
92) Benzo(g,h,i)perylene	26.36	276	2050962	22.002	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP082620\
Data File : BP003125.D
Acq On : 26 Aug 2020 21:26
Operator : CG/JU
Sample : L3786-08DL 4X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
1994DL

Quant Time: Aug 27 00:29:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 24 15:32:08 2020
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.66 min Scan# 812
Delta R.T. -0.01 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26

Instrument :
BNA_P
ClientSampleId :
1994DL




Tot Ion	Ratio	Lower	Upper
152	100		
150	157.5	123.1	184.7
115	53.4	42.8	64.2




#5

2-Fluorophenol
Concen: 13.901 ng
RT: 5.27 min Scan# 405
Delta R.T. -0.00 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26

Tgt Ion	Ratio	Lower	Upper
112	100		
64	40.1	33.0	49.6
63	20.8	16.9	25.3





#7

Phenol-d6

Concen: 13.531 ng

RT: 6.84 min Scan# 672

Delta R.T. -0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Instrument :


BNA_P

ClientSampleId :

1994DL

Tot Ion: 99 Resp: 281482


Ion	Ratio	Lower	Upper
99	100		
42	10.6	8.5	12.7
71	29.3	22.8	34.2




Abundance Ion 99.00 (98.70 to 99.70): BP0

Ion 42.00 (41.70 to 42.70): BP0

Ion 71.00 (70.70 to 71.70): BP0



Time-->



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.44 min Scan# 1284


Delta R.T. -0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Tgt Ion: 136 Resp: 1179051

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	4.5	3.5	5.3
68	4.4	3.2	4.8




Abundance Ion 136.00 (135.70 to 136.70): E


Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BP0

Ion 68.00 (67.70 to 68.70): BP0







Time-->



#23
 Nitrobenzene-d5
 Concen: 9.785 ng
 RT: 8.80 min Scan# 1006
 Delta R.T. -0.01 min
 Lab File: BP003125.D
 Acq: 26 Aug 2020 21:26




Instrument :
 BNA_P
 ClientSampleID :
 1994DL


Tot Ion	82	Resp	160034
Ion Ratio	Lower	Upper	
82	100		
128	53.5	44.9	67.3
54	38.1	29.4	44.2



#32
 Benzoic acid
 Concen: 9.054 ng
 RT: 9.74 min Scan# 1165
 Delta R.T. -0.04 min
 Lab File: BP003125.D
 Acq: 26 Aug 2020 21:26

Tgt Ion	122	Resp	783
Ion Ratio	Lower	Upper	
122	100		
105	107.0	98.1	138.1
77	81.4	55.3	95.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.30 min Scan# 1941

Delta R.T. -0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Instrument :

BNA_P

ClientSampleId :

1994DL


Tot Ion:164 Resp: 700604

Ion Ratio Lower Upper

164 100

162 100.2 81.2 121.8




160 44.5 36.0 54.0



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#42

2,4,6-Tribromophenol

Concen: 13.792 ng

RT: 15.80 min Scan# 2195

Delta R.T. -0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26


Tgt Ion:330 Resp: 130686

Ion Ratio Lower Upper

330 100

332 97.0 77.3 115.9



141 38.1 24.4 36.6#




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#45
2-Fluorobiphenyl
Concen: 10.980 ng
RT: 12.92 min Scan# 1706
Delta R.T. -0.01 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26




Instrument :
BNA_P
ClientSampleId :
1994DL


Tot Ion:172 Resp: 525274
Ion Ratio Lower Upper
172 100
171 35.5 28.7 43.1
170 23.6 19.0 28.4



#49
Acenaphthylene
Concen: 4.101 ng
RT: 14.02 min Scan# 1893
Delta R.T. -0.01 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26

Tgt Ion:152 Resp: 266738
Ion Ratio Lower Upper
152 100
151 23.4 16.2 24.4
153 14.6 10.5 15.7





#51

2,6-Dinitrotoluene

Concen: 4.206 ng

RT: 13.86 min Scan# 1865

Delta R.T. -0.02 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Instrument :

BNA_P

ClientSampled :

1994DL


Tot Ion:165 Resp: 46705

Ion Ratio Lower Upper

165 100

63 3.0 27.5 41.3#




89 2.4 33.6 50.4#



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BP0

Ion 89.00 (88.70 to 89.70): BP0



#52

Acenaphthene

Concen: 3.187 ng

RT: 14.37 min Scan# 1952

Delta R.T. -0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26


Tgt Ion:154 Resp: 127358

Ion Ratio Lower Upper

154 100

153 111.7 89.2 133.8



152 56.4 42.5 63.7




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#54
2,4-Dinitrophenol
Concen: 7.759 ng
RT: 14.41 min Scan# 1959
Delta R.T. -0.01 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26




Instrument :
BNA_P
ClientSampleId :
1994DL


Tot Ion:184 Resp: 27
Ion Ratio Lower Upper
184 100
63 877.5 32.0 48.0#
154 1824.7 48.2 72.2#



#58
Fluorene
Concen: 4.091 ng
RT: 15.36 min Scan# 2120
Delta R.T. -0.01 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26

Tgt Ion:166 Resp: 197755
Ion Ratio Lower Upper
166 100
165 102.0 78.3 117.5
167 14.7 11.0 16.4









#62
4-Nitroaniline
Concen: 2.758 ng
RT: 15.20 min Scan# 2093
Delta R.T. -0.18 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26




Instrument :
BNA_P
ClientSampleId :
1994DL


Tot Ion:138 Resp: 35666
Ion Ratio Lower Upper
138 100
92 27.4 21.1 61.1
108 81.7 46.6 86.6



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.06 min Scan# 2410
Delta R.T. -0.00 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26

Tgt Ion:188 Resp: 1036494
Ion Ratio Lower Upper
188 100
94 10.6 6.2 9.4#
80 9.1 6.4 9.6









#69
Atrazine
Concen: 2.187 ng
RT: 16.54 min Scan# 2321
Delta R.T. 0.01 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26




Instrument :
BNA_P
ClientSampleId :
1994DL


Tot Ion:200 Resp: 22359
Ion Ratio Lower Upper
200 100
173 117.8 6.3 46.3#
215 120.0 27.9 67.9#



#71
Phenanthrene
Concen: 36.770 ng
RT: 17.11 min Scan# 2418
Delta R.T. 0.01 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26

Tgt Ion:178 Resp: 2065145
Ion Ratio Lower Upper
178 100
176 20.1 15.4 23.2
179 19.1 12.4 18.6#





#72

Anthracene

Concen: 11.675 ng

RT: 17.20 min Scan# 2433

Delta R.T. 0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Instrument :

BNA_P

ClientSampleId :

1994DL


Tot Ion:178 Resp: 627529

Ion Ratio Lower Upper

178 100

176 21.3 15.2 22.8




179 22.1 12.6 19.0#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#73

Carbazole

Concen: 4.176 ng

RT: 17.47 min Scan# 2479

Delta R.T. 0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26


Tgt Ion:167 Resp: 215425

Ion Ratio Lower Upper

167 100

166 32.5 17.6 26.4#



139 30.1 10.3 15.5#




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#75

Fluoranthene

Concen: 57.745 ng

RT: 19.09 min Scan# 2755

Delta R.T. 0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Instrument :


BNA_P

ClientSampleId :

1994DL

Tot Ion:202 Resp: 3748381




Ion	Ratio	Lower	Upper
202	100		
101	6.1	0.0	30.0
203	19.0	0.0	37.7



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.16 min Scan# 3106


Delta R.T. -0.00 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Tgt Ion:240 Resp: 1375378



Ion	Ratio	Lower	Upper
240	100		
120	10.3	8.4	12.6
236	26.2	20.6	31.0




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#78

Pvrene

Concen: 40.243 ng

RT: 19.44 min Scan# 2815

Delta R.T. 0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Instrument :


BNA_P

ClientSampleId :

1994DL

Tot Ion:202 Resp: 3592308




Ion	Ratio	Lower	Upper
202	100		
200	22.0	17.0	25.4
203	19.1	14.0	21.0



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#79

Terphenyl-d14

Concen: 9.256 ng

RT: 19.64 min Scan# 2848


Delta R.T. -0.00 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Tgt Ion:244 Resp: 578574



Ion	Ratio	Lower	Upper
244	100		
212	9.4	5.8	8.8#
122	13.7	7.7	11.5#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#81

Benzo(a)anthracene

Concen: 29.036 ng

RT: 21.14 min Scan# 3104

Delta R.T. 0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Instrument :

BNA_P

ClientSampled :

1994DL


Tot Ion:228 Resp: 2501505

Ion Ratio Lower Upper

228 100

226 29.1 22.0 33.0




229 21.2 16.2 24.2



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#83

Chrysene

Concen: 30.527 ng

RT: 21.19 min Scan# 3112

Delta R.T. -0.00 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26


Tgt Ion:228 Resp: 2518055

Ion Ratio Lower Upper

228 100

226 30.8 24.2 36.4



229 22.7 16.0 24.0




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#84

Bis(2-ethylhexyl)phthalate

Concen: 6.851 ng

RT: 21.09 min Scan# 3094

Delta R.T. -0.00 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26

Instrument :

BNA_P

ClientSampleId :

1994DL


Tot Ion:149 Resp: 363654

Ion Ratio Lower Upper

149 100

167 30.1 22.5 33.7




279 5.6 3.8 5.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 23.39 min Scan# 3486

Delta R.T. 0.01 min

Lab File: BP003125.D

Acq: 26 Aug 2020 21:26


Tgt Ion:264 Resp: 1516218

Ion Ratio Lower Upper

264 100

260 23.9 19.0 28.6



265 22.8 17.7 26.5




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E









#87
 Indeno(1,2,3-cd)pyrene
 Concen: 18.483 ng
 RT: 25.66 min Scan# 3872
 Delta R.T. -0.01 min
 Lab File: BP003125.D
 Acq: 26 Aug 2020 21:26

Instrument :
 BNA_P
 ClientSampleID :
 1994DL

Tot Ion:276 Resp: 2089898




Ion	Ratio	Lower	Upper
276	100		
138	20.4	20.2	30.4
277	24.9	20.4	30.6




#88
 Benzo(b)fluoranthene
 Concen: 38.263 ng
 RT: 22.72 min Scan# 3372
 Delta R.T. -0.00 min
 Lab File: BP003125.D
 Acq: 26 Aug 2020 21:26

Tgt Ion:252 Resp: 3605782

Ion	Ratio	Lower	Upper
252	100		
253	23.1	17.5	26.3
125	10.9	7.6	11.4









#89
Benzo(k)fluoranthene
Concen: 40.047 ng
RT: 22.72 min Scan# 3372
Delta R.T. -0.04 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26




Instrument :
BNA_P
ClientSampleId :
1994DL


Tot Ion:252 Resp: 3605782
Ion Ratio Lower Upper
252 100
253 23.1 17.8 26.6
125 10.9 8.0 12.0



#90
Benzo(a)pyrene
Concen: 28.985 ng
RT: 23.30 min Scan# 3470
Delta R.T. 0.01 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26

Tgt Ion:252 Resp: 2534082
Ion Ratio Lower Upper
252 100
253 23.1 17.7 26.5
125 10.9 8.8 13.2









#91
Dibenzo(a,h)anthracene
Concen: 4.117 ng
RT: 25.92 min Scan# 3916
Delta R.T. 0.24 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26

Instrument :
BNA_P
ClientSampleId :
1994DL

Tot Ion:278 Resp: 382391
Ion Ratio Lower Upper
278 100
139 16.7 12.8 19.2
279 26.8 19.0 28.6



#92
Benzo(g,h,i)perylene
Concen: 22.002 ng
RT: 26.36 min Scan# 3991
Delta R.T. -0.00 min
Lab File: BP003125.D
Acq: 26 Aug 2020 21:26

Tgt Ion:276 Resp: 2050962
Ion Ratio Lower Upper
276 100
277 24.1 19.4 29.2
138 21.1 17.0 25.6

