

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091922\
 Data File : BP011738.D
 Acq On : 19 Sep 2022 16:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 20 04:44:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 04:42:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	39.358	1.6	103	0.00
3	Pyridine	40.000	39.156	2.1	100	0.00
4	n-Nitrosodimethylamine	40.000	39.860	0.4	102	0.00
5 S	2-Fluorophenol	80.000	80.762	-1.0	104	0.00
6	Aniline	40.000	39.075	2.3	98	0.00
7 S	Phenol-d6	80.000	79.153	1.1	100	0.00
8	2-Chlorophenol	40.000	39.970	0.1	102	0.00
9	Benzaldehyde	40.000	37.149	7.1	98	0.00
10 C	Phenol	40.000	39.176	2.1	99	0.00
11	bis(2-Chloroethyl)ether	40.000	38.474	3.8	98	0.00
12	1,3-Dichlorobenzene	40.000	40.401	-1.0	104	0.00
13 C	1,4-Dichlorobenzene	40.000	40.224	-0.6	103	0.00
14	1,2-Dichlorobenzene	40.000	40.253	-0.6	103	0.00
15	Benzyl Alcohol	40.000	38.593	3.5	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.894	2.8	99	0.00
17	2-Methylphenol	40.000	39.360	1.6	99	0.00
18	Hexachloroethane	40.000	40.248	-0.6	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.845	5.4	94	0.00
20	3+4-Methylphenols	40.000	38.779	3.1	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	40.847	-2.1	98	0.00
23 S	Nitrobenzene-d5	80.000	81.852	-2.3	97	0.00
24	Nitrobenzene	40.000	40.873	-2.2	98	0.00
25	Isophorone	40.000	39.912	0.2	94	0.00
26 C	2-Nitrophenol	40.000	36.440	8.9	92	0.00
27	2,4-Dimethylphenol	40.000	41.272	-3.2	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.949	0.1	96	0.00
29 C	2,4-Dichlorophenol	40.000	43.117	-7.8	102	0.00
30	1,2,4-Trichlorobenzene	40.000	43.527	-8.8	106	0.00
31	Naphthalene	40.000	41.211	-3.0	100	0.00
32	Benzoic acid	40.000	34.216	14.5	85	0.00
33	4-Chloroaniline	40.000	41.362	-3.4	98	0.00
34 C	Hexachlorobutadiene	40.000	45.357	-13.4	110	0.00
35	Caprolactam	40.000	38.258	4.4	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.180	-0.4	94	0.00
37	2-Methylnaphthalene	40.000	41.130	-2.8	99	0.00
38	1-Methylnaphthalene	40.000	41.173	-2.9	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.770	-9.4	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.250	-13.1	107	0.00
42 S	2,4,6-Tribromophenol	80.000	105.574	-32.0#	124	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.481	-8.7	102	0.00
44	2,4,5-Trichlorophenol	40.000	43.668	-9.2	104	0.00
45 S	2-Fluorobiphenyl	80.000	86.161	-7.7	103	0.00
46	1,1'-Biphenyl	40.000	41.363	-3.4	100	0.00
47	2-Chloronaphthalene	40.000	41.539	-3.8	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091922\
 Data File : BP011738.D
 Acq On : 19 Sep 2022 16:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 20 04:44:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 04:42:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.546	-1.4	92	0.00
49	Acenaphthylene	40.000	41.920	-4.8	99	0.00
50	Dimethylphthalate	40.000	41.152	-2.9	99	0.00
51	2,6-Dinitrotoluene	40.000	42.035	-5.1	97	0.00
52 C	Acenaphthene	40.000	41.188	-3.0	98	0.00
53	3-Nitroaniline	40.000	38.626	3.4	96	0.00
54 P	2,4-Dinitrophenol	40.000	35.355	11.6	90	0.00
55	Dibenzofuran	40.000	41.806	-4.5	101	0.00
56 P	4-Nitrophenol	40.000	39.775	0.6	94	0.00
57	2,4-Dinitrotoluene	40.000	38.416	4.0	96	0.00
58	Fluorene	40.000	42.850	-7.1	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.737	-11.8	106	0.00
60	Diethylphthalate	40.000	40.956	-2.4	97	0.00
61	4-Chlorophenyl-phenylether	40.000	43.677	-9.2	105	0.00
62	4-Nitroaniline	40.000	39.212	2.0	97	0.00
63	Azobenzene	40.000	40.407	-1.0	94	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	36.347	9.1	93	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.366	-3.4	100	0.00
67	4-Bromophenyl-phenylether	40.000	44.265	-10.7	107	0.00
68	Hexachlorobenzene	40.000	46.944	-17.4	115	0.00
69	Atrazine	40.000	43.085	-7.7	102	0.00
70 C	Pentachlorophenol	40.000	44.658	-11.6	109	0.00
71	Phenanthrene	40.000	41.896	-4.7	102	0.00
72	Anthracene	40.000	42.329	-5.8	101	0.00
73	Carbazole	40.000	42.086	-5.2	100	0.00
74	Di-n-butylphthalate	40.000	40.866	-2.2	95	0.00
75 C	Fluoranthene	40.000	44.028	-10.1	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzidine	40.000	35.043	12.4	83	0.00
78	Pyrene	40.000	39.078	2.3	106	0.00
79 S	Terphenyl-d14	80.000	84.082	-5.1	114	0.00
80	Butylbenzylphthalate	40.000	37.084	7.3	97	0.00
81	Benzo(a)anthracene	40.000	41.352	-3.4	113	0.00
82	3,3'-Dichlorobenzidine	40.000	44.177	-10.4	115	0.00
83	Chrysene	40.000	41.355	-3.4	113	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.694	3.3	98	0.00
85 c	Di-n-octyl phthalate	40.000	37.017	7.5	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	116	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	44.617	-11.5	127	0.00
88	Benzo(b)fluoranthene	40.000	42.871	-7.2	122	0.00
89	Benzo(k)fluoranthene	40.000	41.047	-2.6	117	0.00
90 C	Benzo(a)pyrene	40.000	42.115	-5.3	119	0.00
91	Dibenzo(a,h)anthracene	40.000	45.100	-12.8	128	0.00
92	Benzo(g,h,i)perylene	40.000	44.102	-10.3	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091922\
Data File : BP011738.D
Acq On : 19 Sep 2022 16:51
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 20 04:44:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 20 04:42:52 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0