

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091922\
 Data File : BP011736.D
 Acq On : 19 Sep 2022 15:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo

09/20/2022
 Supervised By :Jagrut
 Upadhyay

09/22/2022

Quant Time: Sep 20 04:40:28 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 14 19:12:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	583416	20.000	ng	0.00
21) Naphthalene-d8	10.746	136	2283914	20.000	ng	#-0.01
39) Acenaphthene-d10	14.575	164	1328545	20.000	ng	-0.01
64) Phenanthrene-d10	17.333	188	2672094	20.000	ng	0.00
76) Chrysene-d12	21.410	240	2649453	20.000	ng	0.00
86) Perylene-d12	23.863	264	2699717	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	2556735	79.768	ng	0.00
7) Phenol-d6	7.099	99	3357286	78.577	ng	0.00
23) Nitrobenzene-d5	9.104	82	3039084	78.917	ng	0.00
42) 2,4,6-Tribromophenol	16.069	330	1077224	103.775	ng	0.00
45) 2-Fluorobiphenyl	13.204	172	6936782	81.162	ng	0.00
79) Terphenyl-d14	19.880	244	9606387	78.846	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.375	88	539309	38.207	ng	# 82
3) Pyridine	3.787	79	1558812	38.306	ng	# 78
4) n-Nitrosodimethylamine	3.693	42	619634	39.719	ng	# 49
6) Aniline	7.263	93	2010881	38.207	ng	89
8) 2-Chlorophenol	7.499	128	1490601	39.220	ng	91
9) Benzaldehyde	7.075	77	1001488	36.619	ng	89
10) Phenol	7.122	94	1858695	38.678	ng	82
11) bis(2-Chloroethyl)ether	7.357	93	1442697	37.686	ng	92
12) 1,3-Dichlorobenzene	7.828	146	1660359	39.394	ng	# 92
13) 1,4-Dichlorobenzene	7.975	146	1685162	39.315	ng	96
14) 1,2-Dichlorobenzene	8.293	146	1604077	39.257	ng	96
15) Benzyl Alcohol	8.181	79	1175846	38.550	ng	# 90
16) 2,2'-oxybis(1-Chloropr...	8.469	45	1955172	38.236	ng	94
17) 2-Methylphenol	8.381	107	1209548	38.765	ng	# 91
18) Hexachloroethane	9.022	117	586038	39.529	ng	96
19) n-Nitroso-di-n-propyla...	8.752	70	1038251	37.531	ng	# 85
20) 3+4-Methylphenols	8.710	107	1673024	39.042	ng	92
22) Acetophenone	8.763	105	2103007	39.279	ng	# 88
24) Nitrobenzene	9.146	77	1568046	39.135	ng	# 89
25) Isophorone	9.675	82	2652485	38.936	ng	94
26) 2-Nitrophenol	9.857	139	655904	36.551	ng	# 81
27) 2,4-Dimethylphenol	9.916	122	1151732	40.438	ng	89
28) bis(2-Chloroethoxy)met...	10.157	93	1675923	38.472	ng	98
29) 2,4-Dichlorophenol	10.393	162	1325311	42.006	ng	95
30) 1,2,4-Trichlorobenzene	10.610	180	1452983	41.598	ng	98
31) Naphthalene	10.798	128	4665362	39.363	ng	99
32) Benzoic acid	10.051	122	777682	34.644	ng	86
33) 4-Chloroaniline	10.910	127	1862038	40.043	ng	# 91
34) Hexachlorobutadiene	11.087	225	841619	43.740	ng	98
35) Caprolactam	11.704	113	391183	38.265	ng	# 75
36) 4-Chloro-3-methylphenol	12.028	107	1340635	39.511	ng	90
37) 2-Methylnaphthalene	12.404	142	3167980	39.940	ng	96
38) 1-Methylnaphthalene	12.622	142	3069649	39.584	ng	100
40) 1,2,4,5-Tetrachloroben...	12.769	216	1471637	41.970	ng	99
41) Hexachlorocyclopentadiene	12.751	237	762550	43.569	ng	99
43) 2,4,6-Trichlorophenol	13.010	196	1010275	42.358	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091922\
 Data File : BP011736.D
 Acq On : 19 Sep 2022 15:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Manual Integrations
 APPROVED

Reviewed By : Christian
 Giraldo

09/20/2022
 Supervised By : Jagrut
 Upadhyay

Quant Time: Sep 20 04:40:28 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 14 19:12:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.081	196	1125503	42.205	ng	97
46) 1,1'-Biphenyl	13.416	154	3798226	39.299	ng	100
47) 2-Chloronaphthalene	13.451	162	3018094	39.690	ng	99
48) 2-Nitroaniline	13.657	65	848856	39.805	ng	# 81
49) Acenaphthylene	14.298	152	4723601	40.145	ng	99
50) Dimethylphthalate	14.040	163	3674444	39.645	ng	100
51) 2,6-Dinitrotoluene	14.151	165	774234	41.194	ng	# 85
52) Acenaphthene	14.639	154	3089871m	39.478	ng	
53) 3-Nitroaniline	14.481	138	891162	37.942	ng	88
54) 2,4-Dinitrophenol	14.687	184	346337	35.855	ng	# 83
55) Dibenzofuran	14.975	168	4505925	39.979	ng	96
56) 4-Nitrophenol	14.781	139	714210	39.109	ng	# 71
57) 2,4-Dinitrotoluene	14.939	165	1028211	38.307	ng	# 81
58) Fluorene	15.628	166	3746174	40.979	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.198	232	957444	43.703	ng	99
60) Diethylphthalate	15.398	149	3642075	39.470	ng	99
61) 4-Chlorophenyl-phenyle...	15.622	204	1772678	41.949	ng	94
62) 4-Nitroaniline	15.645	138	928774	38.680	ng	# 78
63) Azobenzene	15.916	77	3443257	38.831	ng	88
65) 4,6-Dinitro-2-methylph...	15.698	198	496140	36.093	ng	89
66) n-Nitrosodiphenylamine	15.834	169	3110804	39.127	ng	98
67) 4-Bromophenyl-phenylether	16.516	248	1067912	42.692	ng	97
68) Hexachlorobenzene	16.628	284	1206938	44.802	ng	97
69) Atrazine	16.792	200	968698	39.915	ng	99
70) Pentachlorophenol	16.975	266	747796	43.755	ng	99
71) Phenanthrene	17.375	178	5721171	39.349	ng	98
72) Anthracene	17.463	178	5485581	39.891	ng	97
73) Carbazole	17.733	167	5170260	39.854	ng	99
74) Di-n-butylphthalate	18.280	149	6102200	39.025	ng	99
75) Fluoranthene	19.333	202	6501352	41.460	ng	95
77) Benzidine	19.516	184	2346728	37.245	ng	99
78) Pyrene	19.686	202	6751676	37.783	ng	97
80) Butylbenzylphthalate	20.557	149	2746891	36.683	ng	90
81) Benzo(a)anthracene	21.398	228	6744111	39.492	ng	96
82) 3,3'-Dichlorobenzidine	21.327	252	2207946	42.570	ng	98
83) Chrysene	21.451	228	6394831	39.193	ng	97
84) Bis(2-ethylhexyl)phtha...	21.316	149	4030361	37.414	ng	99
85) Di-n-octyl phthalate	22.263	149	6691047	36.196	ng	# 92
87) Indeno(1,2,3-cd)pyrene	26.433	276	8388273	43.128	ng	# 95
88) Benzo(b)fluoranthene	23.115	252	6662912	39.904	ng	98
89) Benzo(k)fluoranthene	23.163	252	6709262	39.874	ng	# 97
90) Benzo(a)pyrene	23.757	252	5587053	40.505	ng	97
91) Dibenzo(a,h)anthracene	26.456	278	6986570	43.260	ng	# 97
92) Benzo(g,h,i)perylene	27.227	276	6714045	42.642	ng	# 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

09/20/2022
Supervised By :Jagrut
Upadhyay

09/22/2022






#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.940 min Scan# 843
 Delta R.T. -0.006 min
 Lab File: BP011736.D
 Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC



Tgt Ion: 152 Resp: 583416
 Ion Ratio Lower Upper
 152 100
 150 153.2 123.4 185.2
 115 56.5 51.3 76.9




Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo


09/20/2022

Supervised By :Jagrut
 Upadhyay



09/22/2022




#2
 1,4-Dioxane
 Concen: 38.207 ng
 RT: 3.375 min Scan# 66
 Delta R.T. -0.000 min
 Lab File: BP011736.D
 Acq: 19 Sep 2022 15:31




Tgt Ion: 88 Resp: 539309
 Ion Ratio Lower Upper
 88 100
 58 77.0 74.5 111.7
 43 26.8 31.5 47.3





#3
Pyridine
Concen: 38.306 ng
RT: 3.787 min Scan# 136
Delta R.T. -0.000 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion: 79 Resp: 1558812
Ion Ratio Lower Upper
79 100
52 70.1 73.5 110.3
51 33.9 38.6 57.8

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#4
n-Nitrosodimethylamine
Concen: 39.719 ng
RT: 3.693 min Scan# 120
Delta R.T. -0.000 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion: 42 Resp: 619634
Ion Ratio Lower Upper
42 100
74 145.0 74.4 111.6
44 8.2 5.7 8.5





#5
2-Fluorophenol
Concen: 79.768 ng
RT: 5.493 min Scan# 41
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion: 112 Resp: 2556735
Ion Ratio Lower Upper
112 100
64 57.3 56.9 85.3
63 28.6 31.2 46.8

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#6
Aniline
Concen: 38.207 ng
RT: 7.263 min Scan# 727
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion: 93 Resp: 2010881
Ion Ratio Lower Upper
93 100
66 37.6 36.7 55.1
65 19.7 18.2 27.4





#7
Phenol-d6
Concen: 78.577 ng
RT: 7.099 min Scan# 699
Delta R.T. -0.000 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion: 99 Resp: 3357286
Ion Ratio Lower Upper
99 100
42 15.9 20.3 30.5
71 29.6 33.9 50.9

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#8
2-Chlorophenol
Concen: 39.220 ng
RT: 7.499 min Scan# 767
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:128 Resp: 1490601
Ion Ratio Lower Upper
128 100
130 32.8 11.8 51.8
64 45.7 35.3 75.3





#9
Benzaldehyde
Concen: 36.619 ng
RT: 7.075 min Scan# 696
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion: 77 Resp: 1001488
Ion Ratio Lower Upper
77 100
105 100.7 71.4 111.4
106 98.0 66.9 106.9

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#10
Phenol
Concen: 38.678 ng
RT: 7.122 min Scan# 703
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion: 94 Resp: 1858695
Ion Ratio Lower Upper
94 100
65 27.0 14.0 54.0
66 34.0 29.2 69.2





#11
bis(2-Chloroethyl)ether
Concen: 37.686 ng
RT: 7.357 min Scan# 744
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion: 93 Resp: 1442697
Ion Ratio Lower Upper
93 100
63 78.0 68.2 108.2
95 32.1 12.2 52.2

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#12
1,3-Dichlorobenzene
Concen: 39.394 ng
RT: 7.828 min Scan# 823
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:146 Resp: 1660359
Ion Ratio Lower Upper
146 100
148 64.3 53.0 79.4
75 27.9 31.3 46.9#





#13

1,4-Dichlorobenzene

Concen: 39.315 ng

RT: 7.975 min Scan# 84

Delta R.T. -0.006 min

Lab File: BP011736.D

Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:146 Resp: 1685162

Ion Ratio Lower Upper

146 100

148 65.7 51.6 77.4

111 39.3 35.2 52.8

Manual Integrations





APPROVED

Reviewed By :Christian
Giraldo

09/20/2022

Supervised By :Jagrut
Upadhyay

09/22/2022



#14

1,2-Dichlorobenzene

Concen: 39.257 ng

RT: 8.293 min Scan# 902

Delta R.T. -0.006 min

Lab File: BP011736.D

Acq: 19 Sep 2022 15:31




Tgt Ion:146 Resp: 1604077


Ion Ratio Lower Upper

146 100

148 63.9 52.2 78.2


111 41.3 36.6 55.0





#15
Benzyl Alcohol
Concen: 38.550 ng
RT: 8.181 min Scan# 884
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion: 79 Resp: 1175846
Ion Ratio Lower Upper
79 100
108 82.9 53.4 80.0
77 64.3 51.5 77.3

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.236 ng
RT: 8.469 min Scan# 932
Delta R.T. -0.012 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion: 45 Resp: 1955172
Ion Ratio Lower Upper
45 100
77 14.3 0.0 31.8
79 10.8 0.0 28.8





#17
2-Methylphenol
Concen: 38.765 ng
RT: 8.381 min Scan# 917
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:107 Resp: 1209548
Ion Ratio Lower Upper
107 100
108 114.1 88.9 133.3
77 43.2 44.7 67.1
79 41.6 41.7 62.5

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#18
Hexachloroethane
Concen: 39.529 ng
RT: 9.022 min Scan# 1026
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:117 Resp: 586038
Ion Ratio Lower Upper
117 100
119 96.8 75.6 113.4
201 96.1 80.9 121.3





#19
n-Nitroso-di-n-propylamine
Concen: 37.531 ng
RT: 8.752 min Scan# 981
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC





Tgt Ion: 70 Resp: 103825
Ion Ratio Lower Upper
70 100
42 55.4 58.1 87.1
101 9.9 8.4 12.6
130 22.1 16.1 24.1

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#20
3+4-Methylphenols
Concen: 39.042 ng
RT: 8.710 min Scan# 973
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:107 Resp: 1673024
Ion Ratio Lower Upper
107 100
108 90.1 67.1 107.1
77 31.2 19.2 59.2
79 26.7 14.8 54.8






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.746 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC



Tgt Ion:136 Resp: 2283914
Ion Ratio Lower Upper
136 100
137 11.3 9.0 13.4
54 8.4 8.7 13.1
68 6.4 4.8 7.2




Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#22
Acetophenone
Concen: 39.279 ng
RT: 8.763 min Scan# 982
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:105 Resp: 2103007
Ion Ratio Lower Upper
105 100
71 4.9 3.4 5.0
51 26.2 29.5 44.3
120 23.2 17.0 25.4





#23
Nitrobenzene-d5
Concen: 78.917 ng
RT: 9.104 min Scan# 1040
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion: 82 Resp: 3039084
Ion Ratio Lower Upper
82 100
128 46.0 31.3 46.9
54 49.5 49.8 74.8

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#24
Nitrobenzene
Concen: 39.135 ng
RT: 9.146 min Scan# 1047
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion: 77 Resp: 1568046
Ion Ratio Lower Upper
77 100
123 47.1 31.1 46.7
65 13.2 12.1 18.1





#25
Isophorone
Concen: 38.936 ng
RT: 9.675 min Scan# 1138
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion: 82 Resp: 2652485
Ion Ratio Lower Upper
82 100
95 6.7 6.5 9.7
138 18.3 12.4 18.6

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#26
2-Nitrophenol
Concen: 36.551 ng
RT: 9.857 min Scan# 1168
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:139 Resp: 655904
Ion Ratio Lower Upper
139 100
109 27.9 30.2 45.4#
65 48.5 52.2 78.2#






#27
2,4-Dimethylphenol
Concen: 40.438 ng
RT: 9.916 min Scan# 1178
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC



Tgt Ion:122 Resp: 1151732
Ion Ratio Lower Upper
122 100
107 111.4 103.5 155.3
121 58.4 47.5 71.3




Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#28
bis(2-Chloroethoxy)methane
Concen: 38.472 ng
RT: 10.157 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion: 93 Resp: 1675923
Ion Ratio Lower Upper
93 100
95 32.5 25.0 37.6
123 11.7 9.0 13.6





#29
2,4-Dichlorophenol
Concen: 42.006 ng
RT: 10.393 min Scan# 1259
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:162 Resp: 132531
Ion Ratio Lower Upper
162 100
164 64.8 43.0 83.0
98 34.5 20.4 60.4





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#30
1,2,4-Trichlorobenzene
Concen: 41.598 ng
RT: 10.610 min Scan# 1296
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:180 Resp: 1452983
Ion Ratio Lower Upper
180 100
182 95.7 78.0 117.0
145 30.4 24.6 36.8





#31
Naphthalene
Concen: 39.363 ng
RT: 10.798 min Scan# 1329
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:128 Resp: 4665362
Ion Ratio Lower Upper
128 100
129 11.1 8.6 12.8
127 13.2 10.3 15.5

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#32
Benzoic acid
Concen: 34.644 ng
RT: 10.051 min Scan# 1201
Delta R.T. -0.001 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:122 Resp: 777682
Ion Ratio Lower Upper
122 100
105 119.8 112.6 152.6
77 90.9 90.2 130.2






#33
4-Chloroaniline
Concen: 40.043 ng
RT: 10.910 min Scan# 1347
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC



Tgt Ion:127 Resp: 1862038
Ion Ratio Lower Upper
127 100
129 32.2 25.4 38.2
65 28.5 30.4 45.6
92 17.5 17.9 26.9




Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#34
Hexachlorobutadiene
Concen: 43.740 ng
RT: 11.087 min Scan# 1377
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:225 Resp: 841619
Ion Ratio Lower Upper
225 100
223 62.3 52.2 78.4
227 64.1 50.8 76.2





#35
Caprolactam
Concen: 38.265 ng
RT: 11.704 min Scan# 1483
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:113 Resp: 391183
Ion Ratio Lower Upper
113 100
55 176.4 198.6 238.6
56 130.6 142.7 182.7

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#36
4-Chloro-3-methylphenol
Concen: 39.511 ng
RT: 12.028 min Scan# 1537
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:107 Resp: 1340635
Ion Ratio Lower Upper
107 100
144 27.1 19.8 29.6
142 83.9 59.0 88.4





#37
2-Methylnaphthalene
Concen: 39.940 ng
RT: 12.404 min Scan# 1602
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:142 Resp: 3167980
Ion Ratio Lower Upper
142 100
141 88.0 71.3 106.9
115 32.7 31.4 47.2





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#38
1-Methylnaphthalene
Concen: 39.584 ng
RT: 12.622 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:142 Resp: 3069649
Ion Ratio Lower Upper
142 100
141 91.1 72.6 108.8
116 3.7 3.3 4.9





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.575 min Scan# 1972

Delta R.T. -0.012 min

Lab File: BP011736.D

Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:164 Resp: 132854

Ion	Ratio	Lower	Upper
164	100		
162	100.6	81.9	122.9
160	45.1	34.6	51.8





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022

Supervised By :Jagrut
Upadhyay

09/22/2022



#40

1,2,4,5-Tetrachlorobenzene

Concen: 41.970 ng

RT: 12.769 min Scan# 1663




Delta R.T. -0.006 min


Lab File: BP011736.D

Acq: 19 Sep 2022 15:31

Tgt Ion:216 Resp: 1471637

Ion	Ratio	Lower	Upper
216	100		
214	78.8	61.9	92.9
179	21.1	17.0	25.6
108	19.3	15.5	23.3





#41

Hexachlorocyclopentadiene

Concen: 43.569 ng

RT: 12.751 min Scan# 1660

Delta R.T. -0.006 min

Lab File: BP011736.D

Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:	237	Resp:	762550
Ion Ratio	Lower	Upper	
237	100		
235	64.1	43.3	83.3
272	13.4	0.0	33.0





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022

Supervised By :Jagrut
Upadhyay

09/22/2022



#42

2,4,6-Tribromophenol

Concen: 103.775 ng




RT: 16.069 min Scan# 2224


Delta R.T. -0.006 min

Lab File: BP011736.D

Acq: 19 Sep 2022 15:31


Tgt Ion:	330	Resp:	1077224
Ion Ratio	Lower	Upper	
330	100		
332	96.0	78.1	117.1
141	36.5	31.4	47.2





#43
2,4,6-Trichlorophenol
Concen: 42.358 ng
RT: 13.010 min Scan# 1704
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:196 Resp: 1010275
Ion Ratio Lower Upper
196 100
198 95.1 77.7 116.5
200 30.9 25.5 38.3

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#44
2,4,5-Trichlorophenol
Concen: 42.205 ng
RT: 13.081 min Scan# 1716
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:196 Resp: 1125503
Ion Ratio Lower Upper
196 100
198 95.8 75.7 113.5
97 46.5 42.4 63.6
132 27.0 21.5 32.3





#45
2-Fluorobiphenyl
Concen: 81.162 ng
RT: 13.204 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:172 Resp: 6936782
Ion Ratio Lower Upper
172 100
171 35.6 27.9 41.9
170 24.3 18.2 27.2

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#46
1,1'-Biphenyl
Concen: 39.299 ng
RT: 13.416 min Scan# 1773
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:154 Resp: 3798226
Ion Ratio Lower Upper
154 100
153 40.1 20.1 60.1
76 12.6 0.0 33.2





#47
2-Chloronaphthalene
Concen: 39.690 ng
RT: 13.451 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:162 Resp: 3018094
Ion Ratio Lower Upper
162 100
127 39.0 32.4 48.6
164 32.9 26.3 39.5





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#48
2-Nitroaniline
Concen: 39.805 ng
RT: 13.657 min Scan# 1814
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion: 65 Resp: 848856
Ion Ratio Lower Upper
65 100
92 70.4 50.0 75.0
138 113.0 71.2 106.8#





#49
Acenaphthylene
Concen: 40.145 ng
RT: 14.298 min Scan# 1923
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:152 Resp: 4723601
Ion Ratio Lower Upper
152 100
151 20.1 15.7 23.5
153 13.4 10.4 15.6

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#50
Dimethylphthalate
Concen: 39.645 ng
RT: 14.040 min Scan# 1879
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:163 Resp: 3674444
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.2 8.2 12.4





#51
2,6-Dinitrotoluene
Concen: 41.194 ng
RT: 14.151 min Scan# 1899
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:165 Resp: 774234
Ion Ratio Lower Upper
165 100
63 54.8 57.3 85.9
89 55.9 50.2 75.2

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#52
Acenaphthene
Concen: 39.478 ng m
RT: 14.639 min Scan# 1981
Delta R.T. -0.012 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:154 Resp: 3089871
Ion Ratio Lower Upper
154 100
153 112.4 87.4 131.2
152 54.7 41.7 62.5





#53
3-Nitroaniline
Concen: 37.942 ng
RT: 14.481 min Scan# 1955
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:138 Resp: 891162
Ion Ratio Lower Upper
138 100
108 10.2 9.5 14.3
92 110.4 99.9 149.9





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#54
2,4-Dinitrophenol
Concen: 35.855 ng
RT: 14.687 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:184 Resp: 346337
Ion Ratio Lower Upper
184 100
63 62.1 65.4 98.2#
154 59.7 55.4 83.2





#55
Dibenzofuran
Concen: 39.979 ng
RT: 14.975 min Scan# 2038
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:168 Resp: 4505925
Ion Ratio Lower Upper
168 100
139 36.9 31.5 47.3
169 13.7 10.3 15.5

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#56
4-Nitrophenol
Concen: 39.109 ng
RT: 14.781 min Scan# 2005
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:139 Resp: 714210
Ion Ratio Lower Upper
139 100
109 65.9 70.6 110.6#
65 90.8 106.6 146.6#





#57
2,4-Dinitrotoluene
Concen: 38.307 ng
RT: 14.939 min Scan# 2032
Delta R.T. 0.000 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:165 Resp: 1028211
Ion Ratio Lower Upper
165 100
63 40.3 46.2 69.2
89 67.4 65.8 98.8
182 2.6 2.3 3.5





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#58
Fluorene
Concen: 40.979 ng
RT: 15.628 min Scan# 2149
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:166 Resp: 3746174
Ion Ratio Lower Upper
166 100
165 97.1 77.0 115.4
167 13.4 10.9 16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 43.703 ng
RT: 15.198 min Scan# 2076
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:232 Resp: 95744
Ion Ratio Lower Upper
232 100
131 46.2 37.0 55.6
130 2.3 2.2 3.2
166 30.8 25.4 38.2

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#60
Diethylphthalate
Concen: 39.470 ng
RT: 15.398 min Scan# 2110
Delta R.T. -0.012 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:149 Resp: 3642075
Ion Ratio Lower Upper
149 100
177 22.0 17.8 26.8
150 12.7 9.9 14.9






#61
4-Chlorophenyl-phenylether
Concen: 41.949 ng
RT: 15.622 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC



Tgt Ion: 204 Resp: 1772678
Ion Ratio Lower Upper
204 100
206 33.1 26.3 39.5
141 56.5 51.1 76.7




Manual Integrations
APPROVED

Reviewed By : Christian
Giraldo

09/20/2022




Supervised By : Jagrut
Upadhyay


09/22/2022



#62
4-Nitroaniline
Concen: 38.680 ng
RT: 15.645 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion: 138 Resp: 928774
Ion Ratio Lower Upper
138 100
92 49.6 38.0 78.0
108 70.3 77.6 117.6#





#63
Azobenzene
Concen: 38.831 ng
RT: 15.916 min Scan# 2199
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion: 77 Resp: 344325
Ion Ratio Lower Upper
77 100
182 27.5 3.0 43.0
105 22.4 0.0 38.6
51 27.7 16.7 56.7

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.333 min Scan# 2439
Delta R.T. 0.000 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:188 Resp: 2672094
Ion Ratio Lower Upper
188 100
94 9.3 6.6 10.0
80 9.8 8.0 12.0





#65
4,6-Dinitro-2-methylphenol
Concen: 36.093 ng
RT: 15.698 min Scan# 2161
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC





Tgt Ion:198 Resp: 496140
Ion Ratio Lower Upper
198 100
51 46.3 41.2 81.2
105 45.7 25.3 65.3

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#66
n-Nitrosodiphenylamine
Concen: 39.127 ng
RT: 15.834 min Scan# 2184
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:169 Resp: 3110804
Ion Ratio Lower Upper
169 100
168 66.2 51.8 77.8
167 37.1 28.3 42.5





#67
4-Bromophenyl-phenylether
Concen: 42.692 ng
RT: 16.516 min Scan# 2319
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:248 Resp: 1067912
Ion Ratio Lower Upper
248 100
250 97.3 77.7 116.5
141 72.6 62.0 93.0

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#68
Hexachlorobenzene
Concen: 44.802 ng
RT: 16.628 min Scan# 2319
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:284 Resp: 1206938
Ion Ratio Lower Upper
284 100
142 37.9 30.8 46.2
249 30.8 26.8 40.2





#69
Atrazine
Concen: 39.915 ng
RT: 16.792 min Scan# 2348
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:200 Resp: 968698
Ion Ratio Lower Upper
200 100
173 26.0 5.3 45.3
215 49.0 29.4 69.4

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#70
Pentachlorophenol
Concen: 43.755 ng
RT: 16.975 min Scan# 2378
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:266 Resp: 747796
Ion Ratio Lower Upper
266 100
268 62.4 50.5 75.7
264 62.1 50.4 75.6





#71
Phenanthrene
Concen: 39.349 ng
RT: 17.375 min Scan# 2446
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:178 Resp: 572117
Ion Ratio Lower Upper
178 100
176 20.0 14.7 22.1
179 15.7 12.2 18.2





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#72
Anthracene
Concen: 39.891 ng
RT: 17.463 min Scan# 2461
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31


Tgt Ion:178 Resp: 5485581
Ion Ratio Lower Upper
178 100
176 19.5 14.5 21.7
179 16.1 12.2 18.4





#73
 Carbazole
 Concen: 39.854 ng
 RT: 17.733 min Scan# 2508
 Delta R.T. -0.006 min
 Lab File: BP011736.D
 Acq: 19 Sep 2022 15:31

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC






Tgt Ion:167 Resp: 5170260
 Ion Ratio Lower Upper
 167 100
 166 21.8 16.9 25.3
 139 13.4 10.6 15.8

Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo




09/20/2022
 Supervised By :Jagrut
 Upadhyay


09/22/2022



#74
 Di-n-butylphthalate
 Concen: 39.025 ng
 RT: 18.280 min Scan# 2600
 Delta R.T. -0.012 min
 Lab File: BP011736.D
 Acq: 19 Sep 2022 15:31

Tgt Ion:149 Resp: 6102200
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.5 11.3
 104 5.0 4.8 7.2





#75
Fluoranthene
Concen: 41.460 ng
RT: 19.333 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:202 Resp: 6501352
Ion Ratio Lower Upper
202 100
101 12.5 0.0 29.6
203 18.2 0.0 36.9





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.410 min Scan# 3132
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:240 Resp: 2649453
Ion Ratio Lower Upper
240 100
120 12.1 8.6 12.8
236 26.1 19.8 29.6






#77
Benzidine
Concen: 37.245 ng
RT: 19.516 min Scan# 2810
Delta R.T. -0.000 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC



Tgt Ion:184 Resp: 2346728
Ion Ratio Lower Upper
184 100
185 14.4 10.7 16.1
183 11.4 9.0 13.6




Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo


09/20/2022

Supervised By :Jagrut
Upadhyay



09/22/2022




#78
Pyrene
Concen: 37.783 ng
RT: 19.686 min Scan# 2839
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31



Tgt Ion:202 Resp: 6751676
Ion Ratio Lower Upper
202 100
200 22.2 16.3 24.5
203 18.3 14.2 21.4





#79
 Terphenyl-d14
 Concen: 78.846 ng
 RT: 19.880 min Scan# 2873
 Delta R.T. -0.006 min
 Lab File: BP011736.D
 Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:244 Resp: 960638
 Ion Ratio Lower Upper
 244 100
 212 7.7 5.8 8.8
 122 12.5 7.6 11.4





Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo

09/20/2022




Supervised By :Jagrut
 Upadhyay


09/22/2022



#80
 Butylbenzylphthalate
 Concen: 36.683 ng
 RT: 20.557 min Scan# 2987
 Delta R.T. -0.006 min
 Lab File: BP011736.D
 Acq: 19 Sep 2022 15:31


Tgt Ion:149 Resp: 2746891
 Ion Ratio Lower Upper
 149 100
 91 67.6 62.5 93.7
 206 20.8 17.0 25.4





#81
Benzo(a)anthracene
Concen: 39.492 ng
RT: 21.398 min Scan# 3119
Delta R.T. -0.000 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:228 Resp: 6744111
Ion Ratio Lower Upper
228 100
226 28.4 20.8 31.2
229 20.0 15.2 22.8

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#82
3,3'-Dichlorobenzidine
Concen: 42.570 ng
RT: 21.327 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:252 Resp: 2207946
Ion Ratio Lower Upper
252 100
254 65.0 51.1 76.7
126 11.4 8.4 12.6





#83

Chrysene

Concen: 39.193 ng

RT: 21.451 min Scan# 3140

Delta R.T. -0.006 min

Lab File: BP011736.D

Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:228 Resp: 639483

Ion Ratio Lower Upper

228 100

226 31.2 23.4 35.2

229 20.5 15.7 23.5

Manual Integrations





APPROVED

Reviewed By :Christian
Giraldo

09/20/2022

Supervised By :Jagrut
Upadhyay

09/22/2022



#84

Bis(2-ethylhexyl)phthalate

Concen: 37.414 ng

RT: 21.316 min Scan# 3116

Delta R.T. -0.006 min

Lab File: BP011736.D

Acq: 19 Sep 2022 15:31




Tgt Ion:149 Resp: 4030361


Ion Ratio Lower Upper

149 100

167 27.8 21.8 32.8


279 3.9 3.7 5.5





#85
Di-n-octyl phthalate
Concen: 36.196 ng
RT: 22.263 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC






Tgt Ion:149 Resp: 669104
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 8.8 10.0 15.0

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




09/20/2022
Supervised By :Jagrut
Upadhyay


09/22/2022



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.863 min Scan# 3549
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:264 Resp: 2699717
Ion Ratio Lower Upper
264 100
260 24.4 19.3 28.9
265 22.0 18.0 27.0





#87
Indeno(1,2,3-cd)pyrene
Concen: 43.128 ng
RT: 26.433 min Scan# 3987
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:276 Resp: 8388273
Ion Ratio Lower Upper
276 100
138 26.7 17.5 26.3
277 25.4 20.2 30.4





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#88
Benzo(b)fluoranthene
Concen: 39.904 ng
RT: 23.115 min Scan# 3422
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:252 Resp: 6662912
Ion Ratio Lower Upper
252 100
253 22.6 17.8 26.6
125 10.5 7.0 10.6





#89
Benzo(k)fluoranthene
Concen: 39.874 ng
RT: 23.163 min Scan# 3431
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:252 Resp: 6709262
Ion Ratio Lower Upper
252 100
253 22.5 17.2 25.8
125 11.0 7.0 10.4





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022




Supervised By :Jagrut
Upadhyay


09/22/2022



#90
Benzo(a)pyrene
Concen: 40.505 ng
RT: 23.757 min Scan# 3531
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:252 Resp: 5587053
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.6
125 11.8 7.9 11.9





#91
Dibenzo(a,h)anthracene
Concen: 43.260 ng
RT: 26.456 min Scan# 3991
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC

Tgt Ion:278 Resp: 6986570
Ion Ratio Lower Upper
278 100
139 17.5 11.4 17.0
279 24.3 19.2 28.8





Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

09/20/2022

Supervised By :Jagrut
Upadhyay

09/22/2022



#92
Benzo(g,h,i)perylene
Concen: 42.642 ng
RT: 27.227 min Scan# 4121
Delta R.T. -0.006 min
Lab File: BP011736.D
Acq: 19 Sep 2022 15:31

Tgt Ion:276 Resp: 6714045
Ion Ratio Lower Upper
276 100
277 24.2 18.7 28.1
138 22.9 15.2 22.8

