

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092221\
 Data File : BP007112.D
 Acq On : 22 Sep 2021 23:09
 Operator : CG/JU
 Sample : M3797-07 2X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 26240

Manual Integrations
 APPROVED

mohammad
 9/23/2021 1:21:34 PM

Quant Time: Sep 23 01:08:26 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 16:50:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	329346	20.000	ng	0.00
21) Naphthalene-d8	10.699	136	1195483	20.000	ng	# 0.00
39) Acenaphthene-d10	14.534	164	716533	20.000	ng	0.00
64) Phenanthrene-d10	17.287	188	1415222	20.000	ng	# 0.00
76) Chrysene-d12	21.369	240	1532167	20.000	ng	0.00
86) Perylene-d12	23.798	264	1700406	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.470	112	770789	39.176	ng	0.00
7) Phenol-d6	7.064	99	952353	35.416	ng	0.00
23) Nitrobenzene-d5	9.058	82	571416	27.836	ng	-0.01
42) 2,4,6-Tribromophenol	16.022	330	438667	47.222	ng	0.00
45) 2-Fluorobiphenyl	13.157	172	1355659	27.105	ng	0.00
79) Terphenyl-d14	19.839	244	2114036	26.440	ng	0.00
Target Compounds						Qvalue
52) Acenaphthene	14.593	154	84839	2.202	ng	# 87
66) n-Nitrosodiphenylamine	15.798	169	255341	6.064	ng	# 57
71) Phenanthrene	17.328	178	226981	2.991	ng	# 40
74) Di-n-butylphthalate	18.239	149	169871	2.107	ng	# 54
75) Fluoranthene	19.298	202	942188	10.777	ng	90
78) Pyrene	19.645	202	1054604	10.491	ng	# 93
81) Benzo(a)anthracene	21.351	228	307524	3.092	ng	# 86
83) Chrysene	21.404	228	437471m	4.603	ng	
84) Bis(2-ethylhexyl)phtha...	21.280	149	2836026	49.001	ng	# 92
85) Di-n-octyl phthalate	22.210	149	212636	2.158	ng	# 83
87) Indeno(1,2,3-cd)pyrene	26.327	276	263432	2.156	ng	96
88) Benzo(b)fluoranthene	23.051	252	596285	5.868	ng	# 17
90) Benzo(a)pyrene	23.686	252	271482	3.145	ng	# 1
92) Benzo(g,h,i)perylene	27.121	276	321739	3.219	ng	# 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092221\
Data File : BP007112.D
Acq On : 22 Sep 2021 23:09
Operator : CG/JU
Sample : M3797-07 2X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampled :
26240

Manual Integrations
APPROVED

mohammad
9/23/2021 1:21:34 PM

Quant Time: Sep 23 01:08:26 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 16:50:24 2021
Response via : Initial Calibration

