

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP092821\  
 Data File : BP007241.D  
 Acq On : 28 Sep 2021 21:35  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040EC

Manual Integrations  
 APPROVED

Quant Time: Sep 29 05:27:28 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP092121.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Sep 23 11:17:25 2021  
 Response via : Initial Calibration

| Compound                  | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|---------------------------|--------|------|----------|--------|-------|----------|
| -----                     |        |      |          |        |       |          |
| Internal Standards        |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4 | 7.875  | 152  | 220571   | 20.000 | ng    | -0.02    |
| 21) Naphthalene-d8        | 10.681 | 136  | 837672   | 20.000 | ng    | #-0.01   |
| 39) Acenaphthene-d10      | 14.510 | 164  | 538466   | 20.000 | ng    | -0.02    |
| 64) Phenanthrene-d10      | 17.263 | 188  | 1093483  | 20.000 | ng    | -0.01    |
| 76) Chrysene-d12          | 21.351 | 240  | 1038385  | 20.000 | ng    | 0.00     |
| 86) Perylene-d12          | 23.762 | 264  | 1105567  | 20.000 | ng    | 0.00     |

mohammad

|                             |        |     |         |        |    |       |
|-----------------------------|--------|-----|---------|--------|----|-------|
| System Monitoring Compounds |        |     |         |        |    |       |
| 5) 2-Fluorophenol           | 5.452  | 112 | 973649  | 73.892 | ng | -0.01 |
| 7) Phenol-d6                | 7.052  | 99  | 1331191 | 73.917 | ng | -0.01 |
| 23) Nitrobenzene-d5         | 9.040  | 82  | 1135664 | 78.953 | ng | -0.02 |
| 42) 2,4,6-Tribromophenol    | 16.004 | 330 | 617611  | 88.471 | ng | -0.01 |
| 45) 2-Fluorobiphenyl        | 13.139 | 172 | 2875889 | 76.515 | ng | -0.01 |
| 79) Terphenyl-d14           | 19.821 | 244 | 4304139 | 79.430 | ng | -0.01 |

9/29/2021 12:55:38 PM

|                               |        |     |         |        |        |      |
|-------------------------------|--------|-----|---------|--------|--------|------|
| Target Compounds              |        |     |         |        | Qvalue |      |
| 2) 1,4-Dioxane                | 3.340  | 88  | 182260  | 34.204 | ng     | # 87 |
| 3) Pyridine                   | 3.752  | 79  | 510345  | 35.000 | ng     | # 87 |
| 4) n-Nitrosodimethylamine     | 3.658  | 42  | 184627  | 36.934 | ng     | # 77 |
| 6) Aniline                    | 7.204  | 93  | 727425  | 36.644 | ng     | 99   |
| 8) 2-Chlorophenol             | 7.446  | 128 | 546382  | 38.153 | ng     | 100  |
| 9) Benzaldehyde               | 7.016  | 77  | 381939m | 37.649 | ng     |      |
| 10) Phenol                    | 7.081  | 94  | 638993  | 36.802 | ng     | 92   |
| 11) bis(2-Chloroethyl)ether   | 7.304  | 93  | 491155  | 35.836 | ng     | 95   |
| 12) 1,3-Dichlorobenzene       | 7.763  | 146 | 615608  | 38.110 | ng     | 97   |
| 13) 1,4-Dichlorobenzene       | 7.910  | 146 | 629745  | 38.241 | ng     | 99   |
| 14) 1,2-Dichlorobenzene       | 8.228  | 146 | 610636  | 38.265 | ng     | 97   |
| 15) Benzyl Alcohol            | 8.122  | 79  | 447134  | 38.764 | ng     | 95   |
| 16) 2,2'-oxybis(1-Chloropr... | 8.410  | 45  | 567208  | 36.143 | ng     | 90   |
| 17) 2-Methylphenol            | 8.328  | 107 | 444138  | 37.944 | ng     | 94   |
| 18) Hexachloroethane          | 8.951  | 117 | 205347  | 37.140 | ng     | 96   |
| 19) n-Nitroso-di-n-propyla... | 8.693  | 70  | 349456  | 35.677 | ng     | 96   |
| 20) 3+4-Methylphenols         | 8.657  | 107 | 606500  | 37.473 | ng     | 93   |
| 22) Acetophenone              | 8.704  | 105 | 768379  | 38.069 | ng     | 99   |
| 24) Nitrobenzene              | 9.081  | 77  | 535091  | 39.018 | ng     | 96   |
| 25) Isophorone                | 9.616  | 82  | 947600  | 37.780 | ng     | 99   |
| 26) 2-Nitrophenol             | 9.793  | 139 | 300844  | 42.070 | ng     | 91   |
| 27) 2,4-Dimethylphenol        | 9.857  | 122 | 430753  | 38.257 | ng     | 99   |
| 28) bis(2-Chloroethoxy)met... | 10.093 | 93  | 636617  | 35.976 | ng     | 99   |
| 29) 2,4-Dichlorophenol        | 10.334 | 162 | 529709  | 40.330 | ng     | 99   |
| 30) 1,2,4-Trichlorobenzene    | 10.545 | 180 | 604932  | 40.429 | ng     | 97   |
| 31) Naphthalene               | 10.728 | 128 | 1644715 | 38.471 | ng     | 100  |
| 32) Benzoic acid              | 10.028 | 122 | 363992  | 42.061 | ng     | 95   |
| 33) 4-Chloroaniline           | 10.840 | 127 | 681956  | 38.611 | ng     | 100  |
| 34) Hexachlorobutadiene       | 11.010 | 225 | 383526  | 42.809 | ng     | 99   |
| 35) Caprolactam               | 11.651 | 113 | 162369  | 40.183 | ng     | # 81 |
| 36) 4-Chloro-3-methylphenol   | 11.975 | 107 | 523944  | 39.245 | ng     | 98   |
| 37) 2-Methylnaphthalene       | 12.339 | 142 | 1155490 | 38.600 | ng     | 96   |
| 38) 1-Methylnaphthalene       | 12.557 | 142 | 1124929 | 38.460 | ng     | 98   |
| 40) 1,2,4,5-Tetrachloroben... | 12.710 | 216 | 670548  | 40.665 | ng     | # 97 |
| 41) Hexachlorocyclopentadiene | 12.681 | 237 | 139887  | 15.306 | ng     | 98   |
| 43) 2,4,6-Trichlorophenol     | 12.951 | 196 | 461776  | 41.041 | ng     | 96   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP092821\  
 Data File : BP007241.D  
 Acq On : 28 Sep 2021 21:35  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040EC

Manual Integrations  
 APPROVED

Quant Time: Sep 29 05:27:28 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP092121.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Sep 23 11:17:25 2021  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol     | 13.028 | 196  | 492567   | 40.653 | ng    | 94       |
| 46) 1,1'-Biphenyl             | 13.345 | 154  | 1527271  | 38.132 | ng    | 99       |
| 47) 2-Chloronaphthalene       | 13.392 | 162  | 1199718  | 38.669 | ng    | 97       |
| 48) 2-Nitroaniline            | 13.598 | 65   | 312407   | 41.300 | ng    | 91       |
| 49) Acenaphthylene            | 14.233 | 152  | 1858518  | 38.893 | ng    | 100      |
| 50) Dimethylphthalate         | 13.975 | 163  | 1465624  | 37.844 | ng    | 100      |
| 51) 2,6-Dinitrotoluene        | 14.092 | 165  | 313984   | 39.690 | ng    | 91       |
| 52) Acenaphthene              | 14.575 | 154  | 1102499  | 38.080 | ng    | 99       |
| 53) 3-Nitroaniline            | 14.422 | 138  | 343155   | 40.118 | ng    | 98       |
| 54) 2,4-Dinitrophenol         | 14.633 | 184  | 74617    | 16.374 | ng    | 95       |
| 55) Dibenzofuran              | 14.910 | 168  | 1843812  | 38.171 | ng    | 96       |
| 56) 4-Nitrophenol             | 14.739 | 139  | 273227   | 39.346 | ng    | # 84     |
| 57) 2,4-Dinitrotoluene        | 14.881 | 165  | 428611   | 41.057 | ng    | 93       |
| 58) Fluorene                  | 15.563 | 166  | 1430157  | 38.324 | ng    | 100      |
| 59) 2,3,4,6-Tetrachlorophenol | 15.139 | 232  | 429279   | 40.347 | ng    | 95       |
| 60) Diethylphthalate          | 15.333 | 149  | 1434219  | 38.218 | ng    | 98       |
| 61) 4-Chlorophenyl-phenyle... | 15.551 | 204  | 768866   | 38.990 | ng    | 93       |
| 62) 4-Nitroaniline            | 15.586 | 138  | 357124   | 41.514 | ng    | # 81     |
| 63) Azobenzene                | 15.845 | 77   | 1215803  | 37.387 | ng    | 93       |
| 65) 4,6-Dinitro-2-methylph... | 15.651 | 198  | 130394   | 19.871 | ng    | 97       |
| 66) n-Nitrosodiphenylamine    | 15.769 | 169  | 1257010  | 38.635 | ng    | 98       |
| 67) 4-Bromophenyl-phenylether | 16.451 | 248  | 492658   | 41.108 | ng    | 95       |
| 68) Hexachlorobenzene         | 16.569 | 284  | 553475   | 41.787 | ng    | 96       |
| 69) Atrazine                  | 16.727 | 200  | 443079   | 38.634 | ng    | 98       |
| 70) Pentachlorophenol         | 16.916 | 266  | 325187   | 39.534 | ng    | 99       |
| 71) Phenanthrene              | 17.304 | 178  | 2245996  | 38.307 | ng    | 99       |
| 72) Anthracene                | 17.398 | 178  | 2248766  | 39.207 | ng    | 100      |
| 73) Carbazole                 | 17.669 | 167  | 2159978  | 38.431 | ng    | 98       |
| 74) Di-n-butylphthalate       | 18.216 | 149  | 2424086  | 38.918 | ng    | 98       |
| 75) Fluoranthene              | 19.274 | 202  | 2619872  | 38.784 | ng    | 98       |
| 77) Benzidine                 | 19.457 | 184  | 1142516  | 35.803 | ng    | 100      |
| 78) Pyrene                    | 19.627 | 202  | 2675307  | 39.269 | ng    | 99       |
| 80) Butylbenzylphthalate      | 20.498 | 149  | 1096594  | 40.193 | ng    | 93       |
| 81) Benzo(a)anthracene        | 21.333 | 228  | 2604302  | 38.639 | ng    | 99       |
| 82) 3,3'-Dichlorobenzidine    | 21.263 | 252  | 949752   | 41.030 | ng    | 98       |
| 83) Chrysene                  | 21.386 | 228  | 2455110  | 38.112 | ng    | 99       |
| 84) Bis(2-ethylhexyl)phtha... | 21.257 | 149  | 1513319  | 38.581 | ng    | 98       |
| 85) Di-n-octyl phthalate      | 22.180 | 149  | 2629027  | 39.368 | ng    | 99       |
| 87) Indeno(1,2,3-cd)pyrene    | 26.292 | 276  | 3126831  | 39.355 | ng    | 98       |
| 88) Benzo(b)fluoranthene      | 23.027 | 252  | 2579598  | 39.043 | ng    | 99       |
| 89) Benzo(k)fluoranthene      | 23.074 | 252  | 2457274  | 36.733 | ng    | 99       |
| 90) Benzo(a)pyrene            | 23.657 | 252  | 2158542  | 38.454 | ng    | 99       |
| 91) Dibenzo(a,h)anthracene    | 26.303 | 278  | 2595015  | 38.974 | ng    | 98       |
| 92) Benzo(g,h,i)perylene      | 27.062 | 276  | 2548255  | 39.218 | ng    | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

mohammad

9/29/2021 12:55:38 PM








#1  
1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 7.875 min Scan# 814  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:152 Resp: 220571  
Ion Ratio Lower Upper  
152 100  
150 157.8 125.6 188.4  
115 57.9 48.5 72.7


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#2  
1,4-Dioxane  
Concen: 34.204 ng  
RT: 3.340 min Scan# 43  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion: 88 Resp: 182260  
Ion Ratio Lower Upper  
88 100  
58 71.2 46.7 70.1#  
43 24.0 20.3 30.5








#3  
Pyridine  
Concen: 35.000 ng  
RT: 3.752 min Scan# 113  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion: 79 Resp: 510345  
Ion Ratio Lower Upper  
79 100  
52 65.0 41.7 62.5#  
51 32.0 24.1 36.1

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#4  
n-Nitrosodimethylamine  
Concen: 36.934 ng  
RT: 3.658 min Scan# 97  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion: 42 Resp: 184627  
Ion Ratio Lower Upper  
42 100  
74 159.5 105.4 158.0#  
44 7.6 4.2 6.4#








#5  
2-Fluorophenol  
Concen: 73.892 ng  
RT: 5.452 min Scan# 402  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 112   | Resp: | 973649 |
| Ion      | Ratio | Lower | Upper  |
| 112      | 100   |       |        |
| 64       | 53.4  | 41.4  | 62.2   |
| 63       | 27.3  | 23.8  | 35.8   |

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#6  
Aniline  
Concen: 36.644 ng  
RT: 7.204 min Scan# 700  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 93    | Resp: | 727425 |
| Ion      | Ratio | Lower | Upper  |
| 93       | 100   |       |        |
| 66       | 37.9  | 29.7  | 44.5   |
| 65       | 18.7  | 15.3  | 22.9   |








#7  
Phenol-d6  
Concen: 73.917 ng  
RT: 7.052 min Scan# 674  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Tgt Ion: 99 Resp: 1331191


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 99  | 100   |       |       |
| 42  | 13.9  | 13.4  | 20.0  |
| 71  | 31.8  | 30.6  | 46.0  |

Manual Integrations  
APPROVED



mohammad




9/29/2021 12:55:38 PM




#8  
2-Chlorophenol  
Concen: 38.153 ng  
RT: 7.446 min Scan# 741  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion: 128 Resp: 546382


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 128 | 100   |       |       |
| 130 | 32.5  | 12.4  | 52.4  |
| 64  | 40.5  | 20.3  | 60.3  |







#9  
Benzaldehyde  
Concen: 37.649 ng m  
RT: 7.016 min Scan# 668  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion: 77 Resp: 381939  
Ion Ratio Lower Upper  
77 100  
105 103.2 76.7 116.7  
106 99.3 76.2 116.2

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#10  
Phenol  
Concen: 36.802 ng  
RT: 7.081 min Scan# 679  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35




Tgt Ion: 94 Resp: 638993  
Ion Ratio Lower Upper  
94 100  
65 25.5 8.8 48.8  
66 35.4 20.7 60.7







#11  
bis(2-Chloroethyl)ether  
Concen: 35.836 ng  
RT: 7.304 min Scan# 717  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion: 93 Resp: 491155  
Ion Ratio Lower Upper  
93 100  
63 71.4 46.2 86.2  
95 32.9 12.2 52.2

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#12  
1,3-Dichlorobenzene  
Concen: 38.110 ng  
RT: 7.763 min Scan# 795  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion: 146 Resp: 615608  
Ion Ratio Lower Upper  
146 100  
148 64.5 51.5 77.3  
75 27.6 25.7 38.5








#13  
1,4-Dichlorobenzene  
Concen: 38.241 ng  
RT: 7.910 min Scan# 820  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:146 Resp: 629745  
Ion Ratio Lower Upper  
146 100  
148 64.6 51.3 76.9  
111 39.6 33.0 49.6


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#14  
1,2-Dichlorobenzene  
Concen: 38.265 ng  
RT: 8.228 min Scan# 874  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:146 Resp: 610636  
Ion Ratio Lower Upper  
146 100  
148 64.2 50.3 75.5  
111 39.6 33.7 50.5








#15  
Benzyl Alcohol  
Concen: 38.764 ng  
RT: 8.122 min Scan# 856  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion: 79 Resp: 447134  
Ion Ratio Lower Upper  
79 100  
108 82.3 62.5 93.7  
77 61.8 52.9 79.3

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 36.143 ng  
RT: 8.410 min Scan# 905  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35


Tgt Ion: 45 Resp: 567208  
Ion Ratio Lower Upper  
45 100  
77 16.6 2.0 42.0  
79 13.1 0.0 36.7







#17  
2-Methylphenol  
Concen: 37.944 ng  
RT: 8.328 min Scan# 891  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 444138 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 110.8 | 87.7  | 131.5  |
| 77       | 42.0  | 40.2  | 60.2   |
| 79       | 42.0  | 39.0  | 58.4   |

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#18  
Hexachloroethane  
Concen: 37.140 ng  
RT: 8.951 min Scan# 997  
Delta R.T. -0.024 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 117   | Resp: | 205347 |
| Ion      | Ratio | Lower | Upper  |
| 117      | 100   |       |        |
| 119      | 94.9  | 76.2  | 114.2  |
| 201      | 111.1 | 83.4  | 125.0  |








#19  
n-Nitroso-di-n-propylamine  
Concen: 35.677 ng  
RT: 8.693 min Scan# 953  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 70    | Resp: | 349456 |
| Ion      | Ratio | Lower | Upper  |
| 70       | 100   |       |        |
| 42       | 47.2  | 40.8  | 61.2   |
| 101      | 10.6  | 8.8   | 13.2   |
| 130      | 25.3  | 18.9  | 28.3   |


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#20  
3+4-Methylphenols  
Concen: 37.473 ng  
RT: 8.657 min Scan# 947  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 606500 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 87.7  | 71.0  | 111.0  |
| 77       | 30.7  | 16.6  | 56.6   |
| 79       | 26.4  | 12.7  | 52.7   |








#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.681 min Scan# 1291  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 136   | Resp: | 837672 |
| Ion      | Ratio | Lower | Upper  |
| 136      | 100   |       |        |
| 137      | 10.8  | 8.6   | 12.8   |
| 54       | 7.0   | 4.6   | 6.8#   |
| 68       | 5.4   | 3.6   | 5.4    |

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#22  
Acetophenone  
Concen: 38.069 ng  
RT: 8.704 min Scan# 955  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 105   | Resp: | 768379 |
| Ion      | Ratio | Lower | Upper  |
| 105      | 100   |       |        |
| 71       | 2.9   | 1.9   | 2.9    |
| 51       | 22.8  | 18.4  | 27.6   |
| 120      | 22.6  | 17.9  | 26.9   |








#23  
Nitrobenzene-d5  
Concen: 78.953 ng  
RT: 9.040 min Scan# 1012  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion: 82 Resp: 1135664  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 82  | 100   |       |       |
| 128 | 47.9  | 35.7  | 53.5  |
| 54  | 45.9  | 35.0  | 52.4  |


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED



mohammad




9/29/2021 12:55:38 PM




#24  
Nitrobenzene  
Concen: 39.018 ng  
RT: 9.081 min Scan# 1019  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion: 77 Resp: 535091  


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 77  | 100   |       |       |
| 123 | 50.3  | 38.2  | 57.2  |
| 65  | 13.3  | 11.4  | 17.2  |







#25  
Isophorone  
Concen: 37.780 ng  
RT: 9.616 min Scan# 1110  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion: 82 Resp: 947600  
Ion Ratio Lower Upper  
82 100  
95 7.7 6.5 9.7  
138 19.4 15.0 22.4

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#26  
2-Nitrophenol  
Concen: 42.070 ng  
RT: 9.793 min Scan# 1140  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion: 139 Resp: 300844  
Ion Ratio Lower Upper  
139 100  
109 29.9 28.8 43.2  
65 45.3 41.0 61.6








#27  
2,4-Dimethylphenol  
Concen: 38.257 ng  
RT: 9.857 min Scan# 1151  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:122 Resp: 430753  
Ion Ratio Lower Upper  
122 100  
107 116.3 94.5 141.7  
121 59.4 47.4 71.2

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#28  
bis(2-Chloroethoxy)methane  
Concen: 35.976 ng  
RT: 10.093 min Scan# 1191  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion: 93 Resp: 636617  
Ion Ratio Lower Upper  
93 100  
95 32.8 25.9 38.9  
123 11.8 10.2 15.2








#29  
2,4-Dichlorophenol  
Concen: 40.330 ng  
RT: 10.334 min Scan# 1232  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:162 Resp: 529709  
Ion Ratio Lower Upper  
162 100  
164 64.6 44.9 84.9  
98 32.6 14.3 54.3

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#30  
1,2,4-Trichlorobenzene  
Concen: 40.429 ng  
RT: 10.545 min Scan# 1268  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35


Tgt Ion:180 Resp: 604932  
Ion Ratio Lower Upper  
180 100  
182 94.4 77.9 116.9  
145 28.3 24.7 37.1







#31  
Naphthalene  
Concen: 38.471 ng  
RT: 10.728 min Scan# 1299  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:128 Resp: 1644715  
Ion Ratio Lower Upper  
128 100  
129 11.0 9.0 13.6  
127 13.4 10.7 16.1

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#32  
Benzoic acid  
Concen: 42.061 ng  
RT: 10.028 min Scan# 1180  
Delta R.T. 0.023 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35




Tgt Ion:122 Resp: 363992  
Ion Ratio Lower Upper  
122 100  
105 121.2 103.2 143.2  
77 86.9 76.2 116.2





#33  
4-Chloroaniline  
Concen: 38.611 ng  
RT: 10.840 min Scan# 1318  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC



Tgt Ion:127 Resp: 681956


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 127 | 100   |       |       |
| 129 | 32.1  | 26.1  | 39.1  |
| 65  | 27.9  | 22.5  | 33.7  |
| 92  | 18.8  | 15.0  | 22.6  |

Manual Integrations  
APPROVED




mohammad

9/29/2021 12:55:38 PM






#34  
Hexachlorobutadiene  
Concen: 42.809 ng  
RT: 11.010 min Scan# 1347  
Delta R.T. -0.024 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion:225 Resp: 383526

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 225 | 100   |       |       |
| 223 | 64.2  | 51.0  | 76.6  |
| 227 | 65.1  | 51.4  | 77.2  |








#35  
Caprolactam  
Concen: 40.183 ng  
RT: 11.651 min Scan# 1456  
Delta R.T. 0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:113 Resp: 162369  
Ion Ratio Lower Upper  
113 100  
55 141.4 98.9 138.9#  
56 107.9 71.7 111.7

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#36  
4-Chloro-3-methylphenol  
Concen: 39.245 ng  
RT: 11.975 min Scan# 1511  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:107 Resp: 523944  
Ion Ratio Lower Upper  
107 100  
144 27.4 21.1 31.7  
142 84.2 66.0 99.0








#37  
2-Methylnaphthalene  
Concen: 38.600 ng  
RT: 12.339 min Scan# 1573  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:142 Resp: 1155490  
Ion Ratio Lower Upper  
142 100  
141 88.7 71.8 107.6  
115 33.4 31.6 47.4


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#38  
1-Methylnaphthalene  
Concen: 38.460 ng  
RT: 12.557 min Scan# 1610  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:142 Resp: 1124929  
Ion Ratio Lower Upper  
142 100  
141 93.2 73.3 109.9  
116 3.6 3.5 5.3








#39  
Acenaphthene-d10  
Concen: 20.000 ng  
RT: 14.510 min Scan# 1942  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:164 Resp: 538466  
Ion Ratio Lower Upper  
164 100  
162 100.5 80.2 120.2  
160 44.1 34.5 51.7


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#40  
1,2,4,5-Tetrachlorobenzene  
Concen: 40.665 ng  
RT: 12.710 min Scan# 1636  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:216 Resp: 670548  
Ion Ratio Lower Upper  
216 100  
214 78.5 61.9 92.9  
179 20.2 16.8 25.2  
108 14.7 14.9 22.3#








#41  
Hexachlorocyclopentadiene  
Concen: 15.306 ng  
RT: 12.681 min Scan# 1631  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 237   | Resp: | 139887 |
| Ion Ratio | Lower | Upper |        |
| 237       | 100   |       |        |
| 235       | 62.3  | 44.5  | 84.5   |
| 272       | 12.7  | 0.0   | 32.5   |


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#42  
2,4,6-Tribromophenol  
Concen: 88.471 ng  
RT: 16.004 min Scan# 2196  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 330   | Resp: | 617611 |
| Ion Ratio | Lower | Upper |        |
| 330       | 100   |       |        |
| 332       | 96.9  | 77.8  | 116.8  |
| 141       | 27.9  | 25.4  | 38.2   |








#43  
2,4,6-Trichlorophenol  
Concen: 41.041 ng  
RT: 12.951 min Scan# 1677  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:196 Resp: 461776  
Ion Ratio Lower Upper  
196 100  
198 94.8 79.5 119.3  
200 30.1 25.0 37.6

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#44  
2,4,5-Trichlorophenol  
Concen: 40.653 ng  
RT: 13.028 min Scan# 1690  
Delta R.T. -0.000 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:196 Resp: 492567  
Ion Ratio Lower Upper  
196 100  
198 98.0 74.7 112.1  
97 40.6 37.4 56.0  
132 24.2 20.0 30.0








#45  
2-Fluorobiphenyl  
Concen: 76.515 ng  
RT: 13.139 min Scan# 1709  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:172 Resp: 2875889  
Ion Ratio Lower Upper  
172 100  
171 35.8 27.9 41.9  
170 23.7 18.6 27.8

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#46  
1,1'-Biphenyl  
Concen: 38.132 ng  
RT: 13.345 min Scan# 1744  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35


Tgt Ion:154 Resp: 1527271  
Ion Ratio Lower Upper  
154 100  
153 40.6 20.4 60.4  
76 11.5 0.0 30.8







#47  
2-Chloronaphthalene  
Concen: 38.669 ng  
RT: 13.392 min Scan# 1752  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:162 Resp: 1199718  
Ion Ratio Lower Upper  
162 100  
127 36.7 31.4 47.0  
164 33.6 26.6 39.8

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#48  
2-Nitroaniline  
Concen: 41.300 ng  
RT: 13.598 min Scan# 1787  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion: 65 Resp: 312407  
Ion Ratio Lower Upper  
65 100  
92 76.5 57.7 86.5  
138 126.5 92.1 138.1








#49  
Acenaphthylene  
Concen: 38.893 ng  
RT: 14.233 min Scan# 1895  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:152 Resp: 1858518  
Ion Ratio Lower Upper  
152 100  
151 20.3 16.2 24.2  
153 13.4 10.6 16.0


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#50  
Dimethylphthalate  
Concen: 37.844 ng  
RT: 13.975 min Scan# 1851  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35


Tgt Ion:163 Resp: 1465624  
Ion Ratio Lower Upper  
163 100  
194 4.4 3.4 5.2  
164 10.2 8.2 12.2







#51  
2,6-Dinitrotoluene  
Concen: 39.690 ng  
RT: 14.092 min Scan# 1871  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:165 Resp: 313984  
Ion Ratio Lower Upper  
165 100  
63 46.2 41.3 61.9  
89 48.9 44.8 67.2

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#52  
Acenaphthene  
Concen: 38.080 ng  
RT: 14.575 min Scan# 1953  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion:154 Resp: 1102499  
Ion Ratio Lower Upper  
154 100  
153 112.5 89.4 134.2  
152 53.2 43.0 64.4









#53  
3-Nitroaniline  
Concen: 40.118 ng  
RT: 14.422 min Scan# 1927  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 138   | Resp: | 343155 |
| Ion      | Ratio | Lower | Upper  |
| 138      | 100   |       |        |
| 108      | 10.7  | 8.4   | 12.6   |
| 92       | 107.2 | 87.8  | 131.8  |




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED




#54  
2,4-Dinitrophenol  
Concen: 16.374 ng  
RT: 14.633 min Scan# 1963  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 184   | Resp: | 74617 |
| Ion      | Ratio | Lower | Upper |
| 184      | 100   |       |       |
| 63       | 54.1  | 47.0  | 70.6  |
| 154      | 62.9  | 52.3  | 78.5  |




mohammad

9/29/2021 12:55:38 PM





#55  
Dibenzenofuran  
Concen: 38.171 ng  
RT: 14.910 min Scan# 2010  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:168 Resp: 1843812  
Ion Ratio Lower Upper  
168 100  
139 35.7 31.0 46.6  
169 13.4 10.8 16.2

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#56  
4-Nitrophenol  
Concen: 39.346 ng  
RT: 14.739 min Scan# 1981  
Delta R.T. 0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35




Tgt Ion:139 Resp: 273227  
Ion Ratio Lower Upper  
139 100  
109 69.0 72.3 112.3#  
65 85.0 72.5 112.5








#57  
2,4-Dinitrotoluene  
Concen: 41.057 ng  
RT: 14.881 min Scan# 2005  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35




Tgt Ion:165 Resp: 428611

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 165 | 100   |       |       |
| 63  | 36.3  | 32.3  | 48.5  |
| 89  | 63.3  | 55.9  | 83.9  |
| 182 | 2.9   | 2.2   | 3.4   |





#58  
Fluorene  
Concen: 38.324 ng  
RT: 15.563 min Scan# 2121  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion:166 Resp: 1430157

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 166 | 100   |       |       |
| 165 | 97.3  | 78.0  | 117.0 |
| 167 | 13.3  | 10.5  | 15.7  |




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED


mohammad

9/29/2021 12:55:38 PM





#59  
2,3,4,6-Tetrachlorophenol  
Concen: 40.347 ng  
RT: 15.139 min Scan# 2049  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:232 Resp: 429279  
Ion Ratio Lower Upper  
232 100  
131 38.8 32.2 48.4  
130 2.1 2.0 3.0  
166 28.4 19.4 29.2

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#60  
Diethylphthalate  
Concen: 38.218 ng  
RT: 15.333 min Scan# 2082  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35




Tgt Ion:149 Resp: 1434219  
Ion Ratio Lower Upper  
149 100  
177 23.6 17.5 26.3  
150 12.6 9.9 14.9







#61  
4-Chlorophenyl-phenylether  
Concen: 38.990 ng  
RT: 15.551 min Scan# 2119  
Delta R.T. -0.018 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:204 Resp: 768866  
Ion Ratio Lower Upper  
204 100  
206 33.0 26.6 40.0  
141 52.7 48.6 73.0

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#62  
4-Nitroaniline  
Concen: 41.514 ng  
RT: 15.586 min Scan# 2125  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35




Tgt Ion:138 Resp: 357124  
Ion Ratio Lower Upper  
138 100  
92 47.0 31.2 71.2  
108 73.0 78.6 118.6#







#63  
Azobenzene  
Concen: 37.387 ng  
RT: 15.845 min Scan# 2169  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion: 77 Resp: 1215803  
Ion Ratio Lower Upper  
77 100  
182 30.2 6.0 46.0  
105 22.3 0.0 38.8  
51 26.3 8.8 48.8

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#64  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 17.263 min Scan# 2410  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion: 188 Resp: 1093483  
Ion Ratio Lower Upper  
188 100  
94 8.4 5.6 8.4  
80 9.0 6.3 9.5








#65  
4,6-Dinitro-2-methylphenol  
Concen: 19.871 ng  
RT: 15.651 min Scan# 2136  
Delta R.T. -0.000 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:198 Resp: 130394  
Ion Ratio Lower Upper  
198 100  
51 32.4 14.0 54.0  
105 37.7 19.5 59.5


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#66  
n-Nitrosodiphenylamine  
Concen: 38.635 ng  
RT: 15.769 min Scan# 2156  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:169 Resp: 1257010  
Ion Ratio Lower Upper  
169 100  
168 66.2 54.4 81.6  
167 35.7 29.0 43.6








#67  
4-Bromophenyl-phenylether  
Concen: 41.108 ng  
RT: 16.451 min Scan# 2272  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 248   | Resp: | 492658 |
| Ion Ratio | Lower | Upper |        |
| 248       | 100   |       |        |
| 250       | 99.0  | 79.8  | 119.8  |
| 141       | 62.5  | 57.8  | 86.6   |

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#68  
Hexachlorobenzene  
Concen: 41.787 ng  
RT: 16.569 min Scan# 2292  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 284   | Resp: | 553475 |
| Ion Ratio | Lower | Upper |        |
| 284       | 100   |       |        |
| 142       | 29.9  | 26.9  | 40.3   |
| 249       | 30.2  | 24.5  | 36.7   |








#69  
Atrazine  
Concen: 38.634 ng  
RT: 16.727 min Scan# 2319  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 200   | Resp: | 443079 |
| Ion Ratio | Lower | Upper |        |
| 200       | 100   |       |        |
| 173       | 25.9  | 4.1   | 44.1   |
| 215       | 48.9  | 27.7  | 67.7   |


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#70  
Pentachlorophenol  
Concen: 39.534 ng  
RT: 16.916 min Scan# 2351  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 266   | Resp: | 325187 |
| Ion Ratio | Lower | Upper |        |
| 266       | 100   |       |        |
| 268       | 63.6  | 51.6  | 77.4   |
| 264       | 63.7  | 50.3  | 75.5   |







#71  
Phenanthrene  
Concen: 38.307 ng  
RT: 17.304 min Scan# 2417  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:178 Resp: 2245996  
Ion Ratio Lower Upper  
178 100  
176 19.4 15.1 22.7  
179 15.8 12.4 18.6

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#72  
Anthracene  
Concen: 39.207 ng  
RT: 17.398 min Scan# 2433  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35


Tgt Ion:178 Resp: 2248766  
Ion Ratio Lower Upper  
178 100  
176 18.6 14.9 22.3  
179 15.6 12.2 18.2





#73  
 Carbazole  
 Concen: 38.431 ng  
 RT: 17.669 min Scan# 2479  
 Delta R.T. -0.012 min  
 Lab File: BP007241.D  
 Acq: 28 Sep 2021 21:35



Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040EC



Tgt Ion:167 Resp: 2159978


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 167 | 100   |       |       |
| 166 | 21.8  | 17.1  | 25.7  |
| 139 | 12.3  | 10.7  | 16.1  |

Manual Integrations  
 APPROVED




mohammad

9/29/2021 12:55:38 PM






#74  
 Di-n-butylphthalate  
 Concen: 38.918 ng  
 RT: 18.216 min Scan# 2572  
 Delta R.T. -0.012 min  
 Lab File: BP007241.D  
 Acq: 28 Sep 2021 21:35



Tgt Ion:149 Resp: 2424086


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 149 | 100   |       |       |
| 150 | 9.5   | 7.3   | 10.9  |
| 104 | 4.6   | 4.5   | 6.7   |







#75  
Fluoranthene  
Concen: 38.784 ng  
RT: 19.274 min Scan# 2752  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:202 Resp: 2619872  
Ion Ratio Lower Upper  
202 100  
101 9.6 0.0 27.6  
203 17.7 0.0 37.8

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#76  
Chrysene-d12  
Concen: 20.000 ng  
RT: 21.351 min Scan# 3105  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion:240 Resp: 1038385  
Ion Ratio Lower Upper  
240 100  
120 9.4 6.5 9.7  
236 25.8 20.6 31.0








#77  
Benzidine  
Concen: 35.803 ng  
RT: 19.457 min Scan# 2783  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:184 Resp: 1142516  
Ion Ratio Lower Upper  
184 100  
185 13.6 11.0 16.4  
183 11.6 9.5 14.3

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#78  
Pyrene  
Concen: 39.269 ng  
RT: 19.627 min Scan# 2812  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:202 Resp: 2675307  
Ion Ratio Lower Upper  
202 100  
200 21.4 16.3 24.5  
203 17.6 14.2 21.2








#79  
Terphenyl-d14  
Concen: 79.430 ng  
RT: 19.821 min Scan# 2845  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:244 Resp: 4304139  
Ion Ratio Lower Upper  
244 100  
212 7.2 6.0 9.0  
122 8.4 7.1 10.7


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#80  
Butylbenzylphthalate  
Concen: 40.193 ng  
RT: 20.498 min Scan# 2960  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Tgt Ion:149 Resp: 1096594  
Ion Ratio Lower Upper  
149 100  
91 62.0 54.6 81.8  
206 23.7 16.7 25.1








#81  
Benzo(a)anthracene  
Concen: 38.639 ng  
RT: 21.333 min Scan# 3102  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:228 Resp: 2604302  
Ion Ratio Lower Upper  
228 100  
226 27.4 21.6 32.4  
229 19.9 15.8 23.6

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#82  
3,3'-Dichlorobenzidine  
Concen: 41.030 ng  
RT: 21.263 min Scan# 3090  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35


Tgt Ion:252 Resp: 949752  
Ion Ratio Lower Upper  
252 100  
254 64.7 50.6 76.0  
126 9.5 7.0 10.6







#83  
Chrysene  
Concen: 38.112 ng  
RT: 21.386 min Scan# 3111  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:228 Resp: 2455110  
Ion Ratio Lower Upper  
228 100  
226 30.3 23.8 35.6  
229 20.0 15.4 23.2

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#84  
Bis(2-ethylhexyl)phthalate  
Concen: 38.581 ng  
RT: 21.257 min Scan# 3089  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35




Tgt Ion:149 Resp: 1513319  
Ion Ratio Lower Upper  
149 100  
167 28.6 21.9 32.9  
279 5.5 4.3 6.5







#85  
Di-n-octyl phthalate  
Concen: 39.368 ng  
RT: 22.180 min Scan# 3246  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:149 Resp: 2629027  
Ion Ratio Lower Upper  
149 100  
167 1.7 1.4 2.0  
43 6.7 5.6 8.4

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#86  
Perylene-d12  
Concen: 20.000 ng  
RT: 23.762 min Scan# 3515  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35




Tgt Ion:264 Resp: 1105567  
Ion Ratio Lower Upper  
264 100  
260 23.6 18.6 28.0  
265 22.2 17.6 26.4







#87  
Indeno(1,2,3-cd)pyrene  
Concen: 39.355 ng  
RT: 26.292 min Scan# 3945  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:276 Resp: 3126831  
Ion Ratio Lower Upper  
276 100  
138 21.1 15.3 22.9  
277 25.5 20.4 30.6

Manual Integrations  
APPROVED




mohammad



9/29/2021 12:55:38 PM




#88  
Benzo(b)fluoranthene  
Concen: 39.043 ng  
RT: 23.027 min Scan# 3390  
Delta R.T. -0.006 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion:252 Resp: 2579598  
Ion Ratio Lower Upper  
252 100  
253 22.0 17.5 26.3  
125 8.4 6.2 9.4








#89  
Benzo(k)fluoranthene  
Concen: 36.733 ng  
RT: 23.074 min Scan# 3398  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:252 Resp: 2457274  
Ion Ratio Lower Upper  
252 100  
253 22.4 17.6 26.4  
125 8.4 6.0 9.0

Manual Integrations  
APPROVED






mohammad


9/29/2021 12:55:38 PM



#90  
Benzo(a)pyrene  
Concen: 38.454 ng  
RT: 23.657 min Scan# 3497  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35


Tgt Ion:252 Resp: 2158542  
Ion Ratio Lower Upper  
252 100  
253 22.2 17.9 26.9  
125 9.4 6.5 9.7








#91  
Dibenzo(a,h)anthracene  
Concen: 38.974 ng  
RT: 26.303 min Scan# 3947  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040EC




Tgt Ion:278 Resp: 2595015  
Ion Ratio Lower Upper  
278 100  
139 13.7 9.3 13.9  
279 23.8 18.8 28.2



Manual Integrations  
APPROVED



#92  
Benzo(g,h,i)perylene  
Concen: 39.218 ng  
RT: 27.062 min Scan# 4076  
Delta R.T. -0.012 min  
Lab File: BP007241.D  
Acq: 28 Sep 2021 21:35



Tgt Ion:276 Resp: 2548255  
Ion Ratio Lower Upper  
276 100  
277 23.7 18.9 28.3  
138 18.3 12.6 18.8



mohammad

9/29/2021 12:55:38 PM