

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP093021\
 Data File : BP007281.D
 Acq On : 30 Sep 2021 20:53
 Operator : CG/JU
 Sample : M3984-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 GARAGE-FLOOR

Manual Integrations
 APPROVED

mohammad
 10/1/2021 11:29:18 AM

Quant Time: Oct 01 05:04:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 29 14:45:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	233962	20.000	ng	0.00
21) Naphthalene-d8	10.669	136	841166	20.000	ng	# 0.00
39) Acenaphthene-d10	14.504	164	482520	20.000	ng	0.00
64) Phenanthrene-d10	17.257	188	957606	20.000	ng	# 0.00
76) Chrysene-d12	21.345	240	1017137	20.000	ng	# 0.00
86) Perylene-d12	23.757	264	1124556	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.452	112	1398848	100.084	ng	0.00
7) Phenol-d6	7.058	99	1636393	85.663	ng	0.00
23) Nitrobenzene-d5	9.034	82	990722	68.590	ng	0.00
42) 2,4,6-Tribromophenol	15.998	330	580021	92.720	ng	0.00
45) 2-Fluorobiphenyl	13.128	172	2305106	68.440	ng	0.00
79) Terphenyl-d14	19.816	244	3342722	62.976	ng	0.00
Target Compounds						
74) Di-n-butylphthalate	18.216	149	266520	4.886	ng	# 89
80) Butylbenzylphthalate	20.498	149	1041422	38.968	ng	# 61
84) Bis(2-ethylhexyl)phtha...	21.251	149	547017m	14.237	ng	
85) Di-n-octyl phthalate	22.174	149	138575	2.118	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP093021\
Data File : BP007281.D
Acq On : 30 Sep 2021 20:53
Operator : CG/JU
Sample : M3984-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
GARAGE-FLOOR

Manual Integrations
APPROVED

mohammad
10/1/2021 11:29:18 AM

Quant Time: Oct 01 05:04:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 29 14:45:49 2021
Response via : Initial Calibration

