

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/3/2025	BNA_P	<u>BP025912.D</u>
Compound Name	Response	Retention Time
DDT	2779419	20.765
DDD	54374	20.307
DDE	6332	19.971
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
60706	2840125	2.14