


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100325\
Data File : BP025912.D
Acq On : 03 Oct 2025 13:25
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Sep 11 16:45:04 2025






AutoFind: Scans 2493, 2494, 2495; Background Corrected with Scan 2483

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
68	69	0.00	2	1.6	3234	PASS
69	69	100	100	100.0	196107	PASS
70	69	0.00	2	0.5	898	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	603890	PASS
199	198	5	9	6.8	40978	PASS
365	198	1	100	3.3	19845	PASS
441	443	0.01	150	79.9	81739	PASS
442	442	100	100	100.0	527381	PASS
443	442	15	24	19.4	102321	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100325\
Data File : BP025912.D
Acq On : 03 Oct 2025 13:25
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Oct 03 13:46:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration



TIC: BP025912.D\data.ms

(70) Pentachlorophenol (C)

16.801min (-0.030) 3092542.18 ng




response 1175251

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.30	62.80
264.00	62.70	63.22
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100325\
Data File : BP025912.D
Acq On : 03 Oct 2025 13:25
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Oct 03 13:46:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration



TIC: BP025912.D\data.ms

(77) Benzidine

19.483min (-0.024) 856514.67 ng

response 6735512

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.99
183.00	11.50	11.80
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/3/2025	BNA_P	<u>BP025912.D</u>
Compound Name	Response	Retention Time
DDT	2779419	20.765
DDD	54374	20.307
DDE	6332	19.971
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
60706	2840125	2.14

Instrument :
BNA_P
ClientSampleId :
DFTPP