

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100620\
 Data File : BP003539.D
 Acq On : 07 Oct 2020 03:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Oct 07 05:41:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 05 14:33:14 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.528	152	66	20.00	ng	# 0.03
21) Naphthalene-d8	10.299	136	19	20.00	ng	# 0.02
39) Acenaphthene-d10	13.963	164	18	20.00	ng	-0.19
64) Phenanthrene-d10	16.692	188	18	20.00	ng	#-0.22
76) Chrysene-d12	20.757	240	23	20.00	ng	#-0.27
86) Perylene-d12	22.874	264	128	20.00	ng	#-0.32
System Monitoring Compounds						
5) 2-Fluorophenol	5.116	112	10	2.65	ng	0.00
7) Phenol-d6	6.734	99	15	2.98	ng	0.03
23) Nitrobenzene-d5	8.640	82	55	170.06	ng	-0.01
42) 2,4,6-Tribromophenol	15.439	330	6	21.85	ng	-0.22
45) 2-Fluorobiphenyl	12.604	172	13	10.56	ng	-0.17
79) Terphenyl-d14	19.310	244	126	114.14	ng	-0.19
Target Compounds						Qvalue
2) 1,4-Dioxane	3.011	88	86	55.37	ng	# 68
3) Pyridine	3.422	79	158	38.20	ng	# 35
4) n-Nitrosodimethylamine	3.346	42	115	92.96	ng	# 1
6) Aniline	6.875	93	92	15.90	ng	# 85
8) 2-Chlorophenol	7.063	128	12	2.85	ng	88
9) Benzaldehyde	6.669	77	54	19.22	ng	# 33
10) Phenol	6.722	94	26	5.40	ng	85
11) bis(2-Chloroethyl)ether	6.993	93	35	9.18	ng	74
14) 1,2-Dichlorobenzene	7.881	146	15	3.07	ng	# 19
15) Benzyl Alcohol	7.757	79	50	14.93	ng	# 45
16) 2,2'-oxybis(1-Chloropr...	8.063	45	43	15.51	ng	# 1
17) 2-Methylphenol	8.010	107	36	10.35	ng	# 1
18) Hexachloroethane	8.581	117	47	24.77	ng	# 33
19) n-Nitroso-di-n-propyla...	8.352	70	10	3.70	ng	# 34
20) 3+4-Methylphenols	8.340	107	32	6.91	ng	# 22
22) Acetophenone	8.334	105	29	62.15	ng	# 46
24) Nitrobenzene	8.710	77	44	135.97	ng	# 37
25) Isophorone	9.240	82	19	32.06	ng	# 39
26) 2-Nitrophenol	9.422	139	6	33.95	ng	# 1
27) 2,4-Dimethylphenol	9.510	122	5	20.00	ng	# 74
28) bis(2-Chloroethoxy)met...	9.722	93	42	106.46	ng	# 54
29) 2,4-Dichlorophenol	9.975	162	12	38.42	ng	# 21
30) 1,2,4-Trichlorobenzene	10.157	180	16	43.18	ng	# 51
31) Naphthalene	10.351	128	13	12.95	ng	# 1
32) Benzoic acid	9.563	122	23	113.22	ng	97
33) 4-Chloroaniline	10.446	127	54	126.65	ng	# 14
34) Hexachlorobutadiene	10.557	225	12	47.44	ng	99
35) Caprolactam	11.228	113	8	76.75	ng	# 1
36) 4-Chloro-3-methylphenol	11.669	107	11	32.66	ng	# 56
37) 2-Methylnaphthalene	11.998	142	42	57.76	ng	# 54
38) 1-Methylnaphthalene	12.193	142	11	15.93	ng	# 53
40) 1,2,4,5-Tetrachloroben...	12.163	216	5	8.48	ng	# 1
41) Hexachlorocyclopentadiene	12.145	237	7	44.23	ng	90
43) 2,4,6-Trichlorophenol	12.281	196	3	7.87	ng	88
44) 2,4,5-Trichlorophenol	12.504	196	21	53.58	ng	# 1
46) 1,1'-Biphenyl	12.822	154	9	6.70	ng	# 1

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100620\
 Data File : BP003539.D
 Acq On : 07 Oct 2020 03:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Oct 07 05:41:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 05 14:33:14 2020
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
47) 2-Chloronaphthalene	12.851	162	12	10.79	ng	# 36
48) 2-Nitroaniline	13.057	65	19	67.53	ng	# 14
49) Acenaphthylene	13.675	152	20	11.77	ng	# 1
50) Dimethylphthalate	13.434	163	10	6.96	ng	# 1
51) 2,6-Dinitrotoluene	13.569	165	36	116.85	ng	# 82
52) Acenaphthene	14.034	154	21	20.32	ng	# 31
53) 3-Nitroaniline	13.887	138	23	68.89	ng	# 22
54) 2,4-Dinitrophenol	14.128	184	30	194.38	ng	# 1
55) Dibenzofuran	14.363	168	10	6.06	ng	# 42
56) 4-Nitrophenol	14.251	139	42	205.83	ng	# 1
57) 2,4-Dinitrotoluene	14.351	165	38	88.93	ng	# 60
58) Fluorene	15.010	166	31	23.79	ng	# 41
59) 2,3,4,6-Tetrachlorophenol	14.622	232	17	46.13	ng	# 1
60) Diethylphthalate	14.798	149	15	10.23	ng	# 29
61) 4-Chlorophenyl-phenyle...	15.004	204	7	9.84	ng	# 68
62) 4-Nitroaniline	15.151	138	24	67.74	ng	# 96
63) Azobenzene	15.322	77	10	8.95	ng	# 1
65) 4,6-Dinitro-2-methylph...	15.128	198	19	200.03	ng	# 1
66) n-Nitrosodiphenylamine	15.222	169	27	51.07	ng	# 1
67) 4-Bromophenyl-phenylether	15.898	248	19	87.33	ng	# 58
68) Hexachlorobenzene	16.022	284	25	96.72	ng	# 15
69) Atrazine	16.375	200	10	48.22	ng	# 82
70) Pentachlorophenol	16.586	266	103883	857065.94	ng	# 99
71) Phenanthrene	16.710	178	40	40.84	ng	# 1
72) Anthracene	16.851	178	25	25.90	ng	# 84
73) Carbazole	17.069	167	117	127.62	ng	# 53
74) Di-n-butylphthalate	17.680	149	115	98.12	ng	# 56
75) Fluoranthene	18.963	202	814	661.03	ng	# 86
77) Benzidine	19.127	184	315832	442836.21	ng	# 99
78) Pyrene	19.304	202	1523	1061.78	ng	# 96
80) Butylbenzylphthalate	19.904	149	511	778.69	ng	# 76
81) Benzo(a)anthracene	20.857	228	45	30.27	ng	# 56
82) 3,3'-Dichlorobenzidine	20.692	252	65	113.11	ng	# 45
83) Chrysene	21.027	228	2336	1635.36	ng	# 89
84) Bis(2-ethylhexyl)phtha...	20.816	149	1368	1480.72	ng	# 98
85) Di-n-octyl phthalate	21.498	149	100	59.24	ng	# 1
87) Indeno(1,2,3-cd)pyrene	25.045	276	23	2.37	ng	# 41
88) Benzo(b)fluoranthene	22.204	252	32	3.96	ng	# 1
89) Benzo(k)fluoranthene	22.163	252	83	10.71	ng	# 1
90) Benzo(a)pyrene	22.610	252	3732	489.64	ng	# 77
91) Dibenzo(a,h)anthracene	25.068	278	42	5.25	ng	# 1
92) Benzo(g,h,i)perylene	25.715	276	44	5.51	ng	# 32

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100620\
 Data File : BP003539.D
 Acq On : 07 Oct 2020 03:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampled :
 DFTPP


Quant Time: Oct 07 05:41:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 05 14:33:14 2020
 Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100620\
Data File : BP003539.D
Acq On : 07 Oct 2020 03:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Oct 07 05:41:00 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 05 14:33:14 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.528 min Scan# 789
Delta R.T. 0.030 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:152 Resp: 66
Ion Ratio Lower Upper
152 100
150 142.9 123.8 185.6
115 114.3 43.8 65.6#



#2
1,4-Dioxane
Concen: 55.37 ng
RT: 3.011 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 88 Resp: 86
Ion Ratio Lower Upper
88 100
58 34.9 44.2 66.4#
43 0.0 15.2 22.8#









#3
Pyridine
Concen: 38.20 ng
RT: 3.422 min Scan# 91
Delta R.T. 0.024 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 79 Resp: 158
Ion Ratio Lower Upper
79 100
52 29.7 41.3 61.9#
51 93.1 21.2 31.8#



#4
n-Nitrosodimethylamine
Concen: 92.96 ng
RT: 3.346 min Scan# 78
Delta R.T. 0.024 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 42 Resp: 115
Ion Ratio Lower Upper
42 100
74 0.0 148.6 223.0#
44 2378.7 5.3 7.9#









#5
2-Fluorophenol
Concen: 2.65 ng
RT: 5.116 min Scan# 379
Delta R.T. 0.006 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:112 Resp: 10
Ion Ratio Lower Upper
112 100
64 76.9 32.8 49.2#
63 61.5 17.0 25.4#



#6
Aniline
Concen: 15.90 ng
RT: 6.875 min Scan# 678
Delta R.T. 0.041 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 93 Resp: 92
Ion Ratio Lower Upper
93 100
66 19.7 25.4 38.0#
65 15.5 12.5 18.7









#7
Phenol-d6
Concen: 2.98 ng
RT: 6.734 min Scan# 654
Delta R.T. 0.030 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 99 Resp: 15
Ion Ratio Lower Upper
99 100
42 164.3 8.6 13.0#
71 85.7 23.4 35.2#



#8
2-Chlorophenol
Concen: 2.85 ng
RT: 7.063 min Scan# 710
Delta R.T. -0.012 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 128 Resp: 12
Ion Ratio Lower Upper
128 100
130 38.5 13.0 53.0
64 38.5 10.4 50.4









#9
Benzaldehyde
Concen: 19.22 ng
RT: 6.669 min Scan# 643
Delta R.T. 0.024 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	77	Resp:	54
Ion	Ratio	Lower	Upper
77	100		
105	38.0	97.1	137.1#
106	48.0	93.7	133.7#



#10
Phenol
Concen: 5.40 ng
RT: 6.722 min Scan# 652
Delta R.T. -0.012 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	94	Resp:	26
Ion	Ratio	Lower	Upper
94	100		
65	24.3	2.3	42.3
66	43.2	10.9	50.9









#11
bis(2-Chloroethyl)ether
Concen: 9.18 ng
RT: 6.993 min Scan# 698
Delta R.T. 0.059 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 93	Resp: 35
Ion Ratio	Lower Upper
93 100	
63 75.9	39.0 79.0
95 51.7	12.8 52.8



#14
1,2-Dichlorobenzene
Concen: 3.07 ng
RT: 7.881 min Scan# 849
Delta R.T. 0.035 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:146	Resp:	15	
Ion	Ratio	Lower	Upper
146	100		
148	93.3	51.6	77.4#
111	133.3	31.0	46.6#









#15
Benzyl Alcohol
Concen: 14.93 ng
RT: 7.757 min Scan# 828
Delta R.T. 0.006 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 79 Resp: 50
Ion Ratio Lower Upper
79 100
108 8.6 73.5 110.3#
77 56.8 50.3 75.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 15.51 ng
RT: 8.063 min Scan# 880
Delta R.T. 0.024 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 45 Resp: 43
Ion Ratio Lower Upper
45 100
77 75.8 2.3 42.3#
79 115.2 0.0 36.6#









#17
2-Methylphenol
Concen: 10.35 ng
RT: 8.010 min Scan# 871
Delta R.T. 0.041 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	107	Resp:	36
Ion	Ratio	Lower	Upper
107	100		
108	45.0	89.6	134.4#
77	305.0	31.8	47.6#
79	70.0	31.1	46.7#



#18
Hexachloroethane
Concen: 24.77 ng
RT: 8.581 min Scan# 968
Delta R.T. 0.018 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	117	Resp:	47
Ion	Ratio	Lower	Upper
117	100		
119	56.4	78.8	118.2#
201	20.5	92.3	138.5#









#19
n-Nitroso-di-n-propylamine
Concen: 3.70 ng
RT: 8.352 min Scan# 929
Delta R.T. 0.041 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 70	Resp: 10
Ion Ratio	Lower Upper
70 100	
42 86.2	32.8 49.2#
101 62.1	9.6 14.4#
130 48.3	24.3 36.5#



#20
3+4-Methylphenols
Concen: 6.91 ng
RT: 8.340 min Scan# 927
Delta R.T. 0.035 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt	Ion:107	Resp:	32
Ion	Ratio	Lower	Upper
107	100		
108	53.6	69.0	109.0#
77	107.1	8.5	48.5#
79	96.4	4.8	44.8#









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.299 min Scan# 1260
Delta R.T. 0.024 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	136	Resp:	19
Ion	Ratio	Lower	Upper
136	100		
137	75.0	8.6	13.0#
54	130.0	3.6	5.4#
68	95.0	3.3	4.9#



#22
Acetophenone
Concen: 62.15 ng
RT: 8.334 min Scan# 926
Delta R.T. 0.012 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	105	Resp:	29
Ion	Ratio	Lower	Upper
105	100		
71	23.3	2.6	3.8#
51	66.7	13.5	20.3#
120	30.0	19.6	29.4#









#23
Nitrobenzene-d5
Concen: 170.06 ng
RT: 8.640 min Scan# 978
Delta R.T. -0.012 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion: 82 Resp: 55
Ion Ratio Lower Upper
82 100
128 48.6 45.6 68.4
54 67.6 30.1 45.1#



#24
Nitrobenzene
Concen: 135.97 ng
RT: 8.710 min Scan# 990
Delta R.T. 0.018 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 77 Resp: 44
Ion Ratio Lower Upper
77 100
123 18.2 48.6 72.8#
65 50.0 10.0 15.0#









#25
Isophorone
Concen: 32.06 ng
RT: 9.240 min Scan# 1080
Delta R.T. 0.024 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 82 Resp: 19
Ion Ratio Lower Upper
82 100
95 90.3 6.4 9.6#
138 22.6 18.7 28.1



#26
2-Nitrophenol
Concen: 33.95 ng
RT: 9.422 min Scan# 1111
Delta R.T. 0.012 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 139 Resp: 6
Ion Ratio Lower Upper
139 100
109 100.0 22.1 33.1#
65 131.3 26.5 39.7#








#27
2,4-Dimethylphenol
Concen: 20.00 ng
RT: 9.510 min Scan# 1126
Delta R.T. 0.024 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion:122 Resp: 5
Ion Ratio Lower Upper
122 100
107 92.3 87.3 130.9
121 92.3 47.3 70.9#



#28
bis(2-Chloroethoxy)methane
Concen: 106.46 ng
RT: 9.722 min Scan# 1162
Delta R.T. 0.018 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 93 Resp: 42
Ion Ratio Lower Upper
93 100
95 62.5 26.7 40.1#
123 0.0 10.5 15.7#









#29
2,4-Dichlorophenol
Concen: 38.42 ng
RT: 9.975 min Scan# 1205
Delta R.T. 0.018 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:162 Resp: 12
Ion Ratio Lower Upper
162 100
164 70.6 45.8 85.8
98 158.8 10.2 50.2#



#30
1,2,4-Trichlorobenzene
Concen: 43.18 ng
RT: 10.157 min Scan# 1236
Delta R.T. 0.006 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:180 Resp: 16
Ion Ratio Lower Upper
180 100
182 42.1 76.2 114.2#
145 47.4 23.8 35.8#









#31
Naphthalene
Concen: 12.95 ng
RT: 10.351 min Scan# 1269
Delta R.T. 0.030 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:128 Resp: 13
Ion Ratio Lower Upper
128 100
129 15.0 8.8 13.2#
127 85.0 10.6 15.8#



#32
Benzoic acid
Concen: 113.22 ng
RT: 9.563 min Scan# 1135
Delta R.T. -0.129 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:122 Resp: 23
Ion Ratio Lower Upper
122 100
105 116.7 94.1 134.1
77 66.7 50.1 90.1









#33
4-Chloroaniline
Concen: 126.65 ng
RT: 10.446 min Scan# 1285
Delta R.T. -0.006 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	127	Resp:	54
Ion	Ratio	Lower	Upper
127	100		
129	79.2	25.7	38.5#
65	58.3	17.3	25.9#
92	58.3	12.9	19.3#



#34
Hexachlorobutadiene
Concen: 47.44 ng
RT: 10.557 min Scan# 1304
Delta R.T. -0.059 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	225	Resp:	12
Ion	Ratio	Lower	Upper
225	100		
223	62.5	50.1	75.1
227	62.5	51.4	77.2









#35
Caprolactam
Concen: 76.75 ng
RT: 11.228 min Scan# 1418
Delta R.T. 0.012 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	113	Resp:	8
Ion	Ratio	Lower	Upper
113	100		
55	737.5	77.3	117.3#
56	275.0	55.9	95.9#



#36
4-Chloro-3-methylphenol
Concen: 32.66 ng
RT: 11.669 min Scan# 1493
Delta R.T. 0.059 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	107	Resp:	11
Ion	Ratio	Lower	Upper
107	100		
144	5.3	22.7	34.1#
142	47.4	70.7	106.1#









#37
2-Methylnaphthalene
Concen: 57.76 ng
RT: 11.998 min Scan# 1549
Delta R.T. 0.047 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	142	Resp:	42
Ion Ratio	Lower	Upper	
142	100		
141	36.4	71.4	107.2#
115	42.4	25.8	38.6#



#38
1-Methylnaphthalene
Concen: 15.93 ng
RT: 12.193 min Scan# 1582
Delta R.T. 0.018 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	142	Resp:	11
Ion Ratio	Lower	Upper	
142	100		
141	133.3	74.4	111.6#
116	60.0	2.9	4.3#





#39

Acenaphthene-d10

Concen: 20.00 ng

RT: 13.963 min Scan# 1883

Delta R.T. -0.194 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18





Instrument :

BNA_P

ClientSampleId :

DFTPP

Tgt Ion:164	Resp:	18
Ion Ratio	Lower	Upper
164	100	
162	83.3	79.4 119.2
160	38.9	36.1 54.1



#40

1,2,4,5-Tetrachlorobenzene

Concen: 8.48 ng




RT: 12.163 min Scan# 1577


Delta R.T. -0.176 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18

Tgt Ion:216	Resp:	5
Ion Ratio	Lower	Upper
216	100	
214	120.0	63.0 94.6#
179	0.0	16.9 25.3#
108	360.0	12.6 19.0#









#41
Hexachlorocyclopentadiene
Concen: 44.23 ng
RT: 12.145 min Scan# 1574
Delta R.T. -0.170 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	237	Resp:	7
Ion	Ratio	Lower	Upper
237	100		
235	70.0	43.2	83.2
272	20.0	0.0	33.3



#42
2,4,6-Tribromophenol
Concen: 21.85 ng
RT: 15.439 min Scan# 2134
Delta R.T. -0.223 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	330	Resp:	6
Ion	Ratio	Lower	Upper
330	100		
332	166.7	77.1	115.7#
141	583.3	23.4	35.2#









#43
2,4,6-Trichlorophenol
Concen: 7.87 ng
RT: 12.281 min Scan# 1597
Delta R.T. -0.312 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:196	Resp:	3
Ion Ratio	Lower	Upper
196	100	
198	111.1	77.0 115.6
200	33.3	25.1 37.7



#44
2,4,5-Trichlorophenol
Concen: 53.58 ng
RT: 12.504 min Scan# 1635
Delta R.T. -0.182 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:196	Resp:	21
Ion Ratio	Lower	Upper
196	100	
198	0.0	77.3 115.9#
97	221.1	31.0 46.4#
132	121.1	20.9 31.3#









#45
2-Fluorobiphenyl
Concen: 10.56 ng
RT: 12.604 min Scan# 1652
Delta R.T. -0.170 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:172 Resp: 13
Ion Ratio Lower Upper
172 100
171 190.0 28.5 42.7#
170 60.0 19.2 28.8#



#46
1,1'-Biphenyl
Concen: 6.70 ng
RT: 12.822 min Scan# 1689
Delta R.T. -0.165 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:154 Resp: 9
Ion Ratio Lower Upper
154 100
153 0.0 20.4 60.4#
76 300.0 0.0 29.3#









#47
2-Chloronaphthalene
Concen: 10.79 ng
RT: 12.851 min Scan# 1694
Delta R.T. -0.170 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:162	Resp:	12
Ion	Ratio	Lower Upper
162	100	
127	78.9	29.9 44.9#
164	0.0	26.1 39.1#



#48
2-Nitroaniline
Concen: 67.53 ng
RT: 13.057 min Scan# 1729
Delta R.T. -0.188 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 65	Resp:	19
Ion	Ratio	Lower Upper
65	100	
92	72.7	70.6 106.0
138	0.0	133.6 200.4#









#49
Acenaphthylene
Concen: 11.77 ng
RT: 13.675 min Scan# 1834
Delta R.T. -0.200 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:152	Resp:	20
Ion Ratio	Lower	Upper
152	100	
151	161.3	15.9 23.9#
153	58.1	10.5 15.7#



#50
Dimethylphthalate
Concen: 6.96 ng
RT: 13.434 min Scan# 1793
Delta R.T. -0.194 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:163	Resp:	10
Ion Ratio	Lower	Upper
163	100	
194	156.3	3.3 4.9#
164	100.0	8.2 12.2#









#51
2,6-Dinitrotoluene
Concen: 116.85 ng
RT: 13.569 min Scan# 1816
Delta R.T. -0.176 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	165	Resp:	36
Ion	Ratio	Lower	Upper
165	100		
63	24.5	27.8	41.6#
89	55.1	34.6	52.0#



#52
Acenaphthene
Concen: 20.32 ng
RT: 14.034 min Scan# 1895
Delta R.T. -0.188 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	154	Resp:	21
Ion	Ratio	Lower	Upper
154	100		
153	95.0	91.0	136.6
152	180.0	43.0	64.6#









#53
3-Nitroaniline
Concen: 68.89 ng
RT: 13.887 min Scan# 1870
Delta R.T. -0.194 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:138 Resp: 23
Ion Ratio Lower Upper
138 100
108 193.3 7.5 11.3#
92 60.0 75.0 112.6#



#54
2,4-Dinitrophenol
Concen: 194.38 ng
RT: 14.128 min Scan# 1911
Delta R.T. -0.200 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:184 Resp: 30
Ion Ratio Lower Upper
184 100
63 189.5 32.3 48.5#
154 136.8 49.1 73.7#









#55
Dibenzenofuran
Concen: 6.06 ng
RT: 14.363 min Scan# 1951
Delta R.T. -0.200 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:168	Resp:	10
Ion	Ratio	Lower Upper
168	100	
139	50.0	29.1 43.7#
169	75.0	10.4 15.6#



#56
4-Nitrophenol
Concen: 205.83 ng
RT: 14.251 min Scan# 1932
Delta R.T. -0.200 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:139	Resp:	42
Ion	Ratio	Lower Upper
139	100	
109	78.3	44.6 84.6
65	221.7	41.6 81.6#









#57
2,4-Dinitrotoluene
Concen: 88.93 ng
RT: 14.351 min Scan# 1949
Delta R.T. -0.206 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	165	Resp:	38
Ion	Ratio	Lower	Upper
165	100		
63	27.8	20.6	31.0
89	24.1	42.6	63.8#
182	116.7	2.4	3.6#



#58
Fluorene
Concen: 23.79 ng
RT: 15.010 min Scan# 2061
Delta R.T. -0.206 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	166	Resp:	31
Ion	Ratio	Lower	Upper
166	100		
165	67.4	77.8	116.8#
167	118.6	11.0	16.4#









#59
2,3,4,6-Tetrachlorophenol
Concen: 46.13 ng
RT: 14.622 min Scan# 1995
Delta R.T. -0.188 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:232	Resp:	17
Ion	Ratio	Lower Upper
232	100	
131	0.0	34.1 51.1#
130	188.2	1.8 2.8#
166	188.2	24.6 36.8#



#60
Diethylphthalate
Concen: 10.23 ng
RT: 14.798 min Scan# 2025
Delta R.T. -0.206 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:149	Resp:	15
Ion	Ratio	Lower Upper
149	100	
177	0.0	18.0 27.0#
150	58.1	10.0 15.0#









#61
4-Chlorophenyl-phenylether
Concen: 9.84 ng
RT: 15.004 min Scan# 2060
Delta R.T. -0.212 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:204 Resp: 7
Ion Ratio Lower Upper
204 100
206 61.9 25.9 38.9#
141 42.9 46.5 69.7#



#62
4-Nitroaniline
Concen: 67.74 ng
RT: 15.151 min Scan# 2085
Delta R.T. -0.100 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:138 Resp: 24
Ion Ratio Lower Upper
138 100
92 45.0 20.7 60.7
108 70.0 47.8 87.8









#63
Azobenzene
Concen: 8.95 ng
RT: 15.322 min Scan# 2114
Delta R.T. -0.188 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion: 77	Resp: 10
Ion Ratio	Lower Upper
77 100	
182 222.2	16.9 56.9#
105 19.4	5.6 45.6
51 116.7	2.9 42.9#



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 16.692 min Scan# 2347
Delta R.T. -0.223 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:188	Resp:	18	
Ion Ratio	Lower	Upper	
188 100			
94 750.0	6.2	9.2#	
80 664.3	6.2	9.2#	









#65
4,6-Dinitro-2-methylphenol
Concen: 200.03 ng
RT: 15.128 min Scan# 2081
Delta R.T. -0.212 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:198 Resp: 19
Ion Ratio Lower Upper
198 100
51 312.5 6.4 46.4#
105 118.8 18.9 58.9#



#66
n-Nitrosodiphenylamine
Concen: 51.07 ng
RT: 15.222 min Scan# 2097
Delta R.T. -0.212 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:169 Resp: 27
Ion Ratio Lower Upper
169 100
168 27.6 53.8 80.6#
167 331.0 28.9 43.3#









#67
4-Bromophenyl-phenylether
Concen: 87.33 ng
RT: 15.898 min Scan# 2212
Delta R.T. -0.212 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	248	Resp:	19
Ion Ratio	100	Lower	Upper
248	100		
250	28.6	76.7	115.1#
141	67.9	53.0	79.6



#68
Hexachlorobenzene
Concen: 96.72 ng
RT: 16.022 min Scan# 2233
Delta R.T. -0.212 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	284	Resp:	25
Ion Ratio	100	Lower	Upper
284	100		
142	63.2	24.2	36.4#
249	89.5	23.8	35.8#









#69
Atrazine
Concen: 48.22 ng
RT: 16.375 min Scan# 2293
Delta R.T. -0.023 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:200 Resp: 10
Ion Ratio Lower Upper
200 100
173 40.0 5.9 45.9
215 40.0 28.4 68.4



#70
Pentachlorophenol
Concen: 857065.94 ng
RT: 16.586 min Scan# 2329
Delta R.T. -0.006 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:266 Resp: 103883
Ion Ratio Lower Upper
266 100
268 64.6 51.1 76.7
264 63.0 50.6 76.0









#71
Phenanthrene
Concen: 40.84 ng
RT: 16.710 min Scan# 2350
Delta R.T. -0.247 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:178 Resp: 40
Ion Ratio Lower Upper
178 100
176 94.1 15.3 22.9#
179 0.0 12.5 18.7#



#72
Anthracene
Concen: 25.90 ng
RT: 16.851 min Scan# 2374
Delta R.T. -0.200 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:178 Resp: 25
Ion Ratio Lower Upper
178 100
176 25.8 15.0 22.4#
179 22.6 12.5 18.7#





#73

Carbazole

Concen: 127.62 ng

RT: 17.069 min Scan# 2411

Delta R.T. -0.259 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18





Instrument :

BNA_P

ClientSampleId :

DFTPP

Tgt Ion:167	Resp:	117
Ion Ratio	Lower	Upper
167	100	
166	51.7	17.6 26.4#
139	4.5	10.2 15.4#



#74

Di-n-butylphthalate

Concen: 98.12 ng




RT: 17.680 min Scan# 2515


Delta R.T. -0.217 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18

Tgt Ion:149	Resp:	115
Ion Ratio	Lower	Upper
149	100	
150	0.0	7.3 10.9#
104	32.1	3.4 5.0#









#75
Fluoranthene
Concen: 661.03 ng
RT: 18.963 min Scan# 2733
Delta R.T. 0.018 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:202	Resp:	814
Ion Ratio	Lower	Upper
202	100	
101	17.4	0.0 29.4
203	13.7	0.0 37.8



#76
Chrysene-d12
Concen: 20.00 ng
RT: 20.757 min Scan# 3038
Delta R.T. -0.270 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:240	Resp:	23
Ion Ratio	Lower	Upper
240	100	
120	100.0	7.8 11.6#
236	0.0	20.8 31.2#





#77

Benzidine

Concen: 442836.21 ng

RT: 19.127 min Scan# 2761

Delta R.T. -0.006 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18





Instrument :

BNA_P

ClientSampleId :

DFTPP

Tgt Ion:184	Resp: 315832
Ion Ratio	Lower Upper
184	100
185	14.0 11.5 17.3
183	12.0 9.6 14.4



#78

Pyrene

Concen: 1061.78 ng




RT: 19.304 min Scan# 2791


Delta R.T. 0.006 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18

Tgt Ion:202	Resp: 1523
Ion Ratio	Lower Upper
202	100
200	18.9 17.1 25.7
203	18.7 14.2 21.2









#79
 Terphenyl-d14
 Concen: 114.14 ng
 RT: 19.310 min Scan# 2792
 Delta R.T. -0.194 min
 Lab File: BP003539.D
 Acq: 07 Oct 2020 03:18




Instrument :
 BNA_P
 ClientSampleId :
 DFTPP


Tgt Ion:244	Resp:	126
Ion Ratio	Lower	Upper
244	100	
212	121.7	5.8 8.6#
122	56.5	7.8 11.8#



#80
 Butylbenzylphthalate
 Concen: 778.69 ng
 RT: 19.904 min Scan# 2893
 Delta R.T. -0.276 min
 Lab File: BP003539.D
 Acq: 07 Oct 2020 03:18

Tgt Ion:149	Resp:	511
Ion Ratio	Lower	Upper
149	100	
91	45.4	45.4 68.0
206	0.0	17.1 25.7#









#81
Benzo(a)anthracene
Concen: 30.27 ng
RT: 20.857 min Scan# 3055
Delta R.T. -0.153 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	228	Resp:	45
Ion Ratio	Lower	Upper	
228	100		
226	26.9	22.1	33.1
229	67.2	15.9	23.9#



#82
3,3'-Dichlorobenzidine
Concen: 113.11 ng
RT: 20.692 min Scan# 3027
Delta R.T. -0.253 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	252	Resp:	65
Ion Ratio	Lower	Upper	
252	100		
254	24.0	51.0	76.4#
126	39.6	7.9	11.9#





#83

Chrysene

Concen: 1635.36 ng

RT: 21.027 min Scan# 3084

Delta R.T. -0.035 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18





Instrument :

BNA_P

ClientSampleId :

DFTPP

Tgt Ion:228	Resp:	2336
Ion Ratio	Lower	Upper
228	100	
226	24.7	24.4 36.6
229	14.2	15.8 23.6#



#84

Bis(2-ethylhexyl)phthalate

Concen: 1480.72 ng




RT: 20.816 min Scan# 3048


Delta R.T. -0.135 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18

Tgt Ion:149	Resp:	1368
Ion Ratio	Lower	Upper
149	100	
167	27.4	22.6 34.0
279	5.1	4.2 6.2









#85
Di-n-octyl phthalate
Concen: 59.24 ng
RT: 21.498 min Scan# 3164
Delta R.T. -0.300 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	149	Resp:	100
Ion Ratio	Lower	Upper	
149	100		
167	61.0	1.3	1.9#
43	144.0	3.7	5.5#



#86
Perylene-d12
Concen: 20.00 ng
RT: 22.874 min Scan# 3398
Delta R.T. -0.323 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion:	264	Resp:	128
Ion Ratio	Lower	Upper	
264	100		
260	13.6	18.8	28.2#
265	595.5	17.4	26.2#





#87

Indeno(1,2,3-cd)pyrene

Concen: 2.37 ng

RT: 25.045 min Scan# 3767





Delta R.T. -0.353 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18

Tgt Ion:	276	Resp:	23
Ion Ratio	Lower	Upper	
276	100		
138	0.0	18.7	28.1#
277	60.9	20.4	30.6#

Instrument :
BNA_P
ClientSampleId :
DFTPP



#88

Benzo(b)fluoranthene

Concen: 3.96 ng




RT: 22.204 min Scan# 3284


Delta R.T. -0.341 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18

Tgt Ion:	252	Resp:	32
Ion Ratio	Lower	Upper	
252	100		
253	146.7	17.6	26.4#
125	140.0	7.3	10.9#





#89

Benzo(k)fluoranthene

Concen: 10.71 ng

RT: 22.163 min Scan# 3277

Delta R.T. -0.429 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18





Instrument :

BNA_P

ClientSampleId :

DFTPP

Tgt Ion:252	Resp:	83
Ion Ratio	Lower	Upper
252	100	
253	118.5	17.8 26.8#
125	133.1	7.4 11.2#



#90

Benzo(a)pyrene

Concen: 489.64 ng




RT: 22.610 min Scan# 3353


Delta R.T. -0.494 min

Lab File: BP003539.D

Acq: 07 Oct 2020 03:18

Tgt Ion:252	Resp:	3732
Ion Ratio	Lower	Upper
252	100	
253	31.3	17.4 26.0#
125	20.9	8.2 12.4#









#91
Dibenzo(a,h)anthracene
Concen: 5.25 ng
RT: 25.068 min Scan# 3771
Delta R.T. -0.335 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Instrument :
BNA_P
ClientSampleId :
DFTPP

Tgt Ion: 278 Resp: 42
Ion Ratio Lower Upper
278 100
139 160.5 12.4 18.6#
279 121.1 19.0 28.6#



#92
Benzo(g,h,i)perylene
Concen: 5.51 ng
RT: 25.715 min Scan# 3881
Delta R.T. -0.353 min
Lab File: BP003539.D
Acq: 07 Oct 2020 03:18

Tgt Ion: 276 Resp: 44
Ion Ratio Lower Upper
276 100
277 32.8 19.0 28.6#
138 78.7 16.6 24.8#

