

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100620\
 Data File : BP003540.D
 Acq On : 07 Oct 2020 03:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 07 05:41:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 05 14:33:14 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	106	-0.01
2	1,4-Dioxane	40.000	40.700	-1.8	115	-0.01
3	Pyridine	40.000	42.615	-6.5	118	-0.01
4	n-Nitrosodimethylamine	40.000	45.650	-14.1	125	-0.01
5 S	2-Fluorophenol	80.000	84.514	-5.6	117	0.00
6	Aniline	40.000	45.074	-12.7	124	-0.01
7 S	Phenol-d6	80.000	90.738	-13.4	124	0.00
8	2-Chlorophenol	40.000	40.906	-2.3	114	-0.01
9	Benzaldehyde	40.000	41.853	-4.6	118	0.00
10 C	Phenol	40.000	45.501	-13.8	125	-0.01
11	bis(2-Chloroethyl)ether	40.000	44.823	-12.1	123	-0.01
12	1,3-Dichlorobenzene	40.000	38.997	2.5	107	-0.01
13 C	1,4-Dichlorobenzene	40.000	39.176	2.1	109	-0.01
14	1,2-Dichlorobenzene	40.000	38.860	2.9	108	0.00
15	Benzyl Alcohol	40.000	47.598	-19.0	127	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	48.949	-22.4	136	-0.01
17	2-Methylphenol	40.000	43.294	-8.2	119	0.00
18	Hexachloroethane	40.000	40.507	-1.3	112	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	46.451	-16.1	127	-0.01
20	3+4-Methylphenols	40.000	42.915	-7.3	116	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	109	-0.01
22	Acetophenone	40.000	40.719	-1.8	115	-0.01
23 S	Nitrobenzene-d5	80.000	91.940	-14.9	132	0.00
24	Nitrobenzene	40.000	46.873	-17.2	135	0.00
25	Isophorone	40.000	44.908	-12.3	128	0.00
26 C	2-Nitrophenol	40.000	40.109	-0.3	114	-0.01
27	2,4-Dimethylphenol	40.000	40.376	-0.9	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.232	-8.1	125	-0.01
29 C	2,4-Dichlorophenol	40.000	38.199	4.5	109	-0.01
30	1,2,4-Trichlorobenzene	40.000	36.576	8.6	106	-0.01
31	Naphthalene	40.000	39.354	1.6	114	0.00
32	Benzoic acid	40.000	27.178	32.1#	72	0.01
33	4-Chloroaniline	40.000	41.260	-3.1	117	-0.01
34 C	Hexachlorobutadiene	40.000	34.858	12.9	101	0.00
35	Caprolactam	40.000	42.034	-5.1	118	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.071	-5.2	120	0.00
37	2-Methylnaphthalene	40.000	38.664	3.3	113	-0.01
38	1-Methylnaphthalene	40.000	38.603	3.5	113	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.358	6.6	102	-0.01
41 P	Hexachlorocyclopentadiene	40.000	44.114	-10.3	127	-0.01
42 S	2,4,6-Tribromophenol	80.000	73.361	8.3	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.513	6.2	102	0.00
44	2,4,5-Trichlorophenol	40.000	37.615	6.0	101	-0.01
45 S	2-Fluorobiphenyl	80.000	76.816	4.0	107	-0.01
46	1,1'-Biphenyl	40.000	39.874	0.3	110	-0.01
47	2-Chloronaphthalene	40.000	39.509	1.2	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100620\
 Data File : BP003540.D
 Acq On : 07 Oct 2020 03:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 07 05:41:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 05 14:33:14 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	51.977	-29.9#	138	-0.01
49	Acenaphthylene	40.000	39.694	0.8	109	0.00
50	Dimethylphthalate	40.000	39.310	1.7	109	0.00
51	2,6-Dinitrotoluene	40.000	40.132	-0.3	109	0.00
52 C	Acenaphthene	40.000	39.923	0.2	110	0.00
53	3-Nitroaniline	40.000	43.211	-8.0	116	0.00
54 P	2,4-Dinitrophenol	40.000	47.246	-18.1	138	-0.01
55	Dibenzofuran	40.000	38.328	4.2	107	0.00
56 P	4-Nitrophenol	40.000	56.796	-42.0#	147	-0.02
57	2,4-Dinitrotoluene	40.000	40.753	-1.9	108	0.00
58	Fluorene	40.000	38.815	3.0	109	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.224	6.9	101	0.00
60	Diethylphthalate	40.000	39.886	0.3	110	-0.01
61	4-Chlorophenyl-phenylether	40.000	37.239	6.9	105	-0.01
62	4-Nitroaniline	40.000	42.064	-5.2	110	0.00
63	Azobenzene	40.000	47.711	-19.3	131	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.543	1.1	102	-0.01
66 c	n-Nitrosodiphenylamine	40.000	39.203	2.0	108	-0.01
67	4-Bromophenyl-phenylether	40.000	37.145	7.1	101	-0.01
68	Hexachlorobenzene	40.000	36.620	8.5	101	0.00
69	Atrazine	40.000	37.180	7.1	101	0.00
70 C	Pentachlorophenol	40.000	42.684	-6.7	122	0.00
71	Phenanthrene	40.000	39.749	0.6	109	0.00
72	Anthracene	40.000	39.589	1.0	107	-0.01
73	Carbazole	40.000	40.570	-1.4	110	0.00
74	Di-n-butylphthalate	40.000	38.983	2.5	107	-0.01
75 C	Fluoranthene	40.000	39.532	1.2	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	-0.01
77	Benzydine	40.000	33.781	15.5	85	0.00
78	Pyrene	40.000	40.936	-2.3	108	0.00
79 S	Terphenyl-d14	80.000	74.938	6.3	101	-0.01
80	Butylbenzylphthalate	40.000	41.647	-4.1	109	0.00
81	Benzo(a)anthracene	40.000	39.077	2.3	103	0.00
82	3,3'-Dichlorobenzidine	40.000	40.124	-0.3	105	-0.01
83	Chrysene	40.000	38.588	3.5	103	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	37.048	7.4	100	-0.01
85 c	Di-n-octyl phthalate	40.000	39.700	0.7	104	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	111	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	38.678	3.3	113	-0.01
88	Benzo(b)fluoranthene	40.000	39.138	2.2	114	-0.01
89	Benzo(k)fluoranthene	40.000	36.907	7.7	107	-0.01
90 C	Benzo(a)pyrene	40.000	38.865	2.8	114	-0.01
91	Dibenzo(a,h)anthracene	40.000	38.707	3.2	113	-0.02
92	Benzo(g,h,i)perylene	40.000	39.678	0.8	117	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100620\
Data File : BP003540.D
Acq On : 07 Oct 2020 03:52
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Oct 07 05:41:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 05 14:33:14 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0