

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102022\
 Data File : BP012235.D
 Acq On : 21 Oct 2022 09:13
 Operator : CG/JU
 Sample : PB148441BL
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Oct 21 22:29:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 20 00:51:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	387892	20.000	ng/ul	0.00
20) Naphthalene-d8	10.634	136	1695282	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.481	164	1072441	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.245	188	2146134	20.000	ng/ul	0.00
79) Chrysene-d12	21.333	240	1710044	20.000	ng/ul	0.00
88) Perylene-d12	23.745	264	1664301	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	57595	6.009	ng/uL	0.00
4) Pyridine-d5	3.681	84	795027	28.786	ng/ul	0.00
7) Phenol-d5	6.999	99	1020643	30.478	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.163	67	613387	30.203	ng/ul	0.00
11) 2-Chlorophenol-d4	7.357	132	804709	31.570	ng/ul	0.00
15) 4-Methylphenol-d8	8.546	113	802603	30.838	ng/ul	0.00
21) Nitrobenzene-d5	8.999	128	400147	31.888	ng/ul	0.00
24) 2-Nitrophenol-d4	9.722	143	447355	33.042	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.257	165	771639	30.385	ng/ul	0.00
31) 4-Chloroaniline-d4	10.781	131	1138469	31.552	ng/ul	0.00
46) Dimethylphthalate-d6	13.898	166	2339569	31.361	ng/ul	0.00
49) Acenaphthylene-d8	14.175	160	2666962	31.320	ng/ul	0.00
54) 4-Nitrophenol-d4	14.698	143	371656	26.592	ng/ul	0.00
60) Fluorene-d10	15.481	176	2010847	31.332	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.610	200	233209	19.101	ng/ul	0.00
73) Anthracene-d10	17.345	188	3070320	32.638	ng/ul	0.00
81) Pyrene-d10	19.586	212	3215044	33.022	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.592	264	2754747	33.006	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102022\
Data File : BP012235.D
Acq On : 21 Oct 2022 09:13
Operator : CG/JU
Sample : PB148441BL
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Oct 21 22:29:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101922.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 20 00:51:41 2022
Response via : Initial Calibration

