

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007639.D
 Acq On : 25 Oct 2021 12:44
 Operator : CG/JU
 Sample : SST00529
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005629

Quant Time: Oct 25 14:36:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 14:35:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	187574	20.000	ng/ul	0.00
20) Naphthalene-d8	10.746	136	736391	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.575	164	486403	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.334	188	1071768	20.000	ng/ul	0.00
79) Chrysene-d12	21.416	240	1196446	20.000	ng/ul	0.00
88) Perylene-d12	23.863	264	1279553	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.376	96	8798	1.853	ng/uL	0.00
4) Pyridine-d5	3.793	84	61165	4.705	ng/ul	0.00
7) Phenol-d5	7.105	99	66106	4.513	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	38984	4.462	ng/ul	0.00
11) 2-Chlorophenol-d4	7.475	132	52741	4.577	ng/ul	0.00
15) 4-Methylphenol-d8	8.652	113	50490	4.475	ng/ul	0.00
21) Nitrobenzene-d5	9.099	128	22891	4.912	ng/ul	0.00
24) 2-Nitrophenol-d4	9.822	143	23517	5.403	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	52027	4.900	ng/ul	0.00
31) 4-Chloroaniline-d4	10.881	131	70016	4.877	ng/ul	0.00
46) Dimethylphthalate-d6	13.987	166	161407	4.796	ng/ul	0.00
49) Acenaphthylene-d8	14.269	160	188748	4.463	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	24182	4.478	ng/ul	0.00
60) Fluorene-d10	15.569	176	139165	4.942	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.687	200	19569	4.557	ng/ul	0.00
73) Anthracene-d10	17.428	188	218109	4.792	ng/ul	0.00
81) Pyrene-d10	19.663	212	267410	4.471	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.704	264	306371	4.812	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	9614	1.861	ng/uL#	83
5) Pyridine	3.811	79	58629	4.663	ng/ul	96
6) Benzaldehyde	7.075	77	44158	5.365	ng/ul	94
8) Phenol	7.134	94	68535	4.576	ng/ul	94
10) Bis(2-Chloroethyl)ether	7.364	93	55762	4.613	ng/ul	91
12) 2-Chlorophenol	7.505	128	54289	4.589	ng/ul	97
13) 2-Methylphenol	8.387	108	49526	4.390	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.475	45	62436	4.053	ng/ul	97
16) Acetophenone	8.763	105	84634	4.903	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.752	70	39197	4.389	ng/ul	92
18) 4-Methylphenol	8.711	108	54058	4.499	ng/ul	97
19) Hexachloroethane	9.028	117	22651	4.573	ng/ul#	81
22) Nitrobenzene	9.140	77	55752	4.659	ng/ul#	93
23) Isophorone	9.675	82	106168	4.238	ng/ul	97
25) 2-Nitrophenol	9.852	139	25491	5.052	ng/ul	96
26) 2,4-Dimethylphenol	9.922	107	59236	4.598	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.158	93	71839	4.628	ng/ul	99
29) 2,4-Dichlorophenol	10.393	162	52305	5.012	ng/ul	97
30) Naphthalene	10.793	128	177995	4.970	ng/ul	99
32) 4-Chloroaniline	10.905	127	73535	5.062	ng/ul	98
33) Hexachlorobutadiene	11.093	225	40257	5.495	ng/ul	95
34) Caprolactam	11.669	113	14782	5.400	ng/ul	90
35) 4-Chloro-3-methylphenol	12.028	107	53185	4.692	ng/ul	97
36) 2-Methylnaphthalene	12.404	142	119619	4.887	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007639.D
 Acq On : 25 Oct 2021 12:44
 Operator : CG/JU
 Sample : SST00529
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005629

Quant Time: Oct 25 14:36:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 14:35:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.622	142	121609	4.908	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.775	216	71934	5.116	ng/ul	97
40) Hexachlorocyclopentadiene	12.757	237	34938	3.941	ng/ul	96
41) 2,4,6-Trichlorophenol	13.010	196	42211	5.106	ng/ul#	94
42) 2,4,5-Trichlorophenol	13.081	196	44689	5.192	ng/ul	98
43) 1,1'-Biphenyl	13.410	154	168765	4.868	ng/ul	99
44) 2-Chloronaphthalene	13.451	162	130033	4.870	ng/ul	97
45) 2-Nitroaniline	13.651	65	27871	4.489	ng/ul	92
47) Dimethylphthalate	14.034	163	160279	4.795	ng/ul	99
48) 2,6-Dinitrotoluene	14.146	165	25872	4.668	ng/ul	95
50) Acenaphthylene	14.298	152	200242	4.696	ng/ul	98
51) 3-Nitroaniline	14.475	138	27294	4.419	ng/ul	99
52) Acenaphthene	14.640	153	137730	4.988	ng/ul	97
53) 2,4-Dinitrophenol	14.681	184	11804	4.610	ng/ul	96
55) 4-Nitrophenol	14.781	109	19556	4.009	ng/ul	95
56) Dibenzofuran	14.975	168	202802	5.059	ng/ul	98
57) 2,4-Dinitrotoluene	14.934	165	37115	4.767	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.204	232	36744	5.144	ng/ul	91
59) Diethylphthalate	15.398	149	157290	4.720	ng/ul	97
61) Fluorene	15.628	166	158686	5.224	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.622	204	86151	5.570	ng/ul	98
63) 4-Nitroaniline	15.634	138	27157	4.960	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.698	198	19397	4.366	ng/ul#	86
67) N-Nitrosodiphenylamine	15.834	169	137677	4.771	ng/ul	98
68) 4-Bromophenyl-phenylether	16.522	248	51056	5.288	ng/ul	92
69) Hexachlorobenzene	16.640	284	59527	5.126	ng/ul	95
70) Atrazine	16.792	200	50022	4.579	ng/ul	96
71) Pentachlorophenol	16.981	266	25769	4.008	ng/ul	97
72) Phenanthrene	17.375	178	263033	4.909	ng/ul	98
74) Anthracene	17.463	178	259909	4.850	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.375	216	75122	4.832	ng/uL	99
76) Pentachlorobenzene	14.898	250	74516	5.142	ng/uL	97
77) Carbazole	17.734	167	233344	4.957	ng/ul	99
78) Di-n-butylphthalate	18.292	149	255232	4.414	ng/ul	99
80) Fluoranthene	19.339	202	326312	4.397	ng/ul	99
82) Pyrene	19.692	202	344977	4.610	ng/ul	97
83) Butylbenzylphthalate	20.569	149	117700	4.284	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.333	252	82205	3.611	ng/ul#	98
85) Benzo(a)anthracene	21.404	228	355408	4.835	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.333	149	181622	4.324	ng/ul	100
87) Chrysene	21.457	228	348410	4.893	ng/ul	98
89) Di-n-octyl phthalate	22.280	149	304717	4.220	ng/ul	100
90) Benzo(b)fluoranthene	23.116	252	365882	4.698	ng/ul	100
91) Benzo(k)fluoranthene	23.169	252	362207	4.988	ng/ul	98
93) Benzo(a)pyrene	23.757	252	361533	4.853	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.415	276	421482	4.884	ng/ul	97
95) Dibenzo(a,h)anthracene	26.439	278	368123	5.009	ng/ul	97
96) Benzo(g,h,i)perylene	27.198	276	353842	4.761	ng/ul	95





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 1.861 ng/uL
RT: 3.405 min Scan# 20
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:	88	Resp:	9614
Ion	Ratio	Lower	Upper
88	100		
43	31.5	24.2	36.2
58	57.8	62.5	93.7#



#5
Pyridine
Concen: 4.663 ng/uL
RT: 3.811 min Scan# 89
Delta R.T. 0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:	79	Resp:	58629
Ion	Ratio	Lower	Upper
79	100		
52	73.4	57.0	85.4
51	38.2	27.0	40.6









#6
Benzaldehyde
Concen: 5.365 ng/ul
RT: 7.075 min Scan# 644
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion: 77 Resp: 44158
Ion Ratio Lower Upper
77 100
105 104.9 78.2 117.4
106 98.1 74.6 112.0



#8
Phenol
Concen: 4.576 ng/ul
RT: 7.134 min Scan# 654
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion: 94 Resp: 68535
Ion Ratio Lower Upper
94 100
65 22.9 21.9 32.9
66 35.4 29.9 44.9





#10

Bis(2-Chloroethyl)ether

Concen: 4.613 ng/ul

RT: 7.364 min Scan# 693

Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44





Instrument :

BNA_P

ClientSampleId :

SSTD005629

Tgt Ion:	93	Resp:	55762
Ion	Ratio	Lower	Upper
93	100		
63	64.3	59.0	88.4
95	30.5	26.4	39.6



#12

2-Chlorophenol

Concen: 4.589 ng/ul




RT: 7.505 min Scan# 717


Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Tgt Ion:	128	Resp:	54289
Ion	Ratio	Lower	Upper
128	100		
64	43.4	36.9	55.3
130	32.5	25.2	37.8





#13

2-Methylphenol

Concen: 4.390 ng/ul

RT: 8.387 min Scan# 867

Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Instrument :

BNA_P

ClientSampleId :





SSTD005629

Tgt Ion:108 Resp: 49526

Ion Ratio Lower Upper

108 100

107 86.2 72.2 108.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 4.053 ng/ul

RT: 8.475 min Scan# 882

Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44




Tgt Ion: 45 Resp: 62436


Ion Ratio Lower Upper

45 100

77 17.0 12.3 18.5

79 12.8 9.4 14.0









#16
Acetophenone
Concen: 4.903 ng/ul
RT: 8.763 min Scan# 931
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion	Ratio	Lower	Upper
105	100		
77	74.5	61.8	92.6
51	24.1	20.8	31.2



#17
N-Nitroso-di-n-propylamine
Concen: 4.389 ng/ul
RT: 8.752 min Scan# 929
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion	Ratio	Lower	Upper
70	100		
42	49.1	46.3	69.5
101	9.5	7.4	11.0
130	19.7	16.2	24.2









#18
4-Methylphenol
Concen: 4.499 ng/ul
RT: 8.711 min Scan# 922
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:108 Resp: 54058
Ion Ratio Lower Upper
108 100
107 110.1 90.9 136.3



#19
Hexachloroethane
Concen: 4.573 ng/ul
RT: 9.028 min Scan# 976
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:117 Resp: 22651
Ion Ratio Lower Upper
117 100
201 120.6 82.3 123.5
199 81.0 51.8 77.6#





#22

Nitrobenzene

Concen: 4.659 ng/ul

RT: 9.140 min Scan# 995

Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44





Instrument :

BNA_P

ClientSampleId :

SSTD005629

Tgt Ion:	77	Resp:	55752
Ion	Ratio	Lower	Upper
77	100		
123	48.0	34.9	52.3
65	15.9	10.5	15.7#



#23

Isophorone

Concen: 4.238 ng/ul




RT: 9.675 min Scan# 1086


Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Tgt Ion:	82	Resp:	106168
Ion	Ratio	Lower	Upper
82	100		
95	7.8	5.6	8.4
138	18.2	13.4	20.0









#25
2-Nitrophenol
Concen: 5.052 ng/ul
RT: 9.852 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:139	Resp:	25491
Ion Ratio	Lower	Upper
139	100	
65	49.8	40.3 60.5
109	34.0	23.7 35.5



#26
2,4-Dimethylphenol
Concen: 4.598 ng/ul
RT: 9.922 min Scan# 1128
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:107	Resp:	59236
Ion Ratio	Lower	Upper
107	100	
121	51.4	38.8 58.2
122	83.6	64.0 96.0









#27
Bis(2-Chloroethoxy)methane
Concen: 4.628 ng/ul
RT: 10.158 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion: 93	Resp: 71839
Ion Ratio	Lower Upper
93 100	
95 33.0	26.2 39.4
123 11.8	8.3 12.5



#29
2,4-Dichlorophenol
Concen: 5.012 ng/ul
RT: 10.393 min Scan# 1208
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt	Ion:162	Resp:	52305
Ion	Ratio	Lower	Upper
162	100		
164	64.8	52.9	79.3
98	35.5	30.6	45.8









#30
Naphthalene
Concen: 4.970 ng/ul
RT: 10.793 min Scan# 1276
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:128 Resp: 177995
Ion Ratio Lower Upper
128 100
129 10.8 8.8 13.2
127 12.7 10.5 15.7



#32
4-Chloroaniline
Concen: 5.062 ng/ul
RT: 10.905 min Scan# 1295
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:127 Resp: 73535
Ion Ratio Lower Upper
127 100
129 33.2 25.8 38.6









#33
Hexachlorobutadiene
Concen: 5.495 ng/ul
RT: 11.093 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:225 Resp: 40257
Ion Ratio Lower Upper
225 100
223 60.7 50.6 76.0
227 59.0 51.0 76.6



#34
Caprolactam
Concen: 5.400 ng/ul
RT: 11.669 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:113 Resp: 14782
Ion Ratio Lower Upper
113 100
55 145.5 130.1 195.1
56 112.1 96.2 144.4









#35
4-Chloro-3-methylphenol
Concen: 4.692 ng/ul
RT: 12.028 min Scan# 1486
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:107 Resp: 53185
Ion Ratio Lower Upper
107 100
144 24.5 19.3 28.9
142 74.6 62.2 93.4



#36
2-Methylnaphthalene
Concen: 4.887 ng/ul
RT: 12.404 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:142 Resp: 119619
Ion Ratio Lower Upper
142 100
141 89.7 71.8 107.6









#37
1-Methylnaphthalene
Concen: 4.908 ng/ul
RT: 12.622 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:142 Resp: 121609
Ion Ratio Lower Upper
142 100
141 91.9 73.5 110.3
116 4.5 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 5.116 ng/ul
RT: 12.775 min Scan# 1613
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:216 Resp: 71934
Ion Ratio Lower Upper
216 100
214 77.2 63.1 94.7
179 20.7 17.4 26.0
108 15.9 15.0 22.6









#40
Hexachlorocyclopentadiene
Concen: 3.941 ng/ul
RT: 12.757 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:237 Resp: 34938
Ion Ratio Lower Upper
237 100
235 58.9 49.4 74.2
272 10.7 9.8 14.6



#41
2,4,6-Trichlorophenol
Concen: 5.106 ng/ul
RT: 13.010 min Scan# 1653
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:196 Resp: 42211
Ion Ratio Lower Upper
196 100
198 99.2 75.8 113.6
200 34.8 23.0 34.4#





#42

2,4,5-Trichlorophenol

Concen: 5.192 ng/ul

RT: 13.081 min Scan# 1665

Delta R.T. -0.006 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Instrument :

BNA_P

ClientSampleId :

SSTD005629





Tgt Ion:196 Resp: 44689

Ion Ratio Lower Upper

196 100

198 93.9 75.1 112.7

200 28.4 25.8 38.8



#43

1,1'-Biphenyl

Concen: 4.868 ng/ul

RT: 13.410 min Scan# 1721

Delta R.T. -0.006 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44




Tgt Ion:154 Resp: 168765


Ion Ratio Lower Upper

154 100

153 40.3 32.0 48.0

76 13.2 10.3 15.5









#44
2-Chloronaphthalene
Concen: 4.870 ng/ul
RT: 13.451 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:162 Resp: 130033
Ion Ratio Lower Upper
162 100
164 31.7 26.5 39.7
127 37.6 31.8 47.6



#45
2-Nitroaniline
Concen: 4.489 ng/ul
RT: 13.651 min Scan# 1762
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion: 65 Resp: 27871
Ion Ratio Lower Upper
65 100
92 73.3 55.9 83.9
138 115.0 83.4 125.2









#47
Dimethylphthalate
Concen: 4.795 ng/ul
RT: 14.034 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:163 Resp: 160279
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 9.7 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 4.668 ng/ul
RT: 14.146 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:165 Resp: 25872
Ion Ratio Lower Upper
165 100
89 56.2 49.0 73.4
121 21.9 18.4 27.6









#50
Acenaphthylene
Concen: 4.696 ng/ul
RT: 14.298 min Scan# 1872
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:152	Resp: 200242
Ion Ratio	Lower Upper
152	100
151	19.3 15.8 23.8
153	14.2 10.6 15.8



#51
3-Nitroaniline
Concen: 4.419 ng/ul
RT: 14.475 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:138	Resp: 27294
Ion Ratio	Lower Upper
138	100
108	16.2 13.0 19.4
92	120.8 95.7 143.5









#52
Acenaphthene
Concen: 4.988 ng/ul
RT: 14.640 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:153 Resp: 137730
Ion Ratio Lower Upper
153 100
152 46.4 37.9 56.9
154 84.2 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 4.610 ng/ul
RT: 14.681 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:184 Resp: 11804
Ion Ratio Lower Upper
184 100
63 58.3 48.8 73.2
154 66.7 56.2 84.4









#55
4-Nitrophenol
Concen: 4.009 ng/ul
RT: 14.781 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:109 Resp: 19556
Ion Ratio Lower Upper
109 100
139 129.2 100.2 150.2
65 131.7 98.8 148.2



#56
Dibenzofuran
Concen: 5.059 ng/ul
RT: 14.975 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:168 Resp: 202802
Ion Ratio Lower Upper
168 100
139 38.6 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 4.767 ng/ul
RT: 14.934 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:	165	Resp:	37115
Ion Ratio	Lower	Upper	
165	100		
63	39.7	35.0	52.6
182	2.4	2.3	3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 5.144 ng/ul
RT: 15.204 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:	232	Resp:	36744
Ion Ratio	Lower	Upper	
232	100		
131	37.8	35.1	52.7
130	2.5	2.0	3.0
166	26.7	25.8	38.8









#59
Diethylphthalate
Concen: 4.720 ng/ul
RT: 15.398 min Scan# 2059
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:149 Resp: 157290
Ion Ratio Lower Upper
149 100
177 23.8 17.4 26.2
150 13.2 10.0 15.0



#61
Fluorene
Concen: 5.224 ng/ul
RT: 15.628 min Scan# 2098
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:166 Resp: 158686
Ion Ratio Lower Upper
166 100
165 97.7 78.4 117.6
167 13.8 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 5.570 ng/ul
RT: 15.622 min Scan# 2097
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:204 Resp: 86151
Ion Ratio Lower Upper
204 100
206 31.8 26.5 39.7
141 57.9 47.4 71.0



#63
4-Nitroaniline
Concen: 4.960 ng/ul
RT: 15.634 min Scan# 2099
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:138 Resp: 27157
Ion Ratio Lower Upper
138 100
92 54.1 42.1 63.1
108 82.6 70.3 105.5









#66
4,6-Dinitro-2-methylphenol
Concen: 4.366 ng/ul
RT: 15.698 min Scan# 2110
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:198 Resp: 19397
Ion Ratio Lower Upper
198 100
182 7.1 3.3 4.9#
77 30.2 18.4 27.6#



#67
N-Nitrosodiphenylamine
Concen: 4.771 ng/ul
RT: 15.834 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:169 Resp: 137677
Ion Ratio Lower Upper
169 100
168 62.9 52.5 78.7
167 35.3 28.4 42.6









#68
4-Bromophenyl-phenylether
Concen: 5.288 ng/ul
RT: 16.522 min Scan# 2250
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:	248	Resp:	51056
Ion Ratio	Lower	Upper	
248	100		
250	95.2	78.9	118.3
141	68.6	64.3	96.5



#69
Hexachlorobenzene
Concen: 5.126 ng/ul
RT: 16.640 min Scan# 2270
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:	284	Resp:	59527
Ion Ratio	Lower	Upper	
284	100		
142	32.4	29.8	44.8
249	34.0	26.6	39.8





#70

Atrazine

Concen: 4.579 ng/ul

RT: 16.792 min Scan# 2296

Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44





Instrument :

BNA_P

ClientSampleId :

SSTD005629

Tgt Ion:200	Resp:	50022
Ion Ratio	Lower	Upper
200	100	
173	27.2	20.9 31.3
215	50.7	38.1 57.1



#71

Pentachlorophenol

Concen: 4.008 ng/ul




RT: 16.981 min Scan# 2328


Delta R.T. -0.006 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Tgt Ion:266	Resp:	25769
Ion Ratio	Lower	Upper
266	100	
264	59.1	50.6 75.8
268	63.0	51.3 76.9









#72
Phenanthrene
Concen: 4.909 ng/ul
RT: 17.375 min Scan# 2395
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:178 Resp: 263033
Ion Ratio Lower Upper
178 100
179 16.0 12.1 18.1
176 19.6 15.2 22.8



#74
Anthracene
Concen: 4.850 ng/ul
RT: 17.463 min Scan# 2410
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:178 Resp: 259909
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.4
176 18.5 14.6 22.0









#75

1,2,3,4-Tetrachlorobenzene
Concen: 4.832 ng/uL
RT: 13.375 min Scan# 1715
Delta R.T. -0.006 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Instrument :
BNA_P
ClientSampleId :
SSTD005629




Tgt Ion:216 Resp: 75122
Ion Ratio Lower Upper
216 100
214 77.2 62.0 93.0
218 48.5 37.6 56.4




#76

Pentachlorobenzene
Concen: 5.142 ng/uL
RT: 14.898 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:250 Resp: 74516
Ion Ratio Lower Upper
250 100
248 64.1 50.1 75.1
252 65.9 50.6 76.0









#77
 Carbazole
 Concen: 4.957 ng/ul
 RT: 17.734 min Scan# 2456
 Delta R.T. 0.000 min
 Lab File: BP007639.D
 Acq: 25 Oct 2021 12:44




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005629


Tgt Ion:167	Resp:	233344
Ion Ratio	Lower	Upper
167	100	
166	20.9	17.3 25.9
139	12.4	10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 4.414 ng/ul
 RT: 18.292 min Scan# 2551
 Delta R.T. 0.000 min
 Lab File: BP007639.D
 Acq: 25 Oct 2021 12:44

Tgt Ion:149	Resp:	255232
Ion Ratio	Lower	Upper
149	100	
150	8.9	7.3 10.9
104	5.1	4.2 6.4









#80
Fluoranthene
Concen: 4.397 ng/ul
RT: 19.339 min Scan# 2729
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:202 Resp: 326312
Ion Ratio Lower Upper
202 100
101 10.1 8.4 12.6
100 7.7 6.3 9.5



#82
Pyrene
Concen: 4.610 ng/ul
RT: 19.692 min Scan# 2789
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:202 Resp: 344977
Ion Ratio Lower Upper
202 100
101 10.4 9.9 14.9
100 9.4 7.9 11.9









#83
Butylbenzylphthalate
Concen: 4.284 ng/ul
RT: 20.569 min Scan# 2938
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44




Instrument :
BNA_P
ClientSampleId :
SSTD005629


Tgt Ion:149 Resp: 117700
Ion Ratio Lower Upper
149 100
91 67.6 55.9 83.9
206 23.2 17.4 26.0



#84
3,3'-Dichlorobenzidine
Concen: 3.611 ng/ul
RT: 21.333 min Scan# 3068
Delta R.T. 0.000 min
Lab File: BP007639.D
Acq: 25 Oct 2021 12:44

Tgt Ion:252 Resp: 82205
Ion Ratio Lower Upper
252 100
254 63.3 51.1 76.7
126 9.9 10.0 15.0#





#85

Benzo(a)anthracene

Concen: 4.835 ng/ul

RT: 21.404 min Scan# 3080

Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Tgt Ion:228 Resp: 355408

Ion Ratio Lower Upper

228 100

229 18.9 15.7 23.5





226 25.9 21.4 32.0

Instrument :

BNA_P

ClientSampleId :

SSTD005629



#86

Bis(2-ethylhexyl)phthalate

Concen: 4.324 ng/ul

RT: 21.333 min Scan# 3068

Delta R.T. -0.006 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44




Tgt Ion:149 Resp: 181622


Ion Ratio Lower Upper

149 100

167 28.2 22.6 33.8

279 4.2 3.3 4.9





#87

Chrysene

Concen: 4.893 ng/ul

RT: 21.457 min Scan# 3089

Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Instrument :

BNA_P

ClientSampleId :

SSTD005629





Tgt Ion:228 Resp: 348410

Ion Ratio Lower Upper

228 100

226 28.6 23.7 35.5

229 19.1 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 4.220 ng/ul




RT: 22.280 min Scan# 3229


Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Tgt Ion:149 Resp: 304717





#90

Benzo(b)fluoranthene

Concen: 4.698 ng/ul

RT: 23.116 min Scan# 3371

Delta R.T. -0.006 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Instrument :

BNA_P

ClientSampleId :

SSTD005629





Tgt Ion:252 Resp: 365882

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 9.2 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 4.988 ng/ul

RT: 23.169 min Scan# 3380

Delta R.T. 0.000 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44




Tgt Ion:252 Resp: 362207


Ion Ratio Lower Upper

252 100

253 21.7 17.6 26.4

125 8.7 8.0 12.0





#93

Benzo(a)pyrene

Concen: 4.853 ng/ul

RT: 23.757 min Scan# 3480

Delta R.T. -0.006 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Instrument :

BNA_P

ClientSampleId :

SSTD005629





Tgt Ion:252 Resp: 361533

Ion Ratio Lower Upper

252 100

253 21.7 17.3 25.9

125 9.8 8.3 12.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.884 ng/ul

RT: 26.415 min Scan# 3932

Delta R.T. -0.012 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44




Tgt Ion:276 Resp: 421482


Ion Ratio Lower Upper

276 100

138 20.5 17.8 26.8

277 23.9 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 5.009 ng/ul

RT: 26.439 min Scan# 3936

Delta R.T. -0.006 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Instrument :

BNA_P

ClientSampleId :

SSTD005629





Tgt Ion:278 Resp: 368123

Ion Ratio Lower Upper

278 100

139 13.8 12.8 19.2

279 23.3 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 4.761 ng/ul

RT: 27.198 min Scan# 4065

Delta R.T. -0.012 min

Lab File: BP007639.D

Acq: 25 Oct 2021 12:44

Tgt Ion:276 Resp: 353842

Ion Ratio Lower Upper

276 100

138 20.2 17.5 26.3

277 25.7 18.5 27.7

