

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007644.D
 Acq On : 25 Oct 2021 16:09
 Operator : CG/JU
 Sample : SSTD16034
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160634

Quant Time: Oct 25 16:34:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 14:35:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	162808	20.000	ng/ul	0.00
20) Naphthalene-d8	10.751	136	628789	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.581	164	388535	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.339	188	829761	20.000	ng/ul	0.00
79) Chrysene-d12	21.427	240	831339	20.000	ng/ul	0.00
88) Perylene-d12	23.874	264	940479	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	262410	63.658	ng/uL	0.00
4) Pyridine-d5	3.787	84	1901542	168.518	ng/ul	0.00
7) Phenol-d5	7.122	99	2202114	173.198	ng/ul	0.02
9) Bis-(2-Chloroethyl)eth...	7.275	67	1174378	154.862	ng/ul	0.00
11) 2-Chlorophenol-d4	7.481	132	1792265	179.186	ng/ul	0.00
15) 4-Methylphenol-d8	8.669	113	1708558	174.456	ng/ul	0.02
21) Nitrobenzene-d5	9.110	128	821537	206.455	ng/ul	0.01
24) 2-Nitrophenol-d4	9.834	143	957004	257.474	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	1794448	197.935	ng/ul	0.01
31) 4-Chloroaniline-d4	10.893	131	2183740	178.125	ng/ul	0.01
46) Dimethylphthalate-d6	13.998	166	4699137	174.816	ng/ul	0.00
49) Acenaphthylene-d8	14.281	160	5761476	170.551	ng/ul	0.01
54) 4-Nitrophenol-d4	14.804	143	884971	205.178	ng/ul	0.03
60) Fluorene-d10	15.581	176	3869915	172.034	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.710	200	907418	272.924	ng/ul	0.02
73) Anthracene-d10	17.439	188	5957383	169.058	ng/ul	0.00
81) Pyrene-d10	19.674	212	6904955	166.144	ng/ul	0.01
92) Benzo(a)pyrene-d12	23.733	264	8435816	180.266	ng/ul	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	290747	64.841	ng/uL	91
5) Pyridine	3.805	79	1846751	169.219	ng/ul	99
6) Benzaldehyde	7.081	77	592088	82.882	ng/ul	95
8) Phenol	7.152	94	2185123	168.093	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.369	93	1678156	159.959	ng/ul	98
12) 2-Chlorophenol	7.516	128	1764921	171.873	ng/ul	96
13) 2-Methylphenol	8.399	108	1660723	169.600	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.481	45	1899400	142.071	ng/ul	97
16) Acetophenone	8.775	105	2390695	159.576	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.787	70	1154766	148.986	ng/ul	93
18) 4-Methylphenol	8.740	108	1727334	165.619	ng/ul	98
19) Hexachloroethane	9.028	117	723480	168.277	ng/ul	91
22) Nitrobenzene	9.157	77	1833070	179.415	ng/ul	94
23) Isophorone	9.699	82	3524908	164.801	ng/ul	96
25) 2-Nitrophenol	9.869	139	988099	229.341	ng/ul#	92
26) 2,4-Dimethylphenol	9.934	107	1924323	174.932	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.169	93	2130555	160.756	ng/ul	99
29) 2,4-Dichlorophenol	10.404	162	1699035	190.672	ng/ul	97
30) Naphthalene	10.804	128	5051263	165.180	ng/ul	99
32) 4-Chloroaniline	10.916	127	2186768	176.278	ng/ul	100
33) Hexachlorobutadiene	11.098	225	1244701	198.959	ng/ul	99
34) Caprolactam	11.728	113	312528	133.706	ng/ul	93
35) 4-Chloro-3-methylphenol	12.051	107	1777504	183.628	ng/ul	98
36) 2-Methylnaphthalene	12.410	142	3500987	167.510	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007644.D
 Acq On : 25 Oct 2021 16:09
 Operator : CG/JU
 Sample : SSTD16034
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160634

Quant Time: Oct 25 16:34:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 14:35:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.634	142	3505327	165.675	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.781	216	2141140	190.650	ng/ul	99
40) Hexachlorocyclopentadiene	12.763	237	1546214	218.342	ng/ul	98
41) 2,4,6-Trichlorophenol	13.022	196	1435138	217.315	ng/ul	98
42) 2,4,5-Trichlorophenol	13.092	196	1551427	225.650	ng/ul	98
43) 1,1'-Biphenyl	13.422	154	4604659	166.276	ng/ul	97
44) 2-Chloronaphthalene	13.463	162	3634381	170.389	ng/ul	100
45) 2-Nitroaniline	13.669	65	1063777	214.511	ng/ul	92
47) Dimethylphthalate	14.051	163	4555436	170.618	ng/ul	99
48) 2,6-Dinitrotoluene	14.169	165	1030457	232.738	ng/ul	92
50) Acenaphthylene	14.310	152	5591606	164.171	ng/ul	99
51) 3-Nitroaniline	14.492	138	899859	182.400	ng/ul#	94
52) Acenaphthene	14.651	153	3685619	167.108	ng/ul	99
53) 2,4-Dinitrophenol	14.698	184	690366	337.542	ng/ul	96
55) 4-Nitrophenol	14.816	109	800154	205.345	ng/ul	87
56) Dibenzofuran	14.987	168	5291959	165.248	ng/ul	97
57) 2,4-Dinitrotoluene	14.957	165	1436817	231.046	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.216	232	1299343	227.742	ng/ul	95
59) Diethylphthalate	15.416	149	4521963	169.860	ng/ul	98
61) Fluorene	15.639	166	4020830	165.713	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.628	204	2182145	176.630	ng/ul	100
63) 4-Nitroaniline	15.669	138	763506	174.573	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.722	198	884493	257.176	ng/ul	98
67) N-Nitrosodiphenylamine	15.845	169	3711209	166.128	ng/ul	98
68) 4-Bromophenyl-phenylether	16.528	248	1499313	200.588	ng/ul	94
69) Hexachlorobenzene	16.645	284	1735673	193.045	ng/ul	95
70) Atrazine	16.810	200	1589000	187.890	ng/ul	98
71) Pentachlorophenol	16.986	266	1165143	234.100	ng/ul	100
72) Phenanthrene	17.386	178	6819393	164.397	ng/ul	98
74) Anthracene	17.480	178	6768039	163.133	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.387	216	2272101	188.790	ng/uL	98
76) Pentachlorobenzene	14.910	250	2134304	190.244	ng/uL	99
77) Carbazole	17.745	167	6199384	170.111	ng/ul	98
78) Di-n-butylphthalate	18.298	149	7429072	165.940	ng/ul	99
80) Fluoranthene	19.351	202	8378485	162.476	ng/ul#	94
82) Pyrene	19.704	202	8203552	157.759	ng/ul#	93
83) Butylbenzylphthalate	20.574	149	3558632	186.396	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.339	252	2347107	148.389	ng/ul#	97
85) Benzo(a)anthracene	21.416	228	8601633	168.399	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.339	149	4863434	166.622	ng/ul#	98
87) Chrysene	21.469	228	8140908	164.524	ng/ul	97
89) Di-n-octyl phthalate	22.286	149	9049308	170.521	ng/ul	100
90) Benzo(b)fluoranthene	23.139	252	10065330	175.841	ng/ul#	97
91) Benzo(k)fluoranthene	23.192	252	8728843	163.543	ng/ul#	97
93) Benzo(a)pyrene	23.786	252	9373959	171.202	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.462	276	11973874	188.772	ng/ul	96
95) Dibenzo(a,h)anthracene	26.492	278	9962685	184.451	ng/ul#	96
96) Benzo(g,h,i)perylene	27.262	276	10362011	189.681	ng/ul	94





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 64.841 ng/uL
RT: 3.405 min Scan# 20
Delta R.T. -0.000 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion: 88 Resp: 290747
Ion Ratio Lower Upper
88 100
43 25.4 24.2 36.2
58 70.7 62.5 93.7



#5
Pyridine
Concen: 169.219 ng/uL
RT: 3.805 min Scan# 88
Delta R.T. -0.000 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion: 79 Resp: 1846751
Ion Ratio Lower Upper
79 100
52 72.1 57.0 85.4
51 34.0 27.0 40.6









#6
Benzaldehyde
Concen: 82.882 ng/ul
RT: 7.081 min Scan# 645
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion: 77 Resp: 592088
Ion Ratio Lower Upper
77 100
105 101.7 78.2 117.4
106 98.7 74.6 112.0



#8
Phenol
Concen: 168.093 ng/ul
RT: 7.152 min Scan# 657
Delta R.T. 0.018 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion: 94 Resp: 2185123
Ion Ratio Lower Upper
94 100
65 27.7 21.9 32.9
66 40.1 29.9 44.9





#10

Bis(2-Chloroethyl)ether

Concen: 159.959 ng/ul

RT: 7.369 min Scan# 694

Delta R.T. 0.006 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD160634





Tgt Ion: 93 Resp: 1678156

Ion Ratio Lower Upper

93 100

63 71.3 59.0 88.4

95 32.3 26.4 39.6



#12

2-Chlorophenol

Concen: 171.873 ng/ul

RT: 7.516 min Scan# 719

Delta R.T. 0.012 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09




Tgt Ion:128 Resp: 1764921


Ion Ratio Lower Upper

128 100

64 43.2 36.9 55.3

130 32.8 25.2 37.8





#13

2-Methylphenol

Concen: 169.600 ng/ul

RT: 8.399 min Scan# 869

Delta R.T. 0.012 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Instrument :

BNA_P

ClientSampleId :





SSTD160634

Tgt Ion:108 Resp: 1660723

Ion Ratio Lower Upper

108 100

107 89.8 72.2 108.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 142.071 ng/ul

RT: 8.481 min Scan# 883

Delta R.T. 0.006 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09




Tgt Ion: 45 Resp: 1899400


Ion Ratio Lower Upper

45 100

77 16.8 12.3 18.5

79 12.7 9.4 14.0









#16
Acetophenone
Concen: 159.576 ng/ul
RT: 8.775 min Scan# 933
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:105 Resp: 2390695
Ion Ratio Lower Upper
105 100
77 75.0 61.8 92.6
51 24.4 20.8 31.2



#17
N-Nitroso-di-n-propylamine
Concen: 148.986 ng/ul
RT: 8.787 min Scan# 935
Delta R.T. 0.035 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion: 70 Resp: 1154766
Ion Ratio Lower Upper
70 100
42 52.8 46.3 69.5
101 10.4 7.4 11.0
130 23.5 16.2 24.2





#18

4-Methylphenol

Concen: 165.619 ng/ul

RT: 8.740 min Scan# 927

Delta R.T. 0.023 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Instrument :

BNA_P

ClientSampleId :





SSTD160634

Tgt Ion:108 Resp: 1727334

Ion Ratio Lower Upper

108 100

107 111.5 90.9 136.3



#19

Hexachloroethane

Concen: 168.277 ng/ul

RT: 9.028 min Scan# 976

Delta R.T. -0.000 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09




Tgt Ion:117 Resp: 723480


Ion Ratio Lower Upper

117 100

201 112.9 82.3 123.5

199 71.3 51.8 77.6





#22

Nitrobenzene

Concen: 179.415 ng/ul

RT: 9.157 min Scan# 998

Delta R.T. 0.018 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Instrument :





BNA_P

ClientSampleId :

SSTD160634

Tgt Ion: 77 Resp: 1833070

Ion	Ratio	Lower	Upper
77	100		
123	48.1	34.9	52.3
65	13.7	10.5	15.7



#23

Isophorone

Concen: 164.801 ng/ul

RT: 9.699 min Scan# 1090




Delta R.T. 0.023 min


Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Tgt Ion: 82 Resp: 3524908

Ion	Ratio	Lower	Upper
82	100		
95	8.1	5.6	8.4
138	18.8	13.4	20.0









#25
2-Nitrophenol
Concen: 229.341 ng/ul
RT: 9.869 min Scan# 1119
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:139 Resp: 988099
Ion Ratio Lower Upper
139 100
65 47.0 40.3 60.5
109 36.0 23.7 35.5#



#26
2,4-Dimethylphenol
Concen: 174.932 ng/ul
RT: 9.934 min Scan# 1130
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:107 Resp: 1924323
Ion Ratio Lower Upper
107 100
121 49.6 38.8 58.2
122 82.8 64.0 96.0









#27
Bis(2-Chloroethoxy)methane
Concen: 160.756 ng/ul
RT: 10.169 min Scan# 1170
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion: 93 Resp: 2130555
Ion Ratio Lower Upper
93 100
95 32.5 26.2 39.4
123 11.5 8.3 12.5



#29
2,4-Dichlorophenol
Concen: 190.672 ng/ul
RT: 10.404 min Scan# 1210
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:162 Resp: 1699035
Ion Ratio Lower Upper
162 100
164 66.1 52.9 79.3
98 34.0 30.6 45.8









#30
Naphthalene
Concen: 165.180 ng/ul
RT: 10.804 min Scan# 1278
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:128 Resp: 5051263
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.6 10.5 15.7



#32
4-Chloroaniline
Concen: 176.278 ng/ul
RT: 10.916 min Scan# 1297
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:127 Resp: 2186768
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.6









#33
Hexachlorobutadiene
Concen: 198.959 ng/ul
RT: 11.098 min Scan# 1328
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:225 Resp: 1244701
Ion Ratio Lower Upper
225 100
223 62.0 50.6 76.0
227 64.3 51.0 76.6



#34
Caprolactam
Concen: 133.706 ng/ul
RT: 11.728 min Scan# 1435
Delta R.T. 0.059 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:113 Resp: 312528
Ion Ratio Lower Upper
113 100
55 149.4 130.1 195.1
56 117.5 96.2 144.4









#35
4-Chloro-3-methylphenol
Concen: 183.628 ng/ul
RT: 12.051 min Scan# 1490
Delta R.T. 0.018 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:107 Resp: 1777504
Ion Ratio Lower Upper
107 100
144 25.7 19.3 28.9
142 79.3 62.2 93.4



#36
2-Methylnaphthalene
Concen: 167.510 ng/ul
RT: 12.410 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:142 Resp: 3500987
Ion Ratio Lower Upper
142 100
141 89.9 71.8 107.6









#37
1-Methylnaphthalene
Concen: 165.675 ng/ul
RT: 12.634 min Scan# 1589
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:142 Resp: 3505327
Ion Ratio Lower Upper
142 100
141 91.8 73.5 110.3
116 4.2 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 190.650 ng/ul
RT: 12.781 min Scan# 1614
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:216 Resp: 2141140
Ion Ratio Lower Upper
216 100
214 78.9 63.1 94.7
179 21.3 17.4 26.0
108 15.7 15.0 22.6









#40
Hexachlorocyclopentadiene
Concen: 218.342 ng/ul
RT: 12.763 min Scan# 1611
Delta R.T. -0.000 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:237 Resp: 1546214
Ion Ratio Lower Upper
237 100
235 63.8 49.4 74.2
272 12.5 9.8 14.6



#41
2,4,6-Trichlorophenol
Concen: 217.315 ng/ul
RT: 13.022 min Scan# 1655
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:196 Resp: 1435138
Ion Ratio Lower Upper
196 100
198 95.9 75.8 113.6
200 31.5 23.0 34.4









#42
2,4,5-Trichlorophenol
Concen: 225.650 ng/ul
RT: 13.092 min Scan# 1667
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:196 Resp: 1551427
Ion Ratio Lower Upper
196 100
198 95.8 75.1 112.7
200 31.3 25.8 38.8



#43
1,1'-Biphenyl
Concen: 166.276 ng/ul
RT: 13.422 min Scan# 1723
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:154 Resp: 4604659
Ion Ratio Lower Upper
154 100
153 40.8 32.0 48.0
76 10.4 10.3 15.5









#44
2-Chloronaphthalene
Concen: 170.389 ng/ul
RT: 13.463 min Scan# 1730
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:162 Resp: 3634381
Ion Ratio Lower Upper
162 100
164 32.8 26.5 39.7
127 39.5 31.8 47.6



#45
2-Nitroaniline
Concen: 214.511 ng/ul
RT: 13.669 min Scan# 1765
Delta R.T. 0.018 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion: 65 Resp: 1063777
Ion Ratio Lower Upper
65 100
92 72.0 55.9 83.9
138 116.7 83.4 125.2









#47
Dimethylphthalate
Concen: 170.618 ng/ul
RT: 14.051 min Scan# 1830
Delta R.T. 0.018 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:163 Resp: 4555436
Ion Ratio Lower Upper
163 100
194 4.4 3.2 4.8
164 10.2 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 232.738 ng/ul
RT: 14.169 min Scan# 1850
Delta R.T. 0.018 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:165 Resp: 1030457
Ion Ratio Lower Upper
165 100
89 54.7 49.0 73.4
121 19.9 18.4 27.6









#50
Acenaphthylene
Concen: 164.171 ng/ul
RT: 14.310 min Scan# 1874
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:152 Resp: 5591606
Ion Ratio Lower Upper
152 100
151 20.5 15.8 23.8
153 13.5 10.6 15.8



#51
3-Nitroaniline
Concen: 182.400 ng/ul
RT: 14.492 min Scan# 1905
Delta R.T. 0.018 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:138 Resp: 899859
Ion Ratio Lower Upper
138 100
108 11.5 13.0 19.4#
92 113.8 95.7 143.5









#52
Acenaphthene
Concen: 167.108 ng/ul
RT: 14.651 min Scan# 1932
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:153 Resp: 3685619
Ion Ratio Lower Upper
153 100
152 47.7 37.9 56.9
154 88.2 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 337.542 ng/ul
RT: 14.698 min Scan# 1940
Delta R.T. 0.018 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:184 Resp: 690366
Ion Ratio Lower Upper
184 100
63 55.8 48.8 73.2
154 69.8 56.2 84.4









#55
4-Nitrophenol
Concen: 205.345 ng/ul
RT: 14.816 min Scan# 1960
Delta R.T. 0.029 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:109 Resp: 800154
Ion Ratio Lower Upper
109 100
139 112.3 100.2 150.2
65 106.1 98.8 148.2



#56
Dibenzofuran
Concen: 165.248 ng/ul
RT: 14.987 min Scan# 1989
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:168 Resp: 5291959
Ion Ratio Lower Upper
168 100
139 39.6 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 231.046 ng/ul
RT: 14.957 min Scan# 1984
Delta R.T. 0.023 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:165 Resp: 1436817
Ion Ratio Lower Upper
165 100
63 37.0 35.0 52.6
182 3.0 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 227.742 ng/ul
RT: 15.216 min Scan# 2028
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:232 Resp: 1299343
Ion Ratio Lower Upper
232 100
131 40.6 35.1 52.7
130 2.3 2.0 3.0
166 29.5 25.8 38.8









#59
Diethylphthalate
Concen: 169.860 ng/ul
RT: 15.416 min Scan# 2062
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:149 Resp: 4521963
Ion Ratio Lower Upper
149 100
177 22.9 17.4 26.2
150 12.9 10.0 15.0



#61
Fluorene
Concen: 165.713 ng/ul
RT: 15.639 min Scan# 2100
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:166 Resp: 4020830
Ion Ratio Lower Upper
166 100
165 98.4 78.4 117.6
167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 176.630 ng/ul
RT: 15.628 min Scan# 2098
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:204 Resp: 2182145
Ion Ratio Lower Upper
204 100
206 33.1 26.5 39.7
141 58.9 47.4 71.0



#63
4-Nitroaniline
Concen: 174.573 ng/ul
RT: 15.669 min Scan# 2105
Delta R.T. 0.029 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:138 Resp: 763506
Ion Ratio Lower Upper
138 100
92 50.0 42.1 63.1
108 94.2 70.3 105.5









#66
4,6-Dinitro-2-methylphenol
Concen: 257.176 ng/ul
RT: 15.722 min Scan# 2114
Delta R.T. 0.018 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:198 Resp: 884493
Ion Ratio Lower Upper
198 100
182 3.4 3.3 4.9
77 23.7 18.4 27.6



#67
N-Nitrosodiphenylamine
Concen: 166.128 ng/ul
RT: 15.845 min Scan# 2135
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:169 Resp: 3711209
Ion Ratio Lower Upper
169 100
168 67.1 52.5 78.7
167 36.4 28.4 42.6









#68
4-Bromophenyl-phenylether
Concen: 200.588 ng/ul
RT: 16.528 min Scan# 2251
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:248 Resp: 1499313
Ion Ratio Lower Upper
248 100
250 99.6 78.9 118.3
141 70.2 64.3 96.5



#69
Hexachlorobenzene
Concen: 193.045 ng/ul
RT: 16.645 min Scan# 2271
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:284 Resp: 1735673
Ion Ratio Lower Upper
284 100
142 32.4 29.8 44.8
249 32.6 26.6 39.8





#70

Atrazine

Concen: 187.890 ng/ul

RT: 16.810 min Scan# 2299

Delta R.T. 0.018 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD160634





Tgt Ion:200 Resp: 1589000

Ion Ratio Lower Upper

200 100

173 24.8 20.9 31.3

215 48.4 38.1 57.1



#71

Pentachlorophenol

Concen: 234.100 ng/ul

RT: 16.986 min Scan# 2329

Delta R.T. -0.000 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09




Tgt Ion:266 Resp: 1165143


Ion Ratio Lower Upper

266 100

264 63.2 50.6 75.8

268 64.4 51.3 76.9









#72
Phenanthrene
Concen: 164.397 ng/ul
RT: 17.386 min Scan# 2397
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:178 Resp: 6819393
Ion Ratio Lower Upper
178 100
179 15.6 12.1 18.1
176 19.9 15.2 22.8



#74
Anthracene
Concen: 163.133 ng/ul
RT: 17.480 min Scan# 2413
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:178 Resp: 6768039
Ion Ratio Lower Upper
178 100
179 15.9 12.2 18.4
176 19.3 14.6 22.0









#75

1,2,3,4-Tetrachlorobenzene
Concen: 188.790 ng/uL
RT: 13.387 min Scan# 1717
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Instrument :
BNA_P
ClientSampleId :
SSTD160634




Tgt Ion:216 Resp: 2272101
Ion Ratio Lower Upper
216 100
214 76.6 62.0 93.0
218 48.5 37.6 56.4




#76

Pentachlorobenzene
Concen: 190.244 ng/uL
RT: 14.910 min Scan# 1976
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:250 Resp: 2134304
Ion Ratio Lower Upper
250 100
248 63.2 50.1 75.1
252 64.1 50.6 76.0





#77

Carbazole

Concen: 170.111 ng/ul

RT: 17.745 min Scan# 2458

Delta R.T. 0.012 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD160634





Tgt Ion:167 Resp: 6199384

Ion Ratio Lower Upper

167 100

166 22.4 17.3 25.9

139 13.6 10.4 15.6



#78

Di-n-butylphthalate

Concen: 165.940 ng/ul

RT: 18.298 min Scan# 2552

Delta R.T. 0.006 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09




Tgt Ion:149 Resp: 7429072


Ion Ratio Lower Upper

149 100

150 9.7 7.3 10.9

104 5.6 4.2 6.4









#80
Fluoranthene
Concen: 162.476 ng/ul
RT: 19.351 min Scan# 2731
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:202 Resp: 8378485
Ion Ratio Lower Upper
202 100
101 8.1 8.4 12.6#
100 6.1 6.3 9.5#



#82
Pyrene
Concen: 157.759 ng/ul
RT: 19.704 min Scan# 2791
Delta R.T. 0.012 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:202 Resp: 8203552
Ion Ratio Lower Upper
202 100
101 9.4 9.9 14.9#
100 7.6 7.9 11.9#









#83
Butylbenzylphthalate
Concen: 186.396 ng/ul
RT: 20.574 min Scan# 2939
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:149 Resp: 3558632
Ion Ratio Lower Upper
149 100
91 65.2 55.9 83.9
206 23.4 17.4 26.0



#84
3,3'-Dichlorobenzidine
Concen: 148.389 ng/ul
RT: 21.339 min Scan# 3069
Delta R.T. 0.006 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:252 Resp: 2347107
Ion Ratio Lower Upper
252 100
254 65.1 51.1 76.7
126 8.2 10.0 15.0#





#85

Benzo(a)anthracene

Concen: 168.399 ng/ul

RT: 21.416 min Scan# 3082

Delta R.T. 0.012 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09





Tgt Ion:228 Resp: 8601633

Ion Ratio Lower Upper

228 100

229 20.7 15.7 23.5

226 28.6 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 166.622 ng/ul

RT: 21.339 min Scan# 3069

Delta R.T. -0.000 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09




Tgt Ion:149 Resp: 4863434


Ion Ratio Lower Upper

149 100

167 29.1 22.6 33.8

279 5.6 3.3 4.9#





#87

Chrysene

Concen: 164.524 ng/ul

RT: 21.469 min Scan# 3091

Delta R.T. 0.012 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD160634





Tgt Ion:228 Resp: 8140908

Ion Ratio Lower Upper

228 100

226 31.7 23.7 35.5

229 20.8 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 170.521 ng/ul




RT: 22.286 min Scan# 3230


Delta R.T. 0.006 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Tgt Ion:149 Resp: 9049308





#90

Benzo(b)fluoranthene

Concen: 175.841 ng/ul

RT: 23.139 min Scan# 3375

Delta R.T. 0.018 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD160634





Tgt Ion:252 Resp:10065330

Ion Ratio Lower Upper

252 100

253 22.5 17.4 26.2

125 7.4 7.8 11.6#



#91

Benzo(k)fluoranthene

Concen: 163.543 ng/ul

RT: 23.192 min Scan# 3384

Delta R.T. 0.023 min

Lab File: BP007644.D

Acq: 25 Oct 2021 16:09




Tgt Ion:252 Resp: 8728843


Ion Ratio Lower Upper

252 100

253 22.5 17.6 26.4

125 7.1 8.0 12.0#









#93
Benzo(a)pyrene
Concen: 171.202 ng/ul
RT: 23.786 min Scan# 3485
Delta R.T. 0.023 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD160634


Tgt Ion:252 Resp: 9373959
Ion Ratio Lower Upper
252 100
253 22.4 17.3 25.9
125 8.1 8.3 12.5#



#94
Indeno(1,2,3-cd)pyrene
Concen: 188.772 ng/ul
RT: 26.462 min Scan# 3940
Delta R.T. 0.035 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:276 Resp:11973874
Ion Ratio Lower Upper
276 100
138 18.6 17.8 26.8
277 25.4 19.8 29.6









#95
Dibenzo(a,h)anthracene
Concen: 184.451 ng/ul
RT: 26.492 min Scan# 3945
Delta R.T. 0.047 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Instrument :
BNA_P
ClientSampleId :
SSTD160634

Tgt Ion:278 Resp: 9962685
Ion Ratio Lower Upper
278 100
139 12.0 12.8 19.2#
279 24.5 19.1 28.7



#96
Benzo(g,h,i)perylene
Concen: 189.681 ng/ul
RT: 27.262 min Scan# 4076
Delta R.T. 0.053 min
Lab File: BP007644.D
Acq: 25 Oct 2021 16:09

Tgt Ion:276 Resp:10362011
Ion Ratio Lower Upper
276 100
138 17.6 17.5 26.3
277 24.2 18.5 27.7

