

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007673.D
 Acq On : 26 Oct 2021 10:05
 Operator : CG/JU
 Sample : PB140137BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS137

Manual Integrations
 APPROVED

mohammad
 10/27/2021 11:50:27 AM

Quant Time: Oct 26 11:00:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 16:57:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	195718	20.000	ng/u1	0.00
20) Naphthalene-d8	10.740	136	770480	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.575	164	490990	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	1049020	20.000	ng/u1	0.00
79) Chrysene-d12	21.422	240	1125544	20.000	ng/u1	0.00
88) Perylene-d12	23.869	264	1205151	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	33036	7.668	ng/uL	0.00
4) Pyridine-d5	3.781	84	478675	36.739	ng/u1	0.00
7) Phenol-d5	7.099	99	571313	38.431	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.264	67	325188	39.400	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	459311	39.726	ng/u1	0.00
15) 4-Methylphenol-d8	8.646	113	446062	38.320	ng/u1	0.00
21) Nitrobenzene-d5	9.093	128	215235	40.992	ng/u1	0.00
24) 2-Nitrophenol-d4	9.816	143	236344	41.213	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.363	165	462229	39.599	ng/u1	0.00
31) 4-Chloroaniline-d4	10.875	131	561823	36.401	ng/u1	0.00
46) Dimethylphthalate-d6	13.987	166	1314412	39.521	ng/u1	0.00
49) Acenaphthylene-d8	14.269	160	1607017	39.423	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	244016	40.026	ng/u1	0.00
60) Fluorene-d10	15.569	176	1100486	39.250	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.687	200	206509	36.803	ng/u1	0.00
73) Anthracene-d10	17.428	188	1705705	39.566	ng/u1	0.00
81) Pyrene-d10	19.663	212	2141376	40.865	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.710	264	2296505	38.791	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	67500	14.013	ng/uL	90
5) Pyridine	3.799	79	460488	36.081	ng/u1	98
6) Benzaldehyde	7.069	77	341800	47.827	ng/u1	98
8) Phenol	7.128	94	574774	38.268	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.358	93	450566	38.051	ng/u1	98
12) 2-Chlorophenol	7.499	128	461603	39.381	ng/u1	98
13) 2-Methylphenol	8.381	108	431501	38.032	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.469	45	524804	39.021	ng/u1	98
16) Acetophenone	8.758	105	657889	37.868	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.752	70	320266	38.759	ng/u1	94
18) 4-Methylphenol	8.710	108	458461	37.892	ng/u1	99
19) Hexachloroethane	9.022	117	188546	39.305	ng/u1	92
22) Nitrobenzene	9.134	77	490475	39.589	ng/u1	98
23) Isophorone	9.669	82	927006	39.114	ng/u1	97
25) 2-Nitrophenol	9.852	139	250562	40.594	ng/u1	95
26) 2,4-Dimethylphenol	9.916	107	502367	38.228	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.152	93	588438	38.730	ng/u1	99
29) 2,4-Dichlorophenol	10.387	162	446112	39.009	ng/u1	97
30) Naphthalene	10.793	128	1371456	37.703	ng/u1	100
32) 4-Chloroaniline	10.899	127	557956	35.516	ng/u1	100
33) Hexachlorobutadiene	11.087	225	320038	38.385	ng/u1	99
34) Caprolactam	11.669	113	136220m	36.327	ng/u1	
35) 4-Chloro-3-methylphenol	12.028	107	469209	39.631	ng/u1	97
36) 2-Methylnaphthalene	12.399	142	955293	38.200	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007673.D
 Acq On : 26 Oct 2021 10:05
 Operator : CG/JU
 Sample : PB140137BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS137


Manual Integrations
 APPROVED

mohammad
 10/27/2021 11:50:27 AM

Quant Time: Oct 26 11:00:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 16:57:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	959194	37.988	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.769	216	567187	38.084	ng/ul	98
40) Hexachlorocyclopentadiene	12.757	237	318208	32.947	ng/ul	99
41) 2,4,6-Trichlorophenol	13.010	196	371079	39.600	ng/ul	98
42) 2,4,5-Trichlorophenol	13.081	196	404883	40.128	ng/ul	99
43) 1,1'-Biphenyl	13.410	154	1293933	38.148	ng/ul	98
44) 2-Chloronaphthalene	13.451	162	993833	37.809	ng/ul	98
45) 2-Nitroaniline	13.651	65	277858	41.900	ng/ul	91
47) Dimethylphthalate	14.034	163	1277424	38.859	ng/ul	100
48) 2,6-Dinitrotoluene	14.146	165	263045	41.895	ng/ul	93
50) Acenaphthylene	14.293	152	1583777	38.375	ng/ul	100
51) 3-Nitroaniline	14.475	138	262088	39.357	ng/ul#	94
52) Acenaphthene	14.640	153	1035542	38.045	ng/ul	98
53) 2,4-Dinitrophenol	14.681	184	131965	33.674	ng/ul	93
55) 4-Nitrophenol	14.787	109	209280	38.830	ng/ul	93
56) Dibenzofuran	14.975	168	1502488	37.761	ng/ul	100
57) 2,4-Dinitrotoluene	14.934	165	379950	42.135	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.198	232	331949	40.362	ng/ul	99
59) Diethylphthalate	15.404	149	1271364	39.434	ng/ul	98
61) Fluorene	15.622	166	1169849	37.719	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.622	204	613595	37.306	ng/ul	99
63) 4-Nitroaniline	15.640	138	261682	42.327	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.704	198	206856	36.624	ng/ul	96
67) N-Nitrosodiphenylamine	15.834	169	1049661	38.784	ng/ul	100
68) 4-Bromophenyl-phenylether	16.516	248	400652	39.170	ng/ul	96
69) Hexachlorobenzene	16.634	284	460936	38.915	ng/ul	99
70) Atrazine	16.792	200	418952	37.334	ng/ul	99
71) Pentachlorophenol	16.981	266	276094	37.533	ng/ul	99
72) Phenanthrene	17.375	178	1961045	38.677	ng/ul	99
74) Anthracene	17.463	178	1923048	38.131	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.375	216	574519	37.459	ng/ul	99
76) Pentachlorobenzene	14.898	250	543537	36.752	ng/ul	99
77) Carbazole	17.734	167	1801174	38.687	ng/ul	99
78) Di-n-butylphthalate	18.292	149	2157405	40.474	ng/ul	100
80) Fluoranthene	19.339	202	2473919	38.522	ng/ul	98
82) Pyrene	19.692	202	2480629	38.164	ng/ul	97
83) Butylbenzylphthalate	20.569	149	1008852	39.890	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.333	252	848460	40.674	ng/ul	97
85) Benzo(a)anthracene	21.404	228	2593849	38.808	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.339	149	1476798	39.297	ng/ul#	100
87) Chrysene	21.457	228	2487709	38.495	ng/ul	99
89) Di-n-octyl phthalate	22.280	149	2641825	39.353	ng/ul	100
90) Benzo(b)fluoranthene	23.122	252	2841391	40.013	ng/ul	98
91) Benzo(k)fluoranthene	23.169	252	2623690	39.482	ng/ul	99
93) Benzo(a)pyrene	23.763	252	2733692	39.733	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.427	276	3305040	40.328	ng/ul	98
95) Dibenzo(a,h)anthracene	26.451	278	2833963	40.353	ng/ul	98
96) Benzo(g,h,i)perylene	27.210	276	2846474	40.835	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
1,4-Dioxane
Concen: 14.013 ng/uL
RT: 3.399 min Scan# 19
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05

Tgt Ion: 88 Resp: 67500
Ion Ratio Lower Upper
88 100
43 25.8 24.2 36.2
58 68.6 62.5 93.7

Instrument :
BNA_P
ClientSampleId :
SLCS137




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:27 AM



#5
Pyridine
Concen: 36.081 ng/uL
RT: 3.799 min Scan# 87
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05


Tgt Ion: 79 Resp: 460488
Ion Ratio Lower Upper
79 100
52 73.1 57.0 85.4
51 33.8 27.0 40.6





#6
Benzaldehyde
Concen: 47.827 ng/ul
RT: 7.069 min Scan# 643
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion: 77 Resp: 341800
Ion Ratio Lower Upper
77 100
105 99.6 78.2 117.4
106 95.9 74.6 112.0

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#8
Phenol
Concen: 38.268 ng/ul
RT: 7.128 min Scan# 653
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05



Tgt Ion: 94 Resp: 574774
Ion Ratio Lower Upper
94 100
65 26.2 21.9 32.9
66 37.9 29.9 44.9






#10

Bis(2-Chloroethyl)ether
Concen: 38.051 ng/ul
RT: 7.358 min Scan# 692
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137



Tgt Ion: 93	Resp: 450566
Ion Ratio	Lower Upper
93 100	
63 71.9	59.0 88.4
95 32.9	26.4 39.6

Manual Integrations
APPROVED




mohammad
10/27/2021 11:50:27 AM




#12

2-Chlorophenol
Concen: 39.381 ng/ul
RT: 7.499 min Scan# 716
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05

Tgt Ion:128	Resp: 461603		
Ion Ratio	Lower	Upper	
128 100			
64 44.7	36.9	55.3	
130 32.1	25.2	37.8	









#13
2-Methylphenol
Concen: 38.032 ng/ul
RT: 8.381 min Scan# 866
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05

Instrument :
BNA_P
ClientSampleId :
SLCS137

Tgt Ion:108 Resp: 431501
Ion Ratio Lower Upper
108 100
107 90.1 72.2 108.2




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:27 AM



#14
2,2'-oxybis(1-Chloropropane)
Concen: 39.021 ng/ul
RT: 8.469 min Scan# 881
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05


Tgt Ion: 45 Resp: 524804
Ion Ratio Lower Upper
45 100
77 16.0 12.3 18.5
79 12.4 9.4 14.0





#16
Acetophenone
Concen: 37.868 ng/ul
RT: 8.758 min Scan# 930
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:105 Resp: 657889
Ion Ratio Lower Upper
105 100
77 77.1 61.8 92.6
51 25.8 20.8 31.2

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#17
N-Nitroso-di-n-propylamine
Concen: 38.759 ng/ul
RT: 8.752 min Scan# 929
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion: 70 Resp: 320266
Ion Ratio Lower Upper
70 100
42 53.0 46.3 69.5
101 10.3 7.4 11.0
130 23.3 16.2 24.2





#18
4-Methylphenol
Concen: 37.892 ng/ul
RT: 8.710 min Scan# 922
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:108 Resp: 458461
Ion Ratio Lower Upper
108 100
107 115.0 90.9 136.3

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#19
Hexachloroethane
Concen: 39.305 ng/ul
RT: 9.022 min Scan# 975
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:117 Resp: 188546
Ion Ratio Lower Upper
117 100
201 111.4 82.3 123.5
199 70.4 51.8 77.6





#22
Nitrobenzene
Concen: 39.589 ng/ul
RT: 9.134 min Scan# 994
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion: 77 Resp: 490475
Ion Ratio Lower Upper
77 100
123 45.3 34.9 52.3
65 13.2 10.5 15.7

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#23
Isophorone
Concen: 39.114 ng/ul
RT: 9.669 min Scan# 1085
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion: 82 Resp: 927006
Ion Ratio Lower Upper
82 100
95 8.0 5.6 8.4
138 18.2 13.4 20.0





#25
2-Nitrophenol
Concen: 40.594 ng/ul
RT: 9.852 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampled :
SLCS137




Tgt Ion:139 Resp: 250562
Ion Ratio Lower Upper
139 100
65 47.5 40.3 60.5
109 32.6 23.7 35.5

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#26
2,4-Dimethylphenol
Concen: 38.228 ng/ul
RT: 9.916 min Scan# 1127
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:107 Resp: 502367
Ion Ratio Lower Upper
107 100
121 49.1 38.8 58.2
122 82.7 64.0 96.0





#27
Bis(2-Chloroethoxy)methane
Concen: 38.730 ng/ul
RT: 10.152 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampled :
SLCS137




Tgt Ion: 93 Resp: 588438
Ion Ratio Lower Upper
93 100
95 32.4 26.2 39.4
123 11.5 8.3 12.5

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#29
2,4-Dichlorophenol
Concen: 39.009 ng/ul
RT: 10.387 min Scan# 1207
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion: 162 Resp: 446112
Ion Ratio Lower Upper
162 100
164 64.7 52.9 79.3
98 34.8 30.6 45.8





#30
Naphthalene
Concen: 37.703 ng/ul
RT: 10.793 min Scan# 1276
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:128 Resp: 1371456
Ion Ratio Lower Upper
128 100
129 10.8 8.8 13.2
127 13.1 10.5 15.7

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#32
4-Chloroaniline
Concen: 35.516 ng/ul
RT: 10.899 min Scan# 1294
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:127 Resp: 557956
Ion Ratio Lower Upper
127 100
129 32.0 25.8 38.6





#33
Hexachlorobutadiene
Concen: 38.385 ng/ul
RT: 11.087 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:225 Resp: 320038
Ion Ratio Lower Upper
225 100
223 62.0 50.6 76.0
227 62.9 51.0 76.6

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#34
Caprolactam
Concen: 36.327 ng/ul m
RT: 11.669 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:113 Resp: 136220
Ion Ratio Lower Upper
113 100
55 156.4 130.1 195.1
56 117.0 96.2 144.4





#35
4-Chloro-3-methylphenol
Concen: 39.631 ng/ul
RT: 12.028 min Scan# 1486
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:107 Resp: 469209
Ion Ratio Lower Upper
107 100
144 26.3 19.3 28.9
142 79.7 62.2 93.4

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#36
2-Methylnaphthalene
Concen: 38.200 ng/ul
RT: 12.399 min Scan# 1549
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:142 Resp: 955293
Ion Ratio Lower Upper
142 100
141 89.6 71.8 107.6





#37
1-Methylnaphthalene
Concen: 37.988 ng/ul
RT: 12.616 min Scan# 1586
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:142 Resp: 959194
Ion Ratio Lower Upper
142 100
141 92.6 73.5 110.3
116 4.2 3.1 4.7

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 38.084 ng/ul
RT: 12.769 min Scan# 1612
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:216 Resp: 567187
Ion Ratio Lower Upper
216 100
214 77.6 63.1 94.7
179 20.8 17.4 26.0
108 15.8 15.0 22.6





#40
Hexachlorocyclopentadiene
Concen: 32.947 ng/ul
RT: 12.757 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:237 Resp: 318208
Ion Ratio Lower Upper
237 100
235 61.1 49.4 74.2
272 11.9 9.8 14.6

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#41
2,4,6-Trichlorophenol
Concen: 39.600 ng/ul
RT: 13.010 min Scan# 1653
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05



Tgt Ion:196 Resp: 371079
Ion Ratio Lower Upper
196 100
198 96.2 75.8 113.6
200 31.8 23.0 34.4









#42
2,4,5-Trichlorophenol
Concen: 40.128 ng/ul
RT: 13.081 min Scan# 1665
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05

Tgt Ion:196 Resp: 404883
Ion Ratio Lower Upper
196 100
198 95.0 75.1 112.7
200 31.3 25.8 38.8

Instrument :
BNA_P
ClientSampleId :
SLCS137




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:27 AM



#43
1,1'-Biphenyl
Concen: 38.148 ng/ul
RT: 13.410 min Scan# 1721
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05


Tgt Ion:154 Resp: 1293933
Ion Ratio Lower Upper
154 100
153 40.7 32.0 48.0
76 11.3 10.3 15.5





#44
2-Chloronaphthalene
Concen: 37.809 ng/ul
RT: 13.451 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:162 Resp: 993833
Ion Ratio Lower Upper
162 100
164 33.6 26.5 39.7
127 38.2 31.8 47.6

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#45
2-Nitroaniline
Concen: 41.900 ng/ul
RT: 13.651 min Scan# 1762
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion: 65 Resp: 277858
Ion Ratio Lower Upper
65 100
92 73.9 55.9 83.9
138 116.2 83.4 125.2





#47
Dimethylphthalate
Concen: 38.859 ng/ul
RT: 14.034 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:163 Resp: 1277424
Ion Ratio Lower Upper
163 100
194 4.2 3.2 4.8
164 9.9 7.9 11.9

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#48
2,6-Dinitrotoluene
Concen: 41.895 ng/ul
RT: 14.146 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:165 Resp: 263045
Ion Ratio Lower Upper
165 100
89 55.8 49.0 73.4
121 20.5 18.4 27.6





#50
Acenaphthylene
Concen: 38.375 ng/ul
RT: 14.293 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05

Instrument :
BNA_P
ClientSampleId :
SLCS137






Tgt Ion:152 Resp: 1583777


Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.8	23.8
153	13.2	10.6	15.8

Manual Integrations
APPROVED

mohammad
10/27/2021 11:50:27 AM






#51
3-Nitroaniline
Concen: 39.357 ng/ul
RT: 14.475 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05



Tgt Ion:138 Resp: 262088


Ion	Ratio	Lower	Upper
138	100		
108	12.6	13.0	19.4
92	113.0	95.7	143.5





#52
Acenaphthene
Concen: 38.045 ng/ul
RT: 14.640 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:153 Resp: 1035542
Ion Ratio Lower Upper
153 100
152 47.9 37.9 56.9
154 89.7 70.1 105.1

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#53
2,4-Dinitrophenol
Concen: 33.674 ng/ul
RT: 14.681 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:184 Resp: 131965
Ion Ratio Lower Upper
184 100
63 55.6 48.8 73.2
154 63.9 56.2 84.4





#55
4-Nitrophenol
Concen: 38.830 ng/ul
RT: 14.787 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:109 Resp: 209280
Ion Ratio Lower Upper
109 100
139 121.9 100.2 150.2
65 110.7 98.8 148.2

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#56
Dibenzofuran
Concen: 37.761 ng/ul
RT: 14.975 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:168 Resp: 1502488
Ion Ratio Lower Upper
168 100
139 37.5 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 42.135 ng/ul
RT: 14.934 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:165 Resp: 379950
Ion Ratio Lower Upper
165 100
63 38.8 35.0 52.6
182 2.7 2.3 3.5

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#58
2,3,4,6-Tetrachlorophenol
Concen: 40.362 ng/ul
RT: 15.198 min Scan# 2025
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:232 Resp: 331949
Ion Ratio Lower Upper
232 100
131 43.2 35.1 52.7
130 2.3 2.0 3.0
166 31.4 25.8 38.8





#59
Diethylphthalate
Concen: 39.434 ng/ul
RT: 15.404 min Scan# 2060
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampled :
SLCS137




Tgt Ion:149 Resp: 1271364
Ion Ratio Lower Upper
149 100
177 22.9 17.4 26.2
150 12.3 10.0 15.0

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#61
Fluorene
Concen: 37.719 ng/ul
RT: 15.622 min Scan# 2097
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:166 Resp: 1169849
Ion Ratio Lower Upper
166 100
165 98.1 78.4 117.6
167 13.4 10.9 16.3





#62
4-Chlorophenyl-phenylether
Concen: 37.306 ng/ul
RT: 15.622 min Scan# 2097
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:204 Resp: 613595
Ion Ratio Lower Upper
204 100
206 33.1 26.5 39.7
141 60.3 47.4 71.0

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#63
4-Nitroaniline
Concen: 42.327 ng/ul
RT: 15.640 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05



Tgt Ion:138 Resp: 261682
Ion Ratio Lower Upper
138 100
92 50.9 42.1 63.1
108 88.3 70.3 105.5









#66
4,6-Dinitro-2-methylphenol
Concen: 36.624 ng/ul
RT: 15.704 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05

Tgt Ion:198 Resp: 206856
Ion Ratio Lower Upper
198 100
182 3.8 3.3 4.9
77 20.9 18.4 27.6

Instrument :
BNA_P
ClientSampleId :
SLCS137




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:27 AM



#67
N-Nitrosodiphenylamine
Concen: 38.784 ng/ul
RT: 15.834 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05


Tgt Ion:169 Resp: 1049661
Ion Ratio Lower Upper
169 100
168 65.6 52.5 78.7
167 36.1 28.4 42.6





#68
4-Bromophenyl-phenylether
Concen: 39.170 ng/ul
RT: 16.516 min Scan# 2249
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:248 Resp: 400652
Ion Ratio Lower Upper
248 100
250 99.8 78.9 118.3
141 74.6 64.3 96.5

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#69
Hexachlorobenzene
Concen: 38.915 ng/ul
RT: 16.634 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:284 Resp: 460936
Ion Ratio Lower Upper
284 100
142 35.7 29.8 44.8
249 33.3 26.6 39.8





#70
Atrazine
Concen: 37.334 ng/ul
RT: 16.792 min Scan# 2296
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampled :
SLCS137




Tgt Ion:200 Resp: 418952
Ion Ratio Lower Upper
200 100
173 24.5 20.9 31.3
215 47.5 38.1 57.1

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#71
Pentachlorophenol
Concen: 37.533 ng/ul
RT: 16.981 min Scan# 2328
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:266 Resp: 276094
Ion Ratio Lower Upper
266 100
264 62.6 50.6 75.8
268 64.5 51.3 76.9





#72
Phenanthrene
Concen: 38.677 ng/u1
RT: 17.375 min Scan# 2395
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:178 Resp: 1961045
Ion Ratio Lower Upper
178 100
179 14.8 12.1 18.1
176 19.4 15.2 22.8

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#74
Anthracene
Concen: 38.131 ng/u1
RT: 17.463 min Scan# 2410
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:178 Resp: 1923048
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 18.8 14.6 22.0





#75
1,2,3,4-Tetrachlorobenzene
Concen: 37.459 ng/uL
RT: 13.375 min Scan# 1715
Delta R.T. -0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:216 Resp: 574519
Ion Ratio Lower Upper
216 100
214 76.9 62.0 93.0
218 48.5 37.6 56.4

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#76
Pentachlorobenzene
Concen: 36.752 ng/uL
RT: 14.898 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:250 Resp: 543537
Ion Ratio Lower Upper
250 100
248 62.1 50.1 75.1
252 64.1 50.6 76.0





#77
Carbazole
Concen: 38.687 ng/ul
RT: 17.734 min Scan# 2456
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:167 Resp: 1801174
Ion Ratio Lower Upper
167 100
166 21.8 17.3 25.9
139 12.6 10.4 15.6

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#78
Di-n-butylphthalate
Concen: 40.474 ng/ul
RT: 18.292 min Scan# 2551
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:149 Resp: 2157405
Ion Ratio Lower Upper
149 100
150 9.0 7.3 10.9
104 5.0 4.2 6.4





#80
Fluoranthene
Concen: 38.522 ng/u1
RT: 19.339 min Scan# 2729
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:202 Resp: 2473919
Ion Ratio Lower Upper
202 100
101 9.6 8.4 12.6
100 7.1 6.3 9.5

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#82
Pyrene
Concen: 38.164 ng/u1
RT: 19.692 min Scan# 2789
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:202 Resp: 2480629
Ion Ratio Lower Upper
202 100
101 10.9 9.9 14.9
100 9.0 7.9 11.9





#83
Butylbenzylphthalate
Concen: 39.890 ng/ul
RT: 20.569 min Scan# 2938
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137



Tgt Ion:149 Resp: 1008852
Ion Ratio Lower Upper
149 100
91 68.4 55.9 83.9
206 22.3 17.4 26.0




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:27 AM



#84
3,3'-Dichlorobenzidine
Concen: 40.674 ng/ul
RT: 21.333 min Scan# 3068
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05

Tgt Ion:252 Resp: 848460
Ion Ratio Lower Upper
252 100
254 65.5 51.1 76.7
126 10.2 10.0 15.0





#85
Benzo(a)anthracene
Concen: 38.808 ng/ul
RT: 21.404 min Scan# 3080
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05





Tgt Ion:228 Resp: 2593849

Ion	Ratio	Lower	Upper
228	100		
229	19.8	15.7	23.5
226	27.2	21.4	32.0

Instrument :
BNA_P
ClientSampleId :
SLCS137

Manual Integrations
APPROVED




mohammad
10/27/2021 11:50:27 AM




#86
Bis(2-ethylhexyl)phthalate
Concen: 39.297 ng/ul
RT: 21.339 min Scan# 3069
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05

Tgt Ion:149 Resp: 1476798


Ion	Ratio	Lower	Upper
149	100		
167	28.2	22.6	33.8
279	5.0	3.3	4.9





#87
Chrysene
Concen: 38.495 ng/ul
RT: 21.457 min Scan# 3089
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137



Tgt Ion:228 Resp: 2487709
Ion Ratio Lower Upper
228 100
226 30.2 23.7 35.5
229 19.9 16.0 24.0




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:27 AM



#89
Di-n-octyl phthalate
Concen: 39.353 ng/ul
RT: 22.280 min Scan# 3229
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05


Tgt Ion:149 Resp: 2641825





#90
Benzo(b)fluoranthene
Concen: 40.013 ng/ul
RT: 23.122 min Scan# 3372
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137




Tgt Ion:252 Resp: 2841391
Ion Ratio Lower Upper
252 100
253 22.4 17.4 26.2
125 8.4 7.8 11.6

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:27 AM




#91
Benzo(k)fluoranthene
Concen: 39.482 ng/ul
RT: 23.169 min Scan# 3380
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Tgt Ion:252 Resp: 2623690
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 8.5 8.0 12.0





#93
Benzo(a)pyrene
Concen: 39.733 ng/ul
RT: 23.763 min Scan# 3481
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137



Tgt Ion:252 Resp: 2733692
Ion Ratio Lower Upper
252 100
253 21.6 17.3 25.9
125 9.2 8.3 12.5




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:27 AM



#94
Indeno(1,2,3-cd)pyrene
Concen: 40.328 ng/ul
RT: 26.427 min Scan# 3934
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05


Tgt Ion:276 Resp: 3305040
Ion Ratio Lower Upper
276 100
138 21.0 17.8 26.8
277 24.3 19.8 29.6





#95
Dibenzo(a,h)anthracene
Concen: 40.353 ng/ul
RT: 26.451 min Scan# 3938
Delta R.T. 0.006 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05




Instrument :
BNA_P
ClientSampleId :
SLCS137



Tgt Ion:278 Resp: 2833963
Ion Ratio Lower Upper
278 100
139 14.0 12.8 19.2
279 23.7 19.1 28.7

Manual Integrations
APPROVED

mohammad
10/27/2021 11:50:27 AM



#96
Benzo(g,h,i)perylene
Concen: 40.835 ng/ul
RT: 27.210 min Scan# 4067
Delta R.T. 0.000 min
Lab File: BP007673.D
Acq: 26 Oct 2021 10:05

Tgt Ion:276 Resp: 2846474
Ion Ratio Lower Upper
276 100
138 19.4 17.5 26.3
277 24.1 18.5 27.7

