

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007686.D
 Acq On : 26 Oct 2021 19:35
 Operator : CG/JU
 Sample : PB140178BS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS178

Manual Integrations
 APPROVED

mohammad
 10/27/2021 11:50:35 AM

Quant Time: Oct 27 00:55:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 16:57:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	176651	20.000	ng/u1	0.00
20) Naphthalene-d8	10.740	136	714829	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.575	164	457795	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	976882	20.000	ng/u1	0.00
79) Chrysene-d12	21.422	240	1026420	20.000	ng/u1	0.00
88) Perylene-d12	23.868	264	1091228	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	28627m	7.361	ng/uL	0.00
4) Pyridine-d5	3.781	84	420602	35.766	ng/u1	0.00
7) Phenol-d5	7.099	99	522565	38.946	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.263	67	295974	39.731	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	419674	40.215	ng/u1	0.00
15) 4-Methylphenol-d8	8.646	113	413048	39.314	ng/u1	0.00
21) Nitrobenzene-d5	9.093	128	199190	40.890	ng/u1	0.00
24) 2-Nitrophenol-d4	9.816	143	214792	40.371	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.357	165	429043	39.617	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.875	131	538284	37.591	ng/u1	0.00
46) Dimethylphthalate-d6	13.987	166	1248984	40.277	ng/u1	0.00
49) Acenaphthylene-d8	14.263	160	1508448	39.688	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	231141	40.663	ng/u1	0.00
60) Fluorene-d10	15.569	176	1037779	39.697	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.687	200	191505	36.649	ng/u1	0.00
73) Anthracene-d10	17.428	188	1589720	39.599	ng/u1	0.00
81) Pyrene-d10	19.663	212	2002983	41.915	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.710	264	2100638	39.187	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	62206	14.308	ng/uL#	89
5) Pyridine	3.799	79	429045	37.246	ng/u1	97
6) Benzaldehyde	7.069	77	326334	50.591	ng/u1	97
8) Phenol	7.128	94	545152	40.213	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.358	93	429425	40.180	ng/u1	98
12) 2-Chlorophenol	7.499	128	440753	41.661	ng/u1	96
13) 2-Methylphenol	8.381	108	414436	40.470	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.469	45	502795	41.419	ng/u1	99
16) Acetophenone	8.757	105	627517	40.019	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.752	70	307995	41.297	ng/u1	94
18) 4-Methylphenol	8.710	108	445590	40.803	ng/u1	98
19) Hexachloroethane	9.022	117	176622	40.793	ng/u1	94
22) Nitrobenzene	9.134	77	466564	40.591	ng/u1	96
23) Isophorone	9.669	82	886684	40.325	ng/u1	98
25) 2-Nitrophenol	9.852	139	240074	41.923	ng/u1	95
26) 2,4-Dimethylphenol	9.916	107	456691	37.458	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.152	93	565531	40.120	ng/u1	98
29) 2,4-Dichlorophenol	10.387	162	429689	40.499	ng/u1	98
30) Naphthalene	10.793	128	1315786	38.989	ng/u1	99
32) 4-Chloroaniline	10.899	127	550657	37.780	ng/u1	99
33) Hexachlorobutadiene	11.087	225	299037	38.658	ng/u1	100
34) Caprolactam	11.669	113	133581m	38.397	ng/u1	
35) 4-Chloro-3-methylphenol	12.022	107	458517	41.743	ng/u1	98
36) 2-Methylnaphthalene	12.398	142	915100	39.442	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007686.D
 Acq On : 26 Oct 2021 19:35
 Operator : CG/JU
 Sample : PB140178BS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS178

Manual Integrations
 APPROVED

mohammad
 10/27/2021 11:50:35 AM

Quant Time: Oct 27 00:55:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 16:57:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	925811	39.520	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.769	216	551931	39.746	ng/ul#	97
40) Hexachlorocyclopentadiene	12.751	237	279514	31.039	ng/ul	97
41) 2,4,6-Trichlorophenol	13.004	196	360004	41.204	ng/ul	99
42) 2,4,5-Trichlorophenol	13.075	196	389322	41.384	ng/ul	99
43) 1,1'-Biphenyl	13.410	154	1250444	39.539	ng/ul	99
44) 2-Chloronaphthalene	13.445	162	958045	39.091	ng/ul	99
45) 2-Nitroaniline	13.645	65	270666	43.775	ng/ul	96
47) Dimethylphthalate	14.034	163	1250128	40.786	ng/ul	100
48) 2,6-Dinitrotoluene	14.145	165	255079	43.572	ng/ul	93
50) Acenaphthylene	14.292	152	1543595	40.114	ng/ul	100
51) 3-Nitroaniline	14.469	138	269600	43.420	ng/ul#	93
52) Acenaphthene	14.640	153	1008954	39.756	ng/ul	100
53) 2,4-Dinitrophenol	14.681	184	131623	36.022	ng/ul	96
55) 4-Nitrophenol	14.781	109	201237	40.045	ng/ul	95
56) Dibenzofuran	14.975	168	1464487	39.475	ng/ul	99
57) 2,4-Dinitrotoluene	14.934	165	371857	44.227	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	15.198	232	322236	42.022	ng/ul	96
59) Diethylphthalate	15.398	149	1231109	40.954	ng/ul	98
61) Fluorene	15.622	166	1142550	39.510	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.622	204	603694	39.365	ng/ul	97
63) 4-Nitroaniline	15.639	138	260425	45.178	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.698	198	200854	38.188	ng/ul	99
67) N-Nitrosodiphenylamine	15.834	169	1020143	40.477	ng/ul	98
68) 4-Bromophenyl-phenylether	16.516	248	387319	40.662	ng/ul	97
69) Hexachlorobenzene	16.634	284	446728	40.500	ng/ul	96
70) Atrazine	16.792	200	402734	38.539	ng/ul	98
71) Pentachlorophenol	16.981	266	267810	39.095	ng/ul	100
72) Phenanthrene	17.369	178	1916717	40.594	ng/ul	99
74) Anthracene	17.463	178	1872730	39.876	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.375	216	548869	38.430	ng/uL	99
76) Pentachlorobenzene	14.898	250	529973	38.481	ng/uL	100
77) Carbazole	17.733	167	1755483	40.490	ng/ul	99
78) Di-n-butylphthalate	18.292	149	2080158	41.906	ng/ul	99
80) Fluoranthene	19.339	202	2401835	41.011	ng/ul	97
82) Pyrene	19.692	202	2415575	40.752	ng/ul	97
83) Butylbenzylphthalate	20.569	149	965994	41.884	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.333	252	780494	41.029	ng/ul#	99
85) Benzo(a)anthracene	21.404	228	2490871	40.866	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.333	149	1397648	40.783	ng/ul	98
87) Chrysene	21.457	228	2381107	40.404	ng/ul	99
89) Di-n-octyl phthalate	22.280	149	2530016	41.622	ng/ul	100
90) Benzo(b)fluoranthene	23.121	252	2728132	42.429	ng/ul	99
91) Benzo(k)fluoranthene	23.168	252	2512099	41.749	ng/ul	98
93) Benzo(a)pyrene	23.757	252	2614408	41.966	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.427	276	3165910	42.663	ng/ul	99
95) Dibenzo(a,h)anthracene	26.445	278	2718445	42.749	ng/ul	98
96) Benzo(g,h,i)perylene	27.209	276	2730987	43.269	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007686.D
 Acq On : 26 Oct 2021 19:35
 Operator : CG/JU
 Sample : PB140178BS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS178

Manual Integrations
 APPROVED

mohammad
 10/27/2021 11:50:35 AM


Quant Time: Oct 27 00:55:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 16:57:16 2021
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 14.308 ng/uL
RT: 3.405 min Scan# 20
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion: 88 Resp: 62206
Ion Ratio Lower Upper
88 100
43 24.1 24.2 36.2#
58 68.1 62.5 93.7

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#5
Pyridine
Concen: 37.246 ng/uL
RT: 3.799 min Scan# 87
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion: 79 Resp: 429045
Ion Ratio Lower Upper
79 100
52 74.8 57.0 85.4
51 34.3 27.0 40.6





#6
Benzaldehyde
Concen: 50.591 ng/ul
RT: 7.069 min Scan# 643
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion: 77 Resp: 326334
Ion Ratio Lower Upper
77 100
105 100.6 78.2 117.4
106 97.1 74.6 112.0

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#8
Phenol
Concen: 40.213 ng/ul
RT: 7.128 min Scan# 653
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion: 94 Resp: 545152
Ion Ratio Lower Upper
94 100
65 27.4 21.9 32.9
66 37.9 29.9 44.9





#10
Bis(2-Chloroethyl)ether
Concen: 40.180 ng/ul
RT: 7.358 min Scan# 692
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178



Tgt Ion:	93	Resp:	429425
Ion	Ratio	Lower	Upper
93	100		
63	71.3	59.0	88.4
95	32.4	26.4	39.6




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#12
2-Chlorophenol
Concen: 41.661 ng/ul
RT: 7.499 min Scan# 716
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35

Tgt Ion:	128	Resp:	440753
Ion	Ratio	Lower	Upper
128	100		
64	42.8	36.9	55.3
130	32.5	25.2	37.8









#13
2-Methylphenol
Concen: 40.470 ng/ul
RT: 8.381 min Scan# 866
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35

Instrument :
BNA_P
ClientSampleId :
SLCS178

Tgt Ion:108 Resp: 414436
Ion Ratio Lower Upper
108 100
107 88.0 72.2 108.2




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#14
2,2'-oxybis(1-Chloropropane)
Concen: 41.419 ng/ul
RT: 8.469 min Scan# 881
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion: 45 Resp: 502795
Ion Ratio Lower Upper
45 100
77 15.7 12.3 18.5
79 12.1 9.4 14.0





#16
Acetophenone
Concen: 40.019 ng/ul
RT: 8.757 min Scan# 930
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion:105 Resp: 627517
Ion Ratio Lower Upper
105 100
77 76.9 61.8 92.6
51 24.9 20.8 31.2

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#17
N-Nitroso-di-n-propylamine
Concen: 41.297 ng/ul
RT: 8.752 min Scan# 929
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion: 70 Resp: 307995
Ion Ratio Lower Upper
70 100
42 53.4 46.3 69.5
101 10.4 7.4 11.0
130 22.5 16.2 24.2





#18
4-Methylphenol
Concen: 40.803 ng/ul
RT: 8.710 min Scan# 922
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion:108 Resp: 445590
Ion Ratio Lower Upper
108 100
107 111.6 90.9 136.3

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#19
Hexachloroethane
Concen: 40.793 ng/ul
RT: 9.022 min Scan# 975
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion:117 Resp: 176622
Ion Ratio Lower Upper
117 100
201 110.0 82.3 123.5
199 69.2 51.8 77.6





#22
Nitrobenzene
Concen: 40.591 ng/ul
RT: 9.134 min Scan# 994
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion: 77 Resp: 466564
Ion Ratio Lower Upper
77 100
123 46.4 34.9 52.3
65 13.4 10.5 15.7

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#23
Isophorone
Concen: 40.325 ng/ul
RT: 9.669 min Scan# 1085
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35



Tgt Ion: 82 Resp: 886684
Ion Ratio Lower Upper
82 100
95 7.5 5.6 8.4
138 17.8 13.4 20.0









#25
2-Nitrophenol
Concen: 41.923 ng/ul
RT: 9.852 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35

Tgt Ion:139 Resp: 240074
Ion Ratio Lower Upper
139 100
65 47.3 40.3 60.5
109 33.4 23.7 35.5

Instrument :
BNA_P
ClientSampled :
SLCS178




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#26
2,4-Dimethylphenol
Concen: 37.458 ng/ul
RT: 9.916 min Scan# 1127
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:107 Resp: 456691
Ion Ratio Lower Upper
107 100
121 48.8 38.8 58.2
122 84.6 64.0 96.0





#27
Bis(2-Chloroethoxy)methane
Concen: 40.120 ng/ul
RT: 10.152 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178




Tgt Ion: 93 Resp: 565531
Ion Ratio Lower Upper
93 100
95 32.1 26.2 39.4
123 11.5 8.3 12.5

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#29
2,4-Dichlorophenol
Concen: 40.499 ng/ul
RT: 10.387 min Scan# 1207
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion: 162 Resp: 429689
Ion Ratio Lower Upper
162 100
164 66.2 52.9 79.3
98 34.3 30.6 45.8





#30
Naphthalene
Concen: 38.989 ng/ul
RT: 10.793 min Scan# 1276
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion:128 Resp: 1315786
Ion Ratio Lower Upper
128 100
129 10.6 8.8 13.2
127 13.2 10.5 15.7

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#32
4-Chloroaniline
Concen: 37.780 ng/ul
RT: 10.899 min Scan# 1294
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion:127 Resp: 550657
Ion Ratio Lower Upper
127 100
129 32.6 25.8 38.6





#33
Hexachlorobutadiene
Concen: 38.658 ng/ul
RT: 11.087 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178




Tgt Ion:225 Resp: 299037
Ion Ratio Lower Upper
225 100
223 62.8 50.6 76.0
227 64.1 51.0 76.6

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#34
Caprolactam
Concen: 38.397 ng/ul m
RT: 11.669 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion:113 Resp: 133581
Ion Ratio Lower Upper
113 100
55 160.0 130.1 195.1
56 121.3 96.2 144.4





#35
4-Chloro-3-methylphenol
Concen: 41.743 ng/ul
RT: 12.022 min Scan# 1485
Delta R.T. -0.012 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178



Tgt Ion:107 Resp: 458517
Ion Ratio Lower Upper
107 100
144 25.8 19.3 28.9
142 79.7 62.2 93.4




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#36
2-Methylnaphthalene
Concen: 39.442 ng/ul
RT: 12.398 min Scan# 1549
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:142 Resp: 915100
Ion Ratio Lower Upper
142 100
141 89.2 71.8 107.6





#37
1-Methylnaphthalene
Concen: 39.520 ng/ul
RT: 12.616 min Scan# 1586
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178



Tgt Ion:142 Resp: 925811
Ion Ratio Lower Upper
142 100
141 92.1 73.5 110.3
116 4.2 3.1 4.7




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#39
1,2,4,5-Tetrachlorobenzene
Concen: 39.746 ng/ul
RT: 12.769 min Scan# 1612
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:216 Resp: 551931
Ion Ratio Lower Upper
216 100
214 77.0 63.1 94.7
179 20.9 17.4 26.0
108 14.6 15.0 22.6#





#40
Hexachlorocyclopentadiene
Concen: 31.039 ng/ul
RT: 12.751 min Scan# 1609
Delta R.T. -0.012 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178



Tgt Ion:237 Resp: 279514
Ion Ratio Lower Upper
237 100
235 64.4 49.4 74.2
272 12.4 9.8 14.6




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#41
2,4,6-Trichlorophenol
Concen: 41.204 ng/ul
RT: 13.004 min Scan# 1652
Delta R.T. -0.012 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:196 Resp: 360004
Ion Ratio Lower Upper
196 100
198 95.6 75.8 113.6
200 30.1 23.0 34.4





#42
2,4,5-Trichlorophenol
Concen: 41.384 ng/ul
RT: 13.075 min Scan# 1664
Delta R.T. -0.012 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion:196 Resp: 389322
Ion Ratio Lower Upper
196 100
198 94.1 75.1 112.7
200 30.5 25.8 38.8

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#43
1,1'-Biphenyl
Concen: 39.539 ng/ul
RT: 13.410 min Scan# 1721
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion:154 Resp: 1250444
Ion Ratio Lower Upper
154 100
153 40.1 32.0 48.0
76 11.0 10.3 15.5





#44
2-Chloronaphthalene
Concen: 39.091 ng/ul
RT: 13.445 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion:162 Resp: 958045
Ion Ratio Lower Upper
162 100
164 33.3 26.5 39.7
127 39.2 31.8 47.6

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#45
2-Nitroaniline
Concen: 43.775 ng/ul
RT: 13.645 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35



Tgt Ion: 65 Resp: 270666
Ion Ratio Lower Upper
65 100
92 71.3 55.9 83.9
138 110.7 83.4 125.2






#47
Dimethylphthalate
Concen: 40.786 ng/ul
RT: 14.034 min Scan# 1827
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178

Manual Integrations
APPROVED

mohammad
10/27/2021 11:50:35 AM







Tgt Ion:163 Resp: 1250128
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 10.1 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 43.572 ng/ul
RT: 14.145 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:165 Resp: 255079
Ion Ratio Lower Upper
165 100
89 55.6 49.0 73.4
121 20.7 18.4 27.6





#50
Acenaphthylene
Concen: 40.114 ng/ul
RT: 14.292 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178




Tgt Ion:152 Resp: 1543595
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.8
153 13.1 10.6 15.8

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#51
3-Nitroaniline
Concen: 43.420 ng/ul
RT: 14.469 min Scan# 1901
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion:138 Resp: 269600
Ion Ratio Lower Upper
138 100
108 11.5 13.0 19.4#
92 112.6 95.7 143.5





#52
Acenaphthene
Concen: 39.756 ng/ul
RT: 14.640 min Scan# 1930
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion:153 Resp: 1008954
Ion Ratio Lower Upper
153 100
152 47.1 37.9 56.9
154 88.0 70.1 105.1

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#53
2,4-Dinitrophenol
Concen: 36.022 ng/ul
RT: 14.681 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion:184 Resp: 131623
Ion Ratio Lower Upper
184 100
63 56.1 48.8 73.2
154 69.2 56.2 84.4





#55
4-Nitrophenol
Concen: 40.045 ng/ul
RT: 14.781 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178




Tgt Ion:109 Resp: 201237
Ion Ratio Lower Upper
109 100
139 119.3 100.2 150.2
65 117.2 98.8 148.2

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#56
Dibenzofuran
Concen: 39.475 ng/ul
RT: 14.975 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35



Tgt Ion:168 Resp: 1464487
Ion Ratio Lower Upper
168 100
139 38.0 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 44.227 ng/ul
RT: 14.934 min Scan# 1980
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35

Tgt Ion:165 Resp: 371857
Ion Ratio Lower Upper
165 100
63 38.0 35.0 52.6
182 2.9 2.3 3.5

Instrument :
BNA_P
ClientSampled :
SLCS178




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#58
2,3,4,6-Tetrachlorophenol
Concen: 42.022 ng/ul
RT: 15.198 min Scan# 2025
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:232 Resp: 322236
Ion Ratio Lower Upper
232 100
131 40.9 35.1 52.7
130 2.5 2.0 3.0
166 30.8 25.8 38.8





#59
Diethylphthalate
Concen: 40.954 ng/ul
RT: 15.398 min Scan# 2059
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178



Tgt Ion:149 Resp: 1231109
Ion Ratio Lower Upper
149 100
177 22.8 17.4 26.2
150 12.8 10.0 15.0




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#61
Fluorene
Concen: 39.510 ng/ul
RT: 15.622 min Scan# 2097
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:166 Resp: 1142550
Ion Ratio Lower Upper
166 100
165 97.6 78.4 117.6
167 13.6 10.9 16.3





#62
4-Chlorophenyl-phenylether
Concen: 39.365 ng/ul
RT: 15.622 min Scan# 2097
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178




Tgt Ion:204 Resp: 603694
Ion Ratio Lower Upper
204 100
206 32.2 26.5 39.7
141 56.4 47.4 71.0

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#63
4-Nitroaniline
Concen: 45.178 ng/ul
RT: 15.639 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion:138 Resp: 260425
Ion Ratio Lower Upper
138 100
92 52.0 42.1 63.1
108 89.5 70.3 105.5





#66
4,6-Dinitro-2-methylphenol
Concen: 38.188 ng/ul
RT: 15.698 min Scan# 2110
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178



Tgt Ion:198 Resp: 200854
Ion Ratio Lower Upper
198 100
182 4.6 3.3 4.9
77 23.2 18.4 27.6




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#67
N-Nitrosodiphenylamine
Concen: 40.477 ng/ul
RT: 15.834 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:169 Resp: 1020143
Ion Ratio Lower Upper
169 100
168 66.8 52.5 78.7
167 36.6 28.4 42.6





#68
4-Bromophenyl-phenylether
Concen: 40.662 ng/ul
RT: 16.516 min Scan# 2249
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178



Tgt Ion:248 Resp: 387319
Ion Ratio Lower Upper
248 100
250 97.9 78.9 118.3
141 74.7 64.3 96.5




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#69
Hexachlorobenzene
Concen: 40.500 ng/ul
RT: 16.634 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35

Tgt Ion:284 Resp: 446728
Ion Ratio Lower Upper
284 100
142 33.5 29.8 44.8
249 32.8 26.6 39.8









#70
Atrazine
Concen: 38.539 ng/ul
RT: 16.792 min Scan# 2296
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35

Instrument :
BNA_P
ClientSampleId :
SLCS178

Tgt Ion:200 Resp: 402734
Ion Ratio Lower Upper
200 100
173 25.1 20.9 31.3
215 48.8 38.1 57.1




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#71
Pentachlorophenol
Concen: 39.095 ng/ul
RT: 16.981 min Scan# 2328
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:266 Resp: 267810
Ion Ratio Lower Upper
266 100
264 62.8 50.6 75.8
268 64.3 51.3 76.9





#72
Phenanthrene
Concen: 40.594 ng/u1
RT: 17.369 min Scan# 2394
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178




Tgt Ion:178 Resp: 1916717
Ion Ratio Lower Upper
178 100
179 15.1 12.1 18.1
176 19.4 15.2 22.8

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#74
Anthracene
Concen: 39.876 ng/u1
RT: 17.463 min Scan# 2410
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35



Tgt Ion:178 Resp: 1872730
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 18.6 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 38.430 ng/uL
RT: 13.375 min Scan# 1715
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35

Tgt Ion:216 Resp: 548869
Ion Ratio Lower Upper
216 100
214 78.0 62.0 93.0
218 47.7 37.6 56.4

Instrument :
BNA_P
ClientSampleId :
SLCS178




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#76
Pentachlorobenzene
Concen: 38.481 ng/uL
RT: 14.898 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:250 Resp: 529973
Ion Ratio Lower Upper
250 100
248 62.4 50.1 75.1
252 62.9 50.6 76.0





#77
 Carbazole
 Concen: 40.490 ng/ul
 RT: 17.733 min Scan# 2456
 Delta R.T. -0.000 min
 Lab File: BP007686.D
 Acq: 26 Oct 2021 19:35

Instrument :
 BNA_P
 ClientSampleId :
 SLCS178






Tgt Ion:167 Resp: 1755483


Ion	Ratio	Lower	Upper
167	100		
166	22.0	17.3	25.9
139	12.8	10.4	15.6

Manual Integrations
 APPROVED

mohammad
 10/27/2021 11:50:35 AM






#78
 Di-n-butylphthalate
 Concen: 41.906 ng/ul
 RT: 18.292 min Scan# 2551
 Delta R.T. -0.000 min
 Lab File: BP007686.D
 Acq: 26 Oct 2021 19:35



Tgt Ion:149 Resp: 2080158

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.3	10.9
104	5.1	4.2	6.4






#80
Fluoranthene
Concen: 41.011 ng/u1
RT: 19.339 min Scan# 2729
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178

Manual Integrations
APPROVED

mohammad
10/27/2021 11:50:35 AM







Tgt Ion:202 Resp: 2401835
Ion Ratio Lower Upper
202 100
101 9.4 8.4 12.6
100 7.0 6.3 9.5



#82
Pyrene
Concen: 40.752 ng/u1
RT: 19.692 min Scan# 2789
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:202 Resp: 2415575
Ion Ratio Lower Upper
202 100
101 11.0 9.9 14.9
100 8.8 7.9 11.9





#83
Butylbenzylphthalate
Concen: 41.884 ng/ul
RT: 20.569 min Scan# 2938
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178




Tgt Ion:149 Resp: 965994
Ion Ratio Lower Upper
149 100
91 68.3 55.9 83.9
206 22.2 17.4 26.0

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#84
3,3'-Dichlorobenzidine
Concen: 41.029 ng/ul
RT: 21.333 min Scan# 3068
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35



Tgt Ion:252 Resp: 780494
Ion Ratio Lower Upper
252 100
254 64.1 51.1 76.7
126 9.8 10.0 15.0#





#85
Benzo(a)anthracene
Concen: 40.866 ng/ul
RT: 21.404 min Scan# 3080
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35





Tgt Ion:228 Resp: 2490871

Ion	Ratio	Lower	Upper
228	100		
229	20.0	15.7	23.5
226	27.0	21.4	32.0

Instrument :
BNA_P
ClientSampled :
SLCS178

Manual Integrations
APPROVED




mohammad
10/27/2021 11:50:35 AM




#86
Bis(2-ethylhexyl)phthalate
Concen: 40.783 ng/ul
RT: 21.333 min Scan# 3068
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35

Tgt Ion:149 Resp: 1397648


Ion	Ratio	Lower	Upper
149	100		
167	27.1	22.6	33.8
279	4.1	3.3	4.9





#87
Chrysene
Concen: 40.404 ng/u1
RT: 21.457 min Scan# 3089
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampleId :
SLCS178



Tgt Ion:228 Resp: 2381107
Ion Ratio Lower Upper
228 100
226 30.2 23.7 35.5
229 20.0 16.0 24.0




Manual Integrations
APPROVED


mohammad
10/27/2021 11:50:35 AM



#89
Di-n-octyl phthalate
Concen: 41.622 ng/u1
RT: 22.280 min Scan# 3229
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35


Tgt Ion:149 Resp: 2530016





#90
Benzo(b)fluoranthene
Concen: 42.429 ng/ul
RT: 23.121 min Scan# 3372
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178




Tgt Ion:252 Resp: 2728132
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.2
125 8.8 7.8 11.6

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#91
Benzo(k)fluoranthene
Concen: 41.749 ng/ul
RT: 23.168 min Scan# 3380
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion:252 Resp: 2512099
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 8.5 8.0 12.0





#93
Benzo(a)pyrene
Concen: 41.966 ng/ul
RT: 23.757 min Scan# 3480
Delta R.T. -0.006 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178




Tgt Ion:252 Resp: 2614408
Ion Ratio Lower Upper
252 100
253 21.7 17.3 25.9
125 9.5 8.3 12.5

Manual Integrations
APPROVED



mohammad
10/27/2021 11:50:35 AM




#94
Indeno(1,2,3-cd)pyrene
Concen: 42.663 ng/ul
RT: 26.427 min Scan# 3934
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Tgt Ion:276 Resp: 3165910
Ion Ratio Lower Upper
276 100
138 21.4 17.8 26.8
277 25.0 19.8 29.6





#95
Dibenzo(a,h)anthracene
Concen: 42.749 ng/ul
RT: 26.445 min Scan# 3937
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35




Instrument :
BNA_P
ClientSampled :
SLCS178




Tgt Ion:278 Resp: 2718445
Ion Ratio Lower Upper
278 100
139 14.2 12.8 19.2
279 23.8 19.1 28.7

Manual Integrations
APPROVED

mohammad
10/27/2021 11:50:35 AM



#96
Benzo(g,h,i)perylene
Concen: 43.269 ng/ul
RT: 27.209 min Scan# 4067
Delta R.T. -0.000 min
Lab File: BP007686.D
Acq: 26 Oct 2021 19:35



Tgt Ion:276 Resp: 2730987
Ion Ratio Lower Upper
276 100
138 19.6 17.5 26.3
277 24.0 18.5 27.7

