

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
 Data File : BP007710.D
 Acq On : 27 Oct 2021 09:45
 Operator : CG/JU
 Sample : M4303-15
 Misc :
 ALS Vial : 68 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 P021-TW001-02

Quant Time: Oct 27 10:20:05 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 25 16:57:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	175689	20.000	ng/u1	0.00
20) Naphthalene-d8	10.734	136	712242	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.569	164	458748	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.328	188	1001855	20.000	ng/u1	0.00
79) Chrysene-d12	21.416	240	1091680	20.000	ng/u1	0.00
88) Perylene-d12	23.863	264	1152435	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	23015	5.951	ng/uL	0.00
4) Pyridine-d5	3.787	84	94088	8.045	ng/u1	0.00
7) Phenol-d5	7.093	99	121082	9.074	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.258	67	250865	33.860	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.463	132	286711	27.625	ng/u1	-0.01
15) 4-Methylphenol-d8	8.640	113	199376	19.080	ng/u1	-0.01
21) Nitrobenzene-d5	9.087	128	159282	32.816	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.816	143	161325	30.432	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.352	165	311850	28.900	ng/u1	-0.02
31) 4-Chloroaniline-d4	10.869	131	399187	27.978	ng/u1	-0.01
46) Dimethylphthalate-d6	13.981	166	1128638	36.321	ng/u1	-0.01
49) Acenaphthylene-d8	14.263	160	1250850	32.842	ng/u1	0.00
54) 4-Nitrophenol-d4	14.763	143	43888	7.705	ng/u1	-0.01
60) Fluorene-d10	15.563	176	961187	36.691	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.681	200	142825	26.652	ng/u1	0.00
73) Anthracene-d10	17.428	188	1686872	40.971	ng/u1	0.00
81) Pyrene-d10	19.663	212	2074843	40.823	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.710	264	2153710	38.043	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102521\
Data File : BP007710.D
Acq On : 27 Oct 2021 09:45
Operator : CG/JU
Sample : M4303-15
Misc :
ALS Vial : 68 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
P021-TW001-02

Quant Time: Oct 27 10:20:05 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP102521.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 25 16:57:16 2021
Response via : Initial Calibration

